

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101222\
 Data File : BN022060.D
 Acq On : 12 Oct 2022 11:27
 Operator : CG/JU
 Sample : SST00535
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST005245

Quant Time: Oct 13 02:14:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 13 02:12:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.958	152	117300	20.000	ng/u1	0.00
20) Naphthalene-d8	10.787	136	540747	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.634	164	359091	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.386	188	751771	20.000	ng/u1	0.00
79) Chrysene-d12	21.586	240	763581	20.000	ng/u1	# 0.00
88) Perylene-d12	24.057	264	810986	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.258	96	5750	1.855	ng/uL	0.00
4) Pyridine-d5	3.705	84	37297	3.981	ng/u1	0.01
7) Phenol-d5	7.111	99	43654	3.849	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.281	67	31700	4.366	ng/u1	0.00
11) 2-Chlorophenol-d4	7.481	132	33097	4.231	ng/u1	0.00
15) 4-Methylphenol-d8	8.681	113	33781	3.770	ng/u1	0.00
21) Nitrobenzene-d5	9.140	128	16046	4.181	ng/u1	0.00
24) 2-Nitrophenol-d4	9.869	143	15106	3.861	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.404	165	30639	4.140	ng/u1	0.00
31) 4-Chloroaniline-d4	10.934	131	53671	4.295	ng/u1	0.00
46) Dimethylphthalate-d6	14.045	166	111191	4.406	ng/u1	0.00
49) Acenaphthylene-d8	14.328	160	115874	4.220	ng/u1	0.00
54) 4-Nitrophenol-d4	14.834	143	17868	3.363	ng/u1	0.00
60) Fluorene-d10	15.628	176	97115	4.546	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	12681	3.079	ng/u1	0.00
73) Anthracene-d10	17.486	188	149795	4.614	ng/u1	0.00
81) Pyrene-d10	19.786	212	171304	4.614	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.892	264	177236	4.520	ng/u1	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.293	88	7136	2.125	ng/uL	94
5) Pyridine	3.722	79	34799	3.695	ng/u1	96
6) Benzaldehyde	7.093	77	25573	4.502	ng/u1	95
8) Phenol	7.140	94	47796	3.960	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.381	93	42433	4.409	ng/u1	99
12) 2-Chlorophenol	7.516	128	37505	4.358	ng/u1	96
13) 2-Methylphenol	8.410	108	35081	3.924	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.499	45	59148	4.110	ng/u1	97
16) Acetophenone	8.799	105	62063	4.217	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.775	70	30846	4.005	ng/u1	100
18) 4-Methylphenol	8.740	108	39499	3.937	ng/u1	99
19) Hexachloroethane	9.046	117	15034	4.357	ng/u1	95
22) Nitrobenzene	9.181	77	44208	4.192	ng/u1	96
23) Isophorone	9.710	82	82125	3.984	ng/u1	100
25) 2-Nitrophenol	9.899	139	18175	3.919	ng/u1	96
26) 2,4-Dimethylphenol	9.957	107	42627	4.261	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.199	93	57631	4.416	ng/u1	99
29) 2,4-Dichlorophenol	10.434	162	33886	4.293	ng/u1	97
30) Naphthalene	10.840	128	136900	4.665	ng/u1	99
32) 4-Chloroaniline	10.957	127	57600	4.379	ng/u1	96
33) Hexachlorobutadiene	11.122	225	21091	4.829	ng/u1	93
34) Caprolactam	11.740	113	9119	2.970	ng/u1	88
35) 4-Chloro-3-methylphenol	12.081	107	38044	3.962	ng/u1	96
36) 2-Methylnaphthalene	12.457	142	91063	4.557	ng/u1	95

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101222\
 Data File : BN022060.D
 Acq On : 12 Oct 2022 11:27
 Operator : CG/JU
 Sample : SST00535
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SST005245


Quant Time: Oct 13 02:14:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 13 02:12:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.675	142	91117	4.512	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.822	216	42419	4.628	ng/ul	95
40) Hexachlorocyclopentadiene	12.798	237	17804	3.692	ng/ul	91
41) 2,4,6-Trichlorophenol	13.063	196	25076	4.060	ng/ul	98
42) 2,4,5-Trichlorophenol	13.134	196	28607	4.209	ng/ul	97
43) 1,1'-Biphenyl	13.463	154	122679	4.686	ng/ul	98
44) 2-Chloronaphthalene	13.510	162	95658	4.686	ng/ul	96
45) 2-Nitroaniline	13.722	65	22252	3.539	ng/ul	96
47) Dimethylphthalate	14.087	163	123327	4.599	ng/ul	99
48) 2,6-Dinitrotoluene	14.216	165	17198	3.594	ng/ul	94
50) Acenaphthylene	14.357	152	137850	4.369	ng/ul	98
51) 3-Nitroaniline	14.545	138	19123	3.317	ng/ul	98
52) Acenaphthene	14.698	153	106657	4.646	ng/ul	99
53) 2,4-Dinitrophenol	14.751	184	8494	2.577	ng/ul	93
55) 4-Nitrophenol	14.845	109	15392	3.540	ng/ul	96
56) Dibenzofuran	15.034	168	147533	4.768	ng/ul	98
57) 2,4-Dinitrotoluene	15.004	165	25858	3.573	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.257	232	23823	4.126	ng/ul	96
59) Diethylphthalate	15.451	149	120035	4.318	ng/ul	99
61) Fluorene	15.681	166	119709	4.680	ng/ul	95
62) 4-Chlorophenyl-phenyle...	15.675	204	57510	4.952	ng/ul	98
63) 4-Nitroaniline	15.710	138	18414	3.210	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.757	198	14792	3.348	ng/ul#	90
67) N-Nitrosodiphenylamine	15.892	169	99637	4.580	ng/ul	97
68) 4-Bromophenyl-phenylether	16.575	248	28825	4.265	ng/ul	86
69) Hexachlorobenzene	16.686	284	37723	4.708	ng/ul	97
70) Atrazine	16.845	200	31181	4.240	ng/ul	98
71) Pentachlorophenol	17.034	266	17486	3.596	ng/ul#	87
72) Phenanthrene	17.428	178	197260	4.819	ng/ul	98
74) Anthracene	17.522	178	189058	4.638	ng/ul	97
75) 1,2,3,4-Tetrachloroben...	13.428	216	43248	4.750	ng/ul	97
76) Pentachlorobenzene	14.945	250	47222	4.793	ng/ul	93
77) Carbazole	17.798	167	173414	4.609	ng/ul	98
78) Di-n-butylphthalate	18.357	149	187250	4.064	ng/ul	100
80) Fluoranthene	19.451	202	214525	4.550	ng/ul	97
82) Pyrene	19.816	202	238401	4.791	ng/ul	99
83) Butylbenzylphthalate	20.716	149	76399	3.601	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.504	252	64046	3.858	ng/ul	92
85) Benzo(a)anthracene	21.569	228	229340	4.781	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.492	149	114141	3.705	ng/ul#	99
87) Chrysene	21.622	228	226838	4.806	ng/ul	96
89) Di-n-octyl phthalate	22.439	149	194903	3.446	ng/ul	100
90) Benzo(b)fluoranthene	23.298	252	235799	4.605	ng/ul	98
91) Benzo(k)fluoranthene	23.351	252	238489	4.722	ng/ul	98
93) Benzo(a)pyrene	23.945	252	209159	4.626	ng/ul	96
94) Indeno(1,2,3-cd)pyrene	26.651	276	263290	4.587	ng/ul	96
95) Dibenzo(a,h)anthracene	26.668	278	237892	4.583	ng/ul	96
96) Benzo(g,h,i)perylene	27.451	276	233064	4.505	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 13 02:14:30 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101222.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 13 02:12:15 2022
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 2.125 ng/uL
 RT: 3.293 min Scan# 51
 Delta R.T. -0.000 min
 Lab File: BN022060.D
 Acq: 12 Oct 2022 11:27




Instrument :
 BNA_N
 ClientSampleId :
 SST005245


Tgt Ion: 88 Resp: 7136
 Ion Ratio Lower Upper
 88 100
 43 27.5 22.6 34.0
 58 68.3 59.9 89.9



#5
 Pyridine
 Concen: 3.695 ng/uL
 RT: 3.722 min Scan# 125
 Delta R.T. 0.006 min
 Lab File: BN022060.D
 Acq: 12 Oct 2022 11:27

Tgt Ion: 79 Resp: 34799
 Ion Ratio Lower Upper
 79 100
 52 73.5 61.7 92.5
 51 36.0 27.8 41.6









#6
Benzaldehyde
Concen: 4.502 ng/ul
RT: 7.093 min Scan# 698
Delta R.T. -0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion: 77 Resp: 25573
Ion Ratio Lower Upper
77 100
105 94.9 78.6 118.0
106 87.3 75.4 113.2



#8
Phenol
Concen: 3.960 ng/ul
RT: 7.140 min Scan# 706
Delta R.T. 0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion: 94 Resp: 47796
Ion Ratio Lower Upper
94 100
65 26.7 22.1 33.1
66 35.4 29.0 43.6









#10
Bis(2-Chloroethyl)ether
Concen: 4.409 ng/ul
RT: 7.381 min Scan# 747
Delta R.T. 0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion: 93 Resp: 42433
Ion Ratio Lower Upper
93 100
63 77.6 61.1 91.7
95 32.8 26.4 39.6



#12
2-Chlorophenol
Concen: 4.358 ng/ul
RT: 7.516 min Scan# 770
Delta R.T. 0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion:128 Resp: 37505
Ion Ratio Lower Upper
128 100
64 48.0 39.7 59.5
130 30.4 26.7 40.1









#13
2-Methylphenol
Concen: 3.924 ng/ul
RT: 8.410 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:108 Resp: 35081
Ion Ratio Lower Upper
108 100
107 88.3 72.4 108.6



#14
2,2'-oxybis(1-Chloropropane)
Concen: 4.110 ng/ul
RT: 8.499 min Scan# 937
Delta R.T. -0.006 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion: 45 Resp: 59148
Ion Ratio Lower Upper
45 100
77 16.2 11.7 17.5
79 10.1 8.7 13.1









#16
Acetophenone
Concen: 4.217 ng/ul
RT: 8.799 min Scan# 988
Delta R.T. 0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:105 Resp: 62063
Ion Ratio Lower Upper
105 100
77 79.4 64.9 97.3
51 26.8 21.8 32.6



#17
N-Nitroso-di-n-propylamine
Concen: 4.005 ng/ul
RT: 8.775 min Scan# 984
Delta R.T. -0.006 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion: 70 Resp: 30846
Ion Ratio Lower Upper
70 100
42 57.3 45.6 68.4
101 9.6 7.4 11.2
130 19.4 15.4 23.0









#18
4-Methylphenol
Concen: 3.937 ng/ul
RT: 8.740 min Scan# 91
Delta R.T. -0.006 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:108 Resp: 39499
Ion Ratio Lower Upper
108 100
107 110.2 87.2 130.8



#19
Hexachloroethane
Concen: 4.357 ng/ul
RT: 9.046 min Scan# 1030
Delta R.T. -0.006 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion:117 Resp: 15034
Ion Ratio Lower Upper
117 100
201 93.4 69.8 104.6
199 54.8 45.2 67.8









#22
Nitrobenzene
Concen: 4.192 ng/ul
RT: 9.181 min Scan# 1053
Delta R.T. 0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion: 77 Resp: 44208
Ion Ratio Lower Upper
77 100
123 40.5 34.9 52.3
65 11.4 10.2 15.2



#23
Isophorone
Concen: 3.984 ng/ul
RT: 9.710 min Scan# 1143
Delta R.T. 0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion: 82 Resp: 82125
Ion Ratio Lower Upper
82 100
95 6.5 5.4 8.0
138 17.3 13.8 20.8









#25
2-Nitrophenol
Concen: 3.919 ng/ul
RT: 9.899 min Scan# 1175
Delta R.T. -0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:139 Resp: 18175
Ion Ratio Lower Upper
139 100
65 56.9 42.7 64.1
109 28.7 23.8 35.6



#26
2,4-Dimethylphenol
Concen: 4.261 ng/ul
RT: 9.957 min Scan# 1185
Delta R.T. 0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion:107 Resp: 42627
Ion Ratio Lower Upper
107 100
121 53.1 41.0 61.6
122 87.2 71.5 107.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 4.416 ng/ul
 RT: 10.199 min Scan# 1226
 Delta R.T. 0.000 min
 Lab File: BN022060.D
 Acq: 12 Oct 2022 11:27




Instrument :
 BNA_N
 ClientSampleId :
 SSTD005245


Tgt Ion: 93 Resp: 57631
 Ion Ratio Lower Upper
 93 100
 95 32.2 25.2 37.8
 123 11.2 8.6 12.8



#29
 2,4-Dichlorophenol
 Concen: 4.293 ng/ul
 RT: 10.434 min Scan# 1266
 Delta R.T. -0.000 min
 Lab File: BN022060.D
 Acq: 12 Oct 2022 11:27

Tgt Ion: 162 Resp: 33886
 Ion Ratio Lower Upper
 162 100
 164 64.3 52.9 79.3
 98 36.3 30.5 45.7









#30
Naphthalene
Concen: 4.665 ng/ul
RT: 10.840 min Scan# 1335
Delta R.T. -0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:128 Resp: 136900
Ion Ratio Lower Upper
128 100
129 10.3 8.5 12.7
127 13.4 10.8 16.2



#32
4-Chloroaniline
Concen: 4.379 ng/ul
RT: 10.957 min Scan# 1355
Delta R.T. 0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion:127 Resp: 57600
Ion Ratio Lower Upper
127 100
129 34.9 26.2 39.4









#33
Hexachlorobutadiene
Concen: 4.829 ng/ul
RT: 11.122 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:225 Resp: 21091
Ion Ratio Lower Upper
225 100
223 58.0 47.7 71.5
227 55.8 51.4 77.0



#34
Caprolactam
Concen: 2.970 ng/ul
RT: 11.740 min Scan# 1488
Delta R.T. -0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion:113 Resp: 9119
Ion Ratio Lower Upper
113 100
55 159.4 138.7 208.1
56 110.6 101.4 152.0









#35
4-Chloro-3-methylphenol
Concen: 3.962 ng/ul
RT: 12.081 min Scan# 1546
Delta R.T. 0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:107 Resp: 38044
Ion Ratio Lower Upper
107 100
144 24.7 20.3 30.5
142 76.5 64.4 96.6



#36
2-Methylnaphthalene
Concen: 4.557 ng/ul
RT: 12.457 min Scan# 1610
Delta R.T. 0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion:142 Resp: 91063
Ion Ratio Lower Upper
142 100
141 82.8 69.7 104.5









#37
1-Methylnaphthalene
Concen: 4.512 ng/ul
RT: 12.675 min Scan# 1647
Delta R.T. 0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:142 Resp: 91117
Ion Ratio Lower Upper
142 100
141 93.6 76.7 115.1
116 4.1 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 4.628 ng/ul
RT: 12.822 min Scan# 1672
Delta R.T. 0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion:216 Resp: 42419
Ion Ratio Lower Upper
216 100
214 87.3 65.4 98.2
179 23.2 17.4 26.0
108 22.1 16.4 24.6









#40
Hexachlorocyclopentadiene
Concen: 3.692 ng/ul
RT: 12.798 min Scan# 1668
Delta R.T. -0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:237 Resp: 17804
Ion Ratio Lower Upper
237 100
235 61.8 43.4 65.0
272 11.7 10.2 15.2



#41
2,4,6-Trichlorophenol
Concen: 4.060 ng/ul
RT: 13.063 min Scan# 1713
Delta R.T. 0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion:196 Resp: 25076
Ion Ratio Lower Upper
196 100
198 94.5 73.4 110.2
200 30.6 24.3 36.5









#42
2,4,5-Trichlorophenol
Concen: 4.209 ng/ul
RT: 13.134 min Scan# 1725
Delta R.T. 0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:	196	Resp:	28607
Ion Ratio	Lower	Upper	
196	100		
198	93.4	77.0	115.6
200	28.4	24.7	37.1



#43
1,1'-Biphenyl
Concen: 4.686 ng/ul
RT: 13.463 min Scan# 1781
Delta R.T. -0.006 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion:	154	Resp:	122679
Ion Ratio	Lower	Upper	
154	100		
153	40.0	30.9	46.3
76	15.3	11.6	17.4









#44
2-Chloronaphthalene
Concen: 4.686 ng/ul
RT: 13.510 min Scan# 1789
Delta R.T. 0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:162 Resp: 95658
Ion Ratio Lower Upper
162 100
164 32.4 23.7 35.5
127 39.9 30.8 46.2



#45
2-Nitroaniline
Concen: 3.539 ng/ul
RT: 13.722 min Scan# 1825
Delta R.T. -0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion: 65 Resp: 22252
Ion Ratio Lower Upper
65 100
92 71.2 56.2 84.2
138 100.4 85.0 127.6









#47
Dimethylphthalate
Concen: 4.599 ng/ul
RT: 14.087 min Scan# 1888
Delta R.T. -0.006 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:163 Resp: 123327
Ion Ratio Lower Upper
163 100
194 3.8 3.1 4.7
164 10.7 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 3.594 ng/ul
RT: 14.216 min Scan# 1909
Delta R.T. 0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion:165 Resp: 17198
Ion Ratio Lower Upper
165 100
89 71.0 52.1 78.1
121 24.1 19.0 28.4





#50

Acenaphthylene

Concen: 4.369 ng/ul

RT: 14.357 min Scan# 1933

Delta R.T. 0.000 min

Lab File: BN022060.D

Acq: 12 Oct 2022 11:27

Instrument :

BNA_N

ClientSampleId :

SSTD005245





Tgt Ion:152 Resp: 137850

Ion Ratio Lower Upper

152 100

151 20.3 15.5 23.3

153 13.0 9.8 14.8



#51

3-Nitroaniline

Concen: 3.317 ng/ul

RT: 14.545 min Scan# 1965

Delta R.T. -0.000 min

Lab File: BN022060.D

Acq: 12 Oct 2022 11:27




Tgt Ion:138 Resp: 19123


Ion Ratio Lower Upper

138 100

108 12.4 9.0 13.6

92 122.8 96.4 144.6









#52
Acenaphthene
Concen: 4.646 ng/ul
RT: 14.698 min Scan# 1991
Delta R.T. -0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:153 Resp: 106657
Ion Ratio Lower Upper
153 100
152 46.4 38.0 57.0
154 89.8 71.1 106.7



#53
2,4-Dinitrophenol
Concen: 2.577 ng/ul
RT: 14.751 min Scan# 2000
Delta R.T. 0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion:184 Resp: 8494
Ion Ratio Lower Upper
184 100
63 78.2 56.8 85.2
154 64.5 49.0 73.6









#55
4-Nitrophenol
Concen: 3.540 ng/ul
RT: 14.845 min Scan# 2016
Delta R.T. -0.006 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:109 Resp: 15392
Ion Ratio Lower Upper
109 100
139 133.5 108.3 162.5
65 143.5 107.9 161.9



#56
Dibenzofuran
Concen: 4.768 ng/ul
RT: 15.034 min Scan# 2048
Delta R.T. 0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion:168 Resp: 147533
Ion Ratio Lower Upper
168 100
139 36.9 30.6 45.8









#57
2,4-Dinitrotoluene
Concen: 3.573 ng/ul
RT: 15.004 min Scan# 2043
Delta R.T. -0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:165 Resp: 25858
Ion Ratio Lower Upper
165 100
63 47.9 36.2 54.4
182 2.4 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 4.126 ng/ul
RT: 15.257 min Scan# 2086
Delta R.T. 0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion:232 Resp: 23823
Ion Ratio Lower Upper
232 100
131 52.9 44.4 66.6
130 2.7 2.2 3.4
166 33.3 28.8 43.2









#59
Diethylphthalate
Concen: 4.318 ng/ul
RT: 15.451 min Scan# 2119
Delta R.T. -0.006 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:149 Resp: 120035
Ion Ratio Lower Upper
149 100
177 21.5 17.5 26.3
150 12.0 9.8 14.8



#61
Fluorene
Concen: 4.680 ng/ul
RT: 15.681 min Scan# 2158
Delta R.T. -0.006 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion:166 Resp: 119709
Ion Ratio Lower Upper
166 100
165 96.8 72.8 109.2
167 12.7 10.6 16.0





#62

4-Chlorophenyl-phenylether

Concen: 4.952 ng/ul

RT: 15.675 min Scan# 2157

Delta R.T. -0.006 min

Lab File: BN022060.D

Acq: 12 Oct 2022 11:27

Instrument :

BNA_N

ClientSampleId :

SSTD005245





Tgt Ion:204 Resp: 57510

Ion Ratio Lower Upper

204 100

206 30.9 26.2 39.2

141 64.4 52.0 78.0



#63

4-Nitroaniline

Concen: 3.210 ng/ul

RT: 15.710 min Scan# 2163

Delta R.T. 0.000 min

Lab File: BN022060.D

Acq: 12 Oct 2022 11:27




Tgt Ion:138 Resp: 18414


Ion Ratio Lower Upper

138 100

92 54.5 41.8 62.8

108 75.4 60.4 90.6









#66
 4,6-Dinitro-2-methylphenol
 Concen: 3.348 ng/ul
 RT: 15.757 min Scan# 2171
 Delta R.T. -0.006 min
 Lab File: BN022060.D
 Acq: 12 Oct 2022 11:27




Instrument :
 BNA_N
 ClientSampleId :
 SST005245


Tgt Ion:198 Resp: 14792
 Ion Ratio Lower Upper
 198 100
 182 10.9 3.8 5.8#
 77 35.2 24.4 36.6



#67
 N-Nitrosodiphenylamine
 Concen: 4.580 ng/ul
 RT: 15.892 min Scan# 2194
 Delta R.T. 0.000 min
 Lab File: BN022060.D
 Acq: 12 Oct 2022 11:27

Tgt Ion:169 Resp: 99637
 Ion Ratio Lower Upper
 169 100
 168 66.1 50.9 76.3
 167 37.0 28.2 42.4









#68
4-Bromophenyl-phenylether
Concen: 4.265 ng/ul
RT: 16.575 min Scan# 2310
Delta R.T. 0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:248 Resp: 28825
Ion Ratio Lower Upper
248 100
250 87.7 86.3 129.5
141 86.0 74.2 111.4



#69
Hexachlorobenzene
Concen: 4.708 ng/ul
RT: 16.686 min Scan# 2329
Delta R.T. -0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion:284 Resp: 37723
Ion Ratio Lower Upper
284 100
142 45.6 35.7 53.5
249 30.4 27.0 40.6





#70

Atrazine

Concen: 4.240 ng/ul

RT: 16.845 min Scan# 21

Delta R.T. -0.006 min

Lab File: BN022060.D

Acq: 12 Oct 2022 11:27

Instrument :

BNA_N

ClientSampleId :

SSTD005245





Tgt Ion:200 Resp: 31181

Ion Ratio Lower Upper

200 100

173 26.0 22.8 34.2

215 47.6 38.1 57.1



#71

Pentachlorophenol

Concen: 3.596 ng/ul

RT: 17.034 min Scan# 2388

Delta R.T. 0.000 min

Lab File: BN022060.D

Acq: 12 Oct 2022 11:27




Tgt Ion:266 Resp: 17486


Ion Ratio Lower Upper

266 100

264 52.0 52.7 79.1#

268 60.6 54.6 82.0





#72

Phenanthrene

Concen: 4.819 ng/ul

RT: 17.428 min Scan# 2456

Delta R.T. -0.006 min

Lab File: BN022060.D

Acq: 12 Oct 2022 11:27

Instrument :

BNA_N

ClientSampleId :

SSTD005245





Tgt Ion:178 Resp: 197260

Ion Ratio Lower Upper

178 100

179 15.2 12.3 18.5

176 19.5 16.5 24.7



#74

Anthracene

Concen: 4.638 ng/ul

RT: 17.522 min Scan# 2471

Delta R.T. -0.000 min

Lab File: BN022060.D

Acq: 12 Oct 2022 11:27




Tgt Ion:178 Resp: 189058


Ion Ratio Lower Upper

178 100

179 15.5 13.0 19.6

176 17.8 15.8 23.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 4.750 ng/uL
RT: 13.428 min Scan# 1775
Delta R.T. -0.006 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:216 Resp: 43248
Ion Ratio Lower Upper
216 100
214 79.4 61.2 91.8
218 47.3 39.4 59.0



#76
Pentachlorobenzene
Concen: 4.793 ng/uL
RT: 14.945 min Scan# 2033
Delta R.T. -0.006 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion:250 Resp: 47222
Ion Ratio Lower Upper
250 100
248 64.7 46.8 70.2
252 62.2 53.9 80.9









#77
 Carbazole
 Concen: 4.609 ng/ul
 RT: 17.798 min Scan# 2518
 Delta R.T. -0.000 min
 Lab File: BN022060.D
 Acq: 12 Oct 2022 11:27




Instrument :
 BNA_N
 ClientSampleId :
 SSTD005245


Tgt Ion:167 Resp: 173414
 Ion Ratio Lower Upper
 167 100
 166 20.7 17.3 25.9
 139 13.0 10.7 16.1



#78
 Di-n-butylphthalate
 Concen: 4.064 ng/ul
 RT: 18.357 min Scan# 2613
 Delta R.T. -0.000 min
 Lab File: BN022060.D
 Acq: 12 Oct 2022 11:27

Tgt Ion:149 Resp: 187250
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.1 10.7
 104 5.4 4.2 6.2









#80
Fluoranthene
Concen: 4.550 ng/ul
RT: 19.451 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:202 Resp: 214525
Ion Ratio Lower Upper
202 100
101 14.2 10.3 15.5
100 10.7 7.8 11.8



#82
Pyrene
Concen: 4.791 ng/ul
RT: 19.816 min Scan# 2861
Delta R.T. 0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion:202 Resp: 238401
Ion Ratio Lower Upper
202 100
101 16.5 13.4 20.0
100 13.1 10.9 16.3









#83
Butylbenzylphthalate
Concen: 3.601 ng/ul
RT: 20.716 min Scan# 3014
Delta R.T. 0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:149 Resp: 76399
Ion Ratio Lower Upper
149 100
91 83.7 61.7 92.5
206 19.1 14.7 22.1



#84
3,3'-Dichlorobenzidine
Concen: 3.858 ng/ul
RT: 21.504 min Scan# 3148
Delta R.T. -0.006 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion:252 Resp: 64046
Ion Ratio Lower Upper
252 100
254 73.3 52.9 79.3
126 15.4 11.2 16.8









#85
Benzo(a)anthracene
Concen: 4.781 ng/ul
RT: 21.569 min Scan# 3159
Delta R.T. -0.006 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:228 Resp: 229340
Ion Ratio Lower Upper
228 100
229 18.1 13.5 20.3
226 25.9 21.0 31.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 3.705 ng/ul
RT: 21.492 min Scan# 3146
Delta R.T. 0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion:149 Resp: 114141
Ion Ratio Lower Upper
149 100
167 26.4 21.3 31.9
279 2.8 2.9 4.3#









#87
Chrysene
Concen: 4.806 ng/ul
RT: 21.622 min Scan# 3169
Delta R.T. -0.006 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD005245


Tgt Ion:228 Resp: 226838
Ion Ratio Lower Upper
228 100
226 27.5 23.8 35.8
229 19.4 16.5 24.7



#89
Di-n-octyl phthalate
Concen: 3.446 ng/ul
RT: 22.439 min Scan# 3307
Delta R.T. -0.000 min
Lab File: BN022060.D
Acq: 12 Oct 2022 11:27

Tgt Ion:149 Resp: 194903





#90

Benzo(b)fluoranthene

Concen: 4.605 ng/ul

RT: 23.298 min Scan# 3453

Delta R.T. 0.000 min

Lab File: BN022060.D

Acq: 12 Oct 2022 11:27

Instrument :

BNA_N

ClientSampleId :

SSTD005245





Tgt Ion:252 Resp: 235799

Ion Ratio Lower Upper

252 100

253 21.0 17.6 26.4

125 12.9 10.7 16.1



#91

Benzo(k)fluoranthene

Concen: 4.722 ng/ul

RT: 23.351 min Scan# 3462

Delta R.T. -0.000 min

Lab File: BN022060.D

Acq: 12 Oct 2022 11:27




Tgt Ion:252 Resp: 238489


Ion Ratio Lower Upper

252 100

253 21.2 17.7 26.5

125 12.2 10.6 15.8





#93

Benzo(a)pyrene

Concen: 4.626 ng/ul

RT: 23.945 min Scan# 3564

Delta R.T. -0.006 min

Lab File: BN022060.D

Acq: 12 Oct 2022 11:27

Instrument :

BNA_N

ClientSampleId :

SSTD005245





Tgt Ion:252 Resp: 209159

Ion Ratio Lower Upper

252 100

253 19.0 17.2 25.8

125 13.0 10.8 16.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 4.587 ng/ul

RT: 26.651 min Scan# 4023

Delta R.T. -0.012 min

Lab File: BN022060.D

Acq: 12 Oct 2022 11:27




Tgt Ion:276 Resp: 263290


Ion Ratio Lower Upper

276 100

138 27.8 24.6 36.8

277 21.8 18.4 27.6









#95
 Dibenzo(a,h)anthracene
 Concen: 4.583 ng/ul
 RT: 26.668 min Scan# 4159
 Delta R.T. -0.006 min
 Lab File: BN022060.D
 Acq: 12 Oct 2022 11:27

Instrument :
 BNA_N
 ClientSampleId :
 SSTD005245

Tgt Ion:278 Resp: 237892
 Ion Ratio Lower Upper
 278 100
 139 17.9 15.7 23.5
 279 22.5 16.5 24.7



#96
 Benzo(g,h,i)perylene
 Concen: 4.505 ng/ul
 RT: 27.451 min Scan# 4159
 Delta R.T. -0.006 min
 Lab File: BN022060.D
 Acq: 12 Oct 2022 11:27

Tgt Ion:276 Resp: 233064
 Ion Ratio Lower Upper
 276 100
 138 23.8 21.2 31.8
 277 20.9 17.9 26.9

