

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101222\
 Data File : BN022061.D
 Acq On : 12 Oct 2022 12:42
 Operator : CG/JU
 Sample : SST01036
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST010246

Quant Time: Oct 13 02:15:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 13 02:12:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.957	152	118752	20.000	ng/ul	0.00
20) Naphthalene-d8	10.793	136	575820	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.634	164	389529	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.386	188	826594	20.000	ng/ul	0.00
79) Chrysene-d12	21.592	240	841584	20.000	ng/ul	0.00
88) Perylene-d12	24.057	264	855621	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.258	96	12425	3.960	ng/uL	0.00
4) Pyridine-d5	3.699	84	91207	9.615	ng/ul	0.00
7) Phenol-d5	7.116	99	100628	8.765	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.287	67	71944	9.788	ng/ul	0.00
11) 2-Chlorophenol-d4	7.481	132	73299	9.255	ng/ul	0.00
15) 4-Methylphenol-d8	8.681	113	82340	9.078	ng/ul	0.00
21) Nitrobenzene-d5	9.140	128	39117	9.571	ng/ul	0.00
24) 2-Nitrophenol-d4	9.869	143	38328	9.199	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.410	165	78638	9.979	ng/ul	0.00
31) 4-Chloroaniline-d4	10.934	131	132425	9.951	ng/ul	0.00
46) Dimethylphthalate-d6	14.045	166	272328	9.948	ng/ul	0.00
49) Acenaphthylene-d8	14.328	160	307371	10.320	ng/ul	0.00
54) 4-Nitrophenol-d4	14.834	143	52433	9.096	ng/ul	0.00
60) Fluorene-d10	15.628	176	239628	10.340	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	39581	8.742	ng/ul	0.00
73) Anthracene-d10	17.486	188	372800	10.443	ng/ul	0.00
81) Pyrene-d10	19.786	212	412832	10.089	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.892	264	414064	10.008	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.293	88	13998	4.118	ng/uL	97
5) Pyridine	3.722	79	92549	9.708	ng/ul	97
6) Benzaldehyde	7.093	77	57998	10.085	ng/ul	99
8) Phenol	7.140	94	112632	9.218	ng/ul	98
10) Bis(2-Chloroethyl)ether	7.381	93	95915	9.845	ng/ul	99
12) 2-Chlorophenol	7.516	128	85876	9.856	ng/ul	98
13) 2-Methylphenol	8.410	108	82271	9.091	ng/ul	97
14) 2,2'-oxybis(1-Chloropr...	8.499	45	140339	9.633	ng/ul	99
16) Acetophenone	8.799	105	144547	9.701	ng/ul	97
17) N-Nitroso-di-n-propyla...	8.781	70	76616	9.825	ng/ul	99
18) 4-Methylphenol	8.740	108	95354	9.387	ng/ul	99
19) Hexachloroethane	9.052	117	34142	9.774	ng/ul	97
22) Nitrobenzene	9.181	77	109289	9.732	ng/ul	98
23) Isophorone	9.710	82	216729	9.875	ng/ul	99
25) 2-Nitrophenol	9.899	139	47455	9.609	ng/ul	97
26) 2,4-Dimethylphenol	9.957	107	107336	10.075	ng/ul	97
27) Bis(2-Chloroethoxy)met...	10.199	93	138208	9.946	ng/ul	98
29) 2,4-Dichlorophenol	10.440	162	80022	9.521	ng/ul	98
30) Naphthalene	10.840	128	313051	10.017	ng/ul	99
32) 4-Chloroaniline	10.957	127	142409	10.166	ng/ul	98
33) Hexachlorobutadiene	11.122	225	46450	9.987	ng/ul	89
34) Caprolactam	11.746	113	26740	8.179	ng/ul	99
35) 4-Chloro-3-methylphenol	12.081	107	99118	9.693	ng/ul	99
36) 2-Methylnaphthalene	12.457	142	219705	10.325	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101222\
 Data File : BN022061.D
 Acq On : 12 Oct 2022 12:42
 Operator : CG/JU
 Sample : SSTD01036
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD010246

Quant Time: Oct 13 02:15:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 13 02:12:15 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.675	142	223713	10.403	ng/ul	94
39) 1,2,4,5-Tetrachloroben...	12.822	216	100310	10.089	ng/ul	98
40) Hexachlorocyclopentadiene	12.798	237	47819	9.141	ng/ul	96
41) 2,4,6-Trichlorophenol	13.069	196	67107	10.016	ng/ul	97
42) 2,4,5-Trichlorophenol	13.140	196	75163	10.195	ng/ul	96
43) 1,1'-Biphenyl	13.469	154	292550	10.300	ng/ul	96
44) 2-Chloronaphthalene	13.510	162	227881	10.292	ng/ul	95
45) 2-Nitroaniline	13.722	65	62979	9.234	ng/ul	94
47) Dimethylphthalate	14.092	163	300445	10.329	ng/ul	100
48) 2,6-Dinitrotoluene	14.216	165	50738	9.773	ng/ul	99
50) Acenaphthylene	14.357	152	358120	10.464	ng/ul	99
51) 3-Nitroaniline	14.545	138	53786	8.601	ng/ul	90
52) Acenaphthene	14.698	153	253522	10.181	ng/ul	96
53) 2,4-Dinitrophenol	14.751	184	25802	7.216	ng/ul	94
55) 4-Nitrophenol	14.845	109	44245	9.381	ng/ul	98
56) Dibenzofuran	15.034	168	354581	10.563	ng/ul	100
57) 2,4-Dinitrotoluene	15.004	165	76639	9.763	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.257	232	63620	10.159	ng/ul	95
59) Diethylphthalate	15.451	149	306662	10.169	ng/ul	100
61) Fluorene	15.681	166	290759	10.479	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.675	204	136554	10.840	ng/ul	97
63) 4-Nitroaniline	15.710	138	54917	8.826	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.763	198	45580	9.381	ng/ul#	98
67) N-Nitrosodiphenylamine	15.892	169	242976	10.158	ng/ul	97
68) 4-Bromophenyl-phenylether	16.575	248	74588	10.038	ng/ul	95
69) Hexachlorobenzene	16.686	284	91386	10.373	ng/ul	99
70) Atrazine	16.845	200	80576	9.964	ng/ul	99
71) Pentachlorophenol	17.033	266	46584	8.712	ng/ul	97
72) Phenanthrene	17.433	178	470296	10.448	ng/ul	99
74) Anthracene	17.522	178	464612	10.366	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.434	216	100809	10.069	ng/ul	99
76) Pentachlorobenzene	14.951	250	114666	10.584	ng/ul	96
77) Carbazole	17.798	167	434021	10.491	ng/ul	98
78) Di-n-butylphthalate	18.357	149	506603	10.000	ng/ul	100
80) Fluoranthene	19.451	202	532103	10.239	ng/ul	97
82) Pyrene	19.816	202	567563	10.350	ng/ul	99
83) Butylbenzylphthalate	20.716	149	216610	9.263	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.510	252	174475	9.535	ng/ul	98
85) Benzo(a)anthracene	21.627	228	541415	10.240	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.492	149	322024	9.484	ng/ul	98
87) Chrysene	21.627	228	541415	10.408	ng/ul	94
89) Di-n-octyl phthalate	22.439	149	557576	9.345	ng/ul	100
90) Benzo(b)fluoranthene	23.298	252	553694	10.249	ng/ul	100
91) Benzo(k)fluoranthene	23.351	252	557852	10.470	ng/ul	97
93) Benzo(a)pyrene	23.945	252	488079	10.233	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.651	276	602142	9.944	ng/ul	97
95) Dibenzo(a,h)anthracene	26.674	278	540874	9.877	ng/ul	97
96) Benzo(g,h,i)perylene	27.451	276	536940	9.837	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101222\
 Data File : BN022061.D
 Acq On : 12 Oct 2022 12:42
 Operator : CG/JU
 Sample : SST01036
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST010246

Quant Time: Oct 13 02:15:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 13 02:12:15 2022
 Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 4.118 ng/uL
 RT: 3.293 min Scan# 51
 Delta R.T. -0.000 min
 Lab File: BN022061.D
 Acq: 12 Oct 2022 12:42




Instrument :
 BNA_N
 ClientSampleId :
 SSTD010246


Tgt Ion:	88	Resp:	13998
Ion	Ratio	Lower	Upper
88	100		
43	31.2	22.6	34.0
58	73.0	59.9	89.9



#5
 Pyridine
 Concen: 9.708 ng/uL
 RT: 3.722 min Scan# 125
 Delta R.T. 0.006 min
 Lab File: BN022061.D
 Acq: 12 Oct 2022 12:42

Tgt Ion:	79	Resp:	92549
Ion	Ratio	Lower	Upper
79	100		
52	75.4	61.7	92.5
51	37.3	27.8	41.6









#6
Benzaldehyde
Concen: 10.085 ng/ul
RT: 7.093 min Scan# 698
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion: 77 Resp: 57998
Ion Ratio Lower Upper
77 100
105 97.7 78.6 118.0
106 92.3 75.4 113.2



#8
Phenol
Concen: 9.218 ng/ul
RT: 7.140 min Scan# 706
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42

Tgt Ion: 94 Resp: 112632
Ion Ratio Lower Upper
94 100
65 27.7 22.1 33.1
66 34.4 29.0 43.6









#10
Bis(2-Chloroethyl)ether
Concen: 9.845 ng/ul
RT: 7.381 min Scan# 747
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion: 93 Resp: 95915
Ion Ratio Lower Upper
93 100
63 77.2 61.1 91.7
95 33.2 26.4 39.6



#12
2-Chlorophenol
Concen: 9.856 ng/ul
RT: 7.516 min Scan# 770
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42

Tgt Ion: 128 Resp: 85876
Ion Ratio Lower Upper
128 100
64 50.8 39.7 59.5
130 32.4 26.7 40.1





#13

2-Methylphenol

Concen: 9.091 ng/ul

RT: 8.410 min Scan# 91

Delta R.T. -0.000 min

Lab File: BN022061.D

Acq: 12 Oct 2022 12:42

Instrument :

BNA_N

ClientSampleId :





SSTD010246

Tgt Ion:108 Resp: 82271

Ion Ratio Lower Upper

108 100

107 88.0 72.4 108.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 9.633 ng/ul

RT: 8.499 min Scan# 937

Delta R.T. -0.006 min

Lab File: BN022061.D

Acq: 12 Oct 2022 12:42




Tgt Ion: 45 Resp: 140339


Ion Ratio Lower Upper

45 100

77 14.8 11.7 17.5

79 11.2 8.7 13.1









#16
Acetophenone
Concen: 9.701 ng/ul
RT: 8.799 min Scan# 988
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:105 Resp: 144547
Ion Ratio Lower Upper
105 100
77 83.2 64.9 97.3
51 28.8 21.8 32.6



#17
N-Nitroso-di-n-propylamine
Concen: 9.825 ng/ul
RT: 8.781 min Scan# 985
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42

Tgt Ion: 70 Resp: 76616
Ion Ratio Lower Upper
70 100
42 56.7 45.6 68.4
101 8.5 7.4 11.2
130 20.6 15.4 23.0









#18
4-Methylphenol
Concen: 9.387 ng/ul
RT: 8.740 min Scan# 91
Delta R.T. -0.006 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:108 Resp: 95354
Ion Ratio Lower Upper
108 100
107 110.4 87.2 130.8



#19
Hexachloroethane
Concen: 9.774 ng/ul
RT: 9.052 min Scan# 1031
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42

Tgt Ion:117 Resp: 34142
Ion Ratio Lower Upper
117 100
201 82.8 69.8 104.6
199 56.3 45.2 67.8









#22
Nitrobenzene
Concen: 9.732 ng/ul
RT: 9.181 min Scan# 1053
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion: 77 Resp: 109289
Ion Ratio Lower Upper
77 100
123 42.4 34.9 52.3
65 12.6 10.2 15.2



#23
Isophorone
Concen: 9.875 ng/ul
RT: 9.710 min Scan# 1143
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42

Tgt Ion: 82 Resp: 216729
Ion Ratio Lower Upper
82 100
95 6.4 5.4 8.0
138 16.7 13.8 20.8









#25
2-Nitrophenol
Concen: 9.609 ng/ul
RT: 9.899 min Scan# 1175
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:139 Resp: 47455
Ion Ratio Lower Upper
139 100
65 55.7 42.7 64.1
109 30.7 23.8 35.6



#26
2,4-Dimethylphenol
Concen: 10.075 ng/ul
RT: 9.957 min Scan# 1185
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42

Tgt Ion:107 Resp: 107336
Ion Ratio Lower Upper
107 100
121 51.1 41.0 61.6
122 85.1 71.5 107.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 9.946 ng/ul
 RT: 10.199 min Scan# 1226
 Delta R.T. -0.000 min
 Lab File: BN022061.D
 Acq: 12 Oct 2022 12:42




Instrument :
 BNA_N
 ClientSampleId :
 SSTD010246


Tgt Ion: 93 Resp: 138208
 Ion Ratio Lower Upper
 93 100
 95 32.3 25.2 37.8
 123 11.7 8.6 12.8



#29
 2,4-Dichlorophenol
 Concen: 9.521 ng/ul
 RT: 10.440 min Scan# 1267
 Delta R.T. 0.006 min
 Lab File: BN022061.D
 Acq: 12 Oct 2022 12:42

Tgt Ion: 162 Resp: 80022
 Ion Ratio Lower Upper
 162 100
 164 68.9 52.9 79.3
 98 38.0 30.5 45.7









#30
Naphthalene
Concen: 10.017 ng/ul
RT: 10.840 min Scan# 1335
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:128 Resp: 313051
Ion Ratio Lower Upper
128 100
129 10.9 8.5 12.7
127 13.0 10.8 16.2



#32
4-Chloroaniline
Concen: 10.166 ng/ul
RT: 10.957 min Scan# 1355
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42

Tgt Ion:127 Resp: 142409
Ion Ratio Lower Upper
127 100
129 31.6 26.2 39.4









#33
Hexachlorobutadiene
Concen: 9.987 ng/ul
RT: 11.122 min Scan# 1111
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:225 Resp: 46450
Ion Ratio Lower Upper
225 100
223 68.9 47.7 71.5
227 71.3 51.4 77.0



#34
Caprolactam
Concen: 8.179 ng/ul
RT: 11.746 min Scan# 1489
Delta R.T. 0.006 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42

Tgt Ion:113 Resp: 26740
Ion Ratio Lower Upper
113 100
55 174.6 138.7 208.1
56 128.3 101.4 152.0









#35
 4-Chloro-3-methylphenol
 Concen: 9.693 ng/ul
 RT: 12.081 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: BN022061.D
 Acq: 12 Oct 2022 12:42




Instrument :
 BNA_N
 ClientSampleId :
 SST010246


Tgt Ion:107 Resp: 99118
 Ion Ratio Lower Upper
 107 100
 144 26.2 20.3 30.5
 142 80.3 64.4 96.6



#36
 2-Methylnaphthalene
 Concen: 10.325 ng/ul
 RT: 12.457 min Scan# 1610
 Delta R.T. -0.000 min
 Lab File: BN022061.D
 Acq: 12 Oct 2022 12:42

Tgt Ion:142 Resp: 219705
 Ion Ratio Lower Upper
 142 100
 141 85.8 69.7 104.5





#37

1-Methylnaphthalene

Concen: 10.403 ng/ul

RT: 12.675 min Scan# 1647

Delta R.T. -0.000 min

Lab File: BN022061.D

Acq: 12 Oct 2022 12:42

Instrument :

BNA_N

ClientSampleId :

SSTD010246





Tgt Ion:142 Resp: 223713

Ion Ratio Lower Upper

142 100

141 89.4 76.7 115.1

116 3.8 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 10.089 ng/ul

RT: 12.822 min Scan# 1672

Delta R.T. -0.000 min

Lab File: BN022061.D

Acq: 12 Oct 2022 12:42

Tgt Ion:216 Resp: 100310




Ion Ratio Lower Upper


216 100

214 80.4 65.4 98.2

179 23.3 17.4 26.0


108 22.0 16.4 24.6








#40
Hexachlorocyclopentadiene
Concen: 9.141 ng/ul
RT: 12.798 min Scan# 1668
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42


Instrument :
BNA_N
ClientSampleId :
SSTD010246





Tgt Ion:237 Resp: 47819
Ion Ratio Lower Upper
237 100
235 57.2 43.4 65.0
272 11.3 10.2 15.2




#41
2,4,6-Trichlorophenol
Concen: 10.016 ng/ul
RT: 13.069 min Scan# 1714
Delta R.T. 0.006 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42



Tgt Ion:196 Resp: 67107
Ion Ratio Lower Upper
196 100
198 95.7 73.4 110.2
200 30.7 24.3 36.5





#42

2,4,5-Trichlorophenol

Concen: 10.195 ng/ul

RT: 13.140 min Scan# 1725

Delta R.T. 0.006 min

Lab File: BN022061.D

Acq: 12 Oct 2022 12:42

Instrument :

BNA_N

ClientSampleId :

SSTD010246





Tgt Ion:196 Resp: 75163

Ion Ratio Lower Upper

196 100

198 92.3 77.0 115.6

200 30.0 24.7 37.1



#43

1,1'-Biphenyl

Concen: 10.300 ng/ul

RT: 13.469 min Scan# 1782

Delta R.T. -0.000 min

Lab File: BN022061.D

Acq: 12 Oct 2022 12:42




Tgt Ion:154 Resp: 292550


Ion Ratio Lower Upper

154 100

153 41.6 30.9 46.3

76 14.7 11.6 17.4





#44

2-Chloronaphthalene

Concen: 10.292 ng/ul

RT: 13.510 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN022061.D


Acq: 12 Oct 2022 12:42

Instrument :

BNA_N

ClientSampleId :

SSTD010246






Tgt Ion:162 Resp: 227881

Ion Ratio Lower Upper

162 100

164 33.6 23.7 35.5

127 39.9 30.8 46.2



#45

2-Nitroaniline

Concen: 9.234 ng/ul

RT: 13.722 min Scan# 1825

Delta R.T. -0.000 min

Lab File: BN022061.D

Acq: 12 Oct 2022 12:42




Tgt Ion: 65 Resp: 62979


Ion Ratio Lower Upper

65 100

92 67.4 56.2 84.2

138 97.9 85.0 127.6









#47
Dimethylphthalate
Concen: 10.329 ng/ul
RT: 14.092 min Scan# 1888
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:163 Resp: 300445
Ion Ratio Lower Upper
163 100
194 3.9 3.1 4.7
164 10.4 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 9.773 ng/ul
RT: 14.216 min Scan# 1909
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42

Tgt Ion:165 Resp: 50738
Ion Ratio Lower Upper
165 100
89 63.6 52.1 78.1
121 23.8 19.0 28.4









#50
Acenaphthylene
Concen: 10.464 ng/ul
RT: 14.357 min Scan# 1933
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:152 Resp: 358120
Ion Ratio Lower Upper
152 100
151 19.7 15.5 23.3
153 12.8 9.8 14.8



#51
3-Nitroaniline
Concen: 8.601 ng/ul
RT: 14.545 min Scan# 1965
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42

Tgt Ion:138 Resp: 53786
Ion Ratio Lower Upper
138 100
108 12.4 9.0 13.6
92 131.8 96.4 144.6









#52
Acenaphthene
Concen: 10.181 ng/ul
RT: 14.698 min Scan# 1991
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:153 Resp: 253522
Ion Ratio Lower Upper
153 100
152 44.7 38.0 57.0
154 85.9 71.1 106.7



#53
2,4-Dinitrophenol
Concen: 7.216 ng/ul
RT: 14.751 min Scan# 2000
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42

Tgt Ion:184 Resp: 25802
Ion Ratio Lower Upper
184 100
63 77.9 56.8 85.2
154 64.4 49.0 73.6









#55
4-Nitrophenol
Concen: 9.381 ng/ul
RT: 14.845 min Scan# 2016
Delta R.T. -0.006 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:109 Resp: 44245
Ion Ratio Lower Upper
109 100
139 134.5 108.3 162.5
65 138.8 107.9 161.9



#56
Dibenzofuran
Concen: 10.563 ng/ul
RT: 15.034 min Scan# 2048
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42

Tgt Ion:168 Resp: 354581
Ion Ratio Lower Upper
168 100
139 38.1 30.6 45.8









#57
2,4-Dinitrotoluene
Concen: 9.763 ng/ul
RT: 15.004 min Scan# 2043
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:165 Resp: 76639
Ion Ratio Lower Upper
165 100
63 45.5 36.2 54.4
182 2.5 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 10.159 ng/ul
RT: 15.257 min Scan# 2086
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42

Tgt Ion:232 Resp: 63620
Ion Ratio Lower Upper
232 100
131 54.0 44.4 66.6
130 2.8 2.2 3.4
166 30.5 28.8 43.2









#59
Diethylphthalate
Concen: 10.169 ng/ul
RT: 15.451 min Scan# 2119
Delta R.T. -0.006 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:149 Resp: 306662
Ion Ratio Lower Upper
149 100
177 21.8 17.5 26.3
150 12.2 9.8 14.8



#61
Fluorene
Concen: 10.479 ng/ul
RT: 15.681 min Scan# 2158
Delta R.T. -0.006 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42

Tgt Ion:166 Resp: 290759
Ion Ratio Lower Upper
166 100
165 93.3 72.8 109.2
167 13.9 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 10.840 ng/ul
RT: 15.675 min Scan# 2157
Delta R.T. -0.006 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:204 Resp: 136554
Ion Ratio Lower Upper
204 100
206 34.7 26.2 39.2
141 66.6 52.0 78.0



#63
4-Nitroaniline
Concen: 8.826 ng/ul
RT: 15.710 min Scan# 2163
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42

Tgt Ion:138 Resp: 54917
Ion Ratio Lower Upper
138 100
92 55.4 41.8 62.8
108 77.3 60.4 90.6









#66
4,6-Dinitro-2-methylphenol
Concen: 9.381 ng/ul
RT: 15.763 min Scan# 2172
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:198 Resp: 45580
Ion Ratio Lower Upper
198 100
182 7.0 3.8 5.8#
77 31.3 24.4 36.6



#67
N-Nitrosodiphenylamine
Concen: 10.158 ng/ul
RT: 15.892 min Scan# 2194
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42

Tgt Ion:169 Resp: 242976
Ion Ratio Lower Upper
169 100
168 67.1 50.9 76.3
167 35.5 28.2 42.4









#68
4-Bromophenyl-phenylether
Concen: 10.038 ng/ul
RT: 16.575 min Scan# 2310
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:248 Resp: 74588
Ion Ratio Lower Upper
248 100
250 104.2 86.3 129.5
141 87.0 74.2 111.4



#69
Hexachlorobenzene
Concen: 10.373 ng/ul
RT: 16.686 min Scan# 2329
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42

Tgt Ion:284 Resp: 91386
Ion Ratio Lower Upper
284 100
142 44.8 35.7 53.5
249 32.1 27.0 40.6





#70

Atrazine

Concen: 9.964 ng/ul

RT: 16.845 min Scan# 21

Delta R.T. -0.006 min

Lab File: BN022061.D

Acq: 12 Oct 2022 12:42

Instrument :

BNA_N

ClientSampleId :

SSTD010246





Tgt Ion:200 Resp: 80576

Ion Ratio Lower Upper

200 100

173 27.1 22.8 34.2

215 47.7 38.1 57.1



#71

Pentachlorophenol

Concen: 8.712 ng/ul

RT: 17.033 min Scan# 2388

Delta R.T. -0.000 min

Lab File: BN022061.D

Acq: 12 Oct 2022 12:42




Tgt Ion:266 Resp: 46584


Ion Ratio Lower Upper

266 100

264 63.1 52.7 79.1

268 70.7 54.6 82.0









#72
Phenanthrene
Concen: 10.448 ng/ul
RT: 17.433 min Scan# 2456
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:178 Resp: 470296
Ion Ratio Lower Upper
178 100
179 15.3 12.3 18.5
176 19.7 16.5 24.7



#74
Anthracene
Concen: 10.366 ng/ul
RT: 17.522 min Scan# 2471
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42

Tgt Ion:178 Resp: 464612
Ion Ratio Lower Upper
178 100
179 15.0 13.0 19.6
176 18.8 15.8 23.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 10.069 ng/uL
RT: 13.434 min Scan# 1776
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:216 Resp: 100809
Ion Ratio Lower Upper
216 100
214 76.3 61.2 91.8
218 48.4 39.4 59.0



#76
Pentachlorobenzene
Concen: 10.584 ng/uL
RT: 14.951 min Scan# 2034
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42

Tgt Ion:250 Resp: 114666
Ion Ratio Lower Upper
250 100
248 63.2 46.8 70.2
252 66.4 53.9 80.9





#77

Carbazole

Concen: 10.491 ng/ul

RT: 17.798 min Scan# 2518

Delta R.T. -0.000 min

Lab File: BN022061.D

Acq: 12 Oct 2022 12:42

Instrument :

BNA_N

ClientSampleId :

SSTD010246





Tgt Ion:167 Resp: 434021

Ion Ratio Lower Upper

167 100

166 20.9 17.3 25.9

139 12.8 10.7 16.1



#78

Di-n-butylphthalate

Concen: 10.000 ng/ul

RT: 18.357 min Scan# 2613

Delta R.T. -0.000 min

Lab File: BN022061.D

Acq: 12 Oct 2022 12:42




Tgt Ion:149 Resp: 506603


Ion Ratio Lower Upper

149 100

150 9.0 7.1 10.7

104 5.1 4.2 6.2









#80
Fluoranthene
Concen: 10.239 ng/ul
RT: 19.451 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:202 Resp: 532103
Ion Ratio Lower Upper
202 100
101 13.8 10.3 15.5
100 10.9 7.8 11.8



#82
Pyrene
Concen: 10.350 ng/ul
RT: 19.816 min Scan# 2861
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42

Tgt Ion:202 Resp: 567563
Ion Ratio Lower Upper
202 100
101 16.4 13.4 20.0
100 12.9 10.9 16.3









#83
 Butylbenzylphthalate
 Concen: 9.263 ng/ul
 RT: 20.716 min Scan# 3014
 Delta R.T. -0.000 min
 Lab File: BN022061.D
 Acq: 12 Oct 2022 12:42




Instrument :
 BNA_N
 ClientSampleId :
 SSTD010246


Tgt Ion:149 Resp: 216610
 Ion Ratio Lower Upper
 149 100
 91 78.0 61.7 92.5
 206 18.0 14.7 22.1



#84
 3,3'-Dichlorobenzidine
 Concen: 9.535 ng/ul
 RT: 21.510 min Scan# 3149
 Delta R.T. -0.000 min
 Lab File: BN022061.D
 Acq: 12 Oct 2022 12:42

Tgt Ion:252 Resp: 174475
 Ion Ratio Lower Upper
 252 100
 254 68.1 52.9 79.3
 126 13.5 11.2 16.8









#85
Benzo(a)anthracene
Concen: 10.240 ng/ul
RT: 21.627 min Scan# 3169
Delta R.T. 0.053 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:228 Resp: 541415
Ion Ratio Lower Upper
228 100
229 17.6 13.5 20.3
226 27.2 21.0 31.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 9.484 ng/ul
RT: 21.492 min Scan# 3146
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42

Tgt Ion:149 Resp: 322024
Ion Ratio Lower Upper
149 100
167 25.2 21.3 31.9
279 3.2 2.9 4.3









#87
Chrysene
Concen: 10.408 ng/ul
RT: 21.627 min Scan# 3169
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:228 Resp: 541415
Ion Ratio Lower Upper
228 100
226 27.2 23.8 35.8
229 17.6 16.5 24.7



#89
Di-n-octyl phthalate
Concen: 9.345 ng/ul
RT: 22.439 min Scan# 3307
Delta R.T. -0.000 min
Lab File: BN022061.D
Acq: 12 Oct 2022 12:42

Tgt Ion:149 Resp: 557576





#90

Benzo(b)fluoranthene

Concen: 10.249 ng/ul

RT: 23.298 min Scan# 3453

Delta R.T. -0.000 min

Lab File: BN022061.D

Acq: 12 Oct 2022 12:42

Instrument :

BNA_N

ClientSampleId :

SSTD010246





Tgt Ion:252 Resp: 553694

Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4

125 13.0 10.7 16.1



#91

Benzo(k)fluoranthene

Concen: 10.470 ng/ul

RT: 23.351 min Scan# 3462

Delta R.T. -0.000 min

Lab File: BN022061.D

Acq: 12 Oct 2022 12:42




Tgt Ion:252 Resp: 557852


Ion Ratio Lower Upper

252 100

253 24.3 17.7 26.5

125 13.4 10.6 15.8





#93

Benzo(a)pyrene

Concen: 10.233 ng/ul

RT: 23.945 min Scan# 3563

Delta R.T. -0.006 min

Lab File: BN022061.D

Acq: 12 Oct 2022 12:42

Instrument :

BNA_N

ClientSampleId :

SSTD010246





Tgt Ion:252 Resp: 488079

Ion Ratio Lower Upper

252 100

253 20.0 17.2 25.8

125 13.9 10.8 16.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 9.944 ng/ul

RT: 26.651 min Scan# 4023

Delta R.T. -0.012 min

Lab File: BN022061.D

Acq: 12 Oct 2022 12:42




Tgt Ion:276 Resp: 602142


Ion Ratio Lower Upper

276 100

138 27.8 24.6 36.8

277 22.7 18.4 27.6





#95

Dibenzo(a,h)anthracene

Concen: 9.877 ng/ul

RT: 26.674 min Scan# 4027

Delta R.T. -0.000 min

Lab File: BN022061.D

Acq: 12 Oct 2022 12:42

Instrument :

BNA_N

ClientSampleId :

SSTD010246





Tgt Ion:278 Resp: 540874

Ion Ratio Lower Upper

278 100

139 19.0 15.7 23.5

279 23.1 16.5 24.7



#96

Benzo(g,h,i)perylene

Concen: 9.837 ng/ul

RT: 27.451 min Scan# 4159

Delta R.T. -0.006 min

Lab File: BN022061.D

Acq: 12 Oct 2022 12:42

Tgt Ion:276 Resp: 536940

Ion Ratio Lower Upper

276 100

138 26.4 21.2 31.8

277 24.0 17.9 26.9

