

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101222\
 Data File : BN022062.D
 Acq On : 12 Oct 2022 13:20
 Operator : CG/JU
 Sample : SSTD02037
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020247

Quant Time: Oct 13 02:15:32 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 13 02:12:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.958	152	120734	20.000	ng/u1	0.00
20) Naphthalene-d8	10.793	136	610308	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.634	164	413918	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.392	188	892803	20.000	ng/u1	0.00
79) Chrysene-d12	21.592	240	871744	20.000	ng/u1	0.00
88) Perylene-d12	24.057	264	895760	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.258	96	26569	8.329	ng/uL	0.00
4) Pyridine-d5	3.693	84	205566	21.315	ng/u1	0.00
7) Phenol-d5	7.116	99	240051	20.565	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.287	67	156578	20.952	ng/u1	0.00
11) 2-Chlorophenol-d4	7.481	132	174163	21.630	ng/u1	0.00
15) 4-Methylphenol-d8	8.681	113	195361	21.185	ng/u1	0.00
21) Nitrobenzene-d5	9.140	128	93861	21.667	ng/u1	0.00
24) 2-Nitrophenol-d4	9.869	143	96743	21.906	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.410	165	184591	22.101	ng/u1	0.00
31) 4-Chloroaniline-d4	10.934	131	300476	21.304	ng/u1	0.00
46) Dimethylphthalate-d6	14.045	166	608858	20.931	ng/u1	0.00
49) Acenaphthylene-d8	14.328	160	687537	21.724	ng/u1	0.00
54) 4-Nitrophenol-d4	14.834	143	126817	20.704	ng/u1	0.00
60) Fluorene-d10	15.628	176	507195	20.596	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.751	200	98533	20.148	ng/u1	0.00
73) Anthracene-d10	17.486	188	800348	20.756	ng/u1	0.00
81) Pyrene-d10	19.786	212	923560	21.790	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.898	264	950142	21.936	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.293	88	28347	8.202	ng/uL	100
5) Pyridine	3.717	79	202507	20.893	ng/u1	100
6) Benzaldehyde	7.093	77	143957	24.620	ng/u1	100
8) Phenol	7.140	94	258121	20.779	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.381	93	211273	21.329	ng/u1	100
12) 2-Chlorophenol	7.516	128	192065	21.681	ng/u1	100
13) 2-Methylphenol	8.410	108	195132	21.208	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.505	45	306223	20.675	ng/u1	100
16) Acetophenone	8.799	105	331089	21.855	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.781	70	179216	22.606	ng/u1	100
18) 4-Methylphenol	8.746	108	221170	21.416	ng/u1	100
19) Hexachloroethane	9.052	117	75070	21.138	ng/u1	100
22) Nitrobenzene	9.181	77	251230	21.107	ng/u1	100
23) Isophorone	9.710	82	504256	21.677	ng/u1	100
25) 2-Nitrophenol	9.899	139	116806	22.315	ng/u1	100
26) 2,4-Dimethylphenol	9.957	107	245509	21.743	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.199	93	310154	21.058	ng/u1	100
29) 2,4-Dichlorophenol	10.434	162	193643	21.737	ng/u1	100
30) Naphthalene	10.840	128	684242	20.657	ng/u1	100
32) 4-Chloroaniline	10.957	127	309058	20.816	ng/u1	100
33) Hexachlorobutadiene	11.122	225	102723	20.838	ng/u1	100
34) Caprolactam	11.740	113	70501	20.347	ng/u1	100
35) 4-Chloro-3-methylphenol	12.081	107	236390	21.811	ng/u1	100
36) 2-Methylnaphthalene	12.457	142	478739	21.227	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101222\
 Data File : BN022062.D
 Acq On : 12 Oct 2022 13:20
 Operator : CG/JU
 Sample : SSTD02037
 Misc :
 ALS Vial : 5 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SSTD020247


Quant Time: Oct 13 02:15:32 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 13 02:12:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.675	142	473828	20.788	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.822	216	221132	20.931	ng/ul	100
40) Hexachlorocyclopentadiene	12.798	237	118291	21.279	ng/ul	100
41) 2,4,6-Trichlorophenol	13.063	196	153774	21.598	ng/ul	100
42) 2,4,5-Trichlorophenol	13.134	196	170392	21.749	ng/ul	100
43) 1,1'-Biphenyl	13.469	154	628670	20.831	ng/ul	100
44) 2-Chloronaphthalene	13.510	162	489464	20.803	ng/ul	100
45) 2-Nitroaniline	13.722	65	155957	21.519	ng/ul	100
47) Dimethylphthalate	14.093	163	641671	20.760	ng/ul	100
48) 2,6-Dinitrotoluene	14.216	165	123708	22.425	ng/ul	100
50) Acenaphthylene	14.357	152	780906	21.473	ng/ul	100
51) 3-Nitroaniline	14.545	138	143040	21.526	ng/ul	100
52) Acenaphthene	14.698	153	545078	20.599	ng/ul	100
53) 2,4-Dinitrophenol	14.751	184	73515	19.348	ng/ul	100
55) 4-Nitrophenol	14.851	109	103487	20.649	ng/ul	100
56) Dibenzofuran	15.034	168	749975	21.026	ng/ul	100
57) 2,4-Dinitrotoluene	15.004	165	186616	22.373	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.257	232	141641	21.284	ng/ul	100
59) Diethylphthalate	15.457	149	661285	20.636	ng/ul	100
61) Fluorene	15.687	166	616598	20.912	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.681	204	288213	21.530	ng/ul	100
63) 4-Nitroaniline	15.710	138	144238	21.816	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.763	198	110667	21.088	ng/ul	100
67) N-Nitrosodiphenylamine	15.892	169	538141	20.829	ng/ul	100
68) 4-Bromophenyl-phenylether	16.575	248	160176	19.958	ng/ul	100
69) Hexachlorobenzene	16.686	284	195177	20.512	ng/ul	100
70) Atrazine	16.851	200	181673	20.800	ng/ul	100
71) Pentachlorophenol	17.034	266	120601	20.882	ng/ul	100
72) Phenanthrene	17.434	178	981431	20.187	ng/ul	100
74) Anthracene	17.522	178	985003	20.346	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.434	216	226120	20.911	ng/ul	100
76) Pentachlorobenzene	14.951	250	246507	21.066	ng/ul	100
77) Carbazole	17.798	167	948399	21.224	ng/ul	100
78) Di-n-butylphthalate	18.357	149	1170731	21.396	ng/ul	100
80) Fluoranthene	19.457	202	1135756	21.099	ng/ul	100
82) Pyrene	19.816	202	1216358	21.414	ng/ul	100
83) Butylbenzylphthalate	20.716	149	526998	21.757	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.510	252	403088	21.266	ng/ul	100
85) Benzo(a)anthracene	21.574	228	1182274	21.588	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.492	149	784207	22.297	ng/ul	100
87) Chrysene	21.627	228	1107218	20.548	ng/ul	100
89) Di-n-octyl phthalate	22.439	149	1391204	22.271	ng/ul	100
90) Benzo(b)fluoranthene	23.298	252	1210254	21.399	ng/ul	100
91) Benzo(k)fluoranthene	23.351	252	1168453	20.947	ng/ul	100
93) Benzo(a)pyrene	23.951	252	1092073	21.869	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.662	276	1346071	21.233	ng/ul	100
95) Dibenzo(a,h)anthracene	26.674	278	1204127	21.003	ng/ul	100
96) Benzo(g,h,i)perylene	27.457	276	1201166	21.020	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 13 02:15:32 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101222.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 13 02:12:15 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.202 ng/uL
RT: 3.293 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20




Instrument :
BNA_N
ClientSampleId :
SSTD020247


Tgt Ion: 88 Resp: 28347
Ion Ratio Lower Upper
88 100
43 28.3 22.6 34.0
58 74.9 59.9 89.9



#5
Pyridine
Concen: 20.893 ng/uL
RT: 3.717 min Scan# 124
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20

Tgt Ion: 79 Resp: 202507
Ion Ratio Lower Upper
79 100
52 77.1 61.7 92.5
51 34.7 27.8 41.6









#6
Benzaldehyde
Concen: 24.620 ng/ul
RT: 7.093 min Scan# 698
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20




Instrument :
BNA_N
ClientSampleId :
SSTD020247


Tgt Ion: 77 Resp: 143957
Ion Ratio Lower Upper
77 100
105 98.3 78.6 118.0
106 94.3 75.4 113.2



#8
Phenol
Concen: 20.779 ng/ul
RT: 7.140 min Scan# 706
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20

Tgt Ion: 94 Resp: 258121
Ion Ratio Lower Upper
94 100
65 27.6 22.1 33.1
66 36.3 29.0 43.6









#10
Bis(2-Chloroethyl)ether
Concen: 21.329 ng/ul
RT: 7.381 min Scan# 747
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20




Instrument :
BNA_N
ClientSampleId :
SSTD020247


Tgt Ion: 93 Resp: 211273
Ion Ratio Lower Upper
93 100
63 76.4 61.1 91.7
95 33.0 26.4 39.6



#12
2-Chlorophenol
Concen: 21.681 ng/ul
RT: 7.516 min Scan# 770
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20

Tgt Ion: 128 Resp: 192065
Ion Ratio Lower Upper
128 100
64 49.6 39.7 59.5
130 33.4 26.7 40.1









#13
2-Methylphenol
Concen: 21.208 ng/ul
RT: 8.410 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20




Instrument :
BNA_N
ClientSampleId :
SSTD020247


Tgt Ion:108 Resp: 195132
Ion Ratio Lower Upper
108 100
107 90.5 72.4 108.6



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.675 ng/ul
RT: 8.505 min Scan# 938
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20

Tgt Ion: 45 Resp: 306223
Ion Ratio Lower Upper
45 100
77 14.6 11.7 17.5
79 10.9 8.7 13.1









#16
Acetophenone
Concen: 21.855 ng/ul
RT: 8.799 min Scan# 988
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20




Instrument :
BNA_N
ClientSampleId :
SSTD020247


Tgt Ion:105 Resp: 331089
Ion Ratio Lower Upper
105 100
77 81.1 64.9 97.3
51 27.2 21.8 32.6



#17
N-Nitroso-di-n-propylamine
Concen: 22.606 ng/ul
RT: 8.781 min Scan# 985
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20

Tgt Ion: 70 Resp: 179216
Ion Ratio Lower Upper
70 100
42 57.0 45.6 68.4
101 9.3 7.4 11.2
130 19.2 15.4 23.0









#18
4-Methylphenol
Concen: 21.416 ng/ul
RT: 8.746 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20




Instrument :
BNA_N
ClientSampleId :
SSTD020247


Tgt Ion:108 Resp: 221170
Ion Ratio Lower Upper
108 100
107 109.0 87.2 130.8



#19
Hexachloroethane
Concen: 21.138 ng/ul
RT: 9.052 min Scan# 1031
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20

Tgt Ion:117 Resp: 75070
Ion Ratio Lower Upper
117 100
201 87.2 69.8 104.6
199 56.5 45.2 67.8









#22
Nitrobenzene
Concen: 21.107 ng/ul
RT: 9.181 min Scan# 1053
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20




Instrument :
BNA_N
ClientSampleId :
SSTD020247


Tgt Ion: 77 Resp: 251230
Ion Ratio Lower Upper
77 100
123 43.6 34.9 52.3
65 12.7 10.2 15.2



#23
Isophorone
Concen: 21.677 ng/ul
RT: 9.710 min Scan# 1143
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20

Tgt Ion: 82 Resp: 504256
Ion Ratio Lower Upper
82 100
95 6.7 5.4 8.0
138 17.3 13.8 20.8









#25
2-Nitrophenol
Concen: 22.315 ng/ul
RT: 9.899 min Scan# 1175
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20




Instrument :
BNA_N
ClientSampleId :
SSTD020247


Tgt Ion:139 Resp: 116806
Ion Ratio Lower Upper
139 100
65 53.4 42.7 64.1
109 29.7 23.8 35.6



#26
2,4-Dimethylphenol
Concen: 21.743 ng/ul
RT: 9.957 min Scan# 1185
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20

Tgt Ion:107 Resp: 245509
Ion Ratio Lower Upper
107 100
121 51.3 41.0 61.6
122 89.4 71.5 107.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.058 ng/ul
 RT: 10.199 min Scan# 1226
 Delta R.T. 0.000 min
 Lab File: BN022062.D
 Acq: 12 Oct 2022 13:20




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020247


Tgt Ion: 93 Resp: 310154
 Ion Ratio Lower Upper
 93 100
 95 31.5 25.2 37.8
 123 10.7 8.6 12.8



#29
 2,4-Dichlorophenol
 Concen: 21.737 ng/ul
 RT: 10.434 min Scan# 1266
 Delta R.T. 0.000 min
 Lab File: BN022062.D
 Acq: 12 Oct 2022 13:20

Tgt Ion: 162 Resp: 193643
 Ion Ratio Lower Upper
 162 100
 164 66.1 52.9 79.3
 98 38.1 30.5 45.7









#30
Naphthalene
Concen: 20.657 ng/ul
RT: 10.840 min Scan# 1335
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20




Instrument :
BNA_N
ClientSampleId :
SSTD020247


Tgt Ion:128 Resp: 684242
Ion Ratio Lower Upper
128 100
129 10.6 8.5 12.7
127 13.5 10.8 16.2



#32
4-Chloroaniline
Concen: 20.816 ng/ul
RT: 10.957 min Scan# 1355
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20

Tgt Ion:127 Resp: 309058
Ion Ratio Lower Upper
127 100
129 32.8 26.2 39.4





#33

Hexachlorobutadiene

Concen: 20.838 ng/ul

RT: 11.122 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN022062.D

Acq: 12 Oct 2022 13:20

Instrument :

BNA_N

ClientSampleId :

SSTD020247





Tgt Ion:225 Resp: 102723

Ion Ratio Lower Upper

225 100

223 59.6 47.7 71.5

227 64.2 51.4 77.0



#34

Caprolactam

Concen: 20.347 ng/ul

RT: 11.740 min Scan# 1488

Delta R.T. 0.000 min

Lab File: BN022062.D

Acq: 12 Oct 2022 13:20




Tgt Ion:113 Resp: 70501


Ion Ratio Lower Upper

113 100

55 173.4 138.7 208.1

56 126.7 101.4 152.0









#35
 4-Chloro-3-methylphenol
 Concen: 21.811 ng/ul
 RT: 12.081 min Scan# 1546
 Delta R.T. 0.000 min
 Lab File: BN022062.D
 Acq: 12 Oct 2022 13:20




Instrument :
 BNA_N
 ClientSampleId :
 SST020247


Tgt Ion:107 Resp: 236390
 Ion Ratio Lower Upper
 107 100
 144 25.4 20.3 30.5
 142 80.5 64.4 96.6



#36
 2-Methylnaphthalene
 Concen: 21.227 ng/ul
 RT: 12.457 min Scan# 1610
 Delta R.T. 0.000 min
 Lab File: BN022062.D
 Acq: 12 Oct 2022 13:20

Tgt Ion:142 Resp: 478739
 Ion Ratio Lower Upper
 142 100
 141 87.1 69.7 104.5





#37

1-Methylnaphthalene

Concen: 20.788 ng/ul

RT: 12.675 min Scan# 1647

Delta R.T. 0.000 min

Lab File: BN022062.D

Acq: 12 Oct 2022 13:20

Instrument :

BNA_N

ClientSampleId :

SSTD020247





Tgt Ion:142 Resp: 473828

Ion Ratio Lower Upper

142 100

141 95.9 76.7 115.1

116 3.9 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.931 ng/ul

RT: 12.822 min Scan# 1672

Delta R.T. 0.000 min

Lab File: BN022062.D

Acq: 12 Oct 2022 13:20

Tgt Ion:216 Resp: 221132




Ion Ratio Lower Upper


216 100

214 81.8 65.4 98.2

179 21.7 17.4 26.0

108 20.5 16.4 24.6





#40

Hexachlorocyclopentadiene

Concen: 21.279 ng/ul

RT: 12.798 min Scan# 1668

Delta R.T. 0.000 min

Lab File: BN022062.D


Acq: 12 Oct 2022 13:20

Instrument :

BNA_N

ClientSampleId :

SSTD020247






Tgt Ion:237 Resp: 118291

Ion Ratio Lower Upper

237 100

235 54.2 43.4 65.0

272 12.7 10.2 15.2



#41

2,4,6-Trichlorophenol

Concen: 21.598 ng/ul

RT: 13.063 min Scan# 1713

Delta R.T. 0.000 min

Lab File: BN022062.D

Acq: 12 Oct 2022 13:20




Tgt Ion:196 Resp: 153774


Ion Ratio Lower Upper

196 100

198 91.8 73.4 110.2

200 30.4 24.3 36.5





#42

2,4,5-Trichlorophenol

Concen: 21.749 ng/ul

RT: 13.134 min Scan# 1725

Delta R.T. 0.000 min

Lab File: BN022062.D

Acq: 12 Oct 2022 13:20

Instrument :

BNA_N

ClientSampleId :

SSTD020247





Tgt Ion:196 Resp: 170392

Ion Ratio Lower Upper

196 100

198 96.3 77.0 115.6

200 30.9 24.7 37.1



#43

1,1'-Biphenyl

Concen: 20.831 ng/ul

RT: 13.469 min Scan# 1782

Delta R.T. 0.000 min

Lab File: BN022062.D

Acq: 12 Oct 2022 13:20




Tgt Ion:154 Resp: 628670


Ion Ratio Lower Upper

154 100

153 38.6 30.9 46.3

76 14.5 11.6 17.4









#44
2-Chloronaphthalene
Concen: 20.803 ng/ul
RT: 13.510 min Scan# 1789
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20




Instrument :
BNA_N
ClientSampleId :
SSTD020247


Tgt Ion:162 Resp: 489464
Ion Ratio Lower Upper
162 100
164 29.6 23.7 35.5
127 38.5 30.8 46.2



#45
2-Nitroaniline
Concen: 21.519 ng/ul
RT: 13.722 min Scan# 1825
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20

Tgt Ion: 65 Resp: 155957
Ion Ratio Lower Upper
65 100
92 70.2 56.2 84.2
138 106.3 85.0 127.6









#47
Dimethylphthalate
Concen: 20.760 ng/ul
RT: 14.093 min Scan# 1888
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20




Instrument :
BNA_N
ClientSampleId :
SSTD020247


Tgt Ion:163 Resp: 641671
Ion Ratio Lower Upper
163 100
194 3.9 3.1 4.7
164 10.2 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 22.425 ng/ul
RT: 14.216 min Scan# 1909
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20

Tgt Ion:165 Resp: 123708
Ion Ratio Lower Upper
165 100
89 65.1 52.1 78.1
121 23.7 19.0 28.4









#50
Acenaphthylene
Concen: 21.473 ng/ul
RT: 14.357 min Scan# 1933
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20




Instrument :
BNA_N
ClientSampleId :
SSTD020247


Tgt Ion:152 Resp: 780906
Ion Ratio Lower Upper
152 100
151 19.4 15.5 23.3
153 12.3 9.8 14.8



#51
3-Nitroaniline
Concen: 21.526 ng/ul
RT: 14.545 min Scan# 1965
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20

Tgt Ion:138 Resp: 143040
Ion Ratio Lower Upper
138 100
108 11.3 9.0 13.6
92 120.5 96.4 144.6









#52
Acenaphthene
Concen: 20.599 ng/ul
RT: 14.698 min Scan# 1991
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20




Instrument :
BNA_N
ClientSampleId :
SSTD020247


Tgt Ion:153 Resp: 545078
Ion Ratio Lower Upper
153 100
152 47.5 38.0 57.0
154 88.9 71.1 106.7



#53
2,4-Dinitrophenol
Concen: 19.348 ng/ul
RT: 14.751 min Scan# 2000
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20

Tgt Ion:184 Resp: 73515
Ion Ratio Lower Upper
184 100
63 71.0 56.8 85.2
154 61.3 49.0 73.6





#55

4-Nitrophenol

Concen: 20.649 ng/ul

RT: 14.851 min Scan# 2017

Delta R.T. 0.000 min

Lab File: BN022062.D

Acq: 12 Oct 2022 13:20

Instrument :

BNA_N

ClientSampleId :

SSTD020247





Tgt Ion:109 Resp: 103487

Ion Ratio Lower Upper

109 100

139 135.4 108.3 162.5

65 134.9 107.9 161.9



#56

Dibenzofuran

Concen: 21.026 ng/ul

RT: 15.034 min Scan# 2048

Delta R.T. 0.000 min

Lab File: BN022062.D




Acq: 12 Oct 2022 13:20


Tgt Ion:168 Resp: 749975

Ion Ratio Lower Upper

168 100

139 38.2 30.6 45.8









#57
2,4-Dinitrotoluene
Concen: 22.373 ng/ul
RT: 15.004 min Scan# 2043
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20




Instrument :
BNA_N
ClientSampleId :
SSTD020247


Tgt Ion:165 Resp: 186616
Ion Ratio Lower Upper
165 100
63 45.3 36.2 54.4
182 2.7 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.284 ng/ul
RT: 15.257 min Scan# 2086
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20

Tgt Ion:232 Resp: 141641
Ion Ratio Lower Upper
232 100
131 55.5 44.4 66.6
130 2.8 2.2 3.4
166 36.0 28.8 43.2





#59

Diethylphthalate

Concen: 20.636 ng/ul

RT: 15.457 min Scan# 2120

Delta R.T. 0.000 min

Lab File: BN022062.D

Acq: 12 Oct 2022 13:20

Instrument :

BNA_N

ClientSampleId :

SSTD020247





Tgt Ion:149 Resp: 661285

Ion Ratio Lower Upper

149 100

177 21.9 17.5 26.3

150 12.3 9.8 14.8



#61

Fluorene

Concen: 20.912 ng/ul

RT: 15.687 min Scan# 2159

Delta R.T. 0.000 min

Lab File: BN022062.D

Acq: 12 Oct 2022 13:20




Tgt Ion:166 Resp: 616598


Ion Ratio Lower Upper

166 100

165 91.0 72.8 109.2

167 13.3 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 21.530 ng/ul
RT: 15.681 min Scan# 2158
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20




Instrument :
BNA_N
ClientSampleId :
SSTD020247


Tgt Ion:204 Resp: 288213
Ion Ratio Lower Upper
204 100
206 32.7 26.2 39.2
141 65.0 52.0 78.0



#63
4-Nitroaniline
Concen: 21.816 ng/ul
RT: 15.710 min Scan# 2163
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20

Tgt Ion:138 Resp: 144238
Ion Ratio Lower Upper
138 100
92 52.3 41.8 62.8
108 75.5 60.4 90.6









#66
 4,6-Dinitro-2-methylphenol
 Concen: 21.088 ng/ul
 RT: 15.763 min Scan# 2172
 Delta R.T. 0.000 min
 Lab File: BN022062.D
 Acq: 12 Oct 2022 13:20




Instrument :
 BNA_N
 ClientSampleId :
 SST020247


Tgt Ion:198 Resp: 110667
 Ion Ratio Lower Upper
 198 100
 182 4.8 3.8 5.8
 77 30.5 24.4 36.6



#67
 N-Nitrosodiphenylamine
 Concen: 20.829 ng/ul
 RT: 15.892 min Scan# 2194
 Delta R.T. 0.000 min
 Lab File: BN022062.D
 Acq: 12 Oct 2022 13:20

Tgt Ion:169 Resp: 538141
 Ion Ratio Lower Upper
 169 100
 168 63.6 50.9 76.3
 167 35.3 28.2 42.4









#68
4-Bromophenyl-phenylether
Concen: 19.958 ng/ul
RT: 16.575 min Scan# 2310
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20




Instrument :
BNA_N
ClientSampleId :
SSTD020247


Tgt Ion:248 Resp: 160176
Ion Ratio Lower Upper
248 100
250 107.9 86.3 129.5
141 92.8 74.2 111.4



#69
Hexachlorobenzene
Concen: 20.512 ng/ul
RT: 16.686 min Scan# 2329
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20

Tgt Ion:284 Resp: 195177
Ion Ratio Lower Upper
284 100
142 44.6 35.7 53.5
249 33.8 27.0 40.6





#70

Atrazine

Concen: 20.800 ng/ul

RT: 16.851 min Scan# 2357

Delta R.T. 0.000 min

Lab File: BN022062.D

Acq: 12 Oct 2022 13:20

Instrument :

BNA_N

ClientSampleId :

SSTD020247





Tgt Ion:200 Resp: 181673

Ion Ratio Lower Upper

200 100

173 28.5 22.8 34.2

215 47.6 38.1 57.1



#71

Pentachlorophenol

Concen: 20.882 ng/ul

RT: 17.034 min Scan# 2388

Delta R.T. 0.000 min

Lab File: BN022062.D

Acq: 12 Oct 2022 13:20




Tgt Ion:266 Resp: 120601


Ion Ratio Lower Upper

266 100

264 65.9 52.7 79.1

268 68.3 54.6 82.0





#72

Phenanthrene

Concen: 20.187 ng/ul

RT: 17.434 min Scan# 2456

Delta R.T. 0.000 min

Lab File: BN022062.D

Acq: 12 Oct 2022 13:20

Instrument :

BNA_N

ClientSampleId :

SSTD020247





Tgt Ion:178 Resp: 981431

Ion Ratio Lower Upper

178 100

179 15.4 12.3 18.5

176 20.6 16.5 24.7



#74

Anthracene

Concen: 20.346 ng/ul

RT: 17.522 min Scan# 2471

Delta R.T. 0.000 min

Lab File: BN022062.D

Acq: 12 Oct 2022 13:20




Tgt Ion:178 Resp: 985003


Ion Ratio Lower Upper

178 100

179 16.3 13.0 19.6

176 19.7 15.8 23.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.911 ng/uL
RT: 13.434 min Scan# 1776
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20




Instrument :
BNA_N
ClientSampleId :
SSTD020247


Tgt Ion:216 Resp: 226120
Ion Ratio Lower Upper
216 100
214 76.5 61.2 91.8
218 49.2 39.4 59.0



#76
Pentachlorobenzene
Concen: 21.066 ng/uL
RT: 14.951 min Scan# 2034
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20

Tgt Ion:250 Resp: 246507
Ion Ratio Lower Upper
250 100
248 58.5 46.8 70.2
252 67.4 53.9 80.9









#77
 Carbazole
 Concen: 21.224 ng/ul
 RT: 17.798 min Scan# 2518
 Delta R.T. 0.000 min
 Lab File: BN022062.D
 Acq: 12 Oct 2022 13:20




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020247


Tgt Ion:167 Resp: 948399
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.3 25.9
 139 13.4 10.7 16.1



#78
 Di-n-butylphthalate
 Concen: 21.396 ng/ul
 RT: 18.357 min Scan# 2613
 Delta R.T. 0.000 min
 Lab File: BN022062.D
 Acq: 12 Oct 2022 13:20

Tgt Ion:149 Resp: 1170731
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.1 10.7
 104 5.2 4.2 6.2





#80

Fluoranthene

Concen: 21.099 ng/ul

RT: 19.457 min Scan# 2800

Delta R.T. 0.000 min

Lab File: BN022062.D

Acq: 12 Oct 2022 13:20

Instrument :

BNA_N

ClientSampleId :

SSTD020247





Tgt Ion:202 Resp: 1135756

Ion Ratio Lower Upper

202 100

101 12.9 10.3 15.5

100 9.8 7.8 11.8



#82

Pyrene

Concen: 21.414 ng/ul

RT: 19.816 min Scan# 2861

Delta R.T. 0.000 min

Lab File: BN022062.D

Acq: 12 Oct 2022 13:20




Tgt Ion:202 Resp: 1216358


Ion Ratio Lower Upper

202 100

101 16.7 13.4 20.0

100 13.6 10.9 16.3





#83

Butylbenzylphthalate

Concen: 21.757 ng/ul

RT: 20.716 min Scan# 3014

Delta R.T. 0.000 min

Lab File: BN022062.D

Acq: 12 Oct 2022 13:20

Instrument :

BNA_N

ClientSampleId :

SSTD020247





Tgt Ion:149 Resp: 526998

Ion Ratio Lower Upper

149 100

91 77.1 61.7 92.5

206 18.4 14.7 22.1



#84

3,3'-Dichlorobenzidine

Concen: 21.266 ng/ul

RT: 21.510 min Scan# 3149

Delta R.T. 0.000 min

Lab File: BN022062.D

Acq: 12 Oct 2022 13:20




Tgt Ion:252 Resp: 403088


Ion Ratio Lower Upper

252 100

254 66.1 52.9 79.3

126 14.0 11.2 16.8









#85
Benzo(a)anthracene
Concen: 21.588 ng/ul
RT: 21.574 min Scan# 3160
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20




Instrument :
BNA_N
ClientSampleId :
SSTD020247


Tgt Ion:228 Resp: 1182274
Ion Ratio Lower Upper
228 100
229 16.9 13.5 20.3
226 26.3 21.0 31.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 22.297 ng/ul
RT: 21.492 min Scan# 3146
Delta R.T. 0.000 min
Lab File: BN022062.D
Acq: 12 Oct 2022 13:20

Tgt Ion:149 Resp: 784207
Ion Ratio Lower Upper
149 100
167 26.6 21.3 31.9
279 3.6 2.9 4.3





#87

Chrysene

Concen: 20.548 ng/ul

RT: 21.627 min Scan# 3169

Delta R.T. 0.000 min

Lab File: BN022062.D

Acq: 12 Oct 2022 13:20

Instrument :

BNA_N

ClientSampleId :

SSTD020247





Tgt Ion:228 Resp: 1107218

Ion Ratio Lower Upper

228 100

226 29.8 23.8 35.8

229 20.6 16.5 24.7



#89

Di-n-octyl phthalate

Concen: 22.271 ng/ul




RT: 22.439 min Scan# 3307


Delta R.T. 0.000 min

Lab File: BN022062.D

Acq: 12 Oct 2022 13:20

Tgt Ion:149 Resp: 1391204





#90

Benzo(b)fluoranthene

Concen: 21.399 ng/ul

RT: 23.298 min Scan# 3453

Delta R.T. 0.000 min

Lab File: BN022062.D

Acq: 12 Oct 2022 13:20

Instrument :

BNA_N

ClientSampleId :

SSTD020247





Tgt Ion:252 Resp: 1210254

Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4

125 13.4 10.7 16.1



#91

Benzo(k)fluoranthene

Concen: 20.947 ng/ul

RT: 23.351 min Scan# 3462

Delta R.T. 0.000 min

Lab File: BN022062.D

Acq: 12 Oct 2022 13:20




Tgt Ion:252 Resp: 1168453


Ion Ratio Lower Upper

252 100

253 22.1 17.7 26.5

125 13.2 10.6 15.8





#93

Benzo(a)pyrene

Concen: 21.869 ng/ul

RT: 23.951 min Scan# 3564

Delta R.T. 0.000 min

Lab File: BN022062.D

Acq: 12 Oct 2022 13:20

Instrument :

BNA_N

ClientSampleId :

SSTD020247





Tgt Ion:252 Resp: 1092073

Ion Ratio Lower Upper

252 100

253 21.5 17.2 25.8

125 13.5 10.8 16.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.233 ng/ul

RT: 26.662 min Scan# 4025

Delta R.T. 0.000 min

Lab File: BN022062.D

Acq: 12 Oct 2022 13:20




Tgt Ion:276 Resp: 1346071


Ion Ratio Lower Upper

276 100

138 30.7 24.6 36.8

277 23.0 18.4 27.6









#95
 Dibenzo(a,h)anthracene
 Concen: 21.003 ng/ul
 RT: 26.674 min Scan# 4027
 Delta R.T. 0.000 min
 Lab File: BN022062.D
 Acq: 12 Oct 2022 13:20

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020247

Tgt Ion:278 Resp: 1204127
 Ion Ratio Lower Upper
 278 100
 139 19.6 15.7 23.5
 279 20.6 16.5 24.7



#96
 Benzo(g,h,i)perylene
 Concen: 21.020 ng/ul
 RT: 27.457 min Scan# 4160
 Delta R.T. 0.000 min
 Lab File: BN022062.D
 Acq: 12 Oct 2022 13:20

Tgt Ion:276 Resp: 1201166
 Ion Ratio Lower Upper
 276 100
 138 26.5 21.2 31.8
 277 22.4 17.9 26.9

