

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101222\
 Data File : BN022066.D
 Acq On : 12 Oct 2022 15:52
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SICV202

Quant Time: Oct 13 03:13:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 13 03:08:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.963	152	113396	20.000	ng/u1	0.00
20) Naphthalene-d8	10.793	136	564162	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.634	164	389655	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.392	188	819424	20.000	ng/u1	0.00
79) Chrysene-d12	21.592	240	814974	20.000	ng/u1	0.00
88) Perylene-d12	24.062	264	891931	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.258	96	24512	8.193	ng/uL	0.00
4) Pyridine-d5	3.693	84	187352	20.131	ng/u1	0.00
7) Phenol-d5	7.116	99	223404	20.153	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.287	67	141277	20.337	ng/u1	0.00
11) 2-Chlorophenol-d4	7.481	132	162045	21.082	ng/u1	0.00
15) 4-Methylphenol-d8	8.681	113	182549	20.724	ng/u1	0.00
21) Nitrobenzene-d5	9.140	128	89176	21.557	ng/u1	0.00
24) 2-Nitrophenol-d4	9.869	143	94544	22.298	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.410	165	170833	21.464	ng/u1	0.00
31) 4-Chloroaniline-d4	10.934	131	273848	21.046	ng/u1	0.00
46) Dimethylphthalate-d6	14.045	166	568748	20.939	ng/u1	0.00
49) Acenaphthylene-d8	14.328	160	638973	21.066	ng/u1	0.00
54) 4-Nitrophenol-d4	14.834	143	120439	20.773	ng/u1	0.00
60) Fluorene-d10	15.628	176	478349	20.879	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.751	200	96955	20.740	ng/u1	0.00
73) Anthracene-d10	17.492	188	739459	20.766	ng/u1	0.00
81) Pyrene-d10	19.792	212	826405	20.367	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.904	264	887363	20.286	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.293	88	26281	8.060	ng/uL	99
5) Pyridine	3.716	79	186142	20.161	ng/u1	98
6) Benzaldehyde	7.093	77	133155	24.323	ng/u1	97
8) Phenol	7.146	94	238879	20.302	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.381	93	190870	20.390	ng/u1	99
12) 2-Chlorophenol	7.516	128	176646	20.813	ng/u1	98
13) 2-Methylphenol	8.410	108	178929	20.339	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.499	45	274959	20.527	ng/u1	100
16) Acetophenone	8.799	105	306924	21.808	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.781	70	165741	22.111	ng/u1	98
18) 4-Methylphenol	8.746	108	205064	20.818	ng/u1	100
19) Hexachloroethane	9.052	117	70327	21.224	ng/u1	97
22) Nitrobenzene	9.187	77	230887	21.137	ng/u1	100
23) Isophorone	9.716	82	477937	22.162	ng/u1	99
25) 2-Nitrophenol	9.899	139	109469	21.877	ng/u1	99
26) 2,4-Dimethylphenol	9.963	107	226277	21.447	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.198	93	286100	21.215	ng/u1	99
29) 2,4-Dichlorophenol	10.440	162	178505	21.506	ng/u1	99
30) Naphthalene	10.846	128	631847	20.907	ng/u1	99
32) 4-Chloroaniline	10.957	127	284746	21.109	ng/u1	97
33) Hexachlorobutadiene	11.122	225	98247	21.572	ng/u1	93
34) Caprolactam	11.740	113	69911	21.677	ng/u1	95
35) 4-Chloro-3-methylphenol	12.087	107	222839	22.268	ng/u1	98
36) 2-Methylnaphthalene	12.457	142	448286	21.479	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101222\
 Data File : BN022066.D
 Acq On : 12 Oct 2022 15:52
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SICV202


Quant Time: Oct 13 03:13:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 13 03:08:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.675	142	449725	21.518	ng/ul	96
39) 1,2,4,5-Tetrachloroben...	12.828	216	213597	21.131	ng/ul	94
40) Hexachlorocyclopentadiene	12.804	237	114366	20.239	ng/ul	93
41) 2,4,6-Trichlorophenol	13.069	196	147437	21.391	ng/ul	93
42) 2,4,5-Trichlorophenol	13.140	196	156253	20.357	ng/ul	99
43) 1,1'-Biphenyl	13.469	154	585664	20.589	ng/ul	98
44) 2-Chloronaphthalene	13.510	162	460584	20.723	ng/ul	96
45) 2-Nitroaniline	13.722	65	152394	22.132	ng/ul	96
47) Dimethylphthalate	14.092	163	597328	20.459	ng/ul	98
48) 2,6-Dinitrotoluene	14.216	165	120078	21.956	ng/ul	99
50) Acenaphthylene	14.357	152	715606	20.659	ng/ul	98
51) 3-Nitroaniline	14.545	138	138133	21.391	ng/ul	99
52) Acenaphthene	14.698	153	508434	20.698	ng/ul	98
53) 2,4-Dinitrophenol	14.751	184	72477	19.563	ng/ul	98
55) 4-Nitrophenol	14.851	109	97606	20.763	ng/ul	99
56) Dibenzofuran	15.034	168	699222	20.860	ng/ul	99
57) 2,4-Dinitrotoluene	15.004	165	180158	22.486	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.263	232	134607	21.072	ng/ul	93
59) Diethylphthalate	15.457	149	631031	21.278	ng/ul	99
61) Fluorene	15.686	166	563482	20.911	ng/ul	90
62) 4-Chlorophenyl-phenyle...	15.681	204	264187	20.924	ng/ul	97
63) 4-Nitroaniline	15.710	138	139368	23.131	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.763	198	105423	21.006	ng/ul	98
67) N-Nitrosodiphenylamine	15.892	169	500668	21.128	ng/ul	98
68) 4-Bromophenyl-phenylether	16.575	248	159521	22.037	ng/ul	91
69) Hexachlorobenzene	16.686	284	190259	21.592	ng/ul	92
70) Atrazine	16.851	200	177603	22.255	ng/ul	97
71) Pentachlorophenol	17.033	266	115361	21.172	ng/ul	95
72) Phenanthrene	17.433	178	912032	20.633	ng/ul	98
74) Anthracene	17.522	178	939996	21.409	ng/ul	96
75) 1,2,3,4-Tetrachloroben...	13.434	216	211633	20.534	ng/uL	100
76) Pentachlorobenzene	14.951	250	230130	20.803	ng/uL	97
77) Carbazole	17.798	167	883382	21.753	ng/ul	99
78) Di-n-butylphthalate	18.357	149	1119795	22.249	ng/ul	100
80) Fluoranthene	19.451	202	1051305	20.641	ng/ul	97
82) Pyrene	19.822	202	1115822	20.940	ng/ul	97
83) Butylbenzylphthalate	20.716	149	523504	22.644	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.510	252	387298	21.974	ng/ul	98
85) Benzo(a)anthracene	21.627	228	1029441	19.768	ng/ul	94
86) Bis(2-ethylhexyl)phtha...	21.498	149	759934	22.727	ng/ul	99
87) Chrysene	21.627	228	1029441	20.571	ng/ul	99
89) Di-n-octyl phthalate	22.445	149	1354242	21.430	ng/ul	100
90) Benzo(b)fluoranthene	23.304	252	1171424	20.824	ng/ul	99
91) Benzo(k)fluoranthene	23.351	252	1102627	19.839	ng/ul	100
93) Benzo(a)pyrene	23.951	252	1044497	20.774	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.651	276	1236049	19.669	ng/ul	94
95) Dibenzo(a,h)anthracene	26.680	278	1124231	19.996	ng/ul	95
96) Benzo(g,h,i)perylene	27.462	276	1122527	19.855	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 13 03:13:23 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101222.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 13 03:08:08 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.060 ng/uL
RT: 3.293 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52




Instrument :
BNA_N
ClientSampleId :
SICV202


Tgt Ion: 88 Resp: 26281
Ion Ratio Lower Upper
88 100
43 29.1 22.6 34.0
58 75.9 59.9 89.9



#5
Pyridine
Concen: 20.161 ng/uL
RT: 3.716 min Scan# 124
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52

Tgt Ion: 79 Resp: 186142
Ion Ratio Lower Upper
79 100
52 74.9 61.7 92.5
51 34.9 27.8 41.6









#6
Benzaldehyde
Concen: 24.323 ng/ul
RT: 7.093 min Scan# 698
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52




Instrument :
BNA_N
ClientSampleId :
SICV202


Tgt Ion: 77 Resp: 133155
Ion Ratio Lower Upper
77 100
105 95.7 78.6 118.0
106 91.4 75.4 113.2



#8
Phenol
Concen: 20.302 ng/ul
RT: 7.146 min Scan# 707
Delta R.T. 0.006 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52

Tgt Ion: 94 Resp: 238879
Ion Ratio Lower Upper
94 100
65 27.3 22.1 33.1
66 34.0 29.0 43.6









#10
Bis(2-Chloroethyl)ether
Concen: 20.390 ng/ul
RT: 7.381 min Scan# 747
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52




Instrument :
BNA_N
ClientSampleId :
SICV202


Tgt Ion: 93 Resp: 190870
Ion Ratio Lower Upper
93 100
63 77.4 61.1 91.7
95 32.4 26.4 39.6



#12
2-Chlorophenol
Concen: 20.813 ng/ul
RT: 7.516 min Scan# 770
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52

Tgt Ion:128 Resp: 176646
Ion Ratio Lower Upper
128 100
64 48.0 39.7 59.5
130 32.7 26.7 40.1





#13

2-Methylphenol

Concen: 20.339 ng/ul

RT: 8.410 min Scan# 91

Delta R.T. -0.000 min

Lab File: BN022066.D

Acq: 12 Oct 2022 15:52

Instrument :

BNA_N

ClientSampleId :





SICV202

Tgt Ion:108 Resp: 178929

Ion Ratio Lower Upper

108 100

107 88.8 72.4 108.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.527 ng/ul

RT: 8.499 min Scan# 937

Delta R.T. -0.006 min

Lab File: BN022066.D

Acq: 12 Oct 2022 15:52




Tgt Ion: 45 Resp: 274959


Ion Ratio Lower Upper

45 100

77 14.4 11.7 17.5

79 10.8 8.7 13.1









#16
Acetophenone
Concen: 21.808 ng/ul
RT: 8.799 min Scan# 988
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52




Instrument :
BNA_N
ClientSampleId :
SICV202


Tgt Ion:105 Resp: 306924
Ion Ratio Lower Upper
105 100
77 78.6 64.9 97.3
51 27.0 21.8 32.6



#17
N-Nitroso-di-n-propylamine
Concen: 22.111 ng/ul
RT: 8.781 min Scan# 985
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52

Tgt Ion: 70 Resp: 165741
Ion Ratio Lower Upper
70 100
42 55.3 45.6 68.4
101 8.8 7.4 11.2
130 20.0 15.4 23.0









#18
4-Methylphenol
Concen: 20.818 ng/ul
RT: 8.746 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52




Instrument :
BNA_N
ClientSampleId :
SICV202


Tgt Ion:108 Resp: 205064
Ion Ratio Lower Upper
108 100
107 109.4 87.2 130.8



#19
Hexachloroethane
Concen: 21.224 ng/ul
RT: 9.052 min Scan# 1031
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52

Tgt Ion:117 Resp: 70327
Ion Ratio Lower Upper
117 100
201 90.7 69.8 104.6
199 54.4 45.2 67.8









#22
Nitrobenzene
Concen: 21.137 ng/ul
RT: 9.187 min Scan# 1054
Delta R.T. 0.006 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52




Instrument :
BNA_N
ClientSampleId :
SICV202


Tgt Ion: 77 Resp: 230887
Ion Ratio Lower Upper
77 100
123 43.7 34.9 52.3
65 12.6 10.2 15.2



#23
Isophorone
Concen: 22.162 ng/ul
RT: 9.716 min Scan# 1144
Delta R.T. 0.006 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52

Tgt Ion: 82 Resp: 477937
Ion Ratio Lower Upper
82 100
95 6.6 5.4 8.0
138 16.6 13.8 20.8









#25
2-Nitrophenol
Concen: 21.877 ng/ul
RT: 9.899 min Scan# 1175
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52




Instrument :
BNA_N
ClientSampleId :
SICV202


Tgt Ion:139 Resp: 109469
Ion Ratio Lower Upper
139 100
65 53.3 42.7 64.1
109 28.8 23.8 35.6



#26
2,4-Dimethylphenol
Concen: 21.447 ng/ul
RT: 9.963 min Scan# 1186
Delta R.T. 0.006 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52

Tgt Ion:107 Resp: 226277
Ion Ratio Lower Upper
107 100
121 51.3 41.0 61.6
122 83.7 71.5 107.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.215 ng/ul
 RT: 10.198 min Scan# 1226
 Delta R.T. -0.000 min
 Lab File: BN022066.D
 Acq: 12 Oct 2022 15:52




Instrument :
 BNA_N
 ClientSampleId :
 SICV202


Tgt Ion: 93 Resp: 286100
 Ion Ratio Lower Upper
 93 100
 95 31.9 25.2 37.8
 123 11.2 8.6 12.8



#29
 2,4-Dichlorophenol
 Concen: 21.506 ng/ul
 RT: 10.440 min Scan# 1267
 Delta R.T. 0.006 min
 Lab File: BN022066.D
 Acq: 12 Oct 2022 15:52

Tgt Ion: 162 Resp: 178505
 Ion Ratio Lower Upper
 162 100
 164 64.7 52.9 79.3
 98 38.1 30.5 45.7





#30

Naphthalene

Concen: 20.907 ng/ul

RT: 10.846 min Scan# 11

Delta R.T. 0.006 min

Lab File: BN022066.D

Acq: 12 Oct 2022 15:52

Instrument :

BNA_N

ClientSampleId :

SICV202





Tgt Ion:128 Resp: 631847

Ion Ratio Lower Upper

128 100

129 11.1 8.5 12.7

127 12.9 10.8 16.2



#32

4-Chloroaniline

Concen: 21.109 ng/ul

RT: 10.957 min Scan# 1355

Delta R.T. -0.000 min

Lab File: BN022066.D




Acq: 12 Oct 2022 15:52


Tgt Ion:127 Resp: 284746

Ion Ratio Lower Upper

127 100

129 31.0 26.2 39.4









#33
Hexachlorobutadiene
Concen: 21.572 ng/ul
RT: 11.122 min Scan# 1383
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52




Instrument :
BNA_N
ClientSampleId :
SICV202


Tgt Ion:225 Resp: 98247
Ion Ratio Lower Upper
225 100
223 67.1 47.7 71.5
227 67.4 51.4 77.0



#34
Caprolactam
Concen: 21.677 ng/ul
RT: 11.740 min Scan# 1488
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52

Tgt Ion:113 Resp: 69911
Ion Ratio Lower Upper
113 100
55 165.9 138.7 208.1
56 122.4 101.4 152.0









#35
4-Chloro-3-methylphenol
Concen: 22.268 ng/ul
RT: 12.087 min Scan# 11
Delta R.T. 0.006 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52




Instrument :
BNA_N
ClientSampleId :
SICV202


Tgt Ion:107 Resp: 222839
Ion Ratio Lower Upper
107 100
144 27.3 20.3 30.5
142 79.5 64.4 96.6



#36
2-Methylnaphthalene
Concen: 21.479 ng/ul
RT: 12.457 min Scan# 1610
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52

Tgt Ion:142 Resp: 448286
Ion Ratio Lower Upper
142 100
141 84.6 69.7 104.5









#37
1-Methylnaphthalene
Concen: 21.518 ng/ul
RT: 12.675 min Scan# 1647
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52




Instrument :
BNA_N
ClientSampleId :
SICV202


Tgt Ion:142 Resp: 449725
Ion Ratio Lower Upper
142 100
141 91.3 76.7 115.1
116 3.9 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.131 ng/ul
RT: 12.828 min Scan# 1673
Delta R.T. 0.006 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52

Tgt Ion:216 Resp: 213597
Ion Ratio Lower Upper
216 100
214 74.7 65.4 98.2
179 21.5 17.4 26.0
108 18.6 16.4 24.6









#40
 Hexachlorocyclopentadiene
 Concen: 20.239 ng/ul
 RT: 12.804 min Scan# 1669
 Delta R.T. 0.006 min
 Lab File: BN022066.D
 Acq: 12 Oct 2022 15:52




Instrument :
 BNA_N
 ClientSampleId :
 SICV202


Tgt Ion:237 Resp: 114366
 Ion Ratio Lower Upper
 237 100
 235 60.3 43.4 65.0
 272 11.5 10.2 15.2



#41
 2,4,6-Trichlorophenol
 Concen: 21.391 ng/ul
 RT: 13.069 min Scan# 1714
 Delta R.T. 0.006 min
 Lab File: BN022066.D
 Acq: 12 Oct 2022 15:52

Tgt Ion:196 Resp: 147437
 Ion Ratio Lower Upper
 196 100
 198 100.0 73.4 110.2
 200 31.8 24.3 36.5









#42
2,4,5-Trichlorophenol
Concen: 20.357 ng/ul
RT: 13.140 min Scan# 1726
Delta R.T. 0.006 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52




Instrument :
BNA_N
ClientSampleId :
SICV202


Tgt Ion:196 Resp: 156253
Ion Ratio Lower Upper
196 100
198 97.0 77.0 115.6
200 32.4 24.7 37.1



#43
1,1'-Biphenyl
Concen: 20.589 ng/ul
RT: 13.469 min Scan# 1782
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52

Tgt Ion:154 Resp: 585664
Ion Ratio Lower Upper
154 100
153 40.2 30.9 46.3
76 14.5 11.6 17.4





#44

2-Chloronaphthalene

Concen: 20.723 ng/ul

RT: 13.510 min Scan# 1789

Delta R.T. -0.000 min

Lab File: BN022066.D

Acq: 12 Oct 2022 15:52

Instrument :

BNA_N

ClientSampleId :

SICV202





Tgt Ion:162 Resp: 460584

Ion Ratio Lower Upper

162 100

164 32.2 23.7 35.5

127 40.9 30.8 46.2



#45

2-Nitroaniline

Concen: 22.132 ng/ul

RT: 13.722 min Scan# 1825

Delta R.T. -0.000 min

Lab File: BN022066.D

Acq: 12 Oct 2022 15:52




Tgt Ion: 65 Resp: 152394


Ion Ratio Lower Upper

65 100

92 69.8 56.2 84.2

138 100.2 85.0 127.6





#47

Dimethylphthalate

Concen: 20.459 ng/ul

RT: 14.092 min Scan# 1888

Delta R.T. -0.000 min

Lab File: BN022066.D

Acq: 12 Oct 2022 15:52

Instrument :

BNA_N

ClientSampleId :

SICV202





Tgt Ion:163 Resp: 597328

Ion Ratio Lower Upper

163 100

194 4.5 3.1 4.7

164 10.8 8.2 12.2



#48

2,6-Dinitrotoluene

Concen: 21.956 ng/ul

RT: 14.216 min Scan# 1909

Delta R.T. -0.000 min

Lab File: BN022066.D

Acq: 12 Oct 2022 15:52




Tgt Ion:165 Resp: 120078


Ion Ratio Lower Upper

165 100

89 65.5 52.1 78.1

121 24.4 19.0 28.4









#50
Acenaphthylene
Concen: 20.659 ng/ul
RT: 14.357 min Scan# 1933
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52




Instrument :
BNA_N
ClientSampleId :
SICV202


Tgt Ion:152 Resp: 715606
Ion Ratio Lower Upper
152 100
151 20.6 15.5 23.3
153 13.0 9.8 14.8



#51
3-Nitroaniline
Concen: 21.391 ng/ul
RT: 14.545 min Scan# 1965
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52

Tgt Ion:138 Resp: 138133
Ion Ratio Lower Upper
138 100
108 10.2 9.0 13.6
92 121.0 96.4 144.6









#52
Acenaphthene
Concen: 20.698 ng/ul
RT: 14.698 min Scan# 1991
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52




Instrument :
BNA_N
ClientSampleId :
SICV202


Tgt Ion:153 Resp: 508434
Ion Ratio Lower Upper
153 100
152 48.2 38.0 57.0
154 91.2 71.1 106.7



#53
2,4-Dinitrophenol
Concen: 19.563 ng/ul
RT: 14.751 min Scan# 2000
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52

Tgt Ion:184 Resp: 72477
Ion Ratio Lower Upper
184 100
63 73.8 56.8 85.2
154 62.1 49.0 73.6





#55

4-Nitrophenol

Concen: 20.763 ng/ul

RT: 14.851 min Scan# 2017

Delta R.T. -0.000 min

Lab File: BN022066.D

Acq: 12 Oct 2022 15:52

Instrument :

BNA_N

ClientSampleId :

SICV202





Tgt Ion:109 Resp: 97606

Ion Ratio Lower Upper

109 100

139 135.9 108.3 162.5

65 133.3 107.9 161.9



#56

Dibenzofuran

Concen: 20.860 ng/ul

RT: 15.034 min Scan# 2048

Delta R.T. -0.000 min

Lab File: BN022066.D




Acq: 12 Oct 2022 15:52


Tgt Ion:168 Resp: 699222

Ion Ratio Lower Upper

168 100

139 38.9 30.6 45.8









#57
2,4-Dinitrotoluene
Concen: 22.486 ng/ul
RT: 15.004 min Scan# 2043
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52




Instrument :
BNA_N
ClientSampleId :
SICV202


Tgt Ion:165 Resp: 180158
Ion Ratio Lower Upper
165 100
63 45.2 36.2 54.4
182 2.6 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.072 ng/ul
RT: 15.263 min Scan# 2087
Delta R.T. 0.006 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52

Tgt Ion:232 Resp: 134607
Ion Ratio Lower Upper
232 100
131 49.8 44.4 66.6
130 2.3 2.2 3.4
166 32.0 28.8 43.2









#59
Diethylphthalate
Concen: 21.278 ng/ul
RT: 15.457 min Scan# 2120
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52




Instrument :
BNA_N
ClientSampleId :
SICV202


Tgt Ion:149 Resp: 631031
Ion Ratio Lower Upper
149 100
177 22.2 17.5 26.3
150 12.1 9.8 14.8



#61
Fluorene
Concen: 20.911 ng/ul
RT: 15.686 min Scan# 2159
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52

Tgt Ion:166 Resp: 563482
Ion Ratio Lower Upper
166 100
165 101.7 72.8 109.2
167 14.2 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 20.924 ng/ul
RT: 15.681 min Scan# 2158
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52




Instrument :
BNA_N
ClientSampleId :
SICV202


Tgt Ion:204 Resp: 264187
Ion Ratio Lower Upper
204 100
206 33.7 26.2 39.2
141 61.8 52.0 78.0



#63
4-Nitroaniline
Concen: 23.131 ng/ul
RT: 15.710 min Scan# 2163
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52

Tgt Ion:138 Resp: 139368
Ion Ratio Lower Upper
138 100
92 50.5 41.8 62.8
108 76.4 60.4 90.6









#66
 4,6-Dinitro-2-methylphenol
 Concen: 21.006 ng/ul
 RT: 15.763 min Scan# 2172
 Delta R.T. -0.000 min
 Lab File: BN022066.D
 Acq: 12 Oct 2022 15:52




Instrument :
 BNA_N
 ClientSampleId :
 SICV202


Tgt Ion:198 Resp: 105423
 Ion Ratio Lower Upper
 198 100
 182 5.3 3.8 5.8
 77 31.5 24.4 36.6



#67
 N-Nitrosodiphenylamine
 Concen: 21.128 ng/ul
 RT: 15.892 min Scan# 2194
 Delta R.T. -0.000 min
 Lab File: BN022066.D
 Acq: 12 Oct 2022 15:52

Tgt Ion:169 Resp: 500668
 Ion Ratio Lower Upper
 169 100
 168 62.6 50.9 76.3
 167 33.9 28.2 42.4









#68
4-Bromophenyl-phenylether
Concen: 22.037 ng/ul
RT: 16.575 min Scan# 2310
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52




Instrument :
BNA_N
ClientSampleId :
SICV202


Tgt Ion:248 Resp: 159521
Ion Ratio Lower Upper
248 100
250 99.9 86.3 129.5
141 83.2 74.2 111.4



#69
Hexachlorobenzene
Concen: 21.592 ng/ul
RT: 16.686 min Scan# 2329
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52

Tgt Ion:284 Resp: 190259
Ion Ratio Lower Upper
284 100
142 40.3 35.7 53.5
249 28.4 27.0 40.6





#70

Atrazine

Concen: 22.255 ng/ul

RT: 16.851 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN022066.D

Acq: 12 Oct 2022 15:52

Instrument :

BNA_N

ClientSampleId :

SICV202





Tgt Ion:200 Resp: 177603

Ion Ratio Lower Upper

200 100

173 27.7 22.8 34.2

215 45.1 38.1 57.1



#71

Pentachlorophenol

Concen: 21.172 ng/ul

RT: 17.033 min Scan# 2388

Delta R.T. -0.000 min

Lab File: BN022066.D

Acq: 12 Oct 2022 15:52




Tgt Ion:266 Resp: 115361


Ion Ratio Lower Upper

266 100

264 64.3 52.7 79.1

268 61.4 54.6 82.0





#72

Phenanthrene

Concen: 20.633 ng/u1

RT: 17.433 min Scan# 2456

Delta R.T. -0.000 min

Lab File: BN022066.D

Acq: 12 Oct 2022 15:52

Instrument :

BNA_N

ClientSampleId :

SICV202





Tgt Ion:178 Resp: 912032

Ion Ratio Lower Upper

178 100

179 15.3 12.3 18.5

176 19.4 16.5 24.7



#74

Anthracene

Concen: 21.409 ng/u1

RT: 17.522 min Scan# 2471

Delta R.T. -0.000 min

Lab File: BN022066.D

Acq: 12 Oct 2022 15:52




Tgt Ion:178 Resp: 939996


Ion Ratio Lower Upper

178 100

179 15.2 13.0 19.6

176 17.4 15.8 23.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.534 ng/uL
RT: 13.434 min Scan# 1776
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52




Instrument :
BNA_N
ClientSampleId :
SICV202


Tgt Ion:216 Resp: 211633
Ion Ratio Lower Upper
216 100
214 76.4 61.2 91.8
218 48.9 39.4 59.0



#76
Pentachlorobenzene
Concen: 20.803 ng/uL
RT: 14.951 min Scan# 2034
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52

Tgt Ion:250 Resp: 230130
Ion Ratio Lower Upper
250 100
248 55.2 46.8 70.2
252 65.4 53.9 80.9









#77
 Carbazole
 Concen: 21.753 ng/ul
 RT: 17.798 min Scan# 2518
 Delta R.T. -0.000 min
 Lab File: BN022066.D
 Acq: 12 Oct 2022 15:52




Instrument :
 BNA_N
 ClientSampleId :
 SICV202


Tgt Ion:167 Resp: 883382
 Ion Ratio Lower Upper
 167 100
 166 21.2 17.3 25.9
 139 13.4 10.7 16.1



#78
 Di-n-butylphthalate
 Concen: 22.249 ng/ul
 RT: 18.357 min Scan# 2613
 Delta R.T. -0.000 min
 Lab File: BN022066.D
 Acq: 12 Oct 2022 15:52

Tgt Ion:149 Resp: 1119795
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.1 10.7
 104 5.2 4.2 6.2





#80

Fluoranthene

Concen: 20.641 ng/ul

RT: 19.451 min Scan# 21

Delta R.T. -0.006 min

Lab File: BN022066.D

Acq: 12 Oct 2022 15:52

Instrument :

BNA_N

ClientSampleId :

SICV202





Tgt Ion:202 Resp: 1051305

Ion Ratio Lower Upper

202 100

101 14.3 10.3 15.5

100 11.0 7.8 11.8



#82

Pyrene

Concen: 20.940 ng/ul

RT: 19.822 min Scan# 2862

Delta R.T. 0.006 min

Lab File: BN022066.D

Acq: 12 Oct 2022 15:52




Tgt Ion:202 Resp: 1115822


Ion Ratio Lower Upper

202 100

101 15.4 13.4 20.0

100 12.4 10.9 16.3









#83
Butylbenzylphthalate
Concen: 22.644 ng/ul
RT: 20.716 min Scan# 3014
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52




Instrument :
BNA_N
ClientSampleId :
SICV202


Tgt Ion:149 Resp: 523504
Ion Ratio Lower Upper
149 100
91 75.5 61.7 92.5
206 18.1 14.7 22.1



#84
3,3'-Dichlorobenzidine
Concen: 21.974 ng/ul
RT: 21.510 min Scan# 3149
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52

Tgt Ion:252 Resp: 387298
Ion Ratio Lower Upper
252 100
254 68.2 52.9 79.3
126 14.7 11.2 16.8









#85
Benzo(a)anthracene
Concen: 19.768 ng/ul
RT: 21.627 min Scan# 3169
Delta R.T. 0.053 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52




Instrument :
BNA_N
ClientSampleId :
SICV202


Tgt Ion:228 Resp: 1029441
Ion Ratio Lower Upper
228 100
229 19.3 13.5 20.3
226 29.9 21.0 31.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 22.727 ng/ul
RT: 21.498 min Scan# 3147
Delta R.T. 0.006 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52

Tgt Ion:149 Resp: 759934
Ion Ratio Lower Upper
149 100
167 27.1 21.3 31.9
279 3.8 2.9 4.3









#87
Chrysene
Concen: 20.571 ng/ul
RT: 21.627 min Scan# 3169
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52




Instrument :
BNA_N
ClientSampleId :
SICV202


Tgt Ion:228 Resp: 1029441
Ion Ratio Lower Upper
228 100
226 29.9 23.8 35.8
229 19.3 16.5 24.7



#89
Di-n-octyl phthalate
Concen: 21.430 ng/ul
RT: 22.445 min Scan# 3308
Delta R.T. 0.006 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52

Tgt Ion:149 Resp: 1354242









#90
Benzo(b)fluoranthene
Concen: 20.824 ng/ul
RT: 23.304 min Scan# 3453
Delta R.T. 0.006 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52




Instrument :
BNA_N
ClientSampleId :
SICV202


Tgt Ion:252 Resp: 1171424
Ion Ratio Lower Upper
252 100
253 22.3 17.6 26.4
125 12.5 10.7 16.1



#91
Benzo(k)fluoranthene
Concen: 19.839 ng/ul
RT: 23.351 min Scan# 3462
Delta R.T. -0.000 min
Lab File: BN022066.D
Acq: 12 Oct 2022 15:52

Tgt Ion:252 Resp: 1102627
Ion Ratio Lower Upper
252 100
253 22.3 17.7 26.5
125 13.1 10.6 15.8





#93

Benzo(a)pyrene

Concen: 20.774 ng/ul

RT: 23.951 min Scan# 3564

Delta R.T. -0.000 min

Lab File: BN022066.D

Acq: 12 Oct 2022 15:52

Instrument :

BNA_N

ClientSampleId :

SICV202





Tgt Ion:252 Resp: 1044497

Ion Ratio Lower Upper

252 100

253 21.4 17.2 25.8

125 14.1 10.8 16.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.669 ng/ul

RT: 26.651 min Scan# 4023

Delta R.T. -0.012 min

Lab File: BN022066.D

Acq: 12 Oct 2022 15:52




Tgt Ion:276 Resp: 1236049


Ion Ratio Lower Upper

276 100

138 27.5 24.6 36.8

277 20.3 18.4 27.6









#95
 Dibenzo(a,h)anthracene
 Concen: 19.996 ng/ul
 RT: 26.680 min Scan# 4027
 Delta R.T. 0.006 min
 Lab File: BN022066.D
 Acq: 12 Oct 2022 15:52

Instrument :
 BNA_N
 ClientSampleId :
 SICV202

Tgt Ion:278 Resp: 1124231
 Ion Ratio Lower Upper
 278 100
 139 19.3 15.7 23.5
 279 24.4 16.5 24.7



#96
 Benzo(g,h,i)perylene
 Concen: 19.855 ng/ul
 RT: 27.462 min Scan# 4161
 Delta R.T. 0.006 min
 Lab File: BN022066.D
 Acq: 12 Oct 2022 15:52

Tgt Ion:276 Resp: 1122527
 Ion Ratio Lower Upper
 276 100
 138 25.5 21.2 31.8
 277 24.0 17.9 26.9

