

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101322\
 Data File : BN022080.D
 Acq On : 13 Oct 2022 09:58
 Operator : CG/JU
 Sample : PB148293BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB148293BL

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 10/14/2022
 Supervised By :Jagrut Upadhyay 10/14/2022

Quant Time: Oct 14 08:47:04 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 13 01:51:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.964	152	4165	0.400	ng	0.00
7) Naphthalene-d8	10.795	136	12105	0.400	ng	0.00
13) Acenaphthene-d10	14.632	164	6477	0.400	ng	0.00
19) Phenanthrene-d10	17.387	188	14849	0.400	ng	# 0.00
29) Chrysene-d12	21.591	240	12323	0.400	ng	0.00
35) Perylene-d12	24.055	264	7550	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.494	112	4021	0.417	ng	0.00
5) Phenol-d6	7.126	99	4617	0.369	ng	0.00
8) Nitrobenzene-d5	9.140	82	3492m	0.359	ng	0.00
11) 2-Methylnaphthalene-d10	12.387	152	7336	0.330	ng	0.00
14) 2,4,6-Tribromophenol	16.134	330	889	0.360	ng	0.00
15) 2-Fluorobiphenyl	13.263	172	10329	0.365	ng	0.00
27) Fluoranthene-d10	19.424	212	14670	0.372	ng	0.00
31) Terphenyl-d14	20.015	244	9322	0.376	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101322\
Data File : BN022080.D
Acq On : 13 Oct 2022 09:58
Operator : CG/JU
Sample : PB148293BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB148293BL

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 10/14/2022
Supervised By :Jagrut Upadhyay 10/14/2022

Quant Time: Oct 14 08:47:04 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 13 01:51:34 2022
Response via : Initial Calibration

