

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101322\
 Data File : BN022081.D
 Acq On : 13 Oct 2022 10:35
 Operator : CG/JU
 Sample : N5100-03RE
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-RW5-GW-883-885RE

Quant Time: Oct 14 08:47:14 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 13 01:51:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.964	152	4071	0.400	ng	0.00
7) Naphthalene-d8	10.795	136	13797	0.400	ng	# 0.00
13) Acenaphthene-d10	14.625	164	8258	0.400	ng	0.00
19) Phenanthrene-d10	17.378	188	17037	0.400	ng	# 0.00
29) Chrysene-d12	21.582	240	15472	0.400	ng	# 0.00
35) Perylene-d12	24.052	264	13434	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.494	112	289	0.031	ng	0.00
5) Phenol-d6	7.119	99	453	0.037	ng	0.00
8) Nitrobenzene-d5	9.140	82	1675	0.151	ng	0.00
11) 2-Methylnaphthalene-d10	12.383	152	2932	0.116	ng	0.00
14) 2,4,6-Tribromophenol	16.125	330	127	0.040	ng	0.00
15) 2-Fluorobiphenyl	13.256	172	3978	0.110	ng	0.00
27) Fluoranthene-d10	19.424	212	3892	0.086	ng	0.00
31) Terphenyl-d14	20.014	244	3256	0.105	ng	0.00
Target Compounds						
9) Naphthalene	10.837	128	1828	0.044	ng	# 75
24) Pentachlorophenol	17.031	266	166	0.042	ng	96
25) Phenanthrene	17.428	178	1568	0.027	ng	# 96
34) Bis(2-ethylhexyl)phtha...	21.483	149	15838	0.567	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101322\
Data File : BN022081.D
Acq On : 13 Oct 2022 10:35
Operator : CG/JU
Sample : N5100-03RE
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-RW5-GW-883-885RE

Quant Time: Oct 14 08:47:14 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 13 01:51:34 2022
Response via : Initial Calibration

