

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101322\
 Data File : BN022068.D
 Acq On : 12 Oct 2022 17:12
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 13 07:19:58 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 13 01:51:34 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	58	0.00
2	1,4-Dioxane	0.400	0.393	1.8	60	0.00
3	n-Nitrosodimethylamine	0.400	0.378	5.5	60	0.00
4 S	2-Fluorophenol	0.400	0.422	-5.5	67	0.00
5 S	Phenol-d6	0.400	0.393	1.8	63	0.00
6	bis(2-Chloroethyl)ether	0.400	0.391	2.3	62	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	58	0.00
8 S	Nitrobenzene-d5	0.400	0.413	-3.2	66	0.00
9	Naphthalene	0.400	0.394	1.5	61	0.00
10	Hexachlorobutadiene	0.400	0.391	2.3	61	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.376	6.0	59	0.00
12	2-Methylnaphthalene	0.400	0.491	-22.7	85	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	70	0.00
14 S	2,4,6-Tribromophenol	0.400	0.466	-16.5	96	0.00
15 S	2-Fluorobiphenyl	0.400	0.384	4.0	70	0.00
16	Acenaphthylene	0.400	0.383	4.3	76	0.00
17	Acenaphthene	0.400	0.391	2.3	74	0.00
18	Fluorene	0.400	0.477	-19.2	92	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	81	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.492	-23.0	110	0.00
21	4-Bromophenyl-phenylether	0.400	0.384	4.0	85	0.00
22	Hexachlorobenzene	0.400	0.376	6.0	82	0.00
23	Atrazine	0.400	0.432	-8.0	102	0.00
24	Pentachlorophenol	0.400	0.445	-11.2	112	0.00
25	Phenanthrene	0.400	0.396	1.0	87	0.00
26	Anthracene	0.400	0.388	3.0	89	0.00
27 SURR	Fluoranthene-d10	0.400	0.415	-3.7	94	0.00
28	Fluoranthene	0.400	0.417	-4.2	94	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	103	0.00
30	Pyrene	0.400	0.345	13.8	94	0.00
31 S	Terphenyl-d14	0.400	0.332	17.0	90	0.00
32	Benzo(a)anthracene	0.400	0.396	1.0	113	0.00
33	Chrysene	0.400	0.377	5.8	101	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.447	-11.7	131	0.00
35 I	Perylene-d12	0.400	0.400	0.0	111	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.435	-8.7	134	0.00
37	Benzo(b)fluoranthene	0.400	0.377	5.8	116	0.00
38	Benzo(k)fluoranthene	0.400	0.363	9.3	109	0.00
39 C	Benzo(a)pyrene	0.400	0.386	3.5	120	0.00
40	Dibenzo(a,h)anthracene	0.400	0.441	-10.2	135	0.00
41	Benzo(g,h,i)perylene	0.400	0.425	-6.2	126	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4