




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN101420\
Data File : BN012203.D
Acq On : 14 Oct 2020 20:19
Operator : CG/JU
Sample : L4166-05
Misc :
ALS Vial : 15 Sample Multiplier: 1

Quant Time: Oct 15 01:06:08 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN101420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 15 00:46:15 2020
Response via : Initial Calibration



TIC: BN012203.D

(21) Benzo(b)fluoranthene
22.747min (-0.003) 0.65ng/ul
response 17917

Ion	Exp%	Act%
252.00	100	100
253.00	30.80	29.56
125.00	18.40	19.47
0.00	0.00	0.00