




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN101420\
Data File : BN012204.D
Acq On : 14 Oct 2020 20:55
Operator : CG/JU
Sample : L4166-09
Misc :
ALS Vial : 16 Sample Multiplier: 1

Quant Time: Oct 15 00:53:56 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN101420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 15 00:46:15 2020
Response via : Initial Calibration



TIC: BN012204.D

(21) Benzo(b)fluoranthene
22.744min (-0.006) 0.79ng/ul
response 18097

Ion	Exp%	Act%
252.00	100	100
253.00	30.80	29.22
125.00	18.40	19.32
0.00	0.00	0.00