




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN101420\
Data File : BN012207.D
Acq On : 14 Oct 2020 22:43
Operator : CG/JU
Sample : L4166-08
Misc :
ALS Vial : 19 Sample Multiplier: 1

Quant Time: Oct 15 00:54:21 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN101420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 15 00:46:15 2020
Response via : Initial Calibration



TIC: BN012207.D

(21) Benzo(b)fluoranthene

22.741min (-0.009) 1.40ng/ul m

response 37319

Ion	Exp%	Act%
252.00	100	100
253.00	30.80	25.12
125.00	18.40	15.23
0.00	0.00	0.00