




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN101420\
 Data File : BN012240.D
 Acq On : 15 Oct 2020 20:25
 Operator : CG/JU
 Sample : L4235-02
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Quant Time: Oct 16 01:27:28 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN101420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 16 01:19:50 2020
 Response via : Initial Calibration



TIC: BN012240.D

(21) Benzo(b)fluoranthene

22.752min (+0.006) 0.39ng/ul m

response 13319

Ion	Exp%	Act%
252.00	100	100
253.00	30.80	74.47#
125.00	18.40	247.94#
0.00	0.00	0.00