




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN101420\
Data File : BN012244.D
Acq On : 15 Oct 2020 22:49
Operator : CG/JU
Sample : L4235-10
Misc :
ALS Vial : 56 Sample Multiplier: 1

Quant Time: Oct 16 01:28:11 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN101420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 16 01:19:50 2020
Response via : Initial Calibration



TIC: BN012244.D

(21) Benzo(b)fluoranthene
22.744min (-0.003) 1.30ng/ul
response 39277

Ion	Exp%	Act%
252.00	100	100
253.00	30.80	26.36
125.00	18.40	16.77
0.00	0.00	0.00