




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN101420\
Data File : BN012249.D
Acq On : 16 Oct 2020 02:26
Operator : CG/JU
Sample : L4235-15
Misc :
ALS Vial : 61 Sample Multiplier: 1

Quant Time: Oct 16 06:28:02 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN101420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 16 06:21:01 2020
Response via : Initial Calibration



TIC: BN012249.D

(22) Benzo(k)fluoranthene

22.904min (+0.123) 0.51ng/ul

response 16091

Ion	Exp%	Act%
252.00	100	100
253.00	30.50	32.02
125.00	16.10	18.61
0.00	0.00	0.00