




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN101420\
Data File : BN012251.D
Acq On : 16 Oct 2020 03:38
Operator : CG/JU
Sample : L4235-18
Misc :
ALS Vial : 63 Sample Multiplier: 1

Quant Time: Oct 16 06:28:16 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN101420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 16 06:21:01 2020
Response via : Initial Calibration



TIC: BN012251.D

(21) Benzo(b)fluoranthene

22.738min (-0.003) 0.13ng/ul m

response 4725

Ion	Exp%	Act%
252.00	100	100
253.00	30.80	41.71
125.00	18.40	39.12#
0.00	0.00	0.00