




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN101420\
 Data File : BN012254.D
 Acq On : 16 Oct 2020 05:27
 Operator : CG/JU
 Sample : L4235-17
 Misc :
 ALS Vial : 66 Sample Multiplier: 1

Quant Time: Oct 16 06:28:38 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN101420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 16 06:21:01 2020
 Response via : Initial Calibration



TIC: BN012254.D

(22) Benzo(k)fluoranthene

22.785min (+0.003) 0.04ng/ul m

response 1421

Ion	Exp%	Act%
252.00	100	100
253.00	30.50	61.38#
125.00	16.10	72.12#
0.00	0.00	0.00