

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN101420\  
 Data File : BN012242.D  
 Acq On : 15 Oct 2020 21:37  
 Operator : CG/JU  
 Sample : L4235-09  
 Misc :  
 ALS Vial : 54 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :

Quant Time: Oct 16 02:11:12 2020  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN101420.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Oct 16 01:19:50 2020  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units  | Dev(Min) |
|-----------------------------|-------|------|----------|--------|--------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 7.62  | 152  | 1740     | 0.40   | ng/ul  | 0.00     |
| 2) Naphthalene-d8           | 10.39 | 136  | 7274     | 0.40   | ng/ul  | 0.00     |
| 6) Acenaphthene-d10         | 14.26 | 164  | 4233     | 0.40   | ng/ul  | 0.00     |
| 10) Phenanthrene-d10        | 17.01 | 188  | 8898     | 0.40   | ng/ul  | 0.00     |
| 16) Chrysene-d12            | 21.21 | 240  | 7551     | 0.40   | ng/ul  | 0.00     |
| 20) Perylene-d12            | 23.40 | 264  | 7562     | 0.40   | ng/ul  | 0.00     |
| System Monitoring Compounds |       |      |          |        |        |          |
| 4) 2-Methylnaphthalene-d10  | 11.99 | 152  | 1645     | 0.15   | ng/ul  | 0.00     |
| 14) Fluoranthene-d10        | 19.04 | 212  | 3849     | 0.16   | ng/ul  | 0.00     |
| Target Compounds            |       |      |          | Ovalue |        |          |
| 3) Naphthalene              | 10.44 | 128  | 1094     | 0.052  | ng/ul# | 86       |
| 5) 2-Methylnaphthalene      | 12.06 | 142  | 567      | 0.043  | ng/ul  | 97       |
| 7) Acenaphthylene           | 13.98 | 152  | 2093     | 0.114  | ng/ul# | 88       |
| 8) Acenaphthene             | 14.32 | 153  | 1493     | 0.096  | ng/ul  | 97       |
| 9) Fluorene                 | 15.31 | 166  | 2236     | 0.129  | ng/ul# | 99       |
| 11) Pentachlorophenol       | 16.66 | 266  | 272      | 0.106  | ng/ul  | 97       |
| 12) Phenanthrene            | 17.05 | 178  | 33177    | 1.170  | ng/ul  | 98       |
| 13) Anthracene              | 17.14 | 178  | 7926     | 0.326  | ng/ul  | 100      |
| 15) Fluoranthene            | 19.07 | 202  | 70659    | 2.130  | ng/ul  | 98       |
| 17) Pyrene                  | 19.44 | 202  | 58488    | 1.718  | ng/ul  | 99       |
| 18) Benzo(a)anthracene      | 21.19 | 228  | 36958    | 1.321  | ng/ul  | 99       |
| 19) Chrysene                | 21.25 | 228  | 34444    | 1.011  | ng/ul  | 98       |
| 21) Benzo(b)fluoranthene    | 22.75 | 252  | 43121m   | 1.412  | ng/ul  |          |
| 22) Benzo(k)fluoranthene    | 22.78 | 252  | 17068m   | 0.505  | ng/ul  |          |
| 23) Benzo(a)pyrene          | 23.30 | 252  | 29193    | 0.982  | ng/ul# | 87       |
| 24) Indeno(1,2,3-cd)pyrene  | 25.58 | 276  | 19646    | 0.515  | ng/ul# | 100      |
| 25) Dibenzo(a,h)anthracene  | 25.58 | 278  | 5541     | 0.181  | ng/ul# | 84       |
| 26) Benzo(g,h,i)perylene    | 26.24 | 276  | 16167    | 0.462  | ng/ul  | 97       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN101420\  
Data File : BN012242.D  
Acq On : 15 Oct 2020 21:37  
Operator : CG/JU  
Sample : L4235-09  
Misc :  
ALS Vial : 54 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :

Quant Time: Oct 16 02:11:12 2020  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN101420.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Oct 16 01:19:50 2020  
Response via : Initial Calibration

