


Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN101420\
Data File : BN012224.D
Acq On : 15 Oct 2020 10:47
Operator : CG/JU
Sample : L4201-04
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AM4

Manual Integrations
APPROVED

mohammad
10/19/2020 1:19:09 PM

Quant Time: Oct 15 11:18:42 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN101420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 15 10:47:40 2020
Response via : Initial Calibration



Quantitation Report (Qedit)




Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN101420\
 Data File : BN012224.D
 Acq On : 15 Oct 2020 10:47
 Operator : CG/JU
 Sample : L4201-04
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_N
ClientSampleId :
 C0AM4

Manual Integrations
APPROVED

mohammad
 10/19/2020 1:19:09 PM

Quant Time: Oct 15 11:17:18 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN101420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 15 10:47:40 2020
 Response via : Initial Calibration



TIC: BN012224.D

(10) Phenanthrene-d10 (I)
 17.013min (-17.013) 0.00ng/ul
 response 0

Ion	Exp%	Act%
188.00	100	0.00
94.00	11.20	0.00#
80.00	12.20	0.00#
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN101420\
Data File : BN012224.D
Acq On : 15 Oct 2020 10:47
Operator : CG/JU
Sample : L4201-04
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AM4

Manual Integrations
APPROVED

mohammad
10/19/2020 1:19:09 PM

Quant Time: Oct 15 11:17:18 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN101420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 15 10:47:40 2020
Response via : Initial Calibration



TIC: BN012224.D

(10) Phenanthrene-d10 (I)

17.008min (-0.005) 0.40ng/ul m 10/20/20 JU

response 10019

Ion	Exp%	Act%
188.00	100	100
94.00	11.20	10.92
80.00	12.20	12.96
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN101420\
 Data File : BN012224.D
 Acq On : 15 Oct 2020 10:47
 Operator : CG/JU
 Sample : L4201-04
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AM4

Manual Integrations
 APPROVED

mohammad
 10/19/2020 1:19:09 PM

Quant Time: Oct 15 11:18:42 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN101420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 15 10:47:40 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev (Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	1833	0.40	ng/ul	0.00
2) Naphthalene-d8	10.39	136	7789	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.26	164	4792	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.01	188	10019m>	0.40	ng/ul	> 0.00
16) Chrysene-d12	21.21	240	8128	0.40	ng/ul	0.00
20) Perylene-d12	23.40	264	8699	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev (Min)
4) 2-Methylnaphthalene-d10	12.00	152	4358	0.38	ng/ul	0.00
14) Fluoranthene-d10	19.05	212	9358	0.34	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
11) Pentachlorophenol	16.67	266	112	0.039	ng/ul#	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed