

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101421\
 Data File : BN016958.D
 Acq On : 14 Oct 2021 19:27
 Operator : CG/JU
 Sample : PB139115BS
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB139115BS

Quant Time: Oct 15 01:44:00 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 14 17:01:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.615	152	19989	0.400	ng	0.00
7) Naphthalene-d8	10.381	136	88211	0.400	ng	0.00
13) Acenaphthene-d10	14.243	164	46015	0.400	ng	0.00
19) Phenanthrene-d10	16.993	188	84493	0.400	ng	0.00
29) Chrysene-d12	21.196	240	83001	0.400	ng	0.00
35) Perylene-d12	23.426	264	84677	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.217	112	15713	0.316	ng	0.00
5) Phenol-d6	6.794	99	17155	0.310	ng	0.00
8) Nitrobenzene-d5	8.759	82	16091	0.286	ng	0.00
11) 2-Methylnaphthalene-d10	11.976	152	40864	0.327	ng	0.00
14) 2,4,6-Tribromophenol	15.740	330	5669	0.274	ng	0.00
15) 2-Fluorobiphenyl	12.860	172	59379	0.339	ng	0.00
27) Fluoranthene-d10	19.033	212	68415	0.318	ng	0.00
31) Terphenyl-d14	19.644	244	61588	0.334	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.207	88	4062	0.314	ng	# 52
3) n-Nitrosodimethylamine	3.539	42	5083	0.311	ng	99
6) bis(2-Chloroethyl)ether	7.052	93	18701	0.337	ng	100
9) Naphthalene	10.432	128	80121	0.322	ng	100
10) Hexachlorobutadiene	10.723	225	14985	0.334	ng	# 100
12) 2-Methylnaphthalene	12.047	142	50071	0.338	ng	99
16) Acenaphthylene	13.952	152	58922	0.310	ng	100
17) Acenaphthene	14.307	154	42908	0.334	ng	100
18) Fluorene	15.296	166	50749	0.329	ng	100
20) 4,6-Dinitro-2-methylph...	15.385	198	2142	0.365	ng	98
21) 4-Bromophenyl-phenylether	16.190	248	16482	0.325	ng	97
22) Hexachlorobenzene	16.299	284	19648	0.338	ng	99
23) Atrazine	16.470	200	9877	0.281	ng	100
24) Pentachlorophenol	16.652	266	4099	0.343	ng	100
25) Phenanthrene	17.030	178	82419	0.342	ng	100
26) Anthracene	17.127	178	66074	0.319	ng	100
28) Fluoranthene	19.058	202	89523	0.339	ng	100
30) Pyrene	19.425	202	88440	0.335	ng	100
32) Benzo(a)anthracene	21.174	228	82264	0.325	ng	100
33) Chrysene	21.227	228	94148	0.350	ng	100
34) Bis(2-ethylhexyl)phtha...	21.132	149	28188	0.341	ng	100
36) Indeno(1,2,3-cd)pyrene	25.672	276	105382	0.337	ng	99
37) Benzo(b)fluoranthene	22.755	252	90056	0.321	ng	100
38) Benzo(k)fluoranthene	22.800	252	96457	0.347	ng	99
39) Benzo(a)pyrene	23.327	252	80808	0.326	ng	100
40) Dibenzo(a,h)anthracene	25.696	278	85333	0.337	ng	99
41) Benzo(g,h,i)perylene	26.360	276	92137	0.333	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101421\
Data File : BN016958.D
Acq On : 14 Oct 2021 19:27
Operator : CG/JU
Sample : PB139115BS
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB139115BS

Quant Time: Oct 15 01:44:00 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 14 17:01:23 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.615 min Scan# 492
 Delta R.T. -0.000 min
 Lab File: BN016958.D
 Acq: 14 Oct 2021 19:27




Instrument :
 BNA_N
 ClientSampleId :
 PB139115BS


Tgt Ion:152 Resp: 19989
 Ion Ratio Lower Upper
 152 100
 150 154.7 122.2 183.4
 115 58.0 46.1 69.1



#2
 1,4-Dioxane
 Concen: 0.314 ng
 RT: 3.207 min Scan# 14
 Delta R.T. -0.000 min
 Lab File: BN016958.D
 Acq: 14 Oct 2021 19:27

Tgt Ion: 88 Resp: 4062
 Ion Ratio Lower Upper
 88 100
 43 44.1 107.1 160.7#
 58 78.5 62.4 93.6









#3
n-Nitrosodimethylamine
Concen: 0.311 ng
RT: 3.539 min Scan# 50
Delta R.T. -0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:	42	Resp:	5083
Ion	Ratio	Lower	Upper
42	100		
74	144.4	114.9	172.3
44	6.1	5.4	8.2



#4
2-Fluorophenol
Concen: 0.316 ng
RT: 5.217 min Scan# 232
Delta R.T. -0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27

Tgt Ion:	112	Resp:	15713
Ion	Ratio	Lower	Upper
112	100		
64	56.1	44.7	67.1
63	28.1	22.6	33.8









#5
Phenol-d6
Concen: 0.310 ng
RT: 6.794 min Scan# 403
Delta R.T. -0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:	99	Resp:	17155
Ion	Ratio	Lower	Upper
99	100		
42	17.9	14.2	21.4
71	30.7	24.6	36.8



#6
bis(2-Chloroethyl)ether
Concen: 0.337 ng
RT: 7.052 min Scan# 431
Delta R.T. -0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27

Tgt Ion:	93	Resp:	18701
Ion	Ratio	Lower	Upper
93	100		
63	75.4	60.0	90.0
95	32.9	26.5	39.7









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.381 min Scan# 731
Delta R.T. -0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:	136	Resp:	88211
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.8	13.2
54	5.7	4.3	6.5
68	4.9	3.7	5.5



#8
Nitrobenzene-d5
Concen: 0.286 ng
RT: 8.759 min Scan# 603
Delta R.T. -0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27

Tgt Ion:	82	Resp:	16091
Ion	Ratio	Lower	Upper
82	100		
128	45.0	37.0	55.4
54	46.1	36.2	54.4









#9
Naphthalene
Concen: 0.322 ng
RT: 10.432 min Scan# 735
Delta R.T. -0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:	128	Resp:	80121
Ion	Ratio	Lower	Upper
128	100		
129	11.1	8.9	13.3
127	12.7	10.2	15.2



#10
Hexachlorobutadiene
Concen: 0.334 ng
RT: 10.723 min Scan# 758
Delta R.T. 0.013 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27

Tgt Ion:	225	Resp:	14985
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.7	51.2	76.8









#11
2-Methylnaphthalene-d10
Concen: 0.327 ng
RT: 11.976 min Scan# 925
Delta R.T. 0.004 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:152 Resp: 40864
Ion Ratio Lower Upper
152 100
151 20.4 16.3 24.5



#12
2-Methylnaphthalene
Concen: 0.338 ng
RT: 12.047 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27

Tgt Ion:142 Resp: 50071
Ion Ratio Lower Upper
142 100
141 88.9 70.2 105.4
115 33.2 26.1 39.1









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.243 min Scan# 1181
Delta R.T. -0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:	164	Resp:	46015
Ion Ratio	Lower	Upper	
164	100		
162	100.4	79.1	118.7
160	44.5	34.7	52.1



#14
2,4,6-Tribromophenol
Concen: 0.274 ng
RT: 15.740 min Scan# 1299
Delta R.T. -0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27

Tgt Ion:	330	Resp:	5669
Ion Ratio	Lower	Upper	
330	100		
332	0.0	0.0	0.0
141	30.4	24.2	36.2









#15
2-Fluorobiphenyl
Concen: 0.339 ng
RT: 12.860 min Scan# 1072
Delta R.T. -0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:	172	Resp:	59379
Ion Ratio	Lower	Upper	
172	100		
171	35.2	27.9	41.9
170	23.4	18.6	27.8



#16
Acenaphthylene
Concen: 0.310 ng
RT: 13.952 min Scan# 1158
Delta R.T. -0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27

Tgt Ion:	152	Resp:	58922
Ion Ratio	Lower	Upper	
152	100		
151	19.5	15.4	23.2
153	13.0	10.5	15.7









#17
Acenaphthene
Concen: 0.334 ng
RT: 14.307 min Scan# 1186
Delta R.T. -0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:154 Resp: 42908
Ion Ratio Lower Upper
154 100
153 113.1 90.6 135.8
152 55.0 44.2 66.2



#18
Fluorene
Concen: 0.329 ng
RT: 15.296 min Scan# 1264
Delta R.T. -0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27

Tgt Ion:166 Resp: 50749
Ion Ratio Lower Upper
166 100
165 97.6 78.4 117.6
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.993 min Scan# 1401
Delta R.T. -0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:188	Resp:	84493
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0
80	7.1	5.6



#20
4,6-Dinitro-2-methylphenol
Concen: 0.365 ng
RT: 15.385 min Scan# 1271
Delta R.T. 0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27

Tgt Ion:198	Resp:	2142
Ion Ratio	Lower	Upper
198	100	
182	15.2	11.5
77	0.0	0.0









#21
4-Bromophenyl-phenylether
Concen: 0.325 ng
RT: 16.190 min Scan# 1335
Delta R.T. -0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:248 Resp: 16482
Ion Ratio Lower Upper
248 100
250 93.7 76.1 114.1
141 62.8 46.6 70.0



#22
Hexachlorobenzene
Concen: 0.338 ng
RT: 16.299 min Scan# 1344
Delta R.T. -0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27

Tgt Ion:284 Resp: 19648
Ion Ratio Lower Upper
284 100
142 31.8 25.9 38.9
249 29.2 23.3 34.9









#23
Atrazine
Concen: 0.281 ng
RT: 16.470 min Scan# 1358
Delta R.T. -0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:200 Resp: 9877
Ion Ratio Lower Upper
200 100
173 27.2 21.7 32.5
215 48.2 38.6 57.8



#24
Pentachlorophenol
Concen: 0.343 ng
RT: 16.652 min Scan# 1373
Delta R.T. 0.012 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27

Tgt Ion:266 Resp: 4099
Ion Ratio Lower Upper
266 100
264 63.1 50.3 75.5
268 64.0 51.1 76.7









#25
Phenanthrene
Concen: 0.342 ng
RT: 17.030 min Scan# 1404
Delta R.T. -0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:	178	Resp:	82419
Ion Ratio	Lower	Upper	
178	100		
176	18.8	15.0	22.4
179	15.5	12.2	18.4



#26
Anthracene
Concen: 0.319 ng
RT: 17.127 min Scan# 1412
Delta R.T. 0.012 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27

Tgt Ion:	178	Resp:	66074
Ion Ratio	Lower	Upper	
178	100		
176	18.3	14.6	21.8
179	15.3	12.3	18.5









#27
Fluoranthene-d10
Concen: 0.318 ng
RT: 19.033 min Scan# 1690
Delta R.T. 0.005 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:212 Resp: 68415
Ion Ratio Lower Upper
212 100
106 13.6 10.9 16.3
104 7.5 6.0 9.0



#28
Fluoranthene
Concen: 0.339 ng
RT: 19.058 min Scan# 1695
Delta R.T. -0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27

Tgt Ion:202 Resp: 89523
Ion Ratio Lower Upper
202 100
101 11.5 9.1 13.7
203 17.3 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.196 min Scan# 1998
Delta R.T. -0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:240 Resp: 83001
Ion Ratio Lower Upper
240 100
120 7.5 6.5 9.7
236 26.0 20.8 31.2



#30
Pyrene
Concen: 0.335 ng
RT: 19.425 min Scan# 1769
Delta R.T. -0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27

Tgt Ion:202 Resp: 88440
Ion Ratio Lower Upper
202 100
200 20.5 16.4 24.6
203 17.4 14.0 21.0









#31
 Terphenyl-d14
 Concen: 0.334 ng
 RT: 19.644 min Scan# 1813
 Delta R.T. -0.000 min
 Lab File: BN016958.D
 Acq: 14 Oct 2021 19:27




Instrument :
 BNA_N
 ClientSampleId :
 PB139115BS


Tgt Ion:244 Resp: 61588
 Ion Ratio Lower Upper
 244 100
 212 7.2 5.6 8.4
 122 11.4 9.2 13.8



#32
 Benzo(a)anthracene
 Concen: 0.325 ng
 RT: 21.174 min Scan# 1996
 Delta R.T. -0.000 min
 Lab File: BN016958.D
 Acq: 14 Oct 2021 19:27

Tgt Ion:228 Resp: 82264
 Ion Ratio Lower Upper
 228 100
 226 27.8 21.9 32.9
 229 19.1 15.3 22.9





#33

Chrysene

Concen: 0.350 ng

RT: 21.227 min Scan# 2001

Delta R.T. -0.000 min

Lab File: BN016958.D

Acq: 14 Oct 2021 19:27

Tgt Ion:228 Resp: 94148

Ion Ratio Lower Upper

228 100

226 30.4 24.2 36.4





229 19.2 15.3 22.9

Instrument :

BNA_N

ClientSampleId :

PB139115BS



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.341 ng

RT: 21.132 min Scan# 1992

Delta R.T. -0.000 min

Lab File: BN016958.D

Acq: 14 Oct 2021 19:27




Tgt Ion:149 Resp: 28188


Ion Ratio Lower Upper

149 100

167 27.4 22.0 33.0


279 4.2 3.4 5.2








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.426 min Scan# 2410
Delta R.T. 0.003 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27

Instrument :
BNA_N
ClientSampleId :
PB139115BS



Tgt Ion:264 Resp: 84677




Ion	Ratio	Lower	Upper
264	100		
260	24.3	19.6	29.4
265	23.2	18.8	28.2




#36
Indeno(1,2,3-cd)pyrene
Concen: 0.337 ng
RT: 25.672 min Scan# 3066
Delta R.T. -0.004 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27

Tgt Ion:276 Resp: 105382

Ion	Ratio	Lower	Upper
276	100		
138	24.8	19.4	29.2
277	25.2	20.1	30.1





#37

Benzo(b)fluoranthene

Concen: 0.321 ng

RT: 22.755 min Scan# 2214

Delta R.T. -0.000 min

Lab File: BN016958.D

Acq: 14 Oct 2021 19:27





Tgt Ion:252 Resp: 90056

Ion Ratio Lower Upper

252 100

253 22.3 17.7 26.5

125 10.1 7.9 11.9



#38

Benzo(k)fluoranthene

Concen: 0.347 ng

RT: 22.800 min Scan# 2227

Delta R.T. -0.000 min

Lab File: BN016958.D

Acq: 14 Oct 2021 19:27




Tgt Ion:252 Resp: 96457


Ion Ratio Lower Upper

252 100

253 22.6 17.9 26.9

125 11.3 8.4 12.6









#39
Benzo(a)pyrene
Concen: 0.326 ng
RT: 23.327 min Scan# 2381
Delta R.T. -0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27




Instrument :
BNA_N
ClientSampleId :
PB139115BS


Tgt Ion:252 Resp: 80808
Ion Ratio Lower Upper
252 100
253 22.4 17.8 26.8
125 11.3 9.0 13.6



#40
Dibenzo(a,h)anthracene
Concen: 0.337 ng
RT: 25.696 min Scan# 3073
Delta R.T. 0.003 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27

Tgt Ion:278 Resp: 85333
Ion Ratio Lower Upper
278 100
139 16.8 13.0 19.6
279 24.1 19.1 28.7





#41
Benzo(g,h,i)perylene
Concen: 0.333 ng
RT: 26.360 min Scan# 3267
Delta R.T. -0.000 min
Lab File: BN016958.D
Acq: 14 Oct 2021 19:27

Instrument :
BNA_N
ClientSampleId :
PB139115BS

Tgt Ion:276 Resp: 92137
Ion Ratio Lower Upper
276 100
277 24.0 19.1 28.7
138 20.9 16.8 25.2

