

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101421\
 Data File : BN016965.D
 Acq On : 15 Oct 2021 00:55
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Oct 15 01:49:03 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 14 17:01:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.614	152	20709	0.400	ng	0.00
7) Naphthalene-d8	10.381	136	90594	0.400	ng	0.00
13) Acenaphthene-d10	14.243	164	46932	0.400	ng	0.00
19) Phenanthrene-d10	16.993	188	86646	0.400	ng	0.00
29) Chrysene-d12	21.195	240	87095	0.400	ng	# 0.00
35) Perylene-d12	23.419	264	90559	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.217	112	18583	0.361	ng	0.00
5) Phenol-d6	6.794	99	20537	0.359	ng	0.00
8) Nitrobenzene-d5	8.759	82	20009	0.347	ng	0.00
11) 2-Methylnaphthalene-d10	11.971	152	49320	0.385	ng	0.00
14) 2,4,6-Tribromophenol	15.740	330	7068	0.335	ng	0.00
15) 2-Fluorobiphenyl	12.860	172	71002	0.397	ng	0.00
27) Fluoranthene-d10	19.028	212	84184	0.382	ng	0.00
31) Terphenyl-d14	19.638	244	76157	0.394	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.207	88	4926	0.368	ng	# 46
3) n-Nitrosodimethylamine	3.539	42	6266	0.370	ng	100
6) bis(2-Chloroethyl)ether	7.052	93	22571	0.392	ng	99
9) Naphthalene	10.432	128	94553	0.370	ng	100
10) Hexachlorobutadiene	10.710	225	18082	0.392	ng	# 100
12) 2-Methylnaphthalene	12.047	142	60230	0.396	ng	99
16) Acenaphthylene	13.951	152	71945	0.371	ng	100
17) Acenaphthene	14.306	154	51174	0.390	ng	100
18) Fluorene	15.296	166	62057	0.395	ng	100
20) 4,6-Dinitro-2-methylph...	15.385	198	2885	0.397	ng	99
21) 4-Bromophenyl-phenylether	16.190	248	20061	0.386	ng	98
22) Hexachlorobenzene	16.299	284	23882	0.401	ng	99
23) Atrazine	16.470	200	12175	0.338	ng	99
24) Pentachlorophenol	16.640	266	6069	0.408	ng	100
25) Phenanthrene	17.029	178	100055	0.405	ng	100
26) Anthracene	17.115	178	81565	0.384	ng	100
28) Fluoranthene	19.058	202	110613	0.408	ng	100
30) Pyrene	19.420	202	110613	0.400	ng	100
32) Benzo(a)anthracene	21.174	228	103604	0.390	ng	99
33) Chrysene	21.227	228	115241	0.408	ng	98
34) Bis(2-ethylhexyl)phtha...	21.131	149	35859	0.377	ng	100
36) Indeno(1,2,3-cd)pyrene	25.664	276	135638	0.405	ng	99
37) Benzo(b)fluoranthene	22.748	252	113545	0.379	ng	100
38) Benzo(k)fluoranthene	22.793	252	119891	0.404	ng	98
39) Benzo(a)pyrene	23.320	252	101682	0.384	ng	99
40) Dibenzo(a,h)anthracene	25.688	278	110142	0.406	ng	99
41) Benzo(g,h,i)perylene	26.352	276	117158	0.396	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101421\
Data File : BN016965.D
Acq On : 15 Oct 2021 00:55
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Quant Time: Oct 15 01:49:03 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 14 17:01:23 2021
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.614 min Scan# 492
Delta R.T. -0.001 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:152 Resp: 20709
Ion Ratio Lower Upper
152 100
150 153.4 122.2 183.4
115 58.2 46.1 69.1



#2
1,4-Dioxane
Concen: 0.368 ng
RT: 3.207 min Scan# 14
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55

Tgt Ion: 88 Resp: 4926
Ion Ratio Lower Upper
88 100
43 32.5 107.1 160.7#
58 77.0 62.4 93.6









#3
n-Nitrosodimethylamine
Concen: 0.370 ng
RT: 3.539 min Scan# 50
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 42 Resp: 6266
Ion Ratio Lower Upper
42 100
74 144.0 114.9 172.3
44 7.3 5.4 8.2



#4
2-Fluorophenol
Concen: 0.361 ng
RT: 5.217 min Scan# 232
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55

Tgt Ion: 112 Resp: 18583
Ion Ratio Lower Upper
112 100
64 56.3 44.7 67.1
63 28.0 22.6 33.8









#5
Phenol-d6
Concen: 0.359 ng
RT: 6.794 min Scan# 403
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	99	Resp:	20537
Ion	Ratio	Lower	Upper
99	100		
42	17.9	14.2	21.4
71	30.8	24.6	36.8



#6
bis(2-Chloroethyl)ether
Concen: 0.392 ng
RT: 7.052 min Scan# 431
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55

Tgt Ion:	93	Resp:	22571
Ion	Ratio	Lower	Upper
93	100		
63	75.6	60.0	90.0
95	33.1	26.5	39.7









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.381 min Scan# 731
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	90594
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.8	13.2
54	5.2	4.3	6.5
68	4.7	3.7	5.5



#8
Nitrobenzene-d5
Concen: 0.347 ng
RT: 8.759 min Scan# 603
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55

Tgt Ion:	82	Resp:	20009
Ion	Ratio	Lower	Upper
82	100		
128	46.1	37.0	55.4
54	45.0	36.2	54.4









#9
Naphthalene
Concen: 0.370 ng
RT: 10.432 min Scan# 735
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:128 Resp: 94553
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 12.6 10.2 15.2



#10
Hexachlorobutadiene
Concen: 0.392 ng
RT: 10.710 min Scan# 757
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55

Tgt Ion:225 Resp: 18082
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.2 51.2 76.8









#11
2-Methylnaphthalene-d10
Concen: 0.385 ng
RT: 11.971 min Scan# 924
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 49320
Ion Ratio Lower Upper
152 100
151 20.5 16.3 24.5



#12
2-Methylnaphthalene
Concen: 0.396 ng
RT: 12.047 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55

Tgt Ion:142 Resp: 60230
Ion Ratio Lower Upper
142 100
141 88.4 70.2 105.4
115 32.9 26.1 39.1









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.243 min Scan# 1181
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 46932
Ion Ratio Lower Upper
164 100
162 98.8 79.1 118.7
160 43.2 34.7 52.1



#14
2,4,6-Tribromophenol
Concen: 0.335 ng
RT: 15.740 min Scan# 1299
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55

Tgt Ion:330 Resp: 7068
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 30.2 24.2 36.2









#15
2-Fluorobiphenyl
Concen: 0.397 ng
RT: 12.860 min Scan# 1072
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:172	Resp:	71002
Ion Ratio	Lower	Upper
172	100	
171	35.0	27.9 41.9
170	23.3	18.6 27.8



#16
Acenaphthylene
Concen: 0.371 ng
RT: 13.951 min Scan# 1158
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55

Tgt Ion:152	Resp:	71945
Ion Ratio	Lower	Upper
152	100	
151	19.6	15.4 23.2
153	13.0	10.5 15.7









#17
Acenaphthene
Concen: 0.390 ng
RT: 14.306 min Scan# 1186
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 51174
Ion Ratio Lower Upper
154 100
153 113.7 90.6 135.8
152 55.3 44.2 66.2



#18
Fluorene
Concen: 0.395 ng
RT: 15.296 min Scan# 1264
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55

Tgt Ion:166 Resp: 62057
Ion Ratio Lower Upper
166 100
165 98.2 78.4 117.6
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.993 min Scan# 1401
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188	Resp:	86646
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0
80	6.6	5.6



#20
4,6-Dinitro-2-methylphenol
Concen: 0.397 ng
RT: 15.385 min Scan# 1271
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55

Tgt Ion:198	Resp:	2885
Ion Ratio	Lower	Upper
198	100	
182	14.0	11.5
77	0.0	0.0









#21
4-Bromophenyl-phenylether
Concen: 0.386 ng
RT: 16.190 min Scan# 1335
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	248	Resp:	20061
Ion Ratio	Lower	Upper	
248	100		
250	96.2	76.1	114.1
141	55.3	46.6	70.0



#22
Hexachlorobenzene
Concen: 0.401 ng
RT: 16.299 min Scan# 1344
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55

Tgt Ion:	284	Resp:	23882
Ion Ratio	Lower	Upper	
284	100		
142	31.9	25.9	38.9
249	29.5	23.3	34.9









#23
Atrazine
Concen: 0.338 ng
RT: 16.470 min Scan# 1358
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	200	Resp:	12175
Ion Ratio	Lower	Upper	
200	100		
173	25.9	21.7	32.5
215	48.7	38.6	57.8



#24
Pentachlorophenol
Concen: 0.408 ng
RT: 16.640 min Scan# 1372
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55

Tgt Ion:	266	Resp:	6069
Ion Ratio	Lower	Upper	
266	100		
264	62.9	50.3	75.5
268	64.6	51.1	76.7









#25
Phenanthrene
Concen: 0.405 ng
RT: 17.029 min Scan# 1404
Delta R.T. -0.001 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 100055
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.4
179 15.4 12.2 18.4



#26
Anthracene
Concen: 0.384 ng
RT: 17.115 min Scan# 1411
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55

Tgt Ion:178 Resp: 81565
Ion Ratio Lower Upper
178 100
176 18.2 14.6 21.8
179 15.4 12.3 18.5









#27
Fluoranthene-d10
Concen: 0.382 ng
RT: 19.028 min Scan# 1689
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 84184
Ion Ratio Lower Upper
212 100
106 13.5 10.9 16.3
104 7.4 6.0 9.0



#28
Fluoranthene
Concen: 0.408 ng
RT: 19.058 min Scan# 1695
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55

Tgt Ion:202 Resp: 110613
Ion Ratio Lower Upper
202 100
101 11.5 9.1 13.7
203 17.3 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.195 min Scan# 1998
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:240 Resp: 87095
Ion Ratio Lower Upper
240 100
120 5.9 6.5 9.7#
236 25.2 20.8 31.2



#30
Pyrene
Concen: 0.400 ng
RT: 19.420 min Scan# 1768
Delta R.T. -0.005 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55

Tgt Ion:202 Resp: 110613
Ion Ratio Lower Upper
202 100
200 20.6 16.4 24.6
203 17.4 14.0 21.0









#31
Terphenyl-d14
Concen: 0.394 ng
RT: 19.638 min Scan# 1812
Delta R.T. -0.005 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:244 Resp: 76157
Ion Ratio Lower Upper
244 100
212 7.3 5.6 8.4
122 12.6 9.2 13.8



#32
Benzo(a)anthracene
Concen: 0.390 ng
RT: 21.174 min Scan# 1996
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55

Tgt Ion:228 Resp: 103604
Ion Ratio Lower Upper
228 100
226 27.1 21.9 32.9
229 19.5 15.3 22.9









#33
Chrysene
Concen: 0.408 ng
RT: 21.227 min Scan# 2001
Delta R.T. -0.000 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:228	Resp: 115241
Ion Ratio	Lower Upper
228 100	
226 29.0	24.2 36.4
229 19.5	15.3 22.9



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.377 ng
RT: 21.131 min Scan# 1992
Delta R.T. -0.001 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55


Tgt Ion:149	Resp: 35859
Ion Ratio	Lower Upper
149 100	
167 27.3	22.0 33.0
279 4.2	3.4 5.2








#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.419 min Scan# 2408
 Delta R.T. -0.004 min
 Lab File: BN016965.D
 Acq: 15 Oct 2021 00:55

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4



Tgt Ion:264 Resp: 90559




Ion	Ratio	Lower	Upper
264	100		
260	24.4	19.6	29.4
265	22.8	18.8	28.2




#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.405 ng
 RT: 25.664 min Scan# 3064
 Delta R.T. -0.011 min
 Lab File: BN016965.D
 Acq: 15 Oct 2021 00:55

Tgt Ion:276 Resp: 135638

Ion	Ratio	Lower	Upper
276	100		
138	25.2	19.4	29.2
277	25.3	20.1	30.1









#37
Benzo(b)fluoranthene
Concen: 0.379 ng
RT: 22.748 min Scan# 2212
Delta R.T. -0.007 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 113545
Ion Ratio Lower Upper
252 100
253 22.3 17.7 26.5
125 10.0 7.9 11.9



#38
Benzo(k)fluoranthene
Concen: 0.404 ng
RT: 22.793 min Scan# 2225
Delta R.T. -0.007 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55

Tgt Ion:252 Resp: 119891
Ion Ratio Lower Upper
252 100
253 22.5 17.9 26.9
125 12.4 8.4 12.6









#39
Benzo(a)pyrene
Concen: 0.384 ng
RT: 23.320 min Scan# 2379
Delta R.T. -0.007 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 101682
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.8
125 11.0 9.0 13.6



#40
Dibenzo(a,h)anthracene
Concen: 0.406 ng
RT: 25.688 min Scan# 3071
Delta R.T. -0.004 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55


Tgt Ion:278 Resp: 110142
Ion Ratio Lower Upper
278 100
139 17.0 13.0 19.6
279 23.9 19.1 28.7





#41
Benzo(g,h,i)perylene
Concen: 0.396 ng
RT: 26.352 min Scan# 3265
Delta R.T. -0.007 min
Lab File: BN016965.D
Acq: 15 Oct 2021 00:55

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:276 Resp: 117158
Ion Ratio Lower Upper
276 100
277 23.8 19.1 28.7
138 21.2 16.8 25.2

