

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101421\
 Data File : BN016971.D
 Acq On : 15 Oct 2021 04:32
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Oct 15 05:48:57 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 14 17:01:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.606	152	22266	0.400	ng	0.00
7) Naphthalene-d8	10.381	136	99207	0.400	ng	0.00
13) Acenaphthene-d10	14.231	164	52813	0.400	ng	-0.01
19) Phenanthrene-d10	16.981	188	97728	0.400	ng	#-0.01
29) Chrysene-d12	21.196	240	101435	0.400	ng	# 0.00
35) Perylene-d12	23.420	264	105887	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.218	112	21292	0.384	ng	0.00
5) Phenol-d6	6.794	99	23803	0.387	ng	0.00
8) Nitrobenzene-d5	8.747	82	25445	0.403	ng	-0.01
11) 2-Methylnaphthalene-d10	11.972	152	54985	0.392	ng	0.00
14) 2,4,6-Tribromophenol	15.741	330	9229	0.388	ng	0.00
15) 2-Fluorobiphenyl	12.861	172	78138	0.389	ng	0.00
27) Fluoranthene-d10	19.028	212	100880	0.406	ng	0.00
31) Terphenyl-d14	19.639	244	87908	0.391	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.207	88	4673	0.325	ng	# 45
3) n-Nitrosodimethylamine	3.530	42	7158	0.393	ng	97
6) bis(2-Chloroethyl)ether	7.052	93	24644	0.398	ng	99
9) Naphthalene	10.419	128	102639	0.367	ng	99
10) Hexachlorobutadiene	10.711	225	20070	0.398	ng	# 100
12) 2-Methylnaphthalene	12.047	142	66419	0.398	ng	100
16) Acenaphthylene	13.952	152	86001	0.394	ng	100
17) Acenaphthene	14.294	154	56875	0.385	ng	99
18) Fluorene	15.297	166	69538	0.393	ng	100
20) 4,6-Dinitro-2-methylph...	15.385	198	4061	0.429	ng	98
21) 4-Bromophenyl-phenylether	16.190	248	22930	0.391	ng	93
22) Hexachlorobenzene	16.300	284	26190	0.390	ng	99
23) Atrazine	16.470	200	16346	0.403	ng	98
24) Pentachlorophenol	16.640	266	8279	0.453	ng	100
25) Phenanthrene	17.030	178	110141	0.395	ng	100
26) Anthracene	17.115	178	97205	0.405	ng	100
28) Fluoranthene	19.058	202	127575	0.417	ng	100
30) Pyrene	19.420	202	129596	0.402	ng	100
32) Benzo(a)anthracene	21.174	228	130536	0.422	ng	99
33) Chrysene	21.228	228	132227	0.402	ng	99
34) Bis(2-ethylhexyl)phtha...	21.132	149	78338	0.555	ng	99
36) Indeno(1,2,3-cd)pyrene	25.672	276	158001	0.404	ng	99
37) Benzo(b)fluoranthene	22.752	252	136967	0.391	ng	100
38) Benzo(k)fluoranthene	22.797	252	139573	0.402	ng	98
39) Benzo(a)pyrene	23.324	252	127843	0.413	ng	100
40) Dibenzo(a,h)anthracene	25.692	278	128217	0.404	ng	99
41) Benzo(g,h,i)perylene	26.356	276	135987	0.393	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101421\
Data File : BN016971.D
Acq On : 15 Oct 2021 04:32
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Quant Time: Oct 15 05:48:57 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 14 17:01:23 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.606 min Scan# 491
Delta R.T. -0.009 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 22266
Ion Ratio Lower Upper
152 100
150 157.9 122.2 183.4
115 60.9 46.1 69.1



#2
1,4-Dioxane
Concen: 0.325 ng
RT: 3.207 min Scan# 14
Delta R.T. 0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32

Tgt Ion: 88 Resp: 4673
Ion Ratio Lower Upper
88 100
43 30.6 107.1 160.7#
58 77.4 62.4 93.6









#3
n-Nitrosodimethylamine
Concen: 0.393 ng
RT: 3.530 min Scan# 49
Delta R.T. -0.009 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 42 Resp: 7158
Ion Ratio Lower Upper
42 100
74 147.9 114.9 172.3
44 6.8 5.4 8.2



#4
2-Fluorophenol
Concen: 0.384 ng
RT: 5.218 min Scan# 232
Delta R.T. 0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32

Tgt Ion: 112 Resp: 21292
Ion Ratio Lower Upper
112 100
64 56.5 44.7 67.1
63 28.2 22.6 33.8









#5
Phenol-d6
Concen: 0.387 ng
RT: 6.794 min Scan# 403
Delta R.T. 0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	99	Resp:	23803
Ion	Ratio	Lower	Upper
99	100		
42	17.9	14.2	21.4
71	31.0	24.6	36.8



#6
bis(2-Chloroethyl)ether
Concen: 0.398 ng
RT: 7.052 min Scan# 431
Delta R.T. 0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32

Tgt Ion:	93	Resp:	24644
Ion	Ratio	Lower	Upper
93	100		
63	75.6	60.0	90.0
95	33.2	26.5	39.7









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.381 min Scan# 731
Delta R.T. 0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:136	Resp:	99207
Ion Ratio	Lower	Upper
136	100	
137	11.0	8.8 13.2
54	4.7	4.3 6.5
68	4.2	3.7 5.5



#8
Nitrobenzene-d5
Concen: 0.403 ng
RT: 8.747 min Scan# 602
Delta R.T. -0.013 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32

Tgt Ion: 82	Resp:	25445
Ion Ratio	Lower	Upper
82	100	
128	35.7	37.0 55.4#
54	53.4	36.2 54.4









#9
Naphthalene
Concen: 0.367 ng
RT: 10.419 min Scan# 734
Delta R.T. -0.013 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:128 Resp: 102639
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 13.3 10.2 15.2



#10
Hexachlorobutadiene
Concen: 0.398 ng
RT: 10.711 min Scan# 757
Delta R.T. 0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32

Tgt Ion:225 Resp: 20070
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.7 51.2 76.8









#11
2-Methylnaphthalene-d10
Concen: 0.392 ng
RT: 11.972 min Scan# 924
Delta R.T. 0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 54985
Ion Ratio Lower Upper
152 100
151 20.3 16.3 24.5



#12
2-Methylnaphthalene
Concen: 0.398 ng
RT: 12.047 min Scan# 941
Delta R.T. 0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32

Tgt Ion:142 Resp: 66419
Ion Ratio Lower Upper
142 100
141 87.5 70.2 105.4
115 32.3 26.1 39.1









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.231 min Scan# 1180
Delta R.T. -0.013 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:164	Resp:	52813
Ion Ratio	Lower	Upper
164	100	
162	104.6	79.1 118.7
160	48.6	34.7 52.1



#14
2,4,6-Tribromophenol
Concen: 0.388 ng
RT: 15.741 min Scan# 1299
Delta R.T. 0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32

Tgt Ion:330	Resp:	9229
Ion Ratio	Lower	Upper
330	100	
332	0.0	0.0 0.0
141	29.8	24.2 36.2









#15
2-Fluorobiphenyl
Concen: 0.389 ng
RT: 12.861 min Scan# 1072
Delta R.T. 0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:172 Resp: 78138
Ion Ratio Lower Upper
172 100
171 34.6 27.9 41.9
170 22.8 18.6 27.8



#16
Acenaphthylene
Concen: 0.394 ng
RT: 13.952 min Scan# 1158
Delta R.T. 0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32

Tgt Ion:152 Resp: 86001
Ion Ratio Lower Upper
152 100
151 19.5 15.4 23.2
153 13.0 10.5 15.7









#17
Acenaphthene
Concen: 0.385 ng
RT: 14.294 min Scan# 1185
Delta R.T. -0.013 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 56875
Ion Ratio Lower Upper
154 100
153 113.0 90.6 135.8
152 54.2 44.2 66.2



#18
Fluorene
Concen: 0.393 ng
RT: 15.297 min Scan# 1264
Delta R.T. 0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32

Tgt Ion:166 Resp: 69538
Ion Ratio Lower Upper
166 100
165 97.9 78.4 117.6
167 13.4 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.981 min Scan# 1400
Delta R.T. -0.012 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:188 Resp: 97728
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 14.0 5.6 8.4#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.429 ng
RT: 15.385 min Scan# 1271
Delta R.T. 0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32

Tgt Ion:198 Resp: 4061
Ion Ratio Lower Upper
198 100
182 13.7 11.5 17.3
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.391 ng
RT: 16.190 min Scan# 1335
Delta R.T. 0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:248 Resp: 22930
Ion Ratio Lower Upper
248 100
250 97.9 76.1 114.1
141 48.3 46.6 70.0



#22
Hexachlorobenzene
Concen: 0.390 ng
RT: 16.300 min Scan# 1344
Delta R.T. 0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32

Tgt Ion:284 Resp: 26190
Ion Ratio Lower Upper
284 100
142 32.0 25.9 38.9
249 29.6 23.3 34.9









#23
Atrazine
Concen: 0.403 ng
RT: 16.470 min Scan# 1358
Delta R.T. 0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:200 Resp: 16346
Ion Ratio Lower Upper
200 100
173 24.9 21.7 32.5
215 48.9 38.6 57.8



#24
Pentachlorophenol
Concen: 0.453 ng
RT: 16.640 min Scan# 1372
Delta R.T. 0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32

Tgt Ion:266 Resp: 8279
Ion Ratio Lower Upper
266 100
264 62.9 50.3 75.5
268 64.4 51.1 76.7









#25
Phenanthrene
Concen: 0.395 ng
RT: 17.030 min Scan# 1404
Delta R.T. -0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:178 Resp: 110141
Ion Ratio Lower Upper
178 100
176 18.9 15.0 22.4
179 15.6 12.2 18.4



#26
Anthracene
Concen: 0.405 ng
RT: 17.115 min Scan# 1411
Delta R.T. 0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32

Tgt Ion:178 Resp: 97205
Ion Ratio Lower Upper
178 100
176 18.2 14.6 21.8
179 15.2 12.3 18.5









#27
Fluoranthene-d10
Concen: 0.406 ng
RT: 19.028 min Scan# 1689
Delta R.T. 0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 100880
Ion Ratio Lower Upper
212 100
106 13.6 10.9 16.3
104 7.4 6.0 9.0



#28
Fluoranthene
Concen: 0.417 ng
RT: 19.058 min Scan# 1695
Delta R.T. 0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32

Tgt Ion:202 Resp: 127575
Ion Ratio Lower Upper
202 100
101 11.5 9.1 13.7
203 17.3 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.196 min Scan# 1998
Delta R.T. 0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:240 Resp: 101435
Ion Ratio Lower Upper
240 100
120 6.3 6.5 9.7#
236 25.1 20.8 31.2



#30
Pyrene
Concen: 0.402 ng
RT: 19.420 min Scan# 1768
Delta R.T. -0.005 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32

Tgt Ion:202 Resp: 129596
Ion Ratio Lower Upper
202 100
200 20.5 16.4 24.6
203 17.5 14.0 21.0









#31
 Terphenyl-d14
 Concen: 0.391 ng
 RT: 19.639 min Scan# 1812
 Delta R.T. -0.005 min
 Lab File: BN016971.D
 Acq: 15 Oct 2021 04:32




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:244	Resp:	87908
Ion Ratio	Lower	Upper
244	100	
212	7.1	5.6 8.4
122	12.9	9.2 13.8



#32
 Benzo(a)anthracene
 Concen: 0.422 ng
 RT: 21.174 min Scan# 1996
 Delta R.T. 0.000 min
 Lab File: BN016971.D
 Acq: 15 Oct 2021 04:32

Tgt Ion:228	Resp:	130536
Ion Ratio	Lower	Upper
228	100	
226	26.9	21.9 32.9
229	19.3	15.3 22.9









#33
Chrysene
Concen: 0.402 ng
RT: 21.228 min Scan# 2001
Delta R.T. 0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	228	Resp:	132227
Ion Ratio	Lower	Upper	
228	100		
226	29.7	24.2	36.4
229	19.3	15.3	22.9



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.555 ng
RT: 21.132 min Scan# 1992
Delta R.T. -0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32

Tgt Ion:	149	Resp:	78338
Ion Ratio	Lower	Upper	
149	100		
167	27.8	22.0	33.0
279	4.4	3.4	5.2









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.420 min Scan# 2408
 Delta R.T. -0.003 min
 Lab File: BN016971.D
 Acq: 15 Oct 2021 04:32




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:264 Resp: 105887
 Ion Ratio Lower Upper
 264 100
 260 24.7 19.6 29.4
 265 23.0 18.8 28.2



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.404 ng
 RT: 25.672 min Scan# 3066
 Delta R.T. -0.003 min
 Lab File: BN016971.D
 Acq: 15 Oct 2021 04:32

Tgt Ion:276 Resp: 158001
 Ion Ratio Lower Upper
 276 100
 138 25.3 19.4 29.2
 277 25.2 20.1 30.1









#37
Benzo(b)fluoranthene
Concen: 0.391 ng
RT: 22.752 min Scan# 2213
Delta R.T. -0.003 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 136967
Ion Ratio Lower Upper
252 100
253 22.0 17.7 26.5
125 10.2 7.9 11.9



#38
Benzo(k)fluoranthene
Concen: 0.402 ng
RT: 22.797 min Scan# 2226
Delta R.T. -0.003 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32

Tgt Ion:252 Resp: 139573
Ion Ratio Lower Upper
252 100
253 22.6 17.9 26.9
125 12.4 8.4 12.6









#39
Benzo(a)pyrene
Concen: 0.413 ng
RT: 23.324 min Scan# 2380
Delta R.T. -0.003 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 127843
Ion Ratio Lower Upper
252 100
253 22.3 17.8 26.8
125 11.4 9.0 13.6



#40
Dibenzo(a,h)anthracene
Concen: 0.404 ng
RT: 25.692 min Scan# 3072
Delta R.T. 0.000 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32


Tgt Ion:278 Resp: 128217
Ion Ratio Lower Upper
278 100
139 17.0 13.0 19.6
279 23.8 19.1 28.7





#41
Benzo(g,h,i)perylene
Concen: 0.393 ng
RT: 26.356 min Scan# 3266
Delta R.T. -0.003 min
Lab File: BN016971.D
Acq: 15 Oct 2021 04:32

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:276 Resp: 135987
Ion Ratio Lower Upper
276 100
277 23.6 19.1 28.7
138 21.1 16.8 25.2

