

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101421\
 Data File : BN016970.D
 Acq On : 15 Oct 2021 03:56
 Operator : CG/JU
 Sample : M4190-01
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 P287-DITCH

Manual Integrations
 APPROVED

mohammad
 10/18/2021 8:47:24 AM

Quant Time: Oct 15 04:26:23 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 14 17:01:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.615	152	20169	0.40	ng	0.00
7) Naphthalene-d8	10.381	136	91988	0.40	ng	0.00
13) Acenaphthene-d10	14.231	164	50737	0.40	ng	-0.01
19) Phenanthrene-d10	16.981	188	96547	0.40	ng	#-0.01
29) Chrysene-d12	21.196	240	93321	0.40	ng	0.00
35) Perylene-d12	23.423	264	96762	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.236	112	7909	0.16	ng	0.02
5) Phenol-d6	6.803	99	7105	0.13	ng	0.00
8) Nitrobenzene-d5	8.747	82	17268	0.29	ng	-0.01
11) 2-Methylnaphthalene-d10	11.972	152	42755	0.33	ng	0.00
14) 2,4,6-Tribromophenol	15.741	330	10009	0.44	ng	0.00
15) 2-Fluorobiphenyl	12.861	172	58913	0.31	ng	0.00
27) Fluoranthene-d10	19.028	212	104466	0.43	ng	0.00
31) Terphenyl-d14	19.639	244	91352	0.44	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.217	88	391m	0.03	ng	Qvalue
24) Pentachlorophenol	16.640	266	381	0.21	ng	95
34) Bis(2-ethylhexyl)phtha...	21.132	149	33981	0.35	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101421\
Data File : BN016970.D
Acq On : 15 Oct 2021 03:56
Operator : CG/JU
Sample : M4190-01
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
P287-DITCH

Manual Integrations
APPROVED

mohammad
10/18/2021 8:47:24 AM

Quant Time: Oct 15 04:26:23 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 14 17:01:23 2021
Response via : Initial Calibration

