


4,6-Dinitro-2-methylphenol

Response Ratio



$$\text{Response} = 1.018\text{e-}001 * \text{Amt} - 6.759\text{e-}002$$


Coef of Det (r^2) = 0.991369 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN101421.M

Calibration Table Last Updated: Thu Oct 14 17:01:23 2021

Pentachlorophenol

Response Ratio



$$\text{Response} = 1.314\text{e-}001 * \text{Amt} - 6.408\text{e-}002$$


Coef of Det (r^2) = 0.996494 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN101421.M

Calibration Table Last Updated: Thu Oct 14 17:01:23 2021

Bis(2-ethylhexyl)phthalate

Response Ratio



$$\text{Response} = 8.088\text{e-}001 * \text{Amt} - 3.496\text{e-}001$$

Coef of Det (r^2) = 0.997265 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN101421.M

Calibration Table Last Updated: Thu Oct 14 17:01:23 2021