

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101421\
 Data File : BN016954.D
 Acq On : 14 Oct 2021 16:59
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN101421

Quant Time: Oct 14 17:37:08 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 14 17:01:23 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	110	0.00
2	1,4-Dioxane	0.400	0.366	8.5	104	0.00
3	n-Nitrosodimethylamine	0.400	0.372	7.0	109	0.00
4 S	2-Fluorophenol	0.400	0.383	4.3	111	0.00
5 S	Phenol-d6	0.400	0.381	4.8	111	0.00
6	bis(2-Chloroethyl)ether	0.400	0.396	1.0	110	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	110	0.00
8 S	Nitrobenzene-d5	0.400	0.354	11.5	108	0.00
9	Naphthalene	0.400	0.387	3.3	108	0.00
10	Hexachlorobutadiene	0.400	0.395	1.3	111	0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.383	4.3	109	0.00
12	2-Methylnaphthalene	0.400	0.395	1.3	109	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	108	0.00
14 S	2,4,6-Tribromophenol	0.400	0.336	16.0	106	0.00
15 S	2-Fluorobiphenyl	0.400	0.397	0.8	108	0.00
16	Acenaphthylene	0.400	0.384	4.0	110	0.00
17	Acenaphthene	0.400	0.394	1.5	108	0.00
18	Fluorene	0.400	0.386	3.5	107	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	106	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.400	0.0	99	0.00
21	4-Bromophenyl-phenylether	0.400	0.385	3.8	107	0.00
22	Hexachlorobenzene	0.400	0.402	-0.5	108	0.00
23	Atrazine	0.400	0.354	11.5	112	0.00
24	Pentachlorophenol	0.400	0.382	4.5	102	0.01
25	Phenanthrene	0.400	0.397	0.8	106	0.00
26	Anthracene	0.400	0.386	3.5	108	0.01
27 SURR	Fluoranthene-d10	0.400	0.380	5.0	106	0.00
28	Fluoranthene	0.400	0.399	0.3	106	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	105	0.00
30	Pyrene	0.400	0.405	-1.3	107	0.00
31 S	Terphenyl-d14	0.400	0.401	-0.3	107	0.00
32	Benzo(a)anthracene	0.400	0.390	2.5	106	0.01
33	Chrysene	0.400	0.418	-4.5	107	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.394	1.5	111	0.00
35 I	Perylene-d12	0.400	0.400	0.0	112	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.403	-0.8	115	0.00
37	Benzo(b)fluoranthene	0.400	0.389	2.8	111	0.00
38	Benzo(k)fluoranthene	0.400	0.401	-0.3	111	0.00
39 C	Benzo(a)pyrene	0.400	0.390	2.5	111	0.00
40	Dibenzo(a,h)anthracene	0.400	0.403	-0.8	115	0.00
41	Benzo(g,h,i)perylene	0.400	0.393	1.8	112	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN101421