



Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN101421\  
 Data File : BN016964.D  
 Acq On : 15 Oct 2021 00:19  
 Operator : CG/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN101421.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Thu Oct 14 17:01:23 2021






AutoFind: Scans 2438, 2439, 2440; Background Corrected with Scan 2431

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 32.4         | 66155      | PASS                |
| 68             | 69              | 0.00            | 2               | 1.8          | 1229       | PASS                |
| 69             | 198             | 0.00            | 100             | 34.3         | 69912      | PASS                |
| 70             | 69              | 0.00            | 2               | 0.4          | 298        | PASS                |
| 127            | 198             | 10              | 80              | 45.7         | 93173      | PASS                |
| 197            | 198             | 0.00            | 2               | 0.6          | 1162       | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 203989     | PASS                |
| 199            | 198             | 5               | 9               | 6.9          | 14135      | PASS                |
| 275            | 198             | 10              | 60              | 29.5         | 60224      | PASS                |
| 365            | 198             | 1               | 100             | 3.9          | 7978       | PASS                |
| 441            | 198             | 0.01            | 100             | 20.4         | 41632      | PASS                |
| 442            | 442             | 50              | 100             | 100.0        | 273344     | PASS                |
| 443            | 442             | 15              | 24              | 19.0         | 52048      | PASS                |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN101421\  
Data File : BN016964.D  
Acq On : 15 Oct 2021 00:19  
Operator : CG/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Quant Time: Oct 15 04:45:06 2021  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Oct 13 17:23:48 2021  
Response via : Initial Calibration



TIC: BN016964.D\data.ms

**(70) Pentachlorophenol (C)**

**16.616min (-0.006) 28185.84 ng**




**response 269416**

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 62.20  | 62.10  |
| 264.00 | 61.60  | 64.04  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN101421\  
Data File : BN016964.D  
Acq On : 15 Oct 2021 00:19  
Operator : CG/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Quant Time: Oct 15 04:45:06 2021  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Oct 13 17:23:48 2021  
Response via : Initial Calibration



TIC: BN016964.D\data.ms

**(77) Benzidine**

19.239min (-0.006) 0.00 ng

response 1618115

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.50  | 14.41  |
| 183.00 | 13.20  | 12.49  |
| 0.00   | 0.00   | 0.00   |

DDT Breakdown

| Date          | Instrument Name  | DFTPP Data File    |
|---------------|------------------|--------------------|
| 10/14/2021    | BNA_N            | <u>BN016964.D</u>  |
|               |                  |                    |
|               |                  |                    |
| Compound Name | Response         | Retention Time     |
|               |                  |                    |
| DDT           | 795593           | 20.48              |
| DDD           | 3590             | 20.086             |
| DDE           | 0                | 19.58              |
|               |                  |                    |
|               |                  |                    |
| SUM(DDD+DDE)  | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
|               |                  |                    |
| 3590          | 799183           | 0.45               |
|               |                  |                    |