

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101424\
 Data File : BN034570.D
 Acq On : 14 Oct 2024 14:24
 Operator : RC/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Quant Time: Oct 14 17:37:54 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN101424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 14 16:58:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.358	152	4866	0.400	ng	0.00
7) Naphthalene-d8	10.116	136	14277	0.400	ng	0.00
13) Acenaphthene-d10	14.015	164	8368	0.400	ng	0.00
19) Phenanthrene-d10	16.778	188	17846	0.400	ng	0.00
29) Chrysene-d12	21.006	240	6245	0.400	ng	0.00
35) Perylene-d12	23.108	264	10676	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.004	112	11032	0.836	ng	0.00
5) Phenol-d6	6.564	99	11875	0.869	ng	0.00
8) Nitrobenzene-d5	8.504	82	8061	0.866	ng	-0.01
11) 2-Methylnaphthalene-d10	11.720	152	18424	0.861	ng	0.00
14) 2,4,6-Tribromophenol	15.524	330	3104	0.837	ng	0.00
15) 2-Fluorobiphenyl	12.626	172	24654	0.854	ng	-0.01
27) Fluoranthene-d10	18.822	212	34537	0.796	ng	0.00
31) Terphenyl-d14	19.445	244	22428	0.686	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.018	88	4726	0.805	ng	95
3) n-Nitrosodimethylamine	3.321	42	4605	0.861	ng	99
6) bis(2-Chloroethyl)ether	6.809	93	9885	0.884	ng	97
9) Naphthalene	10.159	128	31656	0.848	ng	100
10) Hexachlorobutadiene	10.458	225	6820	0.837	ng	# 100
12) 2-Methylnaphthalene	11.795	142	20772	0.863	ng	99
16) Acenaphthylene	13.727	152	28966	0.851	ng	100
17) Acenaphthene	14.079	154	20727	0.860	ng	100
18) Fluorene	15.074	166	26989	0.877	ng	100
20) 4,6-Dinitro-2-methylph...	16.269	198	197	0.803	ng	# 69
21) 4-Bromophenyl-phenylether	15.984	248	8112	0.837	ng	99
22) Hexachlorobenzene	16.083	284	10712	0.819	ng	99
23) Atrazine	16.269	200	6518	0.845	ng	95
24) Pentachlorophenol	16.443	266	2818	0.753	ng	97
25) Phenanthrene	16.815	178	39349	0.836	ng	100
26) Anthracene	16.902	178	33715	0.825	ng	99
28) Fluoranthene	18.850	202	43086	0.803	ng	99
30) Pyrene	19.217	202	42900	0.700	ng	100
33) Chrysene	21.033	228	38238	0.683	ng	100
36) Indeno(1,2,3-cd)pyrene	25.160	276	35734	0.862	ng	100
37) Benzo(b)fluoranthene	22.488	252	30588	0.857	ng	94
38) Benzo(k)fluoranthene	22.529	252	34707	0.854	ng	# 94
39) Benzo(a)pyrene	23.017	252	25751	0.831	ng	# 91
40) Dibenzo(a,h)anthracene	25.178	278	27443	0.865	ng	95
41) Benzo(g,h,i)perylene	25.783	276	32710	0.862	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101424\
Data File : BN034570.D
Acq On : 14 Oct 2024 14:24
Operator : RC/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Quant Time: Oct 14 17:37:54 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN101424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 14 16:58:38 2024
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.358 min Scan# 61
Delta R.T. -0.007 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8







Tgt Ion:152 Resp: 4866
Ion Ratio Lower Upper
152 100
150 155.4 122.0 183.0
115 63.0 48.3 72.5



#2
1,4-Dioxane
Concen: 0.805 ng
RT: 3.018 min Scan# 19
Delta R.T. -0.007 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24

Tgt Ion: 88 Resp: 4726
Ion Ratio Lower Upper
88 100
43 29.4 22.8 34.2
58 75.3 56.4 84.6









#3
n-Nitrosodimethylamine
Concen: 0.861 ng
RT: 3.321 min Scan# 61
Delta R.T. -0.007 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion: 42 Resp: 4605
Ion Ratio Lower Upper
42 100
74 124.7 99.0 148.6
44 8.3 6.6 9.8



#4
2-Fluorophenol
Concen: 0.836 ng
RT: 5.004 min Scan# 294
Delta R.T. 0.000 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24

Tgt Ion: 112 Resp: 11032
Ion Ratio Lower Upper
112 100
64 55.7 44.6 66.8
63 29.6 23.8 35.6









#5
Phenol-d6
Concen: 0.869 ng
RT: 6.564 min Scan# 510
Delta R.T. 0.000 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	99	Resp:	11875
Ion	Ratio	Lower	Upper
99	100		
42	19.1	15.0	22.6
71	34.7	28.0	42.0



#6
bis(2-Chloroethyl)ether
Concen: 0.884 ng
RT: 6.809 min Scan# 544
Delta R.T. 0.000 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24

Tgt Ion:	93	Resp:	9885
Ion	Ratio	Lower	Upper
93	100		
63	78.0	64.6	96.8
95	34.0	27.9	41.9









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.116 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:136	Resp:	14277
Ion Ratio	Lower	Upper
136	100	
137	11.5	9.4 14.0
54	6.8	6.2 9.4
68	5.1	4.6 6.8



#8
Nitrobenzene-d5
Concen: 0.866 ng
RT: 8.504 min Scan# 770
Delta R.T. -0.011 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24

Tgt Ion: 82	Resp:	8061
Ion Ratio	Lower	Upper
82	100	
128	42.6	38.0 57.0
54	52.5	40.6 60.8









#9
Naphthalene
Concen: 0.848 ng
RT: 10.159 min Scan# 91
Delta R.T. -0.011 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:128 Resp: 31656
Ion Ratio Lower Upper
128 100
129 11.3 9.3 13.9
127 13.5 10.8 16.2



#10
Hexachlorobutadiene
Concen: 0.837 ng
RT: 10.458 min Scan# 953
Delta R.T. 0.000 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24

Tgt Ion:225 Resp: 6820
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.4 51.4 77.0









#11
2-Methylnaphthalene-d10
Concen: 0.861 ng
RT: 11.720 min Scan# 1108
Delta R.T. -0.004 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:152 Resp: 18424
Ion Ratio Lower Upper
152 100
151 20.8 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.863 ng
RT: 11.795 min Scan# 1128
Delta R.T. -0.004 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24

Tgt Ion:142 Resp: 20772
Ion Ratio Lower Upper
142 100
141 88.7 70.6 105.8
115 35.5 29.2 43.8









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.015 min Scan# 1475
Delta R.T. -0.004 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:164 Resp: 8368
Ion Ratio Lower Upper
164 100
162 101.0 79.7 119.5
160 48.4 37.9 56.9



#14
2,4,6-Tribromophenol
Concen: 0.837 ng
RT: 15.524 min Scan# 1615
Delta R.T. -0.005 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24

Tgt Ion:330 Resp: 3104
Ion Ratio Lower Upper
330 100
332 106.6 104.3 156.5
141 35.0 29.4 44.0





#15

2-Fluorobiphenyl

Concen: 0.854 ng

RT: 12.626 min Scan# 1444

Delta R.T. -0.014 min

Lab File: BN034570.D

Acq: 14 Oct 2024 14:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:172 Resp: 24654

Ion Ratio Lower Upper

172 100

171 35.0 27.1 40.7

170 23.8 18.5 27.7



#16

Acenaphthylene

Concen: 0.851 ng

RT: 13.727 min Scan# 1447

Delta R.T. -0.004 min

Lab File: BN034570.D

Acq: 14 Oct 2024 14:24




Tgt Ion:152 Resp: 28966


Ion Ratio Lower Upper

152 100

151 19.5 15.5 23.3

153 12.9 10.4 15.6





#17

Acenaphthene

Concen: 0.860 ng

RT: 14.079 min Scan# 1480

Delta R.T. 0.007 min

Lab File: BN034570.D

Acq: 14 Oct 2024 14:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:154 Resp: 20727

Ion Ratio Lower Upper

154 100

153 113.9 91.5 137.3

152 56.8 45.8 68.8



#18

Fluorene

Concen: 0.877 ng

RT: 15.074 min Scan# 1573

Delta R.T. -0.004 min

Lab File: BN034570.D

Acq: 14 Oct 2024 14:24




Tgt Ion:166 Resp: 26989


Ion Ratio Lower Upper

166 100

165 98.0 78.6 118.0

167 13.3 10.8 16.2









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.778 min Scan# 1716
Delta R.T. -0.005 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:188 Resp: 17846
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.4 6.3 9.5



#20
4,6-Dinitro-2-methylphenol
Concen: 0.803 ng
RT: 16.269 min Scan# 1675
Delta R.T. -0.006 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24

Tgt Ion:198 Resp: 197
Ion Ratio Lower Upper
198 100
51 130.4 174.8 262.2#
105 220.3 167.9 251.9









#21
4-Bromophenyl-phenylether
Concen: 0.837 ng
RT: 15.984 min Scan# 1652
Delta R.T. -0.005 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:248 Resp: 8112
Ion Ratio Lower Upper
248 100
250 99.6 79.8 119.6
141 53.4 44.3 66.5



#22
Hexachlorobenzene
Concen: 0.819 ng
RT: 16.083 min Scan# 1660
Delta R.T. -0.005 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24

Tgt Ion:284 Resp: 10712
Ion Ratio Lower Upper
284 100
142 34.9 28.4 42.6
249 31.1 25.0 37.4









#23
Atrazine
Concen: 0.845 ng
RT: 16.269 min Scan# 1689
Delta R.T. -0.018 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:200 Resp: 6518
Ion Ratio Lower Upper
200 100
173 30.1 22.9 34.3
215 44.2 38.8 58.2



#24
Pentachlorophenol
Concen: 0.753 ng
RT: 16.443 min Scan# 1689
Delta R.T. -0.018 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24

Tgt Ion:266 Resp: 2818
Ion Ratio Lower Upper
266 100
264 62.2 46.6 69.8
268 64.8 52.0 78.0









#25
Phenanthrene
Concen: 0.836 ng
RT: 16.815 min Scan# 1720
Delta R.T. -0.005 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:178 Resp: 39349
Ion Ratio Lower Upper
178 100
176 19.0 15.3 22.9
179 15.2 12.2 18.4



#26
Anthracene
Concen: 0.825 ng
RT: 16.902 min Scan# 1726
Delta R.T. -0.005 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24

Tgt Ion:178 Resp: 33715
Ion Ratio Lower Upper
178 100
176 18.3 14.3 21.5
179 15.3 12.0 18.0









#27
Fluoranthene-d10
Concen: 0.796 ng
RT: 18.822 min Scan# 1922
Delta R.T. -0.002 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:212 Resp: 34537
Ion Ratio Lower Upper
212 100
106 14.1 11.1 16.7
104 7.9 6.5 9.7



#28
Fluoranthene
Concen: 0.803 ng
RT: 18.850 min Scan# 1928
Delta R.T. -0.002 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24

Tgt Ion:202 Resp: 43086
Ion Ratio Lower Upper
202 100
101 11.0 8.8 13.2
203 17.2 13.4 20.0





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.006 min Scan# 24

Delta R.T. -0.002 min

Lab File: BN034570.D

Acq: 14 Oct 2024 14:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:240 Resp: 6245

Ion Ratio Lower Upper

240 100

120 11.1 10.1 15.1

236 27.8 22.6 34.0



#30

Pyrene

Concen: 0.700 ng

RT: 19.217 min Scan# 2007

Delta R.T. -0.002 min

Lab File: BN034570.D

Acq: 14 Oct 2024 14:24




Tgt Ion:202 Resp: 42900


Ion Ratio Lower Upper

202 100

200 20.8 16.6 24.8

203 17.8 14.3 21.5





#31

Terphenyl-d14

Concen: 0.686 ng

RT: 19.445 min Scan# 2056

Delta R.T. -0.007 min

Lab File: BN034570.D

Acq: 14 Oct 2024 14:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:244 Resp: 22428

Ion Ratio Lower Upper

244 100

212 9.0 7.7 11.5

122 15.5 12.6 19.0



#33

Chrysene

Concen: 0.683 ng

RT: 21.033 min Scan# 2395

Delta R.T. -0.002 min

Lab File: BN034570.D

Acq: 14 Oct 2024 14:24




Tgt Ion:228 Resp: 38238


Ion Ratio Lower Upper

228 100

226 30.5 24.4 36.6

229 19.6 15.8 23.6





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.108 min Scan# 21

Delta R.T. -0.002 min

Lab File: BN034570.D

Acq: 14 Oct 2024 14:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:264 Resp: 10676

Ion Ratio Lower Upper

264 100

260 26.1 20.9 31.3

265 73.9 61.6 92.4



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.862 ng

RT: 25.160 min Scan# 3491

Delta R.T. -0.005 min

Lab File: BN034570.D

Acq: 14 Oct 2024 14:24




Tgt Ion:276 Resp: 35734


Ion Ratio Lower Upper

276 100

138 27.6 22.0 33.0

277 24.9 20.1 30.1





#37

Benzo(b)fluoranthene

Concen: 0.857 ng

RT: 22.488 min Scan# 2577

Delta R.T. -0.002 min

Lab File: BN034570.D

Acq: 14 Oct 2024 14:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:252 Resp: 30588

Ion Ratio Lower Upper

252 100

253 23.9 21.3 31.9

125 13.1 12.7 19.1



#38

Benzo(k)fluoranthene

Concen: 0.854 ng

RT: 22.529 min Scan# 2591

Delta R.T. -0.002 min

Lab File: BN034570.D

Acq: 14 Oct 2024 14:24




Tgt Ion:252 Resp: 34707


Ion Ratio Lower Upper

252 100

253 23.8 21.0 31.6

125 13.3 13.3 19.9#





#39

Benzo(a)pyrene

Concen: 0.831 ng

RT: 23.017 min Scan# 21

Delta R.T. -0.002 min

Lab File: BN034570.D


Acq: 14 Oct 2024 14:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8






Tgt Ion:252 Resp: 25751

Ion Ratio Lower Upper

252 100

253 25.4 23.8 35.8

125 16.2 17.2 25.8#



#40

Dibenzo(a,h)anthracene

Concen: 0.865 ng

RT: 25.178 min Scan# 3497

Delta R.T. -0.002 min

Lab File: BN034570.D

Acq: 14 Oct 2024 14:24




Tgt Ion:278 Resp: 27443


Ion Ratio Lower Upper

278 100

139 19.4 16.0 24.0


279 25.1 23.4 35.0





#41
Benzo(g,h,i)perylene
Concen: 0.862 ng
RT: 25.783 min Scan# 31
Delta R.T. -0.005 min
Lab File: BN034570.D
Acq: 14 Oct 2024 14:24

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8



Tgt Ion:	276	Resp:	32710
Ion	Ratio	Lower	Upper
276	100		
277	23.8	19.8	29.8
138	23.9	19.2	28.8

