

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101424\
 Data File : BN034566.D
 Acq On : 14 Oct 2024 11:55
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Oct 14 23:50:19 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN101424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 14 17:52:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.365	152	3643	0.400	ng	0.00
7) Naphthalene-d8	10.116	136	10140	0.400	ng	0.00
13) Acenaphthene-d10	14.019	164	5771	0.400	ng	0.00
19) Phenanthrene-d10	16.783	188	10982	0.400	ng	0.00
29) Chrysene-d12	21.008	240	5201	0.400	ng	0.00
35) Perylene-d12	23.104	264	10155	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.004	112	3838	0.389	ng	0.00
5) Phenol-d6	6.571	99	3753	0.367	ng	0.00
8) Nitrobenzene-d5	8.514	82	2431	0.368	ng	0.00
11) 2-Methylnaphthalene-d10	11.723	152	5767	0.379	ng	0.00
14) 2,4,6-Tribromophenol	15.530	330	805	0.315	ng	0.00
15) 2-Fluorobiphenyl	12.640	172	7602	0.382	ng	0.00
27) Fluoranthene-d10	18.825	212	11405	0.427	ng	0.00
31) Terphenyl-d14	19.447	244	7379	0.392	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.018	88	1813	0.413	ng	98
3) n-Nitrosodimethylamine	3.321	42	1561	0.390	ng	# 97
6) bis(2-Chloroethyl)ether	6.809	93	3059	0.365	ng	98
9) Naphthalene	10.169	128	10330	0.389	ng	99
10) Hexachlorobutadiene	10.458	225	2318	0.401	ng	# 99
12) 2-Methylnaphthalene	11.803	142	6493	0.380	ng	99
16) Acenaphthylene	13.730	152	8452	0.360	ng	99
17) Acenaphthene	14.083	154	6335	0.381	ng	99
18) Fluorene	15.077	166	7928	0.373	ng	100
20) 4,6-Dinitro-2-methylph...	16.287	198	50	0.331	ng	# 81
21) 4-Bromophenyl-phenylether	15.989	248	2279	0.382	ng	94
22) Hexachlorobenzene	16.088	284	3344	0.416	ng	99
23) Atrazine	16.287	200	1761	0.371	ng	98
24) Pentachlorophenol	16.461	266	804	0.349	ng	96
25) Phenanthrene	16.821	178	11655	0.402	ng	100
26) Anthracene	16.920	178	9762	0.388	ng	99
28) Fluoranthene	18.857	202	13910	0.421	ng	99
30) Pyrene	19.220	202	14398	0.282	ng	100
33) Chrysene	21.035	228	16159	0.347	ng	99
36) Indeno(1,2,3-cd)pyrene	25.159	276	15790	0.400	ng	99
37) Benzo(b)fluoranthene	22.484	252	12873	0.379	ng	99
38) Benzo(k)fluoranthene	22.528	252	14909	0.386	ng	99
39) Benzo(a)pyrene	23.013	252	11432	0.388	ng	99
40) Dibenzo(a,h)anthracene	25.177	278	12018	0.398	ng	100
41) Benzo(g,h,i)perylene	25.779	276	14764	0.409	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101424\
Data File : BN034566.D
Acq On : 14 Oct 2024 11:55
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Oct 14 23:50:19 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN101424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 14 17:52:02 2024
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.365 min Scan# 61
 Delta R.T. 0.000 min
 Lab File: BN034566.D
 Acq: 14 Oct 2024 11:55




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:152 Resp: 3643
 Ion Ratio Lower Upper
 152 100
 150 150.1 122.0 183.0
 115 61.0 48.3 72.5



#2
 1,4-Dioxane
 Concen: 0.413 ng
 RT: 3.018 min Scan# 19
 Delta R.T. -0.007 min
 Lab File: BN034566.D
 Acq: 14 Oct 2024 11:55

Tgt Ion: 88 Resp: 1813
 Ion Ratio Lower Upper
 88 100
 43 30.9 22.8 34.2
 58 70.3 56.4 84.6









#3
 n-Nitrosodimethylamine
 Concen: 0.390 ng
 RT: 3.321 min Scan# 61
 Delta R.T. -0.007 min
 Lab File: BN034566.D
 Acq: 14 Oct 2024 11:55




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion: 42 Resp: 1561
 Ion Ratio Lower Upper
 42 100
 74 126.1 99.0 148.6
 44 11.9 6.6 9.8#



#4
 2-Fluorophenol
 Concen: 0.389 ng
 RT: 5.004 min Scan# 294
 Delta R.T. -0.000 min
 Lab File: BN034566.D
 Acq: 14 Oct 2024 11:55

Tgt Ion: 112 Resp: 3838
 Ion Ratio Lower Upper
 112 100
 64 56.1 44.6 66.8
 63 29.5 23.8 35.6









#5
Phenol-d6
Concen: 0.367 ng
RT: 6.571 min Scan# 511
Delta R.T. 0.007 min
Lab File: BN034566.D
Acq: 14 Oct 2024 11:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 99 Resp: 3753
Ion Ratio Lower Upper
99 100
42 19.5 15.0 22.6
71 34.2 28.0 42.0



#6
bis(2-Chloroethyl)ether
Concen: 0.365 ng
RT: 6.809 min Scan# 544
Delta R.T. -0.000 min
Lab File: BN034566.D
Acq: 14 Oct 2024 11:55

Tgt Ion: 93 Resp: 3059
Ion Ratio Lower Upper
93 100
63 83.1 64.6 96.8
95 36.0 27.9 41.9









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.116 min Scan# 911
Delta R.T. -0.000 min
Lab File: BN034566.D
Acq: 14 Oct 2024 11:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	10140
Ion Ratio	Lower	Upper	
136	100		
137	11.7	9.4	14.0
54	8.6	6.2	9.4
68	6.5	4.6	6.8



#8
Nitrobenzene-d5
Concen: 0.368 ng
RT: 8.514 min Scan# 771
Delta R.T. -0.000 min
Lab File: BN034566.D
Acq: 14 Oct 2024 11:55

Tgt Ion:	82	Resp:	2431
Ion Ratio	Lower	Upper	
82	100		
128	47.3	38.0	57.0
54	54.3	40.6	60.8





#9

Naphthalene

Concen: 0.389 ng

RT: 10.169 min Scan# 91

Delta R.T. -0.000 min

Lab File: BN034566.D

Acq: 14 Oct 2024 11:55

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:128 Resp: 10330

Ion Ratio Lower Upper

128 100

129 12.2 9.3 13.9

127 13.9 10.8 16.2



#10

Hexachlorobutadiene

Concen: 0.401 ng

RT: 10.458 min Scan# 953

Delta R.T. -0.000 min

Lab File: BN034566.D

Acq: 14 Oct 2024 11:55




Tgt Ion:225 Resp: 2318


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 63.5 51.4 77.0









#11
2-Methylnaphthalene-d10
Concen: 0.379 ng
RT: 11.723 min Scan# 1109
Delta R.T. -0.000 min
Lab File: BN034566.D
Acq: 14 Oct 2024 11:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 5767
Ion Ratio Lower Upper
152 100
151 20.5 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.380 ng
RT: 11.803 min Scan# 1130
Delta R.T. 0.004 min
Lab File: BN034566.D
Acq: 14 Oct 2024 11:55

Tgt Ion:142 Resp: 6493
Ion Ratio Lower Upper
142 100
141 89.1 70.6 105.8
115 37.6 29.2 43.8









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.019 min Scan# 1475
Delta R.T. -0.000 min
Lab File: BN034566.D
Acq: 14 Oct 2024 11:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 5771
Ion Ratio Lower Upper
164 100
162 102.4 79.7 119.5
160 49.4 37.9 56.9



#14
2,4,6-Tribromophenol
Concen: 0.315 ng
RT: 15.530 min Scan# 1616
Delta R.T. -0.000 min
Lab File: BN034566.D
Acq: 14 Oct 2024 11:55

Tgt Ion:330 Resp: 805
Ion Ratio Lower Upper
330 100
332 140.7 104.3 156.5
141 37.0 29.4 44.0









#15
2-Fluorobiphenyl
Concen: 0.382 ng
RT: 12.640 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN034566.D
Acq: 14 Oct 2024 11:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:172 Resp: 7602
Ion Ratio Lower Upper
172 100
171 36.0 27.1 40.7
170 24.0 18.5 27.7



#16
Acenaphthylene
Concen: 0.360 ng
RT: 13.730 min Scan# 1448
Delta R.T. -0.000 min
Lab File: BN034566.D
Acq: 14 Oct 2024 11:55

Tgt Ion:152 Resp: 8452
Ion Ratio Lower Upper
152 100
151 19.5 15.5 23.3
153 13.5 10.4 15.6









#17
Acenaphthene
Concen: 0.381 ng
RT: 14.083 min Scan# 1481
Delta R.T. 0.011 min
Lab File: BN034566.D
Acq: 14 Oct 2024 11:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 6335
Ion Ratio Lower Upper
154 100
153 115.7 91.5 137.3
152 58.3 45.8 68.8



#18
Fluorene
Concen: 0.373 ng
RT: 15.077 min Scan# 1574
Delta R.T. -0.000 min
Lab File: BN034566.D
Acq: 14 Oct 2024 11:55

Tgt Ion:166 Resp: 7928
Ion Ratio Lower Upper
166 100
165 98.5 78.6 118.0
167 13.6 10.8 16.2





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.783 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN034566.D

Acq: 14 Oct 2024 11:55

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:188 Resp: 10982

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 9.1 6.3 9.5



#20

4,6-Dinitro-2-methylphenol

Concen: 0.331 ng

RT: 16.287 min Scan# 1677

Delta R.T. 0.012 min

Lab File: BN034566.D

Acq: 14 Oct 2024 11:55




Tgt Ion:198 Resp: 50


Ion Ratio Lower Upper

198 100

51 223.6 174.8 262.2

105 155.6 167.9 251.9#









#21
4-Bromophenyl-phenylether
Concen: 0.382 ng
RT: 15.989 min Scan# 1653
Delta R.T. -0.000 min
Lab File: BN034566.D
Acq: 14 Oct 2024 11:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:248 Resp: 2279
Ion Ratio Lower Upper
248 100
250 102.7 79.8 119.6
141 63.2 44.3 66.5



#22
Hexachlorobenzene
Concen: 0.416 ng
RT: 16.088 min Scan# 1661
Delta R.T. -0.000 min
Lab File: BN034566.D
Acq: 14 Oct 2024 11:55

Tgt Ion:284 Resp: 3344
Ion Ratio Lower Upper
284 100
142 34.5 28.4 42.6
249 30.7 25.0 37.4





#23

Atrazine

Concen: 0.371 ng

RT: 16.287 min Scan# 1677

Delta R.T. -0.000 min

Lab File: BN034566.D

Acq: 14 Oct 2024 11:55

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:200 Resp: 1761

Ion Ratio Lower Upper

200 100

173 29.8 22.9 34.3

215 49.5 38.8 58.2



#24

Pentachlorophenol

Concen: 0.349 ng

RT: 16.461 min Scan# 1691

Delta R.T. -0.000 min

Lab File: BN034566.D

Acq: 14 Oct 2024 11:55




Tgt Ion:266 Resp: 804


Ion Ratio Lower Upper

266 100

264 62.4 46.6 69.8

268 63.3 52.0 78.0









#25
Phenanthrene
Concen: 0.402 ng
RT: 16.821 min Scan# 1720
Delta R.T. -0.000 min
Lab File: BN034566.D
Acq: 14 Oct 2024 11:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 11655
Ion Ratio Lower Upper
178 100
176 19.1 15.3 22.9
179 15.4 12.2 18.4



#26
Anthracene
Concen: 0.388 ng
RT: 16.920 min Scan# 1728
Delta R.T. 0.012 min
Lab File: BN034566.D
Acq: 14 Oct 2024 11:55

Tgt Ion:178 Resp: 9762
Ion Ratio Lower Upper
178 100
176 18.3 14.3 21.5
179 15.2 12.0 18.0









#27
Fluoranthene-d10
Concen: 0.427 ng
RT: 18.825 min Scan# 1924
Delta R.T. -0.000 min
Lab File: BN034566.D
Acq: 14 Oct 2024 11:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 11405
Ion Ratio Lower Upper
212 100
106 14.0 11.1 16.7
104 7.9 6.5 9.7



#28
Fluoranthene
Concen: 0.421 ng
RT: 18.857 min Scan# 1931
Delta R.T. 0.005 min
Lab File: BN034566.D
Acq: 14 Oct 2024 11:55

Tgt Ion:202 Resp: 13910
Ion Ratio Lower Upper
202 100
101 10.8 8.8 13.2
203 17.0 13.4 20.0









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.008 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN034566.D
Acq: 14 Oct 2024 11:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:240 Resp: 5201
Ion Ratio Lower Upper
240 100
120 13.0 10.1 15.1
236 28.5 22.6 34.0



#30
Pyrene
Concen: 0.282 ng
RT: 19.220 min Scan# 2009
Delta R.T. -0.000 min
Lab File: BN034566.D
Acq: 14 Oct 2024 11:55

Tgt Ion:202 Resp: 14398
Ion Ratio Lower Upper
202 100
200 20.6 16.6 24.8
203 17.9 14.3 21.5





#31

Terphenyl-d14

Concen: 0.392 ng

RT: 19.447 min Scan# 2058

Delta R.T. -0.005 min

Lab File: BN034566.D

Acq: 14 Oct 2024 11:55

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:244 Resp: 7379

Ion Ratio Lower Upper

244 100

212 10.2 7.7 11.5

122 17.5 12.6 19.0



#33

Chrysene

Concen: 0.347 ng

RT: 21.035 min Scan# 2397

Delta R.T. -0.000 min

Lab File: BN034566.D

Acq: 14 Oct 2024 11:55




Tgt Ion:228 Resp: 16159


Ion Ratio Lower Upper

228 100

226 29.8 24.4 36.6

229 20.4 15.8 23.6









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.104 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN034566.D
Acq: 14 Oct 2024 11:55

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:264 Resp: 10155




Ion	Ratio	Lower	Upper
264	100		
260	26.1	20.9	31.3
265	76.4	61.6	92.4




#36
Indeno(1,2,3-cd)pyrene
Concen: 0.400 ng
RT: 25.159 min Scan# 3492
Delta R.T. -0.006 min
Lab File: BN034566.D
Acq: 14 Oct 2024 11:55

Tgt Ion:276 Resp: 15790

Ion	Ratio	Lower	Upper
276	100		
138	26.4	22.0	33.0
277	25.0	20.1	30.1





#37

Benzo(b)fluoranthene

Concen: 0.379 ng

RT: 22.484 min Scan# 2577

Delta R.T. -0.006 min

Lab File: BN034566.D

Acq: 14 Oct 2024 11:55

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:252 Resp: 12873

Ion Ratio Lower Upper

252 100

253 26.4 21.3 31.9

125 16.3 12.7 19.1



#38

Benzo(k)fluoranthene

Concen: 0.386 ng

RT: 22.528 min Scan# 2592

Delta R.T. -0.003 min

Lab File: BN034566.D

Acq: 14 Oct 2024 11:55




Tgt Ion:252 Resp: 14909


Ion Ratio Lower Upper

252 100

253 27.2 21.0 31.6

125 17.0 13.3 19.9









#39
Benzo(a)pyrene
Concen: 0.388 ng
RT: 23.013 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN034566.D
Acq: 14 Oct 2024 11:55




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 11432
Ion Ratio Lower Upper
252 100
253 29.7 23.8 35.8
125 21.0 17.2 25.8



#40
Dibenzo(a,h)anthracene
Concen: 0.398 ng
RT: 25.177 min Scan# 3498
Delta R.T. -0.003 min
Lab File: BN034566.D
Acq: 14 Oct 2024 11:55

Tgt Ion:278 Resp: 12018
Ion Ratio Lower Upper
278 100
139 20.3 16.0 24.0
279 29.3 23.4 35.0





#41

Benzo(g,h,i)perylene

Concen: 0.409 ng

RT: 25.779 min Scan# 31

Delta R.T. -0.009 min

Lab File: BN034566.D


Acq: 14 Oct 2024 11:55

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4



Tgt Ion:276 Resp: 14764

Ion Ratio Lower Upper

276 100

277 24.6 19.8 29.8

138 25.6 19.2 28.8

