

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101424\
 Data File : BN034568.D
 Acq On : 14 Oct 2024 13:12
 Operator : RC/JU
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Quant Time: Oct 14 17:37:05 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN101424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 14 16:58:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.365	152	3457	0.400	ng	0.00
7) Naphthalene-d8	10.127	136	9590	0.400	ng	0.01
13) Acenaphthene-d10	14.019	164	5462	0.400	ng	0.00
19) Phenanthrene-d10	16.783	188	10528	0.400	ng	# 0.00
29) Chrysene-d12	21.008	240	6496	0.400	ng	# 0.00
35) Perylene-d12	23.110	264	9616	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.011	112	1876	0.200	ng	0.00
5) Phenol-d6	6.571	99	1549	0.159	ng	0.00
8) Nitrobenzene-d5	8.525	82	995	0.159	ng	0.01
11) 2-Methylnaphthalene-d10	11.746	152	2636	0.183	ng	0.02
14) 2,4,6-Tribromophenol	15.542	330	330	0.136	ng	0.01
15) 2-Fluorobiphenyl	12.651	172	3635	0.193	ng	0.01
27) Fluoranthene-d10	18.829	212	5041	0.197	ng	0.00
31) Terphenyl-d14	19.456	244	3186	0.267	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.025	88	1073	0.257	ng	90
3) n-Nitrosodimethylamine	3.328	42	730	0.192	ng	# 96
6) bis(2-Chloroethyl)ether	6.809	93	1354	0.170	ng	99
9) Naphthalene	10.169	128	4882	0.195	ng	# 94
10) Hexachlorobutadiene	10.458	225	1112	0.203	ng	# 98
12) 2-Methylnaphthalene	11.818	142	2898	0.179	ng	97
16) Acenaphthylene	13.730	152	3904	0.176	ng	100
17) Acenaphthene	14.083	154	2978	0.189	ng	98
18) Fluorene	15.088	166	3607	0.180	ng	100
20) 4,6-Dinitro-2-methylph...	16.287	198	28	0.194	ng	# 61
21) 4-Bromophenyl-phenylether	15.989	248	1055	0.185	ng	# 82
22) Hexachlorobenzene	16.088	284	1556	0.202	ng	99
23) Atrazine	16.287	200	743	0.163	ng	# 88
24) Pentachlorophenol	16.486	266	286	0.130	ng	99
25) Phenanthrene	16.821	178	5268	0.190	ng	98
26) Anthracene	16.920	178	4116	0.171	ng	99
28) Fluoranthene	18.857	202	6127	0.194	ng	99
30) Pyrene	19.224	202	6452	0.101	ng	100
33) Chrysene	21.044	228	8053	0.138	ng	98
36) Indeno(1,2,3-cd)pyrene	25.174	276	6736	0.180	ng	96
37) Benzo(b)fluoranthene	22.490	252	5441	0.169	ng	# 86
38) Benzo(k)fluoranthene	22.531	252	7017	0.192	ng	# 85
39) Benzo(a)pyrene	23.022	252	5216	0.187	ng	# 79
40) Dibenzo(a,h)anthracene	25.189	278	5009	0.175	ng	# 87
41) Benzo(g,h,i)perylene	25.785	276	6554	0.192	ng	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101424\
Data File : BN034568.D
Acq On : 14 Oct 2024 13:12
Operator : RC/JU
Sample : SSTDICC0.2
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2

Quant Time: Oct 14 17:37:05 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN101424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 14 16:58:38 2024
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.365 min Scan# 61
 Delta R.T. 0.000 min
 Lab File: BN034568.D
 Acq: 14 Oct 2024 13:12




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2


Tgt Ion:152 Resp: 3457
 Ion Ratio Lower Upper
 152 100
 150 150.2 122.0 183.0
 115 61.3 48.3 72.5



#2
 1,4-Dioxane
 Concen: 0.257 ng
 RT: 3.025 min Scan# 20
 Delta R.T. -0.000 min
 Lab File: BN034568.D
 Acq: 14 Oct 2024 13:12

Tgt Ion: 88 Resp: 1073
 Ion Ratio Lower Upper
 88 100
 43 30.3 22.8 34.2
 58 60.2 56.4 84.6









#3
n-Nitrosodimethylamine
Concen: 0.192 ng
RT: 3.328 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN034568.D
Acq: 14 Oct 2024 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion: 42 Resp: 730
Ion Ratio Lower Upper
42 100
74 127.9 99.0 148.6
44 5.3 6.6 9.8#



#4
2-Fluorophenol
Concen: 0.200 ng
RT: 5.011 min Scan# 295
Delta R.T. 0.007 min
Lab File: BN034568.D
Acq: 14 Oct 2024 13:12


Tgt Ion: 112 Resp: 1876
Ion Ratio Lower Upper
112 100
64 54.5 44.6 66.8
63 30.2 23.8 35.6








#5
Phenol-d6
Concen: 0.159 ng
RT: 6.571 min Scan# 511
Delta R.T. 0.007 min
Lab File: BN034568.D
Acq: 14 Oct 2024 13:12

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2







Tgt Ion: 99 Resp: 1549
Ion Ratio Lower Upper
99 100
42 20.4 15.0 22.6
71 36.4 28.0 42.0



#6
bis(2-Chloroethyl)ether
Concen: 0.170 ng
RT: 6.809 min Scan# 544
Delta R.T. 0.000 min
Lab File: BN034568.D
Acq: 14 Oct 2024 13:12

Tgt Ion: 93 Resp: 1354
Ion Ratio Lower Upper
93 100
63 80.7 64.6 96.8
95 37.4 27.9 41.9





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.127 min Scan# 91

Delta R.T. 0.011 min

Lab File: BN034568.D

Acq: 14 Oct 2024 13:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2

Tgt Ion:136 Resp: 9590





Ion Ratio Lower Upper

136 100

137 12.3 9.4 14.0

54 9.1 6.2 9.4

68 6.6 4.6 6.8



#8

Nitrobenzene-d5

Concen: 0.159 ng

RT: 8.525 min Scan# 772

Delta R.T. 0.011 min

Lab File: BN034568.D

Acq: 14 Oct 2024 13:12




Tgt Ion: 82 Resp: 995


Ion Ratio Lower Upper

82 100

128 54.7 38.0 57.0

54 59.6 40.6 60.8









#9
Naphthalene
Concen: 0.195 ng
RT: 10.169 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN034568.D
Acq: 14 Oct 2024 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:128 Resp: 4882
Ion Ratio Lower Upper
128 100
129 14.1 9.3 13.9#
127 15.6 10.8 16.2



#10
Hexachlorobutadiene
Concen: 0.203 ng
RT: 10.458 min Scan# 953
Delta R.T. 0.000 min
Lab File: BN034568.D
Acq: 14 Oct 2024 13:12

Tgt Ion:225 Resp: 1112
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 66.1 51.4 77.0









#11
2-Methylnaphthalene-d10
Concen: 0.183 ng
RT: 11.746 min Scan# 1115
Delta R.T. 0.023 min
Lab File: BN034568.D
Acq: 14 Oct 2024 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:152 Resp: 2636
Ion Ratio Lower Upper
152 100
151 19.2 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.179 ng
RT: 11.818 min Scan# 1134
Delta R.T. 0.019 min
Lab File: BN034568.D
Acq: 14 Oct 2024 13:12

Tgt Ion:142 Resp: 2898
Ion Ratio Lower Upper
142 100
141 87.0 70.6 105.8
115 41.4 29.2 43.8





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.019 min Scan# 1475

Delta R.T. 0.000 min

Lab File: BN034568.D

Acq: 14 Oct 2024 13:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:164 Resp: 5462

Ion Ratio Lower Upper

164 100

162 103.3 79.7 119.5

160 50.7 37.9 56.9



#14

2,4,6-Tribromophenol

Concen: 0.136 ng

RT: 15.542 min Scan# 1617

Delta R.T. 0.012 min

Lab File: BN034568.D

Acq: 14 Oct 2024 13:12




Tgt Ion:330 Resp: 330


Ion Ratio Lower Upper

330 100

332 205.2 104.3 156.5#

141 37.0 29.4 44.0









#15
2-Fluorobiphenyl
Concen: 0.193 ng
RT: 12.651 min Scan# 1347
Delta R.T. 0.011 min
Lab File: BN034568.D
Acq: 14 Oct 2024 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:172 Resp: 3635
Ion Ratio Lower Upper
172 100
171 37.9 27.1 40.7
170 27.5 18.5 27.7



#16
Acenaphthylene
Concen: 0.176 ng
RT: 13.730 min Scan# 1448
Delta R.T. 0.000 min
Lab File: BN034568.D
Acq: 14 Oct 2024 13:12

Tgt Ion:152 Resp: 3904
Ion Ratio Lower Upper
152 100
151 19.5 15.5 23.3
153 13.0 10.4 15.6









#17
Acenaphthene
Concen: 0.189 ng
RT: 14.083 min Scan# 1481
Delta R.T. 0.011 min
Lab File: BN034568.D
Acq: 14 Oct 2024 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:154 Resp: 2978
Ion Ratio Lower Upper
154 100
153 117.1 91.5 137.3
152 59.0 45.8 68.8



#18
Fluorene
Concen: 0.180 ng
RT: 15.088 min Scan# 1575
Delta R.T. 0.011 min
Lab File: BN034568.D
Acq: 14 Oct 2024 13:12

Tgt Ion:166 Resp: 3607
Ion Ratio Lower Upper
166 100
165 98.6 78.6 118.0
167 14.0 10.8 16.2









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.783 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN034568.D
Acq: 14 Oct 2024 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:188 Resp: 10528
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.6 6.3 9.5#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.194 ng
RT: 16.287 min Scan# 1677
Delta R.T. 0.012 min
Lab File: BN034568.D
Acq: 14 Oct 2024 13:12

Tgt Ion:198 Resp: 28
Ion Ratio Lower Upper
198 100
51 305.6 174.8 262.2#
105 174.1 167.9 251.9









#21
4-Bromophenyl-phenylether
Concen: 0.185 ng
RT: 15.989 min Scan# 1653
Delta R.T. 0.000 min
Lab File: BN034568.D
Acq: 14 Oct 2024 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:	248	Resp:	1055
Ion Ratio	Lower	Upper	
248	100		
250	96.6	79.8	119.6
141	88.0	44.3	66.5



#22
Hexachlorobenzene
Concen: 0.202 ng
RT: 16.088 min Scan# 1661
Delta R.T. 0.000 min
Lab File: BN034568.D
Acq: 14 Oct 2024 13:12

Tgt Ion:	284	Resp:	1556
Ion Ratio	Lower	Upper	
284	100		
142	34.3	28.4	42.6
249	30.8	25.0	37.4





#23

Atrazine

Concen: 0.163 ng

RT: 16.287 min Scan# 1677

Delta R.T. 0.000 min

Lab File: BN034568.D

Acq: 14 Oct 2024 13:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:200 Resp: 743

Ion Ratio Lower Upper

200 100

173 43.2 22.9 34.3#

215 50.7 38.8 58.2



#24

Pentachlorophenol

Concen: 0.130 ng

RT: 16.486 min Scan# 1693

Delta R.T. 0.025 min

Lab File: BN034568.D

Acq: 14 Oct 2024 13:12




Tgt Ion:266 Resp: 286


Ion Ratio Lower Upper

266 100

264 57.3 46.6 69.8

268 65.4 52.0 78.0









#25
 Phenanthrene
 Concen: 0.190 ng
 RT: 16.821 min Scan# 1720
 Delta R.T. 0.000 min
 Lab File: BN034568.D
 Acq: 14 Oct 2024 13:12

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Tgt Ion:178 Resp: 5268




Ion	Ratio	Lower	Upper
178	100		
176	18.3	15.3	22.9
179	16.0	12.2	18.4




#26
 Anthracene
 Concen: 0.171 ng
 RT: 16.920 min Scan# 1728
 Delta R.T. 0.012 min
 Lab File: BN034568.D
 Acq: 14 Oct 2024 13:12

Tgt Ion:178 Resp: 4116

Ion	Ratio	Lower	Upper
178	100		
176	18.8	14.3	21.5
179	15.1	12.0	18.0









#27
Fluoranthene-d10
Concen: 0.197 ng
RT: 18.829 min Scan# 1924
Delta R.T. 0.005 min
Lab File: BN034568.D
Acq: 14 Oct 2024 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:212 Resp: 5041
Ion Ratio Lower Upper
212 100
106 14.0 11.1 16.7
104 7.6 6.5 9.7



#28
Fluoranthene
Concen: 0.194 ng
RT: 18.857 min Scan# 1931
Delta R.T. 0.005 min
Lab File: BN034568.D
Acq: 14 Oct 2024 13:12

Tgt Ion:202 Resp: 6127
Ion Ratio Lower Upper
202 100
101 10.7 8.8 13.2
203 17.2 13.4 20.0









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.008 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN034568.D
Acq: 14 Oct 2024 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:240 Resp: 6496
Ion Ratio Lower Upper
240 100
120 15.6 10.1 15.1#
236 28.6 22.6 34.0



#30
Pyrene
Concen: 0.101 ng
RT: 19.224 min Scan# 2010
Delta R.T. 0.005 min
Lab File: BN034568.D
Acq: 14 Oct 2024 13:12

Tgt Ion:202 Resp: 6452
Ion Ratio Lower Upper
202 100
200 20.7 16.6 24.8
203 17.8 14.3 21.5









#31
 Terphenyl-d14
 Concen: 0.267 ng
 RT: 19.456 min Scan# 2059
 Delta R.T. 0.004 min
 Lab File: BN034568.D
 Acq: 14 Oct 2024 13:12




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2


Tgt Ion:244 Resp: 3186
 Ion Ratio Lower Upper
 244 100
 212 12.7 7.7 11.5#
 122 19.2 12.6 19.0#



#33
 Chrysene
 Concen: 0.138 ng
 RT: 21.044 min Scan# 2398
 Delta R.T. 0.009 min
 Lab File: BN034568.D
 Acq: 14 Oct 2024 13:12

Tgt Ion:228 Resp: 8053
 Ion Ratio Lower Upper
 228 100
 226 30.9 24.4 36.6
 229 21.0 15.8 23.6





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.110 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN034568.D

Acq: 14 Oct 2024 13:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:264 Resp: 9616

Ion Ratio Lower Upper

264 100

260 27.2 20.9 31.3

265 82.7 61.6 92.4



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.180 ng

RT: 25.174 min Scan# 3497

Delta R.T. 0.009 min

Lab File: BN034568.D

Acq: 14 Oct 2024 13:12




Tgt Ion:276 Resp: 6736


Ion Ratio Lower Upper

276 100

138 24.3 22.0 33.0


277 24.0 20.1 30.1







#37
Benzo(b)fluoranthene
Concen: 0.169 ng
RT: 22.490 min Scan# 2579
Delta R.T. 0.000 min
Lab File: BN034568.D
Acq: 14 Oct 2024 13:12

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2







Tgt Ion:252 Resp: 5441
Ion Ratio Lower Upper
252 100
253 32.5 21.3 31.9#
125 23.7 12.7 19.1#



#38
Benzo(k)fluoranthene
Concen: 0.192 ng
RT: 22.531 min Scan# 2593
Delta R.T. 0.000 min
Lab File: BN034568.D
Acq: 14 Oct 2024 13:12

Tgt Ion:252 Resp: 7017
Ion Ratio Lower Upper
252 100
253 32.8 21.0 31.6#
125 24.4 13.3 19.9#





#39

Benzo(a)pyrene

Concen: 0.187 ng

RT: 23.022 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN034568.D


Acq: 14 Oct 2024 13:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2






Tgt Ion:252 Resp: 5216

Ion Ratio Lower Upper

252 100

253 40.2 23.8 35.8#

125 33.1 17.2 25.8#



#40

Dibenzo(a,h)anthracene

Concen: 0.175 ng

RT: 25.189 min Scan# 3502

Delta R.T. 0.009 min

Lab File: BN034568.D

Acq: 14 Oct 2024 13:12




Tgt Ion:278 Resp: 5009


Ion Ratio Lower Upper

278 100

139 23.9 16.0 24.0

279 37.7 23.4 35.0#





#41
Benzo(g,h,i)perylene
Concen: 0.192 ng
RT: 25.785 min Scan# 31
Delta R.T. -0.003 min
Lab File: BN034568.D
Acq: 14 Oct 2024 13:12

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2

Tgt Ion:	276	Resp:	6554
Ion	Ratio	Lower	Upper
276	100		
277	26.4	19.8	29.8
138	27.1	19.2	28.8

