

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101424\
 Data File : BN034583.D
 Acq On : 14 Oct 2024 22:19
 Operator : RC/JU
 Sample : PB164072BS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB164072BS

Quant Time: Oct 15 00:06:39 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN101424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 14 17:52:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.366	152	4830	0.400	ng	0.00
7) Naphthalene-d8	10.116	136	13753	0.400	ng	0.00
13) Acenaphthene-d10	14.008	164	7271	0.400	ng	-0.01
19) Phenanthrene-d10	16.784	188	12370	0.400	ng	0.00
29) Chrysene-d12	21.008	240	4342	0.400	ng	0.00
35) Perylene-d12	23.101	264	8592	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.011	112	4606	0.352	ng	0.00
5) Phenol-d6	6.578	99	4640	0.342	ng	0.01
8) Nitrobenzene-d5	8.504	82	3601	0.401	ng	-0.01
11) 2-Methylnaphthalene-d10	11.720	152	11456	0.556	ng	0.00
14) 2,4,6-Tribromophenol	15.530	330	924	0.287	ng	0.00
15) 2-Fluorobiphenyl	12.629	172	10019	0.400	ng	-0.01
27) Fluoranthene-d10	18.825	212	11168	0.371	ng	0.00
31) Terphenyl-d14	19.452	244	7006	0.419	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.025	88	1942	0.333	ng	# 17
3) n-Nitrosodimethylamine	3.321	42	2377	0.448	ng	# 91
6) bis(2-Chloroethyl)ether	6.817	93	4473	0.403	ng	97
9) Naphthalene	10.170	128	15517	0.431	ng	100
10) Hexachlorobutadiene	10.458	225	3215	0.410	ng	# 100
12) 2-Methylnaphthalene	11.799	142	9636	0.416	ng	100
16) Acenaphthylene	13.731	152	12249	0.414	ng	99
17) Acenaphthene	14.073	154	8771	0.419	ng	99
18) Fluorene	15.078	166	10297	0.385	ng	99
20) 4,6-Dinitro-2-methylph...	16.275	198	60	0.353	ng	83
21) 4-Bromophenyl-phenylether	15.989	248	2855	0.425	ng	99
22) Hexachlorobenzene	16.089	284	4107	0.453	ng	98
23) Atrazine	16.287	200	2127	0.398	ng	98
24) Pentachlorophenol	16.449	266	1823	0.703	ng	98
25) Phenanthrene	16.821	178	14414	0.442	ng	100
26) Anthracene	16.920	178	12279	0.434	ng	99
28) Fluoranthene	18.853	202	14325	0.385	ng	99
30) Pyrene	19.220	202	14535	0.341	ng	99
33) Chrysene	21.035	228	15141	0.389	ng	99
36) Indeno(1,2,3-cd)pyrene	25.157	276	17461	0.523	ng	100
37) Benzo(b)fluoranthene	22.484	252	12026	0.419	ng	99
38) Benzo(k)fluoranthene	22.525	252	14761	0.451	ng	99
39) Benzo(a)pyrene	23.011	252	11738	0.471	ng	99
40) Dibenzo(a,h)anthracene	25.174	278	13286	0.520	ng	99
41) Benzo(g,h,i)perylene	25.776	276	15239	0.499	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101424\
Data File : BN034583.D
Acq On : 14 Oct 2024 22:19
Operator : RC/JU
Sample : PB164072BS
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB164072BS


Quant Time: Oct 15 00:06:39 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN101424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 14 17:52:02 2024
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.366 min Scan# 61
 Delta R.T. 0.001 min
 Lab File: BN034583.D
 Acq: 14 Oct 2024 22:19

Instrument :
 BNA_N
 ClientSampleId :
 PB164072BS







Tgt Ion:152 Resp: 4830
 Ion Ratio Lower Upper
 152 100
 150 155.0 122.0 183.0
 115 61.9 48.3 72.5



#2
 1,4-Dioxane
 Concen: 0.333 ng
 RT: 3.025 min Scan# 20
 Delta R.T. 0.000 min
 Lab File: BN034583.D
 Acq: 14 Oct 2024 22:19

Tgt Ion: 88 Resp: 1942
 Ion Ratio Lower Upper
 88 100
 43 130.8 22.8 34.2#
 58 102.5 56.4 84.6#









#3
 n-Nitrosodimethylamine
 Concen: 0.448 ng
 RT: 3.321 min Scan# 61
 Delta R.T. -0.007 min
 Lab File: BN034583.D
 Acq: 14 Oct 2024 22:19




Instrument :
 BNA_N
 ClientSampleId :
 PB164072BS


Tgt Ion: 42 Resp: 2377
 Ion Ratio Lower Upper
 42 100
 74 134.5 99.0 148.6
 44 6.0 6.6 9.8#



#4
 2-Fluorophenol
 Concen: 0.352 ng
 RT: 5.011 min Scan# 295
 Delta R.T. 0.007 min
 Lab File: BN034583.D
 Acq: 14 Oct 2024 22:19

Tgt Ion: 112 Resp: 4606
 Ion Ratio Lower Upper
 112 100
 64 55.7 44.6 66.8
 63 29.8 23.8 35.6









#5
Phenol-d6
Concen: 0.342 ng
RT: 6.578 min Scan# 510
Delta R.T. 0.015 min
Lab File: BN034583.D
Acq: 14 Oct 2024 22:19




Instrument :
BNA_N
ClientSampleId :
PB164072BS


Tgt Ion: 99 Resp: 4640
Ion Ratio Lower Upper
99 100
42 18.4 15.0 22.6
71 34.6 28.0 42.0



#6
bis(2-Chloroethyl)ether
Concen: 0.403 ng
RT: 6.817 min Scan# 545
Delta R.T. 0.007 min
Lab File: BN034583.D
Acq: 14 Oct 2024 22:19

Tgt Ion: 93 Resp: 4473
Ion Ratio Lower Upper
93 100
63 77.4 64.6 96.8
95 34.8 27.9 41.9









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.116 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN034583.D
Acq: 14 Oct 2024 22:19




Instrument :
BNA_N
ClientSampleId :
PB164072BS


Tgt Ion:	136	Resp:	13753
Ion	Ratio	Lower	Upper
136	100		
137	11.7	9.4	14.0
54	7.4	6.2	9.4
68	5.5	4.6	6.8



#8
Nitrobenzene-d5
Concen: 0.401 ng
RT: 8.504 min Scan# 770
Delta R.T. -0.010 min
Lab File: BN034583.D
Acq: 14 Oct 2024 22:19

Tgt Ion:	82	Resp:	3601
Ion	Ratio	Lower	Upper
82	100		
128	41.3	38.0	57.0
54	55.2	40.6	60.8









#9
Naphthalene
Concen: 0.431 ng
RT: 10.170 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN034583.D
Acq: 14 Oct 2024 22:19




Instrument :
BNA_N
ClientSampleId :
PB164072BS


Tgt Ion:128 Resp: 15517
Ion Ratio Lower Upper
128 100
129 11.7 9.3 13.9
127 13.5 10.8 16.2



#10
Hexachlorobutadiene
Concen: 0.410 ng
RT: 10.458 min Scan# 953
Delta R.T. 0.000 min
Lab File: BN034583.D
Acq: 14 Oct 2024 22:19

Tgt Ion:225 Resp: 3215
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.0 51.4 77.0









#11
2-Methylnaphthalene-d10
Concen: 0.556 ng
RT: 11.720 min Scan# 1108
Delta R.T. -0.004 min
Lab File: BN034583.D
Acq: 14 Oct 2024 22:19




Instrument :
BNA_N
ClientSampleId :
PB164072BS


Tgt Ion:152 Resp: 11456
Ion Ratio Lower Upper
152 100
151 14.9 16.7 25.1#



#12
2-Methylnaphthalene
Concen: 0.416 ng
RT: 11.799 min Scan# 1129
Delta R.T. 0.000 min
Lab File: BN034583.D
Acq: 14 Oct 2024 22:19

Tgt Ion:142 Resp: 9636
Ion Ratio Lower Upper
142 100
141 88.3 70.6 105.8
115 36.6 29.2 43.8





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.008 min Scan# 1475

Delta R.T. -0.010 min

Lab File: BN034583.D

Acq: 14 Oct 2024 22:19

Instrument :

BNA_N

ClientSampleId :

PB164072BS





Tgt Ion:164 Resp: 7271

Ion Ratio Lower Upper

164 100

162 106.3 79.7 119.5

160 51.5 37.9 56.9



#14

2,4,6-Tribromophenol

Concen: 0.287 ng

RT: 15.530 min Scan# 1616

Delta R.T. 0.000 min

Lab File: BN034583.D

Acq: 14 Oct 2024 22:19




Tgt Ion:330 Resp: 924


Ion Ratio Lower Upper

330 100

332 133.4 104.3 156.5

141 34.7 29.4 44.0









#15
2-Fluorobiphenyl
Concen: 0.400 ng
RT: 12.629 min Scan# 11
Delta R.T. -0.011 min
Lab File: BN034583.D
Acq: 14 Oct 2024 22:19




Instrument :
BNA_N
ClientSampleId :
PB164072BS


Tgt Ion:172 Resp: 10019
Ion Ratio Lower Upper
172 100
171 35.4 27.1 40.7
170 23.9 18.5 27.7



#16
Acenaphthylene
Concen: 0.414 ng
RT: 13.731 min Scan# 1448
Delta R.T. 0.000 min
Lab File: BN034583.D
Acq: 14 Oct 2024 22:19

Tgt Ion:152 Resp: 12249
Ion Ratio Lower Upper
152 100
151 19.7 15.5 23.3
153 13.3 10.4 15.6









#17
Acenaphthene
Concen: 0.419 ng
RT: 14.073 min Scan# 1480
Delta R.T. 0.000 min
Lab File: BN034583.D
Acq: 14 Oct 2024 22:19




Instrument :
BNA_N
ClientSampleId :
PB164072BS


Tgt Ion:154 Resp: 8771
Ion Ratio Lower Upper
154 100
153 113.4 91.5 137.3
152 56.6 45.8 68.8



#18
Fluorene
Concen: 0.385 ng
RT: 15.078 min Scan# 1574
Delta R.T. 0.000 min
Lab File: BN034583.D
Acq: 14 Oct 2024 22:19

Tgt Ion:166 Resp: 10297
Ion Ratio Lower Upper
166 100
165 99.3 78.6 118.0
167 13.3 10.8 16.2





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.784 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN034583.D

Acq: 14 Oct 2024 22:19

Instrument :

BNA_N

ClientSampleId :

PB164072BS





Tgt Ion:188 Resp: 12370

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 8.3 6.3 9.5



#20

4,6-Dinitro-2-methylphenol

Concen: 0.353 ng

RT: 16.275 min Scan# 1676

Delta R.T. -0.000 min

Lab File: BN034583.D

Acq: 14 Oct 2024 22:19




Tgt Ion:198 Resp: 60


Ion Ratio Lower Upper

198 100

51 184.1 174.8 262.2

105 191.3 167.9 251.9









#21
4-Bromophenyl-phenylether
Concen: 0.425 ng
RT: 15.989 min Scan# 1653
Delta R.T. 0.000 min
Lab File: BN034583.D
Acq: 14 Oct 2024 22:19




Instrument :
BNA_N
ClientSampleId :
PB164072BS


Tgt Ion:248 Resp: 2855
Ion Ratio Lower Upper
248 100
250 100.7 79.8 119.6
141 56.8 44.3 66.5



#22
Hexachlorobenzene
Concen: 0.453 ng
RT: 16.089 min Scan# 1661
Delta R.T. 0.000 min
Lab File: BN034583.D
Acq: 14 Oct 2024 22:19

Tgt Ion:284 Resp: 4107
Ion Ratio Lower Upper
284 100
142 34.3 28.4 42.6
249 30.7 25.0 37.4









#23
 Atrazine
 Concen: 0.398 ng
 RT: 16.287 min Scan# 1677
 Delta R.T. 0.000 min
 Lab File: BN034583.D
 Acq: 14 Oct 2024 22:19




Instrument :
 BNA_N
 ClientSampleId :
 PB164072BS


Tgt Ion:200 Resp: 2127
 Ion Ratio Lower Upper
 200 100
 173 30.1 22.9 34.3
 215 50.1 38.8 58.2



#24
 Pentachlorophenol
 Concen: 0.703 ng
 RT: 16.449 min Scan# 1690
 Delta R.T. -0.012 min
 Lab File: BN034583.D
 Acq: 14 Oct 2024 22:19

Tgt Ion:266 Resp: 1823
 Ion Ratio Lower Upper
 266 100
 264 61.9 46.6 69.8
 268 65.1 52.0 78.0









#25
Phenanthrene
Concen: 0.442 ng
RT: 16.821 min Scan# 1720
Delta R.T. 0.000 min
Lab File: BN034583.D
Acq: 14 Oct 2024 22:19




Instrument :
BNA_N
ClientSampleId :
PB164072BS


Tgt Ion:178 Resp: 14414
Ion Ratio Lower Upper
178 100
176 19.0 15.3 22.9
179 15.0 12.2 18.4



#26
Anthracene
Concen: 0.434 ng
RT: 16.920 min Scan# 1728
Delta R.T. 0.013 min
Lab File: BN034583.D
Acq: 14 Oct 2024 22:19

Tgt Ion:178 Resp: 12279
Ion Ratio Lower Upper
178 100
176 18.4 14.3 21.5
179 15.3 12.0 18.0









#27
Fluoranthene-d10
Concen: 0.371 ng
RT: 18.825 min Scan# 1924
Delta R.T. 0.000 min
Lab File: BN034583.D
Acq: 14 Oct 2024 22:19




Instrument :
BNA_N
ClientSampleId :
PB164072BS


Tgt Ion:212 Resp: 11168
Ion Ratio Lower Upper
212 100
106 13.8 11.1 16.7
104 7.7 6.5 9.7



#28
Fluoranthene
Concen: 0.385 ng
RT: 18.853 min Scan# 1930
Delta R.T. 0.000 min
Lab File: BN034583.D
Acq: 14 Oct 2024 22:19

Tgt Ion:202 Resp: 14325
Ion Ratio Lower Upper
202 100
101 10.5 8.8 13.2
203 17.0 13.4 20.0





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.008 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN034583.D

Acq: 14 Oct 2024 22:19

Instrument :

BNA_N

ClientSampleId :

PB164072BS





Tgt Ion:240 Resp: 4342

Ion Ratio Lower Upper

240 100

120 14.8 10.1 15.1

236 29.1 22.6 34.0



#30

Pyrene

Concen: 0.341 ng

RT: 19.220 min Scan# 2009

Delta R.T. 0.000 min

Lab File: BN034583.D

Acq: 14 Oct 2024 22:19




Tgt Ion:202 Resp: 14535


Ion Ratio Lower Upper

202 100

200 20.5 16.6 24.8

203 17.7 14.3 21.5









#31
 Terphenyl-d14
 Concen: 0.419 ng
 RT: 19.452 min Scan# 2059
 Delta R.T. 0.000 min
 Lab File: BN034583.D
 Acq: 14 Oct 2024 22:19




Instrument :
 BNA_N
 ClientSampleId :
 PB164072BS


Tgt Ion:244 Resp: 7006
 Ion Ratio Lower Upper
 244 100
 212 10.5 7.7 11.5
 122 16.8 12.6 19.0



#33
 Chrysene
 Concen: 0.389 ng
 RT: 21.035 min Scan# 2397
 Delta R.T. 0.000 min
 Lab File: BN034583.D
 Acq: 14 Oct 2024 22:19

Tgt Ion:228 Resp: 15141
 Ion Ratio Lower Upper
 228 100
 226 30.3 24.4 36.6
 229 20.1 15.8 23.6





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.101 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN034583.D

Acq: 14 Oct 2024 22:19

Instrument :

BNA_N

ClientSampleId :

PB164072BS





Tgt Ion:264 Resp: 8592

Ion Ratio Lower Upper

264 100

260 26.5 20.9 31.3

265 89.9 61.6 92.4



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.523 ng

RT: 25.157 min Scan# 3491

Delta R.T. -0.009 min

Lab File: BN034583.D

Acq: 14 Oct 2024 22:19




Tgt Ion:276 Resp: 17461


Ion Ratio Lower Upper

276 100

138 27.4 22.0 33.0


277 25.0 20.1 30.1








#37
Benzo(b)fluoranthene
Concen: 0.419 ng
RT: 22.484 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN034583.D
Acq: 14 Oct 2024 22:19

Instrument :
BNA_N
ClientSampleId :
PB164072BS







Tgt Ion:252 Resp: 12026
Ion Ratio Lower Upper
252 100
253 27.3 21.3 31.9
125 16.5 12.7 19.1



#38
Benzo(k)fluoranthene
Concen: 0.451 ng
RT: 22.525 min Scan# 2591
Delta R.T. -0.006 min
Lab File: BN034583.D
Acq: 14 Oct 2024 22:19

Tgt Ion:252 Resp: 14761
Ion Ratio Lower Upper
252 100
253 27.2 21.0 31.6
125 16.8 13.3 19.9





#39

Benzo(a)pyrene

Concen: 0.471 ng

RT: 23.011 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN034583.D

Acq: 14 Oct 2024 22:19

Instrument :

BNA_N

ClientSampleId :

PB164072BS





Tgt Ion:252 Resp: 11738

Ion Ratio Lower Upper

252 100

253 30.6 23.8 35.8

125 21.2 17.2 25.8



#40

Dibenzo(a,h)anthracene

Concen: 0.520 ng

RT: 25.174 min Scan# 3497

Delta R.T. -0.006 min

Lab File: BN034583.D

Acq: 14 Oct 2024 22:19




Tgt Ion:278 Resp: 13286


Ion Ratio Lower Upper

278 100

139 20.4 16.0 24.0


279 28.8 23.4 35.0





#41
Benzo(g,h,i)perylene
Concen: 0.499 ng
RT: 25.776 min Scan# 31
Delta R.T. -0.012 min
Lab File: BN034583.D
Acq: 14 Oct 2024 22:19

Instrument :
BNA_N
ClientSampleId :
PB164072BS



Tgt Ion:	276	Resp:	15239
Ion	Ratio	Lower	Upper
276	100		
277	24.4	19.8	29.8
138	25.0	19.2	28.8

