

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101424\
 Data File : BN034584.D
 Acq On : 14 Oct 2024 22:55
 Operator : RC/JU
 Sample : PB164072BSD
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB164072BSD

Quant Time: Oct 15 00:06:57 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN101424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 14 17:52:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.365	152	5203	0.400	ng	0.00
7) Naphthalene-d8	10.116	136	14775	0.400	ng	0.00
13) Acenaphthene-d10	14.008	164	7913	0.400	ng	-0.01
19) Phenanthrene-d10	16.771	188	13541	0.400	ng	#-0.01
29) Chrysene-d12	21.008	240	5062	0.400	ng	0.00
35) Perylene-d12	23.104	264	9196	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.011	112	4829	0.342	ng	0.00
5) Phenol-d6	6.578	99	4934	0.338	ng	0.01
8) Nitrobenzene-d5	8.504	82	3793	0.394	ng	-0.01
11) 2-Methylnaphthalene-d10	11.719	152	11978	0.541	ng	0.00
14) 2,4,6-Tribromophenol	15.530	330	1004	0.286	ng	0.00
15) 2-Fluorobiphenyl	12.629	172	10600	0.388	ng	-0.01
27) Fluoranthene-d10	18.825	212	11794	0.358	ng	0.00
31) Terphenyl-d14	19.452	244	7375	0.397	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.025	88	1925	0.307	ng	# 19
3) n-Nitrosodimethylamine	3.321	42	2642	0.462	ng	98
6) bis(2-Chloroethyl)ether	6.816	93	4769	0.399	ng	97
9) Naphthalene	10.169	128	16400	0.424	ng	99
10) Hexachlorobutadiene	10.458	225	3376	0.400	ng	# 99
12) 2-Methylnaphthalene	11.795	142	10255	0.412	ng	100
16) Acenaphthylene	13.730	152	13168	0.409	ng	100
17) Acenaphthene	14.072	154	9344	0.410	ng	100
18) Fluorene	15.077	166	10977	0.377	ng	100
20) 4,6-Dinitro-2-methylph...	16.287	198	73	0.392	ng	# 67
21) 4-Bromophenyl-phenylether	15.989	248	3041	0.414	ng	98
22) Hexachlorobenzene	16.088	284	4412	0.445	ng	99
23) Atrazine	16.287	200	2257	0.386	ng	99
24) Pentachlorophenol	16.448	266	1913	0.674	ng	96
25) Phenanthrene	16.821	178	15303	0.428	ng	99
26) Anthracene	16.908	178	12849	0.414	ng	99
28) Fluoranthene	18.853	202	15043	0.370	ng	99
30) Pyrene	19.220	202	15424	0.311	ng	100
33) Chrysene	21.035	228	16224	0.358	ng	99
36) Indeno(1,2,3-cd)pyrene	25.159	276	18763	0.525	ng	98
37) Benzo(b)fluoranthene	22.484	252	12783	0.416	ng	99
38) Benzo(k)fluoranthene	22.528	252	14905	0.426	ng	100
39) Benzo(a)pyrene	23.016	252	12624	0.473	ng	99
40) Dibenzo(a,h)anthracene	25.180	278	14187	0.519	ng	98
41) Benzo(g,h,i)perylene	25.779	276	16135	0.494	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101424\
Data File : BN034584.D
Acq On : 14 Oct 2024 22:55
Operator : RC/JU
Sample : PB164072BSD
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB164072BSD

Quant Time: Oct 15 00:06:57 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN101424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 14 17:52:02 2024
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.365 min Scan# 61
 Delta R.T. 0.000 min
 Lab File: BN034584.D
 Acq: 14 Oct 2024 22:55




Instrument :
 BNA_N
 ClientSampleId :
 PB164072BSD


Tgt Ion:152 Resp: 5203
 Ion Ratio Lower Upper
 152 100
 150 153.1 122.0 183.0
 115 62.5 48.3 72.5



#2
 1,4-Dioxane
 Concen: 0.307 ng
 RT: 3.025 min Scan# 20
 Delta R.T. -0.000 min
 Lab File: BN034584.D
 Acq: 14 Oct 2024 22:55

Tgt Ion: 88 Resp: 1925
 Ion Ratio Lower Upper
 88 100
 43 134.3 22.8 34.2#
 58 97.5 56.4 84.6#









#3
n-Nitrosodimethylamine
Concen: 0.462 ng
RT: 3.321 min Scan# 61
Delta R.T. -0.007 min
Lab File: BN034584.D
Acq: 14 Oct 2024 22:55




Instrument :
BNA_N
ClientSampleId :
PB164072BSD


Tgt Ion: 42 Resp: 2642
Ion Ratio Lower Upper
42 100
74 125.9 99.0 148.6
44 7.3 6.6 9.8



#4
2-Fluorophenol
Concen: 0.342 ng
RT: 5.011 min Scan# 295
Delta R.T. 0.007 min
Lab File: BN034584.D
Acq: 14 Oct 2024 22:55

Tgt Ion: 112 Resp: 4829
Ion Ratio Lower Upper
112 100
64 56.2 44.6 66.8
63 30.0 23.8 35.6





#5

Phenol-d6

Concen: 0.338 ng

RT: 6.578 min Scan# 51

Delta R.T. 0.014 min

Lab File: BN034584.D

Acq: 14 Oct 2024 22:55

Instrument :

BNA_N

ClientSampleId :

PB164072BSD





Tgt Ion: 99 Resp: 4934

Ion Ratio Lower Upper

99 100

42 18.2 15.0 22.6

71 34.6 28.0 42.0



#6

bis(2-Chloroethyl)ether

Concen: 0.399 ng

RT: 6.816 min Scan# 545

Delta R.T. 0.007 min

Lab File: BN034584.D

Acq: 14 Oct 2024 22:55




Tgt Ion: 93 Resp: 4769


Ion Ratio Lower Upper

93 100

63 77.6 64.6 96.8

95 34.0 27.9 41.9









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.116 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN034584.D
Acq: 14 Oct 2024 22:55




Instrument :
BNA_N
ClientSampleId :
PB164072BSD


Tgt Ion:	136	Resp:	14775
Ion	Ratio	Lower	Upper
136	100		
137	11.7	9.4	14.0
54	7.2	6.2	9.4
68	5.1	4.6	6.8



#8
Nitrobenzene-d5
Concen: 0.394 ng
RT: 8.504 min Scan# 770
Delta R.T. -0.011 min
Lab File: BN034584.D
Acq: 14 Oct 2024 22:55

Tgt Ion:	82	Resp:	3793
Ion	Ratio	Lower	Upper
82	100		
128	42.5	38.0	57.0
54	54.0	40.6	60.8









#9
Naphthalene
Concen: 0.424 ng
RT: 10.169 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN034584.D
Acq: 14 Oct 2024 22:55




Instrument :
BNA_N
ClientSampleId :
PB164072BSD


Tgt Ion:128 Resp: 16400
Ion Ratio Lower Upper
128 100
129 11.8 9.3 13.9
127 13.2 10.8 16.2



#10
Hexachlorobutadiene
Concen: 0.400 ng
RT: 10.458 min Scan# 953
Delta R.T. -0.000 min
Lab File: BN034584.D
Acq: 14 Oct 2024 22:55

Tgt Ion:225 Resp: 3376
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 65.1 51.4 77.0









#11
2-Methylnaphthalene-d10
Concen: 0.541 ng
RT: 11.719 min Scan# 1108
Delta R.T. -0.004 min
Lab File: BN034584.D
Acq: 14 Oct 2024 22:55




Instrument :
BNA_N
ClientSampleId :
PB164072BSD


Tgt Ion:152 Resp: 11978
Ion Ratio Lower Upper
152 100
151 14.6 16.7 25.1#



#12
2-Methylnaphthalene
Concen: 0.412 ng
RT: 11.795 min Scan# 1128
Delta R.T. -0.004 min
Lab File: BN034584.D
Acq: 14 Oct 2024 22:55

Tgt Ion:142 Resp: 10255
Ion Ratio Lower Upper
142 100
141 88.4 70.6 105.8
115 36.9 29.2 43.8









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.008 min Scan# 1475
Delta R.T. -0.011 min
Lab File: BN034584.D
Acq: 14 Oct 2024 22:55




Instrument :
BNA_N
ClientSampleId :
PB164072BSD


Tgt Ion:164 Resp: 7913
Ion Ratio Lower Upper
164 100
162 103.9 79.7 119.5
160 49.8 37.9 56.9



#14
2,4,6-Tribromophenol
Concen: 0.286 ng
RT: 15.530 min Scan# 1616
Delta R.T. -0.000 min
Lab File: BN034584.D
Acq: 14 Oct 2024 22:55

Tgt Ion:330 Resp: 1004
Ion Ratio Lower Upper
330 100
332 127.4 104.3 156.5
141 35.5 29.4 44.0









#15
2-Fluorobiphenyl
Concen: 0.388 ng
RT: 12.629 min Scan# 1345
Delta R.T. -0.011 min
Lab File: BN034584.D
Acq: 14 Oct 2024 22:55




Instrument :
BNA_N
ClientSampleId :
PB164072BSD


Tgt Ion:172 Resp: 10600
Ion Ratio Lower Upper
172 100
171 35.5 27.1 40.7
170 24.0 18.5 27.7



#16
Acenaphthylene
Concen: 0.409 ng
RT: 13.730 min Scan# 1448
Delta R.T. -0.000 min
Lab File: BN034584.D
Acq: 14 Oct 2024 22:55

Tgt Ion:152 Resp: 13168
Ion Ratio Lower Upper
152 100
151 19.4 15.5 23.3
153 13.1 10.4 15.6





#17

Acenaphthene

Concen: 0.410 ng

RT: 14.072 min Scan# 1480

Delta R.T. -0.000 min

Lab File: BN034584.D

Acq: 14 Oct 2024 22:55

Instrument :

BNA_N

ClientSampleId :

PB164072BSD





Tgt Ion:154 Resp: 9344

Ion Ratio Lower Upper

154 100

153 114.3 91.5 137.3

152 56.6 45.8 68.8



#18

Fluorene

Concen: 0.377 ng

RT: 15.077 min Scan# 1574

Delta R.T. -0.000 min

Lab File: BN034584.D

Acq: 14 Oct 2024 22:55




Tgt Ion:166 Resp: 10977


Ion Ratio Lower Upper

166 100

165 98.5 78.6 118.0

167 13.6 10.8 16.2









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.771 min Scan# 1716
Delta R.T. -0.012 min
Lab File: BN034584.D
Acq: 14 Oct 2024 22:55




Instrument :
BNA_N
ClientSampleId :
PB164072BSD


Tgt Ion:188 Resp: 13541
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 12.9 6.3 9.5#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.392 ng
RT: 16.287 min Scan# 1677
Delta R.T. 0.012 min
Lab File: BN034584.D
Acq: 14 Oct 2024 22:55

Tgt Ion:198 Resp: 73
Ion Ratio Lower Upper
198 100
51 160.3 174.8 262.2#
105 165.4 167.9 251.9#









#21
4-Bromophenyl-phenylether
Concen: 0.414 ng
RT: 15.989 min Scan# 1653
Delta R.T. -0.000 min
Lab File: BN034584.D
Acq: 14 Oct 2024 22:55




Instrument :
BNA_N
ClientSampleId :
PB164072BSD


Tgt Ion:	248	Resp:	3041
Ion Ratio	Lower	Upper	
248	100		
250	98.9	79.8	119.6
141	58.1	44.3	66.5



#22
Hexachlorobenzene
Concen: 0.445 ng
RT: 16.088 min Scan# 1661
Delta R.T. -0.000 min
Lab File: BN034584.D
Acq: 14 Oct 2024 22:55

Tgt Ion:	284	Resp:	4412
Ion Ratio	Lower	Upper	
284	100		
142	34.5	28.4	42.6
249	31.3	25.0	37.4





#23

Atrazine

Concen: 0.386 ng

RT: 16.287 min Scan# 1677

Delta R.T. -0.000 min

Lab File: BN034584.D

Acq: 14 Oct 2024 22:55

Instrument :

BNA_N

ClientSampleId :

PB164072BSD





Tgt Ion:200 Resp: 2257

Ion Ratio Lower Upper

200 100

173 27.6 22.9 34.3

215 49.0 38.8 58.2



#24

Pentachlorophenol

Concen: 0.674 ng

RT: 16.448 min Scan# 1690

Delta R.T. -0.013 min

Lab File: BN034584.D

Acq: 14 Oct 2024 22:55




Tgt Ion:266 Resp: 1913


Ion Ratio Lower Upper

266 100

264 63.7 46.6 69.8

268 64.5 52.0 78.0









#25
 Phenanthrene
 Concen: 0.428 ng
 RT: 16.821 min Scan# 1720
 Delta R.T. -0.000 min
 Lab File: BN034584.D
 Acq: 14 Oct 2024 22:55

Instrument :
 BNA_N
 ClientSampleId :
 PB164072BSD

Tgt Ion:178 Resp: 15303




Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.3	22.9
179	15.4	12.2	18.4




#26
 Anthracene
 Concen: 0.414 ng
 RT: 16.908 min Scan# 1727
 Delta R.T. -0.000 min
 Lab File: BN034584.D
 Acq: 14 Oct 2024 22:55

Tgt Ion:178 Resp: 12849

Ion	Ratio	Lower	Upper
178	100		
176	18.4	14.3	21.5
179	15.3	12.0	18.0









#27
Fluoranthene-d10
Concen: 0.358 ng
RT: 18.825 min Scan# 1924
Delta R.T. -0.000 min
Lab File: BN034584.D
Acq: 14 Oct 2024 22:55




Instrument :
BNA_N
ClientSampleId :
PB164072BSD


Tgt Ion:212 Resp: 11794
Ion Ratio Lower Upper
212 100
106 13.9 11.1 16.7
104 7.9 6.5 9.7



#28
Fluoranthene
Concen: 0.370 ng
RT: 18.853 min Scan# 1930
Delta R.T. -0.000 min
Lab File: BN034584.D
Acq: 14 Oct 2024 22:55

Tgt Ion:202 Resp: 15043
Ion Ratio Lower Upper
202 100
101 10.6 8.8 13.2
203 17.1 13.4 20.0





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.008 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN034584.D

Acq: 14 Oct 2024 22:55

Instrument :

BNA_N

ClientSampleId :

PB164072BSD





Tgt Ion:240 Resp: 5062

Ion Ratio Lower Upper

240 100

120 14.7 10.1 15.1

236 28.7 22.6 34.0



#30

Pyrene

Concen: 0.311 ng

RT: 19.220 min Scan# 2009

Delta R.T. -0.000 min

Lab File: BN034584.D

Acq: 14 Oct 2024 22:55




Tgt Ion:202 Resp: 15424


Ion Ratio Lower Upper

202 100

200 20.8 16.6 24.8

203 17.7 14.3 21.5





#31

Terphenyl-d14

Concen: 0.397 ng

RT: 19.452 min Scan# 2059

Delta R.T. -0.000 min

Lab File: BN034584.D

Acq: 14 Oct 2024 22:55

Instrument :

BNA_N

ClientSampleId :

PB164072BSD





Tgt Ion:244 Resp: 7375

Ion Ratio Lower Upper

244 100

212 10.2 7.7 11.5

122 16.3 12.6 19.0



#33

Chrysene

Concen: 0.358 ng

RT: 21.035 min Scan# 2397

Delta R.T. -0.000 min

Lab File: BN034584.D

Acq: 14 Oct 2024 22:55




Tgt Ion:228 Resp: 16224


Ion Ratio Lower Upper

228 100

226 30.3 24.4 36.6

229 20.2 15.8 23.6









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.104 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: BN034584.D
 Acq: 14 Oct 2024 22:55




Instrument :
 BNA_N
 ClientSampleId :
 PB164072BSD


Tgt Ion:264 Resp: 9196
 Ion Ratio Lower Upper
 264 100
 260 26.1 20.9 31.3
 265 84.7 61.6 92.4



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.525 ng
 RT: 25.159 min Scan# 3492
 Delta R.T. -0.006 min
 Lab File: BN034584.D
 Acq: 14 Oct 2024 22:55

Tgt Ion:276 Resp: 18763
 Ion Ratio Lower Upper
 276 100
 138 26.2 22.0 33.0
 277 24.6 20.1 30.1





#37

Benzo(b)fluoranthene

Concen: 0.416 ng

RT: 22.484 min Scan# 2577

Delta R.T. -0.006 min

Lab File: BN034584.D


Acq: 14 Oct 2024 22:55

Instrument :

BNA_N

ClientSampleId :

PB164072BSD






Tgt Ion:252 Resp: 12783

Ion Ratio Lower Upper

252 100

253 26.4 21.3 31.9

125 16.5 12.7 19.1



#38

Benzo(k)fluoranthene

Concen: 0.426 ng

RT: 22.528 min Scan# 2592

Delta R.T. -0.003 min

Lab File: BN034584.D

Acq: 14 Oct 2024 22:55




Tgt Ion:252 Resp: 14905


Ion Ratio Lower Upper

252 100

253 26.0 21.0 31.6


125 16.7 13.3 19.9








#39
Benzo(a)pyrene
Concen: 0.473 ng
RT: 23.016 min Scan# 21
Delta R.T. -0.003 min
Lab File: BN034584.D
Acq: 14 Oct 2024 22:55

Instrument :
BNA_N
ClientSampleId :
PB164072BSD







Tgt Ion:252 Resp: 12624
Ion Ratio Lower Upper
252 100
253 30.2 23.8 35.8
125 20.4 17.2 25.8



#40
Dibenzo(a,h)anthracene
Concen: 0.519 ng
RT: 25.180 min Scan# 3499
Delta R.T. -0.000 min
Lab File: BN034584.D
Acq: 14 Oct 2024 22:55


Tgt Ion:278 Resp: 14187
Ion Ratio Lower Upper
278 100
139 20.0 16.0 24.0
279 27.1 23.4 35.0





#41
Benzo(g,h,i)perylene
Concen: 0.494 ng
RT: 25.779 min Scan# 31
Delta R.T. -0.009 min
Lab File: BN034584.D
Acq: 14 Oct 2024 22:55

Instrument :
BNA_N
ClientSampleId :
PB164072BSD



Tgt Ion:	276	Resp:	16135
Ion	Ratio	Lower	Upper
276	100		
277	25.1	19.8	29.8
138	24.9	19.2	28.8

