


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101624\
Data File : BN034605.D
Acq On : 16 Oct 2024 11:21
Operator : RC/JU
Sample : DFTPP362
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP362

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101624.MA.M
Title : SVOA CALIBRATION
Last Update : Wed Oct 16 15:50:29 2024



AutoFind: Scans 2487, 2488, 2489; Background Corrected with Scan 2482

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.5	195021	PASS
68	69	0.00	2	1.6	2994	PASS
69	198	0.00	100	39.5	190041	PASS
70	69	0.00	2	0.5	927	PASS
127	198	10	80	50.1	241389	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	481408	PASS
199	198	5	9	6.9	33133	PASS
275	198	10	60	23.6	113843	PASS
365	198	1	100	2.6	12403	PASS
441	198	0.01	100	11.6	55925	PASS
442	198	50	100	74.5	358656	PASS
443	442	15	24	19.7	70587	PASS

m/z	Abundance
36.00	207.0
39.00	441.0
40.00	4026.0
41.10	868.0
41.90	308.0
43.00	729.0
44.00	9734.0
45.00	328.0
49.00	458.0
50.90	328.0
53.90	223.0
55.10	598.0
56.00	331.0
57.00	837.0
58.00	258.0
59.00	209.0
64.90	153.0
66.90	242.0
68.10	257.0
69.00	423.0
70.10	311.0
71.10	555.0
73.00	574.0
77.00	393.0
78.00	270.0
79.00	180.0
81.00	200.0
81.90	176.0
83.00	324.0
84.00	411.0
85.10	234.0
85.90	255.0
91.10	232.0
95.00	218.0
96.00	395.0
97.10	230.0
98.00	227.0
105.00	204.0
111.10	172.0
115.00	178.0
116.90	201.0
119.10	181.0
127.10	190.0
128.10	164.0
133.00	260.0
147.00	247.0
190.90	309.0
192.90	243.0
207.10	2840.0
207.90	416.0
208.90	432.0
280.90	365.0
281.90	201.0

Instrument :
BNA_N
ClientSampleId :
DFTPP362

m/z	Abundance
35.90	217.0
37.00	704.0
38.00	1912.0
39.00	12624.0
40.00	3239.0
41.00	1363.0
42.00	313.0
43.00	1059.0
44.00	9297.0
44.90	611.0
49.10	1612.0
50.00	43360.0
51.00	174848.0
52.00	8669.0
53.00	476.0
54.00	214.0
55.00	1619.0
56.00	5385.0
57.00	12405.0
58.00	844.0
58.90	249.0
60.00	254.0
61.00	2028.0
62.00	2561.0
63.00	7413.0
64.00	988.0
65.00	3376.0
66.00	375.0
67.00	342.0
68.00	2975.0
69.00	164160.0
70.00	996.0
71.00	493.0
73.00	1614.0
74.00	15340.0
75.00	25480.0
76.00	8834.0
77.00	195200.0
78.00	13412.0
79.00	12152.0
80.00	8988.0
81.00	13123.0
82.00	3471.0
83.00	3382.0
83.90	695.0
85.00	1902.0
86.00	3636.0
87.00	1713.0
88.00	618.0
88.90	384.0
91.00	3038.0
92.00	3324.0
93.00	20200.0
94.00	1697.0
95.00	447.0
96.00	1353.0
97.00	594.0
98.00	15251.0
99.00	12313.0
99.90	1082.0
101.00	6903.0
101.90	397.0
103.00	2225.0
104.00	4159.0
105.00	4530.0
106.00	1299.0
107.00	54920.0
108.00	8614.0
109.10	1589.0
110.00	105280.0
111.00	16277.0
112.00	1928.0
113.00	503.0
114.00	158.0
114.90	309.0
116.00	3011.0
117.00	40928.0
118.00	3168.0
118.90	459.0
120.00	732.0
121.00	317.0
122.00	3779.0
123.00	5680.0
124.00	2412.0
125.00	2298.0
127.00	202048.0
128.00	15350.0
129.00	72320.0
130.00	6762.0
130.90	1305.0
132.00	621.0
132.80	458.0

Instrument :
BNA_N
ClientSampleId :
DFTPP362

133.90	2272.0
135.00	5718.0
136.00	2331.0
137.00	3281.0
138.00	563.0
139.00	358.0
139.90	935.0
141.00	9215.0
142.00	3220.0
143.00	2333.0
143.90	543.0
145.00	617.0
145.90	1770.0
147.00	4778.0
148.00	10340.0
149.00	2487.0
150.00	602.0
151.10	1319.0
152.00	783.0
153.00	2671.0
154.00	2300.0
155.00	4909.0
156.00	7365.0
157.00	1447.0
158.00	1641.0
159.00	1190.0
160.00	2980.0
161.00	4247.0
162.00	1113.0
162.90	385.0
164.00	511.0
165.00	3264.0
165.90	2743.0
167.00	17840.0
168.00	9805.0
169.00	1562.0
170.00	579.0
171.00	859.0
172.00	1488.0
173.00	2168.0
174.00	3511.0
175.00	7483.0
176.00	2009.0
177.00	3388.0
177.90	1181.0
179.00	13259.0
180.00	8951.0
181.00	4014.0
181.90	595.0
182.90	440.0
183.90	1013.0
185.00	5831.0
186.00	49216.0
187.00	13862.0
188.00	1325.0
189.00	2680.0
190.00	536.0
191.00	1687.0
192.00	3876.0
193.00	4867.0
194.00	1025.0
194.90	618.0
196.00	10335.0
198.00	358592.0
198.90	26208.0
200.00	2008.0
201.40	1627.0
202.90	1987.0
204.00	12205.0
205.00	20168.0
206.00	88040.0
207.00	13539.0
208.00	3491.0
209.00	1400.0
211.00	3351.0
211.70	891.0
213.00	212.0
214.90	766.0
216.00	1806.0
217.00	22504.0
218.00	3243.0
219.00	346.0
221.10	14683.0
221.80	3380.0
223.00	4616.0
224.00	44968.0
225.00	11688.0
226.00	1533.0
227.00	18032.0
227.90	2902.0
229.00	4024.0
229.90	454.0
231.10	1515.0
232.00	277.0
233.00	402.0
233.90	1083.0

Instrument :
BNA_N
ClientSampleId :
DFTPP362

235.00	1360.0
236.00	986.0
237.00	1536.0
238.00	171.0
238.90	755.0
240.00	637.0
241.00	909.0
242.00	2631.0
243.00	2552.0
244.00	36176.0
245.00	5091.0
246.00	6191.0
247.00	1510.0
247.90	349.0
249.00	1275.0
250.00	342.0
250.90	300.0
251.90	226.0
253.00	853.0
255.00	177472.0
256.00	27016.0
257.00	1999.0
258.00	8760.0
259.00	1462.0
259.80	263.0
260.90	313.0
264.00	368.0
265.00	3806.0
265.90	588.0
267.20	187.0
270.00	197.0
270.90	393.0
272.10	265.0
273.00	5068.0
274.00	13071.0
275.00	80128.0
276.00	11662.0
277.00	6102.0
278.00	902.0
278.90	273.0
281.00	413.0
281.90	253.0
282.90	697.0
284.00	498.0
285.00	1071.0
286.00	328.0
288.90	295.0
289.90	221.0
292.00	298.0
292.90	1426.0
293.90	363.0
295.10	552.0
296.00	19160.0
297.00	3024.0
298.00	367.0
301.00	272.0
302.00	351.0
303.00	2285.0
304.00	510.0
307.90	449.0
308.80	201.0
310.10	205.0
312.90	272.0
313.90	823.0
315.00	2048.0
316.00	1265.0
321.00	715.0
321.90	261.0
323.10	6434.0
324.00	1544.0
325.00	197.0
326.90	1129.0
328.00	618.0
332.00	341.0
333.00	557.0
334.00	3888.0
335.00	1174.0
340.90	655.0
342.00	219.0
346.00	1434.0
347.00	258.0
352.10	2160.0
353.00	1281.0
354.00	1863.0
355.10	566.0
359.00	151.0
365.00	8304.0
365.90	1390.0
371.00	419.0
372.00	3497.0
373.10	929.0
382.90	848.0
383.90	227.0
390.10	507.0
391.00	468.0
391.90	241.0

Instrument :
BNA_N
ClientSampleId :
DFTPP362

401.00	265.0
401.90	1201.0
403.00	2124.0
404.00	807.0
421.00	1597.0
422.00	1429.0
423.00	10452.0
424.00	2330.0
424.90	178.0
440.00	257.0
441.10	28376.0
442.00	207680.0
443.00	38600.0
444.00	3696.0
444.90	222.0

Instrument :
BNA_N
ClientSampleId :
DFTPP362

m/z	Abundance
36.00	286.0
37.00	1003.0
38.00	2703.0
39.00	15617.0
40.00	3843.0
41.00	1330.0
42.10	313.0
43.00	1091.0
44.00	8985.0
45.00	660.0
48.00	213.0
49.10	1912.0
50.00	56872.0
51.00	227392.0
52.00	12089.0
53.10	576.0
54.00	196.0
55.10	1592.0
56.00	7382.0
57.00	15556.0
58.00	967.0
58.90	386.0
60.00	211.0
61.00	2910.0
62.00	3123.0
63.00	9062.0
64.00	1277.0
65.00	4686.0
66.00	320.0
67.00	620.0
68.00	3616.0
69.00	222912.0
70.00	1407.0
71.00	534.0
72.00	197.0
73.00	1944.0
74.00	21784.0
75.00	35320.0
76.00	12537.0
77.00	262208.0
78.00	18288.0
79.00	17736.0
80.00	12948.0
81.00	17544.0
82.00	5045.0
83.00	4686.0
84.00	851.0
85.00	2747.0
86.00	5047.0
87.00	2343.0
88.00	1100.0
89.00	488.0
89.90	192.0
91.00	3684.0
92.00	4619.0
93.00	28264.0
94.00	1906.0
95.00	695.0
96.00	1602.0
97.10	898.0
98.00	21152.0
99.00	17456.0
100.00	1487.0
101.00	10585.0
102.00	628.0
103.00	3613.0
104.00	6789.0
105.00	6221.0
106.00	1872.0
107.00	77656.0
108.00	12250.0
109.00	2334.0
110.00	145920.0
111.00	22256.0
112.00	2537.0
112.90	850.0
113.90	183.0
115.00	479.0
116.00	4556.0
117.00	57712.0
117.90	4079.0
118.90	686.0
120.00	1015.0
121.00	389.0
122.00	4974.0
123.00	7712.0
124.00	3337.0
125.00	3347.0
126.10	1384.0
127.00	279232.0
128.00	21464.0
129.00	104288.0

Instrument :
BNA_N
ClientSampleId :
DFTPP362

130.00	9787.0
131.00	1752.0
132.10	896.0
133.00	610.0
134.00	3065.0
135.00	8775.0
136.00	3451.0
137.00	4164.0
138.10	928.0
139.00	487.0
140.00	1496.0
141.00	13111.0
142.00	4598.0
143.00	3006.0
143.90	775.0
145.00	752.0
145.90	2525.0
147.00	6312.0
148.00	14652.0
149.00	3050.0
150.00	707.0
151.00	1840.0
151.90	976.0
153.00	4443.0
154.00	3285.0
155.00	8064.0
156.00	11079.0
157.00	2159.0
158.00	2623.0
159.00	1932.0
160.00	4386.0
161.00	6258.0
161.90	2040.0
163.00	512.0
164.00	555.0
165.00	5352.0
166.00	3898.0
167.00	26128.0
168.00	13630.0
169.00	2217.0
170.00	998.0
170.90	1102.0
171.90	2437.0
173.00	3063.0
174.00	5691.0
175.00	10230.0
176.10	3133.0
177.00	4547.0
178.00	1614.0
179.00	19544.0
180.00	12634.0
181.00	6335.0
182.00	985.0
182.80	539.0
184.00	1594.0
185.00	9294.0
186.00	74040.0
187.00	21000.0
188.00	2011.0
189.00	4010.0
190.00	699.0
191.00	2484.0
192.00	6720.0
193.00	5916.0
193.90	1431.0
195.00	937.0
196.00	15763.0
198.00	555456.0
199.00	38400.0
200.00	2466.0
201.60	2326.0
202.90	3352.0
204.00	17280.0
205.00	31976.0
206.00	137152.0
207.10	19664.0
208.00	5111.0
209.00	1657.0
210.20	2227.0
211.00	5010.0
211.80	964.0
213.00	414.0
214.90	1562.0
216.00	2857.0
217.00	32736.0
218.00	4196.0
219.00	469.0
221.00	24144.0
223.00	7627.0
224.00	73864.0
225.00	18640.0
226.00	1929.0
227.00	28144.0
228.00	3564.0
229.00	6214.0
229.90	749.0

Instrument :
BNA_N
ClientSampleId :
DFTPP362

231.00	2446.0
232.00	477.0
232.80	425.0
234.00	1983.0
235.00	2274.0
236.00	1265.0
237.00	2426.0
238.10	295.0
239.00	1299.0
240.00	1024.0
241.00	1923.0
242.00	3574.0
243.10	3962.0
244.00	56440.0
245.10	7299.0
246.00	10089.0
246.90	2226.0
247.90	398.0
249.00	2036.0
249.90	509.0
251.00	572.0
252.00	711.0
252.90	1086.0
255.00	276032.0
256.00	40088.0
257.00	3364.0
258.00	14198.0
259.00	2530.0
260.00	540.0
261.00	478.0
263.00	180.0
264.10	349.0
265.00	6010.0
266.00	740.0
267.10	181.0
269.80	155.0
270.90	754.0
272.00	628.0
273.00	8268.0
274.00	22200.0
275.00	136064.0
276.00	18224.0
277.00	10545.0
278.00	1544.0
279.00	324.0
281.00	349.0
282.00	326.0
283.00	1260.0
284.10	762.0
285.00	1541.0
286.00	173.0
288.90	347.0
290.00	288.0
291.10	288.0
292.00	599.0
293.00	2609.0
293.90	552.0
295.10	684.0
296.00	31952.0
297.00	4585.0
297.90	360.0
301.00	452.0
302.00	686.0
303.00	3659.0
304.00	926.0
307.90	503.0
309.00	332.0
310.00	540.0
310.90	159.0
313.00	381.0
314.00	1563.0
315.00	3166.0
316.00	1974.0
317.00	343.0
320.00	156.0
321.00	995.0
322.00	774.0
323.00	10939.0
324.00	2006.0
325.10	198.0
326.10	171.0
327.00	1965.0
327.90	984.0
328.90	212.0
331.90	757.0
333.00	1058.0
334.00	6905.0
335.00	1701.0
336.00	217.0
338.80	151.0
341.10	933.0
342.00	304.0
345.90	2649.0
347.00	591.0
350.90	280.0
352.00	3505.0

Instrument :
BNA_N
ClientSampleId :
DFTPP362

353.00	2070.0
354.00	3609.0
355.00	728.0
359.00	361.0
365.00	14058.0
365.90	2133.0
367.00	269.0
370.10	381.0
371.00	917.0
372.00	5652.0
373.00	1175.0
373.90	220.0
382.90	1630.0
384.00	637.0
390.00	907.0
391.00	550.0
392.00	527.0
401.00	368.0
402.00	2442.0
403.00	3786.0
404.00	1060.0
405.00	217.0
415.00	249.0
421.00	3436.0
422.00	2496.0
423.00	22752.0
424.00	5347.0
425.00	401.0
438.90	161.0
440.20	726.0
441.00	60752.0
442.00	420672.0
443.00	77856.0
444.00	7297.0
444.90	400.0

Instrument :
BNA_N
ClientSampleId :
DFTPP362

m/z	Abundance
35.00	266.0
37.00	767.0
38.00	2118.0
39.00	12795.0
40.00	3557.0
41.00	964.0
42.00	281.0
43.00	954.0
44.00	9199.0
45.00	667.0
49.00	1633.0
50.00	45136.0
51.00	183808.0
52.00	9202.0
53.00	495.0
55.00	1623.0
56.00	5805.0
57.00	13098.0
58.00	884.0
59.10	289.0
61.00	2177.0
62.00	2453.0
63.00	7280.0
64.00	1047.0
65.00	3846.0
66.00	394.0
67.00	392.0
68.00	3163.0
69.00	184320.0
70.00	1311.0
71.10	510.0
73.00	2023.0
74.00	17744.0
75.00	29352.0
76.00	10006.0
77.00	213312.0
78.00	14864.0
79.00	14382.0
80.00	10279.0
81.00	15225.0
82.00	3769.0
83.00	3513.0
84.00	737.0
85.00	2551.0
86.00	4143.0
87.00	1626.0
88.00	572.0
89.10	458.0
91.00	3510.0
92.00	3602.0
93.00	23408.0
94.00	1779.0
95.00	709.0
96.00	1682.0
97.00	682.0
98.00	17208.0
99.00	13721.0
100.00	1057.0
101.00	8440.0
101.90	446.0
103.00	3102.0
104.00	5435.0
105.00	5448.0
106.00	1592.0
107.00	63792.0
108.00	9375.0
109.10	1754.0
110.00	127528.0
111.00	18232.0
112.00	2073.0
113.00	606.0
114.00	173.0
115.00	380.0
116.00	3792.0
117.00	50848.0
118.00	3326.0
119.00	646.0
120.00	852.0
120.90	335.0
122.00	4343.0
123.00	7032.0
124.00	2893.0
125.00	2557.0
127.00	243456.0
128.00	18992.0
129.00	90112.0
130.00	8606.0
131.00	1414.0
132.00	824.0
133.00	596.0
134.00	2609.0
135.00	7655.0

Instrument :
BNA_N
ClientSampleId :
DFTPP362

136.00	3034.0
137.00	4004.0
138.00	676.0
138.90	625.0
140.00	1440.0
141.00	11831.0
142.00	4066.0
143.00	2273.0
144.00	619.0
144.90	726.0
146.00	2185.0
147.00	5952.0
148.00	13687.0
149.00	2593.0
149.90	791.0
151.10	1480.0
153.00	3475.0
154.00	2736.0
155.00	6029.0
156.00	9488.0
157.00	2069.0
158.00	1877.0
159.00	1376.0
159.90	3631.0
161.00	5939.0
162.00	1581.0
162.90	483.0
164.10	703.0
165.00	4577.0
166.00	3683.0
167.00	23584.0
168.00	11018.0
169.00	2054.0
170.00	807.0
171.00	1041.0
172.00	2255.0
173.00	2625.0
174.00	4808.0
175.00	9476.0
176.00	2730.0
177.00	4392.0
178.00	1476.0
179.00	17192.0
180.00	11039.0
181.00	6339.0
182.00	1006.0
182.90	576.0
184.00	1311.0
185.00	8438.0
186.00	67512.0
187.00	19688.0
188.00	2281.0
189.00	4218.0
190.00	739.0
191.10	2070.0
192.00	5480.0
193.00	6139.0
194.00	1325.0
194.90	759.0
196.00	15354.0
198.00	530176.0
199.00	34792.0
200.00	2187.0
201.40	2176.0
203.00	3210.0
204.00	17304.0
205.10	28256.0
206.00	132032.0
207.00	17912.0
208.00	4501.0
208.90	1546.0
210.20	2188.0
211.00	5029.0
211.70	1392.0
213.00	359.0
214.00	177.0
215.00	1304.0
216.00	2574.0
217.00	31648.0
218.00	4207.0
219.00	473.0
221.00	22144.0
223.00	6482.0
224.00	69584.0
225.00	17576.0
226.00	1940.0
227.00	26928.0
228.00	4147.0
229.00	5462.0
229.90	927.0
231.00	2730.0
231.90	384.0
232.90	454.0
234.00	1847.0
235.00	2006.0
235.90	1441.0

Instrument :
BNA_N
ClientSampleId :
DFTPP362

237.00	2355.0
238.00	393.0
239.00	1086.0
240.00	682.0
241.00	1585.0
242.00	3198.0
243.10	4028.0
244.00	56696.0
245.10	7798.0
246.00	9735.0
247.00	1861.0
247.90	411.0
249.00	2043.0
250.00	428.0
250.90	456.0
252.10	427.0
253.00	1064.0
255.00	284096.0
256.00	40368.0
257.00	2947.0
258.00	14663.0
259.00	2405.0
260.00	416.0
261.00	443.0
262.90	161.0
264.10	411.0
265.00	6238.0
266.00	1052.0
269.90	252.0
271.00	527.0
272.00	704.0
273.00	8123.0
274.00	23272.0
275.00	125336.0
276.00	17552.0
277.00	9254.0
278.00	1814.0
278.90	354.0
281.00	415.0
282.00	352.0
283.00	1262.0
284.00	690.0
285.00	2157.0
285.90	435.0
288.90	406.0
289.90	364.0
290.80	275.0
292.10	357.0
293.00	2542.0
294.00	654.0
295.10	839.0
296.00	33960.0
297.00	4512.0
298.10	220.0
301.00	558.0
301.90	622.0
303.00	3860.0
304.00	1195.0
305.00	174.0
307.90	508.0
308.90	386.0
310.00	571.0
313.00	219.0
314.00	1526.0
315.00	3641.0
316.00	2050.0
317.00	366.0
321.00	1227.0
322.10	534.0
323.10	10992.0
324.00	2412.0
325.00	206.0
325.80	224.0
326.90	2208.0
327.90	1425.0
329.00	215.0
331.90	670.0
333.00	1072.0
334.00	7239.0
335.00	2047.0
335.90	168.0
338.90	211.0
341.00	1459.0
342.00	305.0
346.00	2605.0
346.90	420.0
351.10	198.0
352.00	3402.0
353.00	2687.0
354.00	3635.0
355.00	707.0
359.00	182.0
364.00	209.0
365.00	14846.0
366.00	2322.0
367.00	248.0

Instrument :
BNA_N
ClientSampleId :
DFTPP362

369.90	441.0
371.00	817.0
372.00	6802.0
373.10	1678.0
373.90	226.0
377.00	247.0
383.00	1729.0
383.80	400.0
389.90	871.0
391.00	472.0
392.00	426.0
400.90	559.0
402.00	2750.0
403.00	3376.0
404.00	1577.0
405.00	195.0
414.90	229.0
421.10	3391.0
422.10	3447.0
423.00	26560.0
424.00	5020.0
425.00	727.0
438.90	187.0
439.30	189.0
441.00	78648.0
442.10	447616.0
443.00	95304.0
444.10	7386.0
445.00	589.0

Instrument :
BNA_N
ClientSampleId :
DFTPP362