


Data Path : Z:\SVOASRV\HPCHEM1\BNA\_N\DATA\BN101720\  
 Data File : BN012286.D  
 Acq On : 17 Oct 2020 10:55  
 Operator : CG/JU  
 Sample : DFTPP73  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 DFTPP73

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SOM-EPA-BN101720.M  
 Title : SVOA CALIBRATION  
 Last Update : Sat Oct 17 15:47:34 2020



AutoFind: Scans 2457, 2458, 2459; Background Corrected with Scan 2452

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 38.7      | 147570  | PASS             |
| 68          | 69           | 0.00         | 2            | 0.8       | 1250    | PASS             |
| 69          | 198          | 0.00         | 100          | 41.5      | 158189  | PASS             |
| 70          | 69           | 0.00         | 2            | 0.6       | 994     | PASS             |
| 127         | 198          | 10           | 80           | 45.6      | 173909  | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 381141  | PASS             |
| 199         | 198          | 5            | 9            | 6.4       | 24405   | PASS             |
| 275         | 198          | 10           | 60           | 28.9      | 109992  | PASS             |
| 365         | 198          | 1            | 100          | 4.0       | 15289   | PASS             |
| 441         | 198          | 0.01         | 100          | 15.5      | 59029   | PASS             |
| 442         | 198          | 50           | 100          | 97.3      | 371008  | PASS             |
| 443         | 442          | 15           | 24           | 18.1      | 67285   | PASS             |

| m/z    | Abundance |
|--------|-----------|
| 39.10  | 479.0     |
| 40.00  | 941.0     |
| 40.90  | 606.0     |
| 42.00  | 363.0     |
| 43.10  | 541.0     |
| 44.00  | 3593.0    |
| 48.90  | 335.0     |
| 51.00  | 310.0     |
| 55.10  | 385.0     |
| 57.20  | 499.0     |
| 73.00  | 1059.0    |
| 77.00  | 464.0     |
| 83.80  | 588.0     |
| 133.10 | 378.0     |
| 135.00 | 375.0     |
| 147.10 | 575.0     |
| 190.80 | 335.0     |
| 193.00 | 313.0     |
| 207.00 | 3173.0    |
| 208.10 | 814.0     |
| 209.10 | 363.0     |
| 281.10 | 799.0     |
| 281.90 | 331.0     |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP73

| m/z    | Abundance |
|--------|-----------|
| 37.00  | 714.0     |
| 38.00  | 1649.0    |
| 39.00  | 11903.0   |
| 40.00  | 1353.0    |
| 41.00  | 814.0     |
| 43.00  | 591.0     |
| 44.00  | 2396.0    |
| 45.00  | 464.0     |
| 49.10  | 1919.0    |
| 50.00  | 44312.0   |
| 51.00  | 149120.0  |
| 52.00  | 8063.0    |
| 54.90  | 1123.0    |
| 56.00  | 5462.0    |
| 57.00  | 9924.0    |
| 59.00  | 326.0     |
| 61.20  | 1468.0    |
| 62.00  | 2752.0    |
| 63.00  | 6156.0    |
| 63.90  | 767.0     |
| 65.00  | 2560.0    |
| 65.90  | 341.0     |
| 67.00  | 605.0     |
| 68.00  | 3134.0    |
| 69.00  | 163328.0  |
| 70.00  | 771.0     |
| 71.00  | 314.0     |
| 73.00  | 3101.0    |
| 74.00  | 17488.0   |
| 75.00  | 27080.0   |
| 76.10  | 7153.0    |
| 77.00  | 171136.0  |
| 78.10  | 12686.0   |
| 79.00  | 13639.0   |
| 80.00  | 9730.0    |
| 81.00  | 13990.0   |
| 81.90  | 3156.0    |
| 83.00  | 3401.0    |
| 83.90  | 709.0     |
| 85.00  | 2060.0    |
| 86.00  | 3366.0    |
| 87.10  | 1537.0    |
| 88.00  | 742.0     |
| 90.90  | 2884.0    |
| 91.90  | 3374.0    |
| 93.00  | 21960.0   |
| 93.90  | 1446.0    |
| 95.00  | 476.0     |
| 96.00  | 1014.0    |
| 98.00  | 16960.0   |
| 99.00  | 14486.0   |
| 99.90  | 896.0     |
| 101.00 | 6750.0    |
| 102.00 | 317.0     |
| 103.00 | 2396.0    |
| 103.90 | 5125.0    |
| 105.00 | 4555.0    |
| 105.90 | 1083.0    |
| 107.00 | 51664.0   |
| 108.00 | 8348.0    |
| 109.00 | 2147.0    |
| 110.00 | 94952.0   |
| 111.00 | 14067.0   |
| 112.00 | 1977.0    |
| 113.10 | 915.0     |
| 114.50 | 363.0     |
| 116.00 | 2959.0    |
| 117.00 | 45720.0   |
| 118.00 | 3451.0    |
| 118.90 | 630.0     |
| 120.00 | 542.0     |
| 120.90 | 593.0     |
| 122.00 | 4049.0    |
| 123.00 | 5293.0    |
| 123.90 | 2766.0    |
| 125.00 | 1918.0    |
| 125.90 | 1117.0    |
| 127.00 | 164544.0  |
| 128.00 | 15612.0   |
| 129.00 | 68840.0   |
| 130.00 | 6754.0    |
| 131.10 | 1226.0    |
| 132.00 | 833.0     |
| 133.00 | 593.0     |
| 133.90 | 2530.0    |
| 135.00 | 4877.0    |
| 136.00 | 2593.0    |
| 137.00 | 2393.0    |
| 138.00 | 512.0     |
| 139.00 | 559.0     |
| 140.00 | 815.0     |
| 141.00 | 8685.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP73

|        |          |
|--------|----------|
| 142.00 | 2739.0   |
| 142.90 | 2866.0   |
| 144.00 | 505.0    |
| 144.80 | 765.0    |
| 146.00 | 2240.0   |
| 147.00 | 4887.0   |
| 148.00 | 11437.0  |
| 149.10 | 2247.0   |
| 149.90 | 569.0    |
| 151.00 | 1092.0   |
| 152.10 | 385.0    |
| 153.00 | 3494.0   |
| 153.90 | 2604.0   |
| 155.10 | 5399.0   |
| 156.00 | 7479.0   |
| 157.10 | 1622.0   |
| 157.90 | 2017.0   |
| 159.00 | 1569.0   |
| 160.10 | 2691.0   |
| 161.00 | 3961.0   |
| 161.90 | 1941.0   |
| 163.10 | 393.0    |
| 164.10 | 402.0    |
| 165.00 | 3261.0   |
| 166.10 | 2329.0   |
| 167.00 | 18256.0  |
| 168.00 | 9504.0   |
| 169.00 | 1551.0   |
| 170.00 | 479.0    |
| 170.70 | 517.0    |
| 171.20 | 370.0    |
| 172.00 | 1468.0   |
| 173.00 | 1959.0   |
| 173.90 | 4610.0   |
| 175.00 | 6655.0   |
| 176.00 | 1370.0   |
| 176.90 | 3255.0   |
| 177.90 | 1333.0   |
| 179.00 | 13407.0  |
| 180.00 | 8746.0   |
| 181.00 | 4552.0   |
| 182.10 | 872.0    |
| 183.00 | 578.0    |
| 183.80 | 886.0    |
| 185.00 | 6571.0   |
| 186.00 | 48648.0  |
| 187.00 | 12641.0  |
| 187.90 | 1991.0   |
| 188.90 | 3313.0   |
| 190.00 | 662.0    |
| 191.00 | 1879.0   |
| 192.00 | 4346.0   |
| 193.00 | 5635.0   |
| 194.10 | 1000.0   |
| 195.00 | 456.0    |
| 196.00 | 10535.0  |
| 198.00 | 338240.0 |
| 199.00 | 21624.0  |
| 200.00 | 2171.0   |
| 201.40 | 1477.0   |
| 202.10 | 471.0    |
| 203.00 | 2574.0   |
| 204.00 | 12349.0  |
| 205.00 | 21816.0  |
| 206.00 | 86688.0  |
| 207.00 | 14148.0  |
| 208.00 | 4133.0   |
| 209.10 | 1903.0   |
| 210.00 | 1583.0   |
| 211.00 | 3293.0   |
| 215.00 | 1141.0   |
| 216.00 | 2569.0   |
| 216.90 | 22200.0  |
| 217.90 | 2864.0   |
| 221.00 | 14704.0  |
| 221.80 | 3445.0   |
| 223.00 | 4539.0   |
| 224.10 | 45616.0  |
| 225.00 | 14563.0  |
| 226.10 | 2054.0   |
| 227.00 | 21464.0  |
| 228.00 | 3326.0   |
| 229.00 | 3901.0   |
| 229.90 | 743.0    |
| 231.00 | 1188.0   |
| 231.90 | 404.0    |
| 233.10 | 477.0    |
| 234.00 | 1296.0   |
| 235.00 | 1469.0   |
| 236.00 | 1201.0   |
| 237.00 | 2300.0   |
| 239.00 | 1427.0   |
| 240.10 | 763.0    |
| 241.00 | 861.0    |
| 241.90 | 2516.0   |
| 243.00 | 2471.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP73

|        |          |
|--------|----------|
| 244.00 | 35312.0  |
| 245.00 | 4751.0   |
| 246.00 | 6187.0   |
| 247.00 | 1407.0   |
| 248.00 | 594.0    |
| 249.00 | 1389.0   |
| 250.80 | 408.0    |
| 251.80 | 325.0    |
| 253.00 | 829.0    |
| 255.00 | 185216.0 |
| 256.00 | 27128.0  |
| 257.10 | 2103.0   |
| 258.00 | 10918.0  |
| 259.00 | 1701.0   |
| 259.90 | 404.0    |
| 261.00 | 462.0    |
| 265.00 | 4570.0   |
| 266.00 | 565.0    |
| 267.10 | 370.0    |
| 270.00 | 400.0    |
| 270.90 | 376.0    |
| 272.10 | 532.0    |
| 273.00 | 5558.0   |
| 274.00 | 15185.0  |
| 275.00 | 88008.0  |
| 276.00 | 11981.0  |
| 277.00 | 8065.0   |
| 277.90 | 1433.0   |
| 279.00 | 530.0    |
| 281.20 | 843.0    |
| 283.10 | 1081.0   |
| 283.90 | 546.0    |
| 285.10 | 1552.0   |
| 293.00 | 1771.0   |
| 293.90 | 520.0    |
| 296.00 | 24632.0  |
| 297.00 | 3836.0   |
| 297.80 | 308.0    |
| 303.10 | 2878.0   |
| 303.90 | 723.0    |
| 308.90 | 407.0    |
| 313.10 | 405.0    |
| 314.00 | 1377.0   |
| 315.00 | 2898.0   |
| 316.00 | 2056.0   |
| 316.90 | 471.0    |
| 320.90 | 548.0    |
| 321.80 | 338.0    |
| 322.10 | 318.0    |
| 323.10 | 7886.0   |
| 323.90 | 1635.0   |
| 326.90 | 1587.0   |
| 328.10 | 860.0    |
| 331.90 | 431.0    |
| 333.10 | 918.0    |
| 334.10 | 4477.0   |
| 335.10 | 893.0    |
| 341.00 | 751.0    |
| 342.00 | 347.0    |
| 345.90 | 1812.0   |
| 347.00 | 337.0    |
| 352.00 | 3082.0   |
| 353.00 | 1414.0   |
| 354.10 | 2468.0   |
| 355.00 | 739.0    |
| 365.00 | 11294.0  |
| 366.00 | 1276.0   |
| 367.00 | 360.0    |
| 370.90 | 890.0    |
| 372.00 | 4543.0   |
| 373.00 | 1058.0   |
| 383.00 | 1157.0   |
| 383.90 | 353.0    |
| 390.10 | 766.0    |
| 391.00 | 362.0    |
| 402.10 | 1300.0   |
| 403.00 | 2151.0   |
| 403.90 | 1018.0   |
| 421.00 | 1567.0   |
| 421.90 | 1787.0   |
| 423.00 | 13247.0  |
| 424.00 | 4445.0   |
| 424.80 | 485.0    |
| 441.10 | 39280.0  |
| 442.00 | 244672.0 |
| 443.00 | 45120.0  |
| 444.00 | 4072.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP73

| m/z    | Abundance |
|--------|-----------|
| 36.90  | 1123.0    |
| 38.00  | 2467.0    |
| 39.10  | 14303.0   |
| 39.90  | 1805.0    |
| 41.00  | 742.0     |
| 43.10  | 511.0     |
| 43.90  | 3142.0    |
| 44.90  | 773.0     |
| 49.00  | 1720.0    |
| 50.00  | 48264.0   |
| 51.00  | 184192.0  |
| 52.00  | 9171.0    |
| 53.00  | 424.0     |
| 54.90  | 1261.0    |
| 56.00  | 4791.0    |
| 57.00  | 11236.0   |
| 57.90  | 616.0     |
| 58.80  | 318.0     |
| 61.00  | 2796.0    |
| 62.10  | 3008.0    |
| 63.10  | 7330.0    |
| 64.10  | 1428.0    |
| 65.00  | 4132.0    |
| 66.00  | 328.0     |
| 67.00  | 573.0     |
| 68.20  | 3752.0    |
| 69.00  | 191936.0  |
| 70.00  | 1416.0    |
| 73.10  | 2842.0    |
| 74.00  | 21168.0   |
| 75.00  | 30448.0   |
| 76.10  | 11401.0   |
| 77.00  | 206400.0  |
| 78.10  | 14667.0   |
| 79.00  | 15492.0   |
| 80.00  | 12189.0   |
| 81.00  | 15333.0   |
| 82.00  | 3997.0    |
| 83.00  | 2973.0    |
| 83.90  | 456.0     |
| 85.00  | 2566.0    |
| 86.00  | 4352.0    |
| 87.00  | 2023.0    |
| 87.90  | 566.0     |
| 88.90  | 613.0     |
| 91.00  | 3704.0    |
| 92.00  | 3862.0    |
| 93.00  | 29136.0   |
| 94.00  | 2005.0    |
| 95.20  | 496.0     |
| 96.00  | 1290.0    |
| 97.00  | 598.0     |
| 98.00  | 21296.0   |
| 99.00  | 15542.0   |
| 100.00 | 1198.0    |
| 101.00 | 8613.0    |
| 102.00 | 604.0     |
| 103.10 | 2385.0    |
| 104.00 | 5483.0    |
| 104.90 | 5480.0    |
| 106.10 | 2236.0    |
| 107.00 | 63888.0   |
| 108.00 | 9922.0    |
| 109.00 | 2491.0    |
| 110.00 | 112824.0  |
| 111.00 | 18216.0   |
| 112.00 | 2325.0    |
| 112.90 | 699.0     |
| 115.90 | 3848.0    |
| 117.00 | 59328.0   |
| 118.00 | 5028.0    |
| 118.90 | 680.0     |
| 120.00 | 860.0     |
| 121.00 | 485.0     |
| 122.00 | 4709.0    |
| 123.00 | 6076.0    |
| 124.00 | 3511.0    |
| 125.00 | 2476.0    |
| 127.00 | 209600.0  |
| 128.00 | 16728.0   |
| 129.00 | 87064.0   |
| 130.00 | 7802.0    |
| 131.00 | 1750.0    |
| 132.10 | 756.0     |
| 132.90 | 593.0     |
| 134.00 | 2631.0    |
| 135.00 | 7554.0    |
| 136.00 | 2350.0    |
| 137.00 | 3575.0    |
| 137.90 | 889.0     |
| 138.80 | 553.0     |
| 140.10 | 1347.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP73

|        |          |
|--------|----------|
| 141.00 | 11178.0  |
| 142.00 | 4504.0   |
| 142.90 | 3003.0   |
| 143.90 | 714.0    |
| 145.10 | 561.0    |
| 146.00 | 1743.0   |
| 147.00 | 6280.0   |
| 148.00 | 13543.0  |
| 149.00 | 2574.0   |
| 150.00 | 1216.0   |
| 151.10 | 1556.0   |
| 152.10 | 882.0    |
| 153.00 | 3508.0   |
| 154.00 | 2581.0   |
| 155.00 | 6356.0   |
| 156.00 | 9511.0   |
| 157.10 | 1620.0   |
| 157.90 | 2707.0   |
| 159.00 | 1684.0   |
| 160.10 | 3675.0   |
| 161.00 | 4822.0   |
| 162.00 | 1804.0   |
| 162.90 | 630.0    |
| 164.00 | 615.0    |
| 164.90 | 5323.0   |
| 166.00 | 3945.0   |
| 167.00 | 22080.0  |
| 168.00 | 12621.0  |
| 169.00 | 1551.0   |
| 170.10 | 594.0    |
| 171.00 | 1445.0   |
| 172.00 | 2091.0   |
| 173.00 | 2348.0   |
| 174.00 | 5063.0   |
| 175.00 | 10209.0  |
| 176.00 | 2181.0   |
| 177.00 | 4146.0   |
| 178.00 | 1476.0   |
| 178.90 | 17216.0  |
| 180.00 | 11666.0  |
| 181.00 | 5107.0   |
| 181.90 | 869.0    |
| 182.90 | 480.0    |
| 184.00 | 1588.0   |
| 185.00 | 7537.0   |
| 186.00 | 62096.0  |
| 187.00 | 18920.0  |
| 188.00 | 2557.0   |
| 189.10 | 4272.0   |
| 189.90 | 544.0    |
| 191.00 | 2078.0   |
| 192.00 | 5015.0   |
| 193.00 | 6040.0   |
| 193.90 | 1382.0   |
| 194.80 | 577.0    |
| 196.00 | 12403.0  |
| 198.00 | 461056.0 |
| 199.00 | 31416.0  |
| 200.00 | 2305.0   |
| 201.30 | 1685.0   |
| 203.10 | 3606.0   |
| 204.00 | 16744.0  |
| 205.00 | 27336.0  |
| 206.00 | 111952.0 |
| 207.10 | 17664.0  |
| 208.10 | 4460.0   |
| 209.00 | 1504.0   |
| 210.00 | 2334.0   |
| 211.00 | 5098.0   |
| 213.10 | 385.0    |
| 215.00 | 1536.0   |
| 216.00 | 3125.0   |
| 217.00 | 29456.0  |
| 218.00 | 4322.0   |
| 219.10 | 490.0    |
| 219.90 | 450.0    |
| 221.00 | 17008.0  |
| 223.00 | 8218.0   |
| 224.10 | 63456.0  |
| 225.10 | 16257.0  |
| 226.10 | 2083.0   |
| 227.00 | 29536.0  |
| 227.90 | 3621.0   |
| 229.00 | 6556.0   |
| 230.00 | 932.0    |
| 231.00 | 2141.0   |
| 231.90 | 387.0    |
| 232.80 | 544.0    |
| 233.90 | 1888.0   |
| 235.10 | 1998.0   |
| 236.00 | 1437.0   |
| 237.10 | 2391.0   |
| 239.10 | 759.0    |
| 239.90 | 918.0    |
| 241.00 | 1414.0   |
| 242.00 | 4260.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP73

|        |          |
|--------|----------|
| 243.00 | 3623.0   |
| 244.00 | 48352.0  |
| 245.10 | 7345.0   |
| 246.00 | 11539.0  |
| 247.00 | 2480.0   |
| 248.20 | 426.0    |
| 249.00 | 2547.0   |
| 249.90 | 580.0    |
| 250.80 | 670.0    |
| 252.00 | 697.0    |
| 253.10 | 1028.0   |
| 255.00 | 277440.0 |
| 256.00 | 38960.0  |
| 257.00 | 3638.0   |
| 258.00 | 16432.0  |
| 259.00 | 2162.0   |
| 260.00 | 693.0    |
| 260.90 | 597.0    |
| 265.00 | 5717.0   |
| 266.00 | 1025.0   |
| 267.00 | 328.0    |
| 269.90 | 543.0    |
| 270.90 | 540.0    |
| 272.00 | 857.0    |
| 273.00 | 7527.0   |
| 274.00 | 20392.0  |
| 275.00 | 134912.0 |
| 276.00 | 16084.0  |
| 277.00 | 11415.0  |
| 278.00 | 1900.0   |
| 278.90 | 342.0    |
| 281.00 | 1621.0   |
| 282.00 | 510.0    |
| 283.10 | 1362.0   |
| 284.00 | 675.0    |
| 285.00 | 2326.0   |
| 285.80 | 462.0    |
| 288.90 | 813.0    |
| 289.90 | 481.0    |
| 290.90 | 344.0    |
| 291.80 | 528.0    |
| 293.00 | 2909.0   |
| 294.00 | 449.0    |
| 296.00 | 35200.0  |
| 297.00 | 4915.0   |
| 297.90 | 386.0    |
| 301.00 | 533.0    |
| 301.90 | 849.0    |
| 303.00 | 4006.0   |
| 304.00 | 985.0    |
| 305.10 | 341.0    |
| 308.00 | 595.0    |
| 308.90 | 491.0    |
| 310.00 | 398.0    |
| 313.90 | 1604.0   |
| 315.00 | 3689.0   |
| 316.00 | 1974.0   |
| 316.70 | 370.0    |
| 321.10 | 1023.0   |
| 322.00 | 574.0    |
| 323.00 | 11798.0  |
| 324.00 | 1759.0   |
| 326.90 | 2395.0   |
| 328.00 | 1375.0   |
| 332.10 | 675.0    |
| 333.00 | 1038.0   |
| 334.00 | 7996.0   |
| 334.90 | 2559.0   |
| 341.00 | 1094.0   |
| 342.00 | 444.0    |
| 346.00 | 1710.0   |
| 346.90 | 675.0    |
| 352.00 | 3733.0   |
| 353.00 | 3520.0   |
| 354.00 | 3926.0   |
| 354.90 | 669.0    |
| 358.90 | 342.0    |
| 365.00 | 18344.0  |
| 365.90 | 2365.0   |
| 369.90 | 630.0    |
| 371.10 | 1105.0   |
| 372.00 | 5648.0   |
| 372.90 | 1421.0   |
| 382.90 | 1284.0   |
| 384.00 | 471.0    |
| 389.90 | 583.0    |
| 390.90 | 680.0    |
| 391.90 | 469.0    |
| 401.10 | 538.0    |
| 402.10 | 3021.0   |
| 403.00 | 3736.0   |
| 404.00 | 1350.0   |
| 415.00 | 308.0    |
| 420.90 | 3053.0   |
| 422.00 | 3188.0   |
| 423.00 | 16856.0  |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP73



|        |          |
|--------|----------|
| 424.00 | 5792.0   |
| 424.90 | 828.0    |
| 441.00 | 71928.0  |
| 442.00 | 446592.0 |
| 443.00 | 80864.0  |
| 444.00 | 7558.0   |
| 445.00 | 414.0    |

**Instrument :**  
BNA\_N  
**ClientSampleId :**  
DFTPP73

| m/z    | Abundance |
|--------|-----------|
| 37.10  | 588.0     |
| 38.00  | 1615.0    |
| 39.10  | 7171.0    |
| 40.00  | 2114.0    |
| 41.10  | 547.0     |
| 43.00  | 713.0     |
| 44.00  | 3790.0    |
| 44.90  | 347.0     |
| 49.00  | 1463.0    |
| 50.10  | 31624.0   |
| 51.00  | 110328.0  |
| 52.10  | 5747.0    |
| 52.90  | 382.0     |
| 54.90  | 1279.0    |
| 56.00  | 3783.0    |
| 57.00  | 8076.0    |
| 58.00  | 443.0     |
| 58.80  | 374.0     |
| 61.10  | 1404.0    |
| 62.00  | 1386.0    |
| 63.00  | 5083.0    |
| 64.10  | 657.0     |
| 65.00  | 2548.0    |
| 66.00  | 409.0     |
| 66.90  | 439.0     |
| 69.00  | 119304.0  |
| 69.90  | 797.0     |
| 73.00  | 2458.0    |
| 74.00  | 12822.0   |
| 75.00  | 21368.0   |
| 76.10  | 6903.0    |
| 77.00  | 134528.0  |
| 78.00  | 9539.0    |
| 79.00  | 9633.0    |
| 79.90  | 7189.0    |
| 81.00  | 10021.0   |
| 82.00  | 2818.0    |
| 83.00  | 2156.0    |
| 84.00  | 484.0     |
| 85.00  | 1531.0    |
| 86.00  | 3473.0    |
| 86.90  | 1326.0    |
| 88.00  | 691.0     |
| 89.00  | 643.0     |
| 91.00  | 2762.0    |
| 92.00  | 2735.0    |
| 93.00  | 18480.0   |
| 94.10  | 1105.0    |
| 94.90  | 352.0     |
| 96.00  | 1187.0    |
| 97.20  | 458.0     |
| 98.00  | 13361.0   |
| 99.00  | 10592.0   |
| 100.00 | 984.0     |
| 101.00 | 4847.0    |
| 103.00 | 2339.0    |
| 104.00 | 4265.0    |
| 105.00 | 2567.0    |
| 107.00 | 41616.0   |
| 108.00 | 5757.0    |
| 109.20 | 2412.0    |
| 110.00 | 74592.0   |
| 111.00 | 10795.0   |
| 111.90 | 1169.0    |
| 112.90 | 401.0     |
| 115.00 | 389.0     |
| 116.10 | 3013.0    |
| 117.00 | 37248.0   |
| 118.00 | 2763.0    |
| 119.00 | 439.0     |
| 120.00 | 471.0     |
| 121.10 | 425.0     |
| 122.00 | 2745.0    |
| 123.10 | 4133.0    |
| 124.00 | 2393.0    |
| 125.00 | 1908.0    |
| 127.00 | 147584.0  |
| 128.00 | 12279.0   |
| 129.00 | 59200.0   |
| 130.00 | 4038.0    |
| 131.00 | 1162.0    |
| 132.00 | 707.0     |
| 132.90 | 757.0     |
| 134.00 | 1847.0    |
| 135.00 | 5609.0    |
| 135.90 | 1814.0    |
| 137.00 | 2461.0    |
| 137.90 | 388.0     |
| 140.00 | 616.0     |
| 141.00 | 9019.0    |
| 142.00 | 3197.0    |
| 142.90 | 2266.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP73

|        |          |
|--------|----------|
| 143.80 | 662.0    |
| 144.00 | 630.0    |
| 145.00 | 533.0    |
| 146.00 | 1697.0   |
| 147.00 | 3850.0   |
| 148.00 | 8566.0   |
| 149.00 | 2657.0   |
| 150.00 | 865.0    |
| 151.10 | 983.0    |
| 151.80 | 705.0    |
| 153.00 | 2955.0   |
| 153.90 | 2224.0   |
| 155.10 | 4821.0   |
| 156.00 | 6648.0   |
| 157.00 | 1355.0   |
| 158.00 | 1912.0   |
| 158.90 | 1764.0   |
| 160.00 | 2732.0   |
| 161.00 | 3568.0   |
| 162.10 | 981.0    |
| 163.00 | 435.0    |
| 164.00 | 701.0    |
| 165.10 | 3329.0   |
| 166.00 | 2417.0   |
| 167.00 | 17632.0  |
| 168.00 | 8824.0   |
| 169.10 | 1567.0   |
| 170.00 | 597.0    |
| 171.00 | 775.0    |
| 171.90 | 2359.0   |
| 172.90 | 1776.0   |
| 174.10 | 3304.0   |
| 175.10 | 6666.0   |
| 176.00 | 1515.0   |
| 176.90 | 2977.0   |
| 178.20 | 1027.0   |
| 179.00 | 12982.0  |
| 180.00 | 8184.0   |
| 181.00 | 4011.0   |
| 182.10 | 822.0    |
| 183.90 | 1264.0   |
| 185.00 | 6826.0   |
| 186.00 | 44664.0  |
| 187.00 | 13202.0  |
| 188.10 | 1087.0   |
| 189.00 | 3278.0   |
| 190.00 | 605.0    |
| 191.00 | 1907.0   |
| 192.00 | 4540.0   |
| 193.00 | 4400.0   |
| 194.00 | 718.0    |
| 195.00 | 645.0    |
| 196.00 | 10552.0  |
| 198.00 | 344128.0 |
| 199.00 | 20176.0  |
| 200.10 | 1804.0   |
| 201.50 | 1937.0   |
| 203.00 | 2298.0   |
| 204.00 | 14082.0  |
| 205.00 | 21408.0  |
| 206.00 | 92216.0  |
| 207.00 | 14129.0  |
| 208.00 | 3599.0   |
| 209.00 | 1517.0   |
| 210.00 | 1553.0   |
| 211.10 | 3228.0   |
| 212.00 | 681.0    |
| 215.00 | 1575.0   |
| 215.90 | 1862.0   |
| 217.00 | 25360.0  |
| 218.00 | 2950.0   |
| 220.20 | 752.0    |
| 221.00 | 13934.0  |
| 222.00 | 3676.0   |
| 223.00 | 4942.0   |
| 224.10 | 49256.0  |
| 225.00 | 10887.0  |
| 226.10 | 1051.0   |
| 227.00 | 19656.0  |
| 228.10 | 2402.0   |
| 229.00 | 4063.0   |
| 229.80 | 529.0    |
| 231.10 | 1899.0   |
| 232.00 | 369.0    |
| 233.00 | 497.0    |
| 234.00 | 1324.0   |
| 235.00 | 1666.0   |
| 235.90 | 1135.0   |
| 237.00 | 2006.0   |
| 237.90 | 312.0    |
| 239.00 | 924.0    |
| 240.00 | 842.0    |
| 240.90 | 1724.0   |
| 242.00 | 2572.0   |
| 243.10 | 3381.0   |
| 244.00 | 39480.0  |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP73

|        |          |
|--------|----------|
| 245.10 | 5000.0   |
| 246.00 | 8494.0   |
| 246.90 | 1449.0   |
| 248.00 | 466.0    |
| 249.10 | 1368.0   |
| 249.90 | 348.0    |
| 250.90 | 371.0    |
| 251.90 | 329.0    |
| 253.10 | 839.0    |
| 255.00 | 206336.0 |
| 256.00 | 31048.0  |
| 257.00 | 2858.0   |
| 258.00 | 13289.0  |
| 259.10 | 1399.0   |
| 259.90 | 490.0    |
| 260.90 | 349.0    |
| 264.10 | 329.0    |
| 265.00 | 5603.0   |
| 266.10 | 771.0    |
| 271.00 | 625.0    |
| 271.90 | 534.0    |
| 273.00 | 5262.0   |
| 274.10 | 16408.0  |
| 275.00 | 107056.0 |
| 276.10 | 13124.0  |
| 277.00 | 9527.0   |
| 277.90 | 1382.0   |
| 279.00 | 451.0    |
| 281.10 | 1402.0   |
| 281.90 | 588.0    |
| 283.00 | 1403.0   |
| 284.00 | 850.0    |
| 285.00 | 1257.0   |
| 289.00 | 573.0    |
| 292.00 | 674.0    |
| 293.00 | 2018.0   |
| 294.00 | 983.0    |
| 295.10 | 737.0    |
| 296.00 | 31048.0  |
| 297.00 | 3880.0   |
| 300.90 | 326.0    |
| 302.10 | 704.0    |
| 303.00 | 3622.0   |
| 304.20 | 749.0    |
| 308.90 | 489.0    |
| 312.90 | 341.0    |
| 314.10 | 1854.0   |
| 315.00 | 3280.0   |
| 316.00 | 1275.0   |
| 317.00 | 404.0    |
| 320.90 | 959.0    |
| 322.10 | 693.0    |
| 323.00 | 10182.0  |
| 324.10 | 1999.0   |
| 324.90 | 501.0    |
| 326.90 | 1609.0   |
| 328.10 | 976.0    |
| 331.90 | 682.0    |
| 332.90 | 885.0    |
| 334.00 | 6119.0   |
| 335.00 | 1607.0   |
| 340.20 | 335.0    |
| 341.10 | 1466.0   |
| 346.10 | 1989.0   |
| 352.00 | 3459.0   |
| 353.00 | 1703.0   |
| 354.10 | 2807.0   |
| 355.00 | 627.0    |
| 359.10 | 392.0    |
| 365.00 | 16230.0  |
| 365.90 | 1654.0   |
| 371.00 | 840.0    |
| 372.10 | 5495.0   |
| 373.00 | 1516.0   |
| 382.90 | 1499.0   |
| 390.00 | 635.0    |
| 392.00 | 427.0    |
| 401.00 | 409.0    |
| 402.00 | 2620.0   |
| 403.00 | 3032.0   |
| 404.10 | 1028.0   |
| 420.90 | 2686.0   |
| 422.00 | 3699.0   |
| 423.00 | 17704.0  |
| 424.00 | 4111.0   |
| 424.80 | 389.0    |
| 441.00 | 65880.0  |
| 442.00 | 421760.0 |
| 443.00 | 75872.0  |
| 444.00 | 6576.0   |
| 445.00 | 605.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP73