


Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN101722\  
 Data File : BN022141.D  
 Acq On : 17 Oct 2022 08:16  
 Operator : CG/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN100322.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Thu Oct 13 01:51:34 2022






AutoFind: Scans 2520, 2521, 2522; Background Corrected with Scan 2513

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 41.2         | 123811     | PASS                |
| 68             | 69              | 0.00            | 2               | 1.7          | 2269       | PASS                |
| 69             | 198             | 0.00            | 100             | 43.7         | 131384     | PASS                |
| 70             | 69              | 0.00            | 2               | 0.5          | 708        | PASS                |
| 127            | 198             | 10              | 80              | 54.1         | 162667     | PASS                |
| 197            | 198             | 0.00            | 2               | 0.5          | 1464       | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 300544     | PASS                |
| 199            | 198             | 5               | 9               | 7.3          | 21992      | PASS                |
| 275            | 198             | 10              | 60              | 26.5         | 79579      | PASS                |
| 365            | 198             | 1               | 100             | 3.8          | 11329      | PASS                |
| 441            | 198             | 0.01            | 100             | 13.6         | 40952      | PASS                |
| 442            | 442             | 50              | 100             | 100.0        | 243755     | PASS                |
| 443            | 442             | 15              | 24              | 20.7         | 50571      | PASS                |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN101722\  
Data File : BN022141.D  
Acq On : 17 Oct 2022 08:16  
Operator : CG/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Quant Time: Oct 18 07:19:14 2022  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Oct 18 07:19:09 2022  
Response via : Initial Calibration



TIC: BN022141.D\data.ms

**(70) Pentachlorophenol (C)**

17.028min ( 0.000) 26009.69 ng




response 427422

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 62.20  | 67.44  |
| 264.00 | 61.60  | 64.88  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN101722\  
Data File : BN022141.D  
Acq On : 17 Oct 2022 08:16  
Operator : CG/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Quant Time: Oct 18 07:19:14 2022  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Oct 18 07:19:09 2022  
Response via : Initial Calibration



TIC: BN022141.D\data.ms

(77) Benzidine

19.639min ( 0.000) 222192.98 ng

response 2493102

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.50  | 14.32  |
| 183.00 | 13.20  | 11.62  |
| 0.00   | 0.00   | 0.00   |

### DDT Breakdown

| Date          | Instrument Name  | DFTPP Data File    |
|---------------|------------------|--------------------|
| 10/17/2022    | BNA_N            | <u>BN022141.D</u>  |
|               |                  |                    |
|               |                  |                    |
| Compound Name | Response         | Retention Time     |
|               |                  |                    |
| DDT           | 943307           | 20.867             |
| DDD           | 193722           | 20.427             |
| DDE           | 8617             | 19.922             |
|               |                  |                    |
|               |                  |                    |
| SUM(DDD+DDE)  | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
|               |                  |                    |
| 202339        | 1145646          | 17.66              |
|               |                  |                    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP