

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101724\
 Data File : BN034649.D
 Acq On : 18 Oct 2024 13:41
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Oct 18 22:56:16 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101624.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 16 15:50:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.663	152	325959	20.000	ng/u1	0.00
20) Naphthalene-d8	10.440	136	1382892	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.298	164	751048	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.039	188	1253506	20.000	ng/u1	0.00
79) Chrysene-d12	21.263	240	916827	20.000	ng/u1	0.00
88) Perylene-d12	24.151	264	1078462	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.211	96	64158	7.876	ng/uL	0.00
4) Pyridine-d5	3.599	84	473071	19.890	ng/u1	0.00
7) Phenol-d5	6.834	99	547104	20.452	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.005	67	342836	20.777	ng/u1	0.00
11) 2-Chlorophenol-d4	7.199	132	443692	21.471	ng/u1	0.00
15) 4-Methylphenol-d8	8.369	113	429259	20.590	ng/u1	0.00
21) Nitrobenzene-d5	8.810	128	215241	22.366	ng/u1	0.00
24) 2-Nitrophenol-d4	9.522	143	189602	24.932	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.063	165	384022	21.909	ng/u1	0.00
31) 4-Chloroaniline-d4	10.569	131	593165	19.726	ng/u1	0.00
46) Dimethylphthalate-d6	13.716	166	998651	21.018	ng/u1	0.00
49) Acenaphthylene-d8	13.987	160	1254138	21.250	ng/u1	0.00
54) 4-Nitrophenol-d4	14.487	143	162376	18.345	ng/u1	0.00
60) Fluorene-d10	15.292	176	828246	20.154	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.404	200	71453	17.006	ng/u1	0.00
73) Anthracene-d10	17.139	188	1121282	20.926	ng/u1	0.00
81) Pyrene-d10	19.433	212	1052558	22.359	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.951	264	1049787	21.126	ng/u1	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.246	88	70179	8.018	ng/uL	94
5) Pyridine	3.622	79	481672	19.901	ng/u1	98
6) Benzaldehyde	6.811	77	363580	23.155	ng/u1	99
8) Phenol	6.863	94	563323	20.167	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.099	93	449098	20.390	ng/u1	97
12) 2-Chlorophenol	7.228	128	458302	21.236	ng/u1	98
13) 2-Methylphenol	8.105	108	419536	20.346	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.187	45	680610	20.513	ng/u1	99
16) Acetophenone	8.481	105	682820	20.511	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.475	70	329524	20.326	ng/u1	96
18) 4-Methylphenol	8.428	108	457441	20.463	ng/u1	100
19) Hexachloroethane	8.740	117	186365	20.710	ng/u1	96
22) Nitrobenzene	8.852	77	492278	21.385	ng/u1	99
23) Isophorone	9.375	82	922358	20.269	ng/u1	99
25) 2-Nitrophenol	9.557	139	215748	24.125	ng/u1	98
26) 2,4-Dimethylphenol	9.628	107	454410	20.361	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.863	93	581369	20.712	ng/u1	99
29) 2,4-Dichlorophenol	10.087	162	388253	21.893	ng/u1	98
30) Naphthalene	10.487	128	1428672	20.553	ng/u1	99
32) 4-Chloroaniline	10.593	127	559495	19.623	ng/u1	99
33) Hexachlorobutadiene	10.787	225	223198	20.941	ng/u1	98
34) Caprolactam	11.346	113	128791	18.627	ng/u1	92
35) 4-Chloro-3-methylphenol	11.722	107	410624	20.710	ng/u1	98
36) 2-Methylnaphthalene	12.104	142	933586	20.398	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101724\
 Data File : BN034649.D
 Acq On : 18 Oct 2024 13:41
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 10 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :


Quant Time: Oct 18 22:56:16 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101624.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 16 15:50:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.328	142	955317	20.504	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.481	216	404732	21.931	ng/ul	98
40) Hexachlorocyclopentadiene	12.469	237	172822	17.629	ng/ul	98
41) 2,4,6-Trichlorophenol	12.716	196	255464	23.841	ng/ul	96
42) 2,4,5-Trichlorophenol	12.787	196	278876	23.487	ng/ul	95
43) 1,1'-Biphenyl	13.128	154	1149827	22.294	ng/ul	98
44) 2-Chloronaphthalene	13.163	162	883831	21.925	ng/ul	99
45) 2-Nitroaniline	13.363	65	247542	23.409	ng/ul	96
47) Dimethylphthalate	13.763	163	999679	20.947	ng/ul	99
48) 2,6-Dinitrotoluene	13.869	165	195070	23.596	ng/ul	95
50) Acenaphthylene	14.016	152	1364961	21.344	ng/ul	99
51) 3-Nitroaniline	14.192	138	227442	21.554	ng/ul#	92
52) Acenaphthene	14.363	153	928177	20.891	ng/ul	98
53) 2,4-Dinitrophenol	14.398	184	42895	13.301	ng/ul	97
55) 4-Nitrophenol	14.504	109	124149	19.154	ng/ul	97
56) Dibenzofuran	14.698	168	1212384	20.535	ng/ul	100
57) 2,4-Dinitrotoluene	14.657	165	263441	22.344	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	14.922	232	188093	22.100	ng/ul	100
59) Diethylphthalate	15.139	149	1002465	20.703	ng/ul	100
61) Fluorene	15.345	166	973972	20.473	ng/ul	92
62) 4-Chlorophenyl-phenyle...	15.351	204	442962	20.742	ng/ul	98
63) 4-Nitroaniline	15.357	138	220853	21.007	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.422	198	84225	17.327	ng/ul	93
67) N-Nitrosodiphenylamine	15.563	169	809694	22.922	ng/ul	100
68) 4-Bromophenyl-phenylether	16.245	248	247665	22.695	ng/ul	98
69) Hexachlorobenzene	16.351	284	270106	21.556	ng/ul	96
70) Atrazine	16.522	200	247889	20.949	ng/ul	97
71) Pentachlorophenol	16.692	266	136335	20.300	ng/ul	97
72) Phenanthrene	17.081	178	1316015	21.034	ng/ul	99
74) Anthracene	17.175	178	1371224	21.211	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.087	216	396027	25.233	ng/uL	97
76) Pentachlorobenzene	14.616	250	358242	23.851	ng/uL	98
77) Carbazole	17.439	167	1201981	20.079	ng/ul	100
78) Di-n-butylphthalate	18.033	149	1495399	21.469	ng/ul	100
80) Fluoranthene	19.098	202	1257980	22.275	ng/ul	98
82) Pyrene	19.457	202	1365224	22.694	ng/ul	99
83) Butylbenzylphthalate	20.386	149	533271	23.911	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.174	252	348357	18.889	ng/ul	95
85) Benzo(a)anthracene	21.245	228	1225718	21.043	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.216	149	793797	22.684	ng/ul	97
87) Chrysene	21.304	228	1150633	20.811	ng/ul	97
89) Di-n-octyl phthalate	22.327	149	1299729	20.736	ng/ul	100
90) Benzo(b)fluoranthene	23.245	252	1232524	21.268	ng/ul	100
91) Benzo(k)fluoranthene	23.304	252	1244275	21.012	ng/ul	99
93) Benzo(a)pyrene	24.015	252	1187813	21.170	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	27.380	276	1341676	19.167	ng/ul	94
95) Dibenzo(a,h)anthracene	27.445	278	1223313	21.505	ng/ul	99
96) Benzo(g,h,i)perylene	28.374	276	1171515	21.562	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Oct 18 22:56:16 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101624.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Oct 16 15:50:29 2024
Response via : Initial Calibration








#2
 1,4-Dioxane
 Concen: 8.018 ng/uL
 RT: 3.246 min Scan# 61
 Delta R.T. 0.006 min
 Lab File: BN034649.D
 Acq: 18 Oct 2024 13:41

Instrument :
 BNA_N
 ClientSampleId :







Tgt Ion: 88 Resp: 70179
 Ion Ratio Lower Upper
 88 100
 43 26.2 24.5 36.7
 58 76.4 64.5 96.7



#5
 Pyridine
 Concen: 19.901 ng/uL
 RT: 3.622 min Scan# 125
 Delta R.T. 0.006 min
 Lab File: BN034649.D
 Acq: 18 Oct 2024 13:41

Tgt Ion: 79 Resp: 481672
 Ion Ratio Lower Upper
 79 100
 52 77.9 63.9 95.9
 51 34.9 29.2 43.8









#6
Benzaldehyde
Concen: 23.155 ng/ul
RT: 6.811 min Scan# 667
Delta R.T. 0.000 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41




Instrument :
BNA_N
ClientSampleId :


Tgt Ion: 77 Resp: 363580
Ion Ratio Lower Upper
77 100
105 103.0 81.2 121.8
106 98.4 78.3 117.5



#8
Phenol
Concen: 20.167 ng/ul
RT: 6.863 min Scan# 676
Delta R.T. 0.006 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41


Tgt Ion: 94 Resp: 563323
Ion Ratio Lower Upper
94 100
65 27.0 22.2 33.4
66 33.9 27.4 41.0








#10
Bis(2-Chloroethyl)ether
Concen: 20.390 ng/ul
RT: 7.099 min Scan# 716
Delta R.T. 0.000 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41

Instrument :
BNA_N
ClientSampleId :







Tgt Ion: 93 Resp: 449098
Ion Ratio Lower Upper
93 100
63 78.1 65.0 97.6
95 31.4 26.4 39.6



#12
2-Chlorophenol
Concen: 21.236 ng/ul
RT: 7.228 min Scan# 738
Delta R.T. 0.000 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41

Tgt Ion:128 Resp: 458302
Ion Ratio Lower Upper
128 100
64 47.9 39.8 59.8
130 32.1 26.3 39.5





#13

2-Methylphenol

Concen: 20.346 ng/ul

RT: 8.105 min Scan# 81

Delta R.T. 0.006 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41

Instrument :

BNA_N





ClientSampleId :

Tgt Ion:108 Resp: 419536

Ion Ratio Lower Upper

108 100

107 87.6 70.1 105.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.513 ng/ul

RT: 8.187 min Scan# 901

Delta R.T. -0.006 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41




Tgt Ion: 45 Resp: 680610


Ion Ratio Lower Upper

45 100

77 13.5 10.4 15.6

79 9.7 7.4 11.2









#16
Acetophenone
Concen: 20.511 ng/ul
RT: 8.481 min Scan# 91
Delta R.T. 0.006 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:105 Resp: 682820
Ion Ratio Lower Upper
105 100
77 73.6 61.4 92.0
51 26.2 22.1 33.1



#17
N-Nitroso-di-n-propylamine
Concen: 20.326 ng/ul
RT: 8.475 min Scan# 950
Delta R.T. 0.006 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41

Tgt Ion: 70 Resp: 329524
Ion Ratio Lower Upper
70 100
42 57.5 49.8 74.6
101 10.2 8.2 12.4
130 22.0 17.0 25.6









#18
4-Methylphenol
Concen: 20.463 ng/ul
RT: 8.428 min Scan# 942
Delta R.T. 0.000 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:108 Resp: 457441
Ion Ratio Lower Upper
108 100
107 107.3 85.8 128.6



#19
Hexachloroethane
Concen: 20.710 ng/ul
RT: 8.740 min Scan# 995
Delta R.T. 0.000 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41

Tgt Ion:117 Resp: 186365
Ion Ratio Lower Upper
117 100
201 87.6 67.0 100.4
199 52.9 44.7 67.1









#22
Nitrobenzene
Concen: 21.385 ng/ul
RT: 8.852 min Scan# 1014
Delta R.T. 0.000 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41




Instrument :
BNA_N
ClientSampleId :


Tgt Ion: 77 Resp: 492278
Ion Ratio Lower Upper
77 100
123 48.0 37.5 56.3
65 13.4 10.8 16.2



#23
Isophorone
Concen: 20.269 ng/ul
RT: 9.375 min Scan# 1103
Delta R.T. 0.000 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41

Tgt Ion: 82 Resp: 922358
Ion Ratio Lower Upper
82 100
95 6.8 5.3 7.9
138 17.7 13.8 20.6









#25
2-Nitrophenol
Concen: 24.125 ng/ul
RT: 9.557 min Scan# 1134
Delta R.T. 0.000 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:139 Resp: 215748
Ion Ratio Lower Upper
139 100
65 50.3 39.2 58.8
109 28.7 21.8 32.6



#26
2,4-Dimethylphenol
Concen: 20.361 ng/ul
RT: 9.628 min Scan# 1146
Delta R.T. 0.006 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41

Tgt Ion:107 Resp: 454410
Ion Ratio Lower Upper
107 100
121 51.9 41.8 62.6
122 89.8 70.2 105.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.712 ng/ul
 RT: 9.863 min Scan# 1186
 Delta R.T. 0.000 min
 Lab File: BN034649.D
 Acq: 18 Oct 2024 13:41




Instrument :
 BNA_N
 ClientSampleId :


Tgt Ion: 93 Resp: 581369
 Ion Ratio Lower Upper
 93 100
 95 31.9 26.0 39.0
 123 11.3 9.1 13.7



#29
 2,4-Dichlorophenol
 Concen: 21.893 ng/ul
 RT: 10.087 min Scan# 1224
 Delta R.T. 0.000 min
 Lab File: BN034649.D
 Acq: 18 Oct 2024 13:41

Tgt Ion: 162 Resp: 388253
 Ion Ratio Lower Upper
 162 100
 164 66.0 51.9 77.9
 98 37.2 30.6 46.0





#30

Naphthalene

Concen: 20.553 ng/ul

RT: 10.487 min Scan# 1292

Delta R.T. -0.006 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:128 Resp: 1428672

Ion Ratio Lower Upper

128 100

129 11.1 8.2 12.4

127 12.9 10.3 15.5



#32

4-Chloroaniline

Concen: 19.623 ng/ul

RT: 10.593 min Scan# 1310

Delta R.T. 0.000 min

Lab File: BN034649.D




Acq: 18 Oct 2024 13:41


Tgt Ion:127 Resp: 559495

Ion Ratio Lower Upper

127 100


129 31.9 26.0 39.0








#33
Hexachlorobutadiene
Concen: 20.941 ng/ul
RT: 10.787 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41

Instrument :
BNA_N
ClientSampleId :







Tgt Ion:225 Resp: 223198
Ion Ratio Lower Upper
225 100
223 61.1 50.4 75.6
227 63.4 49.8 74.6



#34
Caprolactam
Concen: 18.627 ng/ul
RT: 11.346 min Scan# 1438
Delta R.T. 0.012 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41

Tgt Ion:113 Resp: 128791
Ion Ratio Lower Upper
113 100
55 179.4 152.5 228.7
56 130.8 112.8 169.2









#35
 4-Chloro-3-methylphenol
 Concen: 20.710 ng/ul
 RT: 11.722 min Scan# 1502
 Delta R.T. 0.000 min
 Lab File: BN034649.D
 Acq: 18 Oct 2024 13:41




Instrument :
 BNA_N
 ClientSampleId :


Tgt Ion:107 Resp: 410624
 Ion Ratio Lower Upper
 107 100
 144 25.2 20.7 31.1
 142 78.3 64.1 96.1



#36
 2-Methylnaphthalene
 Concen: 20.398 ng/ul
 RT: 12.104 min Scan# 1567
 Delta R.T. -0.006 min
 Lab File: BN034649.D
 Acq: 18 Oct 2024 13:41

Tgt Ion:142 Resp: 933586
 Ion Ratio Lower Upper
 142 100
 141 89.4 72.3 108.5





#37

1-Methylnaphthalene

Concen: 20.504 ng/ul

RT: 12.328 min Scan# 1605

Delta R.T. 0.000 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:142 Resp: 955317

Ion Ratio Lower Upper

142 100

141 90.9 75.4 113.0

116 3.5 2.8 4.2



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.931 ng/ul

RT: 12.481 min Scan# 1631

Delta R.T. 0.000 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41

Tgt Ion:216 Resp: 404732




Ion Ratio Lower Upper


216 100

214 76.5 63.0 94.6

179 20.6 17.5 26.3

108 18.7 15.5 23.3





#40

Hexachlorocyclopentadiene

Concen: 17.629 ng/ul

RT: 12.469 min Scan# 1629

Delta R.T. 0.000 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:237 Resp: 172822

Ion Ratio Lower Upper

237 100

235 64.0 52.6 78.8

272 13.6 10.1 15.1



#41

2,4,6-Trichlorophenol

Concen: 23.841 ng/ul

RT: 12.716 min Scan# 1671

Delta R.T. -0.006 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41




Tgt Ion:196 Resp: 255464


Ion Ratio Lower Upper

196 100

198 93.4 71.0 106.4

200 30.1 25.3 37.9









#42
2,4,5-Trichlorophenol
Concen: 23.487 ng/ul
RT: 12.787 min Scan# 1683
Delta R.T. 0.000 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:196 Resp: 278876
Ion Ratio Lower Upper
196 100
198 91.4 78.1 117.1
200 29.4 24.8 37.2



#43
1,1'-Biphenyl
Concen: 22.294 ng/ul
RT: 13.128 min Scan# 1741
Delta R.T. -0.006 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41

Tgt Ion:154 Resp: 1149827
Ion Ratio Lower Upper
154 100
153 39.3 32.4 48.6
76 13.6 11.2 16.8





#44

2-Chloronaphthalene

Concen: 21.925 ng/ul

RT: 13.163 min Scan# 1747

Delta R.T. -0.006 min


Lab File: BN034649.D

Acq: 18 Oct 2024 13:41

Instrument :

BNA_N

ClientSampleId :






Tgt Ion:162 Resp: 883831

Ion Ratio Lower Upper

162 100

164 32.6 26.3 39.5

127 40.4 31.9 47.9



#45

2-Nitroaniline

Concen: 23.409 ng/ul

RT: 13.363 min Scan# 1781

Delta R.T. 0.000 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41




Tgt Ion: 65 Resp: 247542


Ion Ratio Lower Upper

65 100

92 68.0 55.1 82.7


138 111.0 84.0 126.0








#47
Dimethylphthalate
Concen: 20.947 ng/ul
RT: 13.763 min Scan# 1849
Delta R.T. -0.006 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41


Instrument :
BNA_N
ClientSampleId :





Tgt Ion:163 Resp: 999679
Ion Ratio Lower Upper
163 100
194 4.0 3.1 4.7
164 10.4 8.0 12.0




#48
2,6-Dinitrotoluene
Concen: 23.596 ng/ul
RT: 13.869 min Scan# 1867
Delta R.T. 0.000 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41



Tgt Ion:165 Resp: 195070
Ion Ratio Lower Upper
165 100
89 58.6 49.7 74.5
121 22.5 20.1 30.1





#50

Acenaphthylene

Concen: 21.344 ng/ul

RT: 14.016 min Scan# 1892

Delta R.T. 0.000 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:152 Resp: 1364961

Ion Ratio Lower Upper

152 100

151 19.2 15.3 22.9

153 12.7 9.9 14.9



#51

3-Nitroaniline

Concen: 21.554 ng/ul

RT: 14.192 min Scan# 1922

Delta R.T. 0.000 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41




Tgt Ion:138 Resp: 227442


Ion Ratio Lower Upper

138 100

108 13.2 13.4 20.0#

92 115.5 100.0 150.0









#52
 Acenaphthene
 Concen: 20.891 ng/ul
 RT: 14.363 min Scan# 1951
 Delta R.T. 0.000 min
 Lab File: BN034649.D
 Acq: 18 Oct 2024 13:41




Instrument :
 BNA_N
 ClientSampleId :


Tgt Ion:	153	Resp:	928177
Ion Ratio	Lower	Upper	
153	100		
152	46.7	36.5	54.7
154	89.0	69.3	103.9



#53
 2,4-Dinitrophenol
 Concen: 13.301 ng/ul
 RT: 14.398 min Scan# 1957
 Delta R.T. 0.000 min
 Lab File: BN034649.D
 Acq: 18 Oct 2024 13:41

Tgt Ion:	184	Resp:	42895
Ion Ratio	Lower	Upper	
184	100		
63	71.1	58.0	87.0
154	64.5	54.7	82.1









#55
4-Nitrophenol
Concen: 19.154 ng/ul
RT: 14.504 min Scan# 1975
Delta R.T. 0.006 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:109 Resp: 124149
Ion Ratio Lower Upper
109 100
139 140.0 113.8 170.6
65 133.5 111.7 167.5



#56
Dibenzofuran
Concen: 20.535 ng/ul
RT: 14.698 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41

Tgt Ion:168 Resp: 1212384
Ion Ratio Lower Upper
168 100
139 37.1 29.7 44.5









#57
 2,4-Dinitrotoluene
 Concen: 22.344 ng/ul
 RT: 14.657 min Scan# 2001
 Delta R.T. 0.000 min
 Lab File: BN034649.D
 Acq: 18 Oct 2024 13:41

Instrument :
 BNA_N
 ClientSampleId :

Tgt Ion:165 Resp: 263441




Ion	Ratio	Lower	Upper
165	100		
63	44.8	37.1	55.7
182	2.5	2.2	3.2




#58
 2,3,4,6-Tetrachlorophenol
 Concen: 22.100 ng/ul
 RT: 14.922 min Scan# 2046
 Delta R.T. 0.000 min
 Lab File: BN034649.D
 Acq: 18 Oct 2024 13:41

Tgt Ion:232 Resp: 188093


Ion	Ratio	Lower	Upper
232	100		
131	54.3	43.5	65.3
130	2.8	2.2	3.4
166	36.9	29.4	44.0








#59
 Diethylphthalate
 Concen: 20.703 ng/ul
 RT: 15.139 min Scan# 2083
 Delta R.T. -0.006 min
 Lab File: BN034649.D
 Acq: 18 Oct 2024 13:41

Instrument :
 BNA_N
 ClientSampleId :



Tgt Ion:149 Resp: 1002465




Ion	Ratio	Lower	Upper
149	100		
177	20.7	16.5	24.7
150	12.3	9.8	14.6




#61
 Fluorene
 Concen: 20.473 ng/ul
 RT: 15.345 min Scan# 2118
 Delta R.T. -0.006 min
 Lab File: BN034649.D
 Acq: 18 Oct 2024 13:41

Tgt Ion:166 Resp: 973972

Ion	Ratio	Lower	Upper
166	100		
165	94.5	82.2	123.4
167	13.1	11.4	17.0









#62
4-Chlorophenyl-phenylether
Concen: 20.742 ng/ul
RT: 15.351 min Scan# 2119
Delta R.T. -0.006 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:204 Resp: 442962
Ion Ratio Lower Upper
204 100
206 31.3 26.9 40.3
141 64.9 51.3 76.9



#63
4-Nitroaniline
Concen: 21.007 ng/ul
RT: 15.357 min Scan# 2120
Delta R.T. 0.000 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41

Tgt Ion:138 Resp: 220853
Ion Ratio Lower Upper
138 100
92 51.0 41.0 61.4
108 72.1 58.5 87.7









#66
4,6-Dinitro-2-methylphenol
Concen: 17.327 ng/ul
RT: 15.422 min Scan# 2131
Delta R.T. 0.000 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:198 Resp: 84225
Ion Ratio Lower Upper
198 100
182 4.4 3.4 5.2
77 27.9 19.3 28.9



#67
N-Nitrosodiphenylamine
Concen: 22.922 ng/ul
RT: 15.563 min Scan# 2155
Delta R.T. 0.000 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41

Tgt Ion:169 Resp: 809694
Ion Ratio Lower Upper
169 100
168 65.9 52.4 78.6
167 35.4 28.3 42.5









#68
4-Bromophenyl-phenylether
Concen: 22.695 ng/ul
RT: 16.245 min Scan# 2271
Delta R.T. 0.000 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:248 Resp: 247665
Ion Ratio Lower Upper
248 100
250 97.8 78.4 117.6
141 82.2 68.3 102.5



#69
Hexachlorobenzene
Concen: 21.556 ng/ul
RT: 16.351 min Scan# 2289
Delta R.T. 0.006 min
Lab File: BN034649.D
Acq: 18 Oct 2024 13:41

Tgt Ion:284 Resp: 270106
Ion Ratio Lower Upper
284 100
142 38.1 33.8 50.8
249 30.1 23.7 35.5





#70

Atrazine

Concen: 20.949 ng/ul

RT: 16.522 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:200 Resp: 247889

Ion Ratio Lower Upper

200 100

173 28.3 23.3 34.9

215 44.9 37.9 56.9



#71

Pentachlorophenol

Concen: 20.300 ng/ul

RT: 16.692 min Scan# 2347

Delta R.T. 0.000 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41




Tgt Ion:266 Resp: 136335


Ion Ratio Lower Upper

266 100

264 63.3 47.8 71.8

268 63.6 51.8 77.6





#72

Phenanthrene

Concen: 21.034 ng/ul

RT: 17.081 min Scan# 2413

Delta R.T. -0.006 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:178 Resp: 1316015

Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.4

176 18.8 15.5 23.3



#74

Anthracene

Concen: 21.211 ng/ul

RT: 17.175 min Scan# 2429

Delta R.T. 0.000 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41




Tgt Ion:178 Resp: 1371224


Ion Ratio Lower Upper

178 100

179 14.9 12.9 19.3

176 18.9 14.7 22.1





#75

1,2,3,4-Tetrachlorobenzene

Concen: 25.233 ng/uL

RT: 13.087 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:216 Resp: 396027

Ion Ratio Lower Upper

216 100

214 79.9 61.5 92.3

218 49.1 38.4 57.6



#76

Pentachlorobenzene

Concen: 23.851 ng/uL

RT: 14.616 min Scan# 1994

Delta R.T. 0.000 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41




Tgt Ion:250 Resp: 358242


Ion Ratio Lower Upper

250 100

248 60.7 49.0 73.4

252 60.7 50.6 75.8





#77

Carbazole

Concen: 20.079 ng/ul

RT: 17.439 min Scan# 2474

Delta R.T. 0.000 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41





Instrument :

BNA_N

ClientSampleId :

Tgt Ion:167 Resp: 1201981

Ion	Ratio	Lower	Upper
167	100		
166	21.4	17.2	25.8
139	13.3	10.6	16.0



#78

Di-n-butylphthalate

Concen: 21.469 ng/ul

RT: 18.033 min Scan# 2575




Delta R.T. -0.006 min


Lab File: BN034649.D

Acq: 18 Oct 2024 13:41

Tgt Ion:149 Resp: 1495399


Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.3	10.9
104	4.9	4.0	6.0








#80
 Fluoranthene
 Concen: 22.275 ng/ul
 RT: 19.098 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: BN034649.D
 Acq: 18 Oct 2024 13:41

Instrument :
 BNA_N
 ClientSampleId :



Tgt Ion:202 Resp: 1257980




Ion	Ratio	Lower	Upper
202	100		
101	13.9	10.6	15.8
100	10.9	8.2	12.2




#82
 Pyrene
 Concen: 22.694 ng/ul
 RT: 19.457 min Scan# 2817
 Delta R.T. -0.006 min
 Lab File: BN034649.D
 Acq: 18 Oct 2024 13:41

Tgt Ion:202 Resp: 1365224

Ion	Ratio	Lower	Upper
202	100		
101	16.3	12.9	19.3
100	13.6	10.7	16.1





#83

Butylbenzylphthalate

Concen: 23.911 ng/ul

RT: 20.386 min Scan# 2975

Delta R.T. -0.012 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:149 Resp: 533271

Ion Ratio Lower Upper

149 100

91 74.8 55.4 83.0

206 18.1 15.2 22.8



#84

3,3'-Dichlorobenzidine

Concen: 18.889 ng/ul

RT: 21.174 min Scan# 3109

Delta R.T. -0.006 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41




Tgt Ion:252 Resp: 348357


Ion Ratio Lower Upper

252 100

254 62.5 54.1 81.1

126 14.5 11.4 17.0





#85

Benzo(a)anthracene

Concen: 21.043 ng/ul

RT: 21.245 min Scan# 3121

Delta R.T. -0.006 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:228 Resp: 1225718

Ion Ratio Lower Upper

228 100

229 18.9 15.7 23.5

226 25.7 21.3 31.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 22.684 ng/ul

RT: 21.216 min Scan# 3116

Delta R.T. -0.012 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41




Tgt Ion:149 Resp: 793797


Ion Ratio Lower Upper

149 100

167 25.6 21.6 32.4

279 3.1 2.9 4.3





#87

Chrysene

Concen: 20.811 ng/ul

RT: 21.304 min Scan# 31

Delta R.T. -0.006 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:228 Resp: 1150633

Ion Ratio Lower Upper

228 100

226 30.2 22.6 34.0

229 20.0 15.2 22.8



#89

Di-n-octyl phthalate

Concen: 20.736 ng/ul




RT: 22.327 min Scan# 3305


Delta R.T. -0.017 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41

Tgt Ion:149 Resp: 1299729





#90

Benzo(b)fluoranthene

Concen: 21.268 ng/ul

RT: 23.245 min Scan# 3461

Delta R.T. -0.012 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:252 Resp: 1232524

Ion Ratio Lower Upper

252 100

253 21.6 17.3 25.9

125 12.4 10.3 15.5



#91

Benzo(k)fluoranthene

Concen: 21.012 ng/ul

RT: 23.304 min Scan# 3471

Delta R.T. -0.012 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41




Tgt Ion:252 Resp: 1244275


Ion Ratio Lower Upper

252 100

253 22.4 17.1 25.7

125 13.0 10.3 15.5





#93

Benzo(a)pyrene

Concen: 21.170 ng/ul

RT: 24.015 min Scan# 3

Delta R.T. -0.012 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:252 Resp: 1187813

Ion Ratio Lower Upper

252 100

253 21.6 17.7 26.5

125 14.4 11.7 17.5



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.167 ng/ul

RT: 27.380 min Scan# 4164

Delta R.T. -0.012 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41




Tgt Ion:276 Resp: 1341676


Ion Ratio Lower Upper

276 100

138 27.2 19.5 29.3

277 25.0 17.7 26.5





#95

Dibenzo(a,h)anthracene

Concen: 21.505 ng/ul

RT: 27.445 min Scan# 4177

Delta R.T. -0.023 min


Lab File: BN034649.D

Acq: 18 Oct 2024 13:41

Instrument :

BNA_N

ClientSampleId :






Tgt Ion:278 Resp: 1223313

Ion Ratio Lower Upper

278 100

139 20.6 15.9 23.9

279 23.9 18.9 28.3



#96

Benzo(g,h,i)perylene

Concen: 21.562 ng/ul

RT: 28.374 min Scan# 4333

Delta R.T. -0.017 min

Lab File: BN034649.D

Acq: 18 Oct 2024 13:41

Tgt Ion:276 Resp: 1171515

Ion Ratio Lower Upper

276 100

138 28.6 21.9 32.9

277 23.9 18.3 27.5

