

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101724\
 Data File : BN034645.D
 Acq On : 18 Oct 2024 11:04
 Operator : RC/JU
 Sample : P4380-02MS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 E27H4MS

Quant Time: Oct 18 22:54:21 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101624.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 16 15:50:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.663	152	347957	20.000	ng/u1	0.00
20) Naphthalene-d8	10.440	136	1347468	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.298	164	674994	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.039	188	1139668	20.000	ng/u1	0.00
79) Chrysene-d12	21.268	240	1018587	20.000	ng/u1	0.00
88) Perylene-d12	24.151	264	1274777	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.211	96	36402	4.186	ng/uL	0.00
4) Pyridine-d5	3.605	84	151901	5.983	ng/u1	0.00
7) Phenol-d5	6.834	99	247489	8.667	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.010	67	568477	32.273	ng/u1	0.00
11) 2-Chlorophenol-d4	7.199	132	679738	30.814	ng/u1	0.00
15) 4-Methylphenol-d8	8.369	113	433314	19.471	ng/u1	0.00
21) Nitrobenzene-d5	8.810	128	348992	37.218	ng/u1	0.00
24) 2-Nitrophenol-d4	9.528	143	333117	44.955	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.063	165	605765	35.468	ng/u1	0.00
31) 4-Chloroaniline-d4	10.569	131	400152	13.657	ng/u1	0.00
46) Dimethylphthalate-d6	13.722	166	1419127	33.233	ng/u1	0.00
49) Acenaphthylene-d8	13.987	160	1822028	34.351	ng/u1	0.00
54) 4-Nitrophenol-d4	14.492	143	81771	10.279	ng/u1	0.00
60) Fluorene-d10	15.292	176	1259708	34.107	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.410	200	149206	39.058	ng/u1	0.00
73) Anthracene-d10	17.139	188	1775888	36.452	ng/u1	0.00
81) Pyrene-d10	19.433	212	1940727	37.108	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.962	264	2195722	37.382	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.246	88	50960	5.454	ng/uL	97
5) Pyridine	3.622	79	214381	8.297	ng/u1	99
6) Benzaldehyde	6.810	77	474273	28.295	ng/u1	98
8) Phenol	6.863	94	315562	10.583	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.099	93	796090	33.858	ng/u1	98
12) 2-Chlorophenol	7.234	128	736711	31.979	ng/u1	97
13) 2-Methylphenol	8.104	108	535223	24.315	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.193	45	1160543	32.766	ng/u1	99
16) Acetophenone	8.481	105	1140498	32.093	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.475	70	542382	31.340	ng/u1	96
18) 4-Methylphenol	8.434	108	505122	21.167	ng/u1	100
19) Hexachloroethane	8.740	117	334820	34.855	ng/u1	97
22) Nitrobenzene	8.851	77	855730	38.152	ng/u1	98
23) Isophorone	9.381	82	1494124	33.697	ng/u1	99
25) 2-Nitrophenol	9.557	139	391268	44.902	ng/u1	97
26) 2,4-Dimethylphenol	9.628	107	711300	32.710	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.869	93	963844	35.241	ng/u1	99
29) 2,4-Dichlorophenol	10.087	162	624811	36.159	ng/u1	98
30) Naphthalene	10.493	128	2367898	34.960	ng/u1	99
32) 4-Chloroaniline	10.593	127	451310	16.245	ng/u1	99
33) Hexachlorobutadiene	10.787	225	373988	36.010	ng/u1	95
34) Caprolactam	11.375	113	27588	4.095	ng/u1	96
35) 4-Chloro-3-methylphenol	11.722	107	618960	32.039	ng/u1	99
36) 2-Methylnaphthalene	12.110	142	1478144	33.144	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101724\
 Data File : BN034645.D
 Acq On : 18 Oct 2024 11:04
 Operator : RC/JU
 Sample : P4380-02MS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 E27H4MS


Quant Time: Oct 18 22:54:21 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101624.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 16 15:50:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.328	142	1467423	32.323	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.481	216	658667	39.711	ng/ul	99
40) Hexachlorocyclopentadiene	12.469	237	295029	33.486	ng/ul	99
41) 2,4,6-Trichlorophenol	12.722	196	418501	43.457	ng/ul	95
42) 2,4,5-Trichlorophenol	12.787	196	446626	41.853	ng/ul	100
43) 1,1'-Biphenyl	13.134	154	1773984	38.271	ng/ul	100
44) 2-Chloronaphthalene	13.169	162	1382109	38.148	ng/ul	99
45) 2-Nitroaniline	13.363	65	408821	43.016	ng/ul	98
47) Dimethylphthalate	13.769	163	1508296	35.165	ng/ul	99
48) 2,6-Dinitrotoluene	13.875	165	324901	43.729	ng/ul	92
50) Acenaphthylene	14.016	152	2113767	36.778	ng/ul	98
51) 3-Nitroaniline	14.198	138	289057	30.480	ng/ul#	88
52) Acenaphthene	14.363	153	1454204	36.418	ng/ul	99
53) 2,4-Dinitrophenol	14.398	184	106030	36.582	ng/ul	98
55) 4-Nitrophenol	14.504	109	66477	11.412	ng/ul	97
56) Dibenzofuran	14.698	168	1897827	35.767	ng/ul	99
57) 2,4-Dinitrotoluene	14.657	165	439945	41.518	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	14.922	232	320555	41.908	ng/ul	98
59) Diethylphthalate	15.145	149	1552952	35.686	ng/ul	97
61) Fluorene	15.351	166	1518974	35.527	ng/ul	92
62) 4-Chlorophenyl-phenyle...	15.357	204	672545	35.042	ng/ul	97
63) 4-Nitroaniline	15.363	138	331692	35.105	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.422	198	175414	39.691	ng/ul#	91
67) N-Nitrosodiphenylamine	15.563	169	1266343	39.430	ng/ul	97
68) 4-Bromophenyl-phenylether	16.245	248	396153	39.928	ng/ul	99
69) Hexachlorobenzene	16.351	284	448944	39.406	ng/ul	96
70) Atrazine	16.528	200	232954	21.653	ng/ul	98
71) Pentachlorophenol	16.692	266	268222	43.927	ng/ul	97
72) Phenanthrene	17.086	178	2217468	38.982	ng/ul	99
74) Anthracene	17.175	178	2241987	38.144	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.087	216	614994	43.098	ng/ul	96
76) Pentachlorobenzene	14.622	250	545201	39.924	ng/ul	99
77) Carbazole	17.439	167	2124586	39.036	ng/ul	99
78) Di-n-butylphthalate	18.039	149	2520599	39.803	ng/ul	99
80) Fluoranthene	19.104	202	2365233	37.697	ng/ul	100
82) Pyrene	19.463	202	2569128	38.439	ng/ul	100
83) Butylbenzylphthalate	20.392	149	1160210	46.825	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.180	252	157403	7.682	ng/ul	96
85) Benzo(a)anthracene	21.251	228	2557634	39.522	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.221	149	1692345	43.530	ng/ul	100
87) Chrysene	21.310	228	2401540	39.097	ng/ul	96
89) Di-n-octyl phthalate	22.333	149	3068230	41.412	ng/ul	100
90) Benzo(b)fluoranthene	23.257	252	2651702	38.711	ng/ul	99
91) Benzo(k)fluoranthene	23.310	252	2697929	38.543	ng/ul	97
93) Benzo(a)pyrene	24.027	252	2487471	37.507	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	27.398	276	3359823	40.606	ng/ul	98
95) Dibenzo(a,h)anthracene	27.456	278	2741142	40.767	ng/ul	99
96) Benzo(g,h,i)perylene	28.398	276	2586998	40.281	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 18 22:54:21 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101624.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Oct 16 15:50:29 2024
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 5.454 ng/uL
RT: 3.246 min Scan# 61
Delta R.T. 0.006 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04




Instrument :
BNA_N
ClientSampleId :
E27H4MS


Tgt Ion: 88 Resp: 50960
Ion Ratio Lower Upper
88 100
43 33.8 24.5 36.7
58 81.9 64.5 96.7



#5
Pyridine
Concen: 8.297 ng/uL
RT: 3.622 min Scan# 125
Delta R.T. 0.006 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04

Tgt Ion: 79 Resp: 214381
Ion Ratio Lower Upper
79 100
52 81.0 63.9 95.9
51 37.6 29.2 43.8





#6

Benzaldehyde

Concen: 28.295 ng/ul

RT: 6.810 min Scan# 6

Delta R.T. 0.000 min

Lab File: BN034645.D

Acq: 18 Oct 2024 11:04

Instrument :

BNA_N

ClientSampleId :

E27H4MS





Tgt Ion: 77 Resp: 474273

Ion Ratio Lower Upper

77 100

105 99.0 81.2 121.8

106 97.4 78.3 117.5



#8

Phenol

Concen: 10.583 ng/ul

RT: 6.863 min Scan# 676

Delta R.T. 0.006 min

Lab File: BN034645.D

Acq: 18 Oct 2024 11:04




Tgt Ion: 94 Resp: 315562


Ion Ratio Lower Upper

94 100

65 25.9 22.2 33.4

66 33.8 27.4 41.0









#10
Bis(2-Chloroethyl)ether
Concen: 33.858 ng/ul
RT: 7.099 min Scan# 716
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04




Instrument :
BNA_N
ClientSampleId :
E27H4MS


Tgt Ion: 93 Resp: 796090
Ion Ratio Lower Upper
93 100
63 79.8 65.0 97.6
95 32.3 26.4 39.6



#12
2-Chlorophenol
Concen: 31.979 ng/ul
RT: 7.234 min Scan# 739
Delta R.T. 0.006 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04

Tgt Ion:128 Resp: 736711
Ion Ratio Lower Upper
128 100
64 46.4 39.8 59.8
130 32.6 26.3 39.5









#13
2-Methylphenol
Concen: 24.315 ng/ul
RT: 8.104 min Scan# 81
Delta R.T. 0.006 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04




Instrument :
BNA_N
ClientSampleId :
E27H4MS


Tgt Ion:108 Resp: 535223
Ion Ratio Lower Upper
108 100
107 90.1 70.1 105.1



#14
2,2'-oxybis(1-Chloropropane)
Concen: 32.766 ng/ul
RT: 8.193 min Scan# 902
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04

Tgt Ion: 45 Resp: 1160543
Ion Ratio Lower Upper
45 100
77 13.2 10.4 15.6
79 9.8 7.4 11.2









#16
Acetophenone
Concen: 32.093 ng/ul
RT: 8.481 min Scan# 91
Delta R.T. 0.006 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04




Instrument :
BNA_N
ClientSampleId :
E27H4MS


Tgt Ion:105 Resp: 1140498
Ion Ratio Lower Upper
105 100
77 73.4 61.4 92.0
51 26.7 22.1 33.1



#17
N-Nitroso-di-n-propylamine
Concen: 31.340 ng/ul
RT: 8.475 min Scan# 950
Delta R.T. 0.006 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04

Tgt Ion: 70 Resp: 542382
Ion Ratio Lower Upper
70 100
42 58.2 49.8 74.6
101 10.6 8.2 12.4
130 23.1 17.0 25.6









#18
4-Methylphenol
Concen: 21.167 ng/ul
RT: 8.434 min Scan# 942
Delta R.T. 0.006 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04




Instrument :
BNA_N
ClientSampleId :
E27H4MS


Tgt Ion:108 Resp: 505122
Ion Ratio Lower Upper
108 100
107 106.8 85.8 128.6



#19
Hexachloroethane
Concen: 34.855 ng/ul
RT: 8.740 min Scan# 995
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04

Tgt Ion:117 Resp: 334820
Ion Ratio Lower Upper
117 100
201 86.5 67.0 100.4
199 53.6 44.7 67.1









#22
Nitrobenzene
Concen: 38.152 ng/ul
RT: 8.851 min Scan# 1104
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04




Instrument :
BNA_N
ClientSampleId :
E27H4MS


Tgt Ion: 77 Resp: 855730
Ion Ratio Lower Upper
77 100
123 48.2 37.5 56.3
65 13.7 10.8 16.2



#23
Isophorone
Concen: 33.697 ng/ul
RT: 9.381 min Scan# 1104
Delta R.T. 0.006 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04

Tgt Ion: 82 Resp: 1494124
Ion Ratio Lower Upper
82 100
95 6.8 5.3 7.9
138 17.7 13.8 20.6









#25
2-Nitrophenol
Concen: 44.902 ng/ul
RT: 9.557 min Scan# 1134
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04




Instrument :
BNA_N
ClientSampleId :
E27H4MS


Tgt Ion:139 Resp: 391268
Ion Ratio Lower Upper
139 100
65 51.1 39.2 58.8
109 29.0 21.8 32.6



#26
2,4-Dimethylphenol
Concen: 32.710 ng/ul
RT: 9.628 min Scan# 1146
Delta R.T. 0.006 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04

Tgt Ion:107 Resp: 711300
Ion Ratio Lower Upper
107 100
121 51.1 41.8 62.6
122 87.2 70.2 105.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 35.241 ng/ul
 RT: 9.869 min Scan# 1186
 Delta R.T. 0.006 min
 Lab File: BN034645.D
 Acq: 18 Oct 2024 11:04




Instrument :
 BNA_N
 ClientSampleId :
 E27H4MS


Tgt Ion: 93 Resp: 963844
 Ion Ratio Lower Upper
 93 100
 95 32.9 26.0 39.0
 123 11.2 9.1 13.7



#29
 2,4-Dichlorophenol
 Concen: 36.159 ng/ul
 RT: 10.087 min Scan# 1224
 Delta R.T. 0.000 min
 Lab File: BN034645.D
 Acq: 18 Oct 2024 11:04

Tgt Ion:162 Resp: 624811
 Ion Ratio Lower Upper
 162 100
 164 63.8 51.9 77.9
 98 39.3 30.6 46.0





#30

Naphthalene

Concen: 34.960 ng/ul

RT: 10.493 min Scan# 1292

Delta R.T. 0.000 min

Lab File: BN034645.D

Acq: 18 Oct 2024 11:04

Instrument :

BNA_N

ClientSampleId :

E27H4MS





Tgt Ion:128 Resp: 2367898

Ion Ratio Lower Upper

128 100

129 10.9 8.2 12.4

127 12.7 10.3 15.5



#32

4-Chloroaniline

Concen: 16.245 ng/ul

RT: 10.593 min Scan# 1310

Delta R.T. 0.000 min

Lab File: BN034645.D




Acq: 18 Oct 2024 11:04


Tgt Ion:127 Resp: 451310

Ion Ratio Lower Upper

127 100

129 32.1 26.0 39.0









#33
Hexachlorobutadiene
Concen: 36.010 ng/ul
RT: 10.787 min Scan# 1343
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04




Instrument :
BNA_N
ClientSampleId :
E27H4MS


Tgt Ion:225 Resp: 373988
Ion Ratio Lower Upper
225 100
223 66.3 50.4 75.6
227 65.9 49.8 74.6



#34
Caprolactam
Concen: 4.095 ng/ul
RT: 11.375 min Scan# 1443
Delta R.T. 0.041 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04

Tgt Ion:113 Resp: 27588
Ion Ratio Lower Upper
113 100
55 192.1 152.5 228.7
56 131.8 112.8 169.2









#35
4-Chloro-3-methylphenol
Concen: 32.039 ng/ul
RT: 11.722 min Scan# 1502
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04




Instrument :
BNA_N
ClientSampleId :
E27H4MS


Tgt Ion:107 Resp: 618960
Ion Ratio Lower Upper
107 100
144 25.3 20.7 31.1
142 79.7 64.1 96.1



#36
2-Methylnaphthalene
Concen: 33.144 ng/ul
RT: 12.110 min Scan# 1568
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04

Tgt Ion:142 Resp: 1478144
Ion Ratio Lower Upper
142 100
141 88.6 72.3 108.5









#37
1-Methylnaphthalene
Concen: 32.323 ng/ul
RT: 12.328 min Scan# 1605
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04




Instrument :
BNA_N
ClientSampleId :
E27H4MS


Tgt Ion:142 Resp: 1467423
Ion Ratio Lower Upper
142 100
141 95.1 75.4 113.0
116 3.6 2.8 4.2



#39
1,2,4,5-Tetrachlorobenzene
Concen: 39.711 ng/ul
RT: 12.481 min Scan# 1631
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04

Tgt Ion:216 Resp: 658667
Ion Ratio Lower Upper
216 100
214 78.2 63.0 94.6
179 22.4 17.5 26.3
108 19.8 15.5 23.3





#40

Hexachlorocyclopentadiene

Concen: 33.486 ng/ul

RT: 12.469 min Scan# 1629

Delta R.T. 0.000 min

Lab File: BN034645.D

Acq: 18 Oct 2024 11:04

Instrument :

BNA_N

ClientSampleId :

E27H4MS





Tgt Ion:237 Resp: 295029

Ion Ratio Lower Upper

237 100

235 64.8 52.6 78.8

272 13.0 10.1 15.1



#41

2,4,6-Trichlorophenol

Concen: 43.457 ng/ul

RT: 12.722 min Scan# 1672

Delta R.T. 0.000 min

Lab File: BN034645.D

Acq: 18 Oct 2024 11:04




Tgt Ion:196 Resp: 418501


Ion Ratio Lower Upper

196 100

198 95.1 71.0 106.4

200 31.3 25.3 37.9









#42
2,4,5-Trichlorophenol
Concen: 41.853 ng/ul
RT: 12.787 min Scan# 1683
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04




Instrument :
BNA_N
ClientSampleId :
E27H4MS


Tgt Ion:196 Resp: 446626
Ion Ratio Lower Upper
196 100
198 98.1 78.1 117.1
200 30.9 24.8 37.2



#43
1,1'-Biphenyl
Concen: 38.271 ng/ul
RT: 13.134 min Scan# 1742
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04

Tgt Ion:154 Resp: 1773984
Ion Ratio Lower Upper
154 100
153 40.4 32.4 48.6
76 14.0 11.2 16.8





#44

2-Chloronaphthalene

Concen: 38.148 ng/ul

RT: 13.169 min Scan# 1747

Delta R.T. 0.000 min

Lab File: BN034645.D

Acq: 18 Oct 2024 11:04

Instrument :

BNA_N

ClientSampleId :

E27H4MS





Tgt Ion:162 Resp: 1382109

Ion Ratio Lower Upper

162 100

164 31.5 26.3 39.5

127 39.9 31.9 47.9



#45

2-Nitroaniline

Concen: 43.016 ng/ul

RT: 13.363 min Scan# 1781

Delta R.T. 0.000 min

Lab File: BN034645.D

Acq: 18 Oct 2024 11:04




Tgt Ion: 65 Resp: 408821


Ion Ratio Lower Upper

65 100

92 70.8 55.1 82.7

138 106.4 84.0 126.0









#47
Dimethylphthalate
Concen: 35.165 ng/ul
RT: 13.769 min Scan# 1849
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04




Instrument :
BNA_N
ClientSampleId :
E27H4MS


Tgt Ion:163 Resp: 1508296
Ion Ratio Lower Upper
163 100
194 4.0 3.1 4.7
164 10.3 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 43.729 ng/ul
RT: 13.875 min Scan# 1868
Delta R.T. 0.006 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04

Tgt Ion:165 Resp: 324901
Ion Ratio Lower Upper
165 100
89 55.4 49.7 74.5
121 21.9 20.1 30.1









#50
Acenaphthylene
Concen: 36.778 ng/ul
RT: 14.016 min Scan# 1892
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04




Instrument :
BNA_N
ClientSampleId :
E27H4MS


Tgt Ion:152 Resp: 2113767
Ion Ratio Lower Upper
152 100
151 20.0 15.3 22.9
153 13.5 9.9 14.9



#51
3-Nitroaniline
Concen: 30.480 ng/ul
RT: 14.198 min Scan# 1923
Delta R.T. 0.006 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04

Tgt Ion:138 Resp: 289057
Ion Ratio Lower Upper
138 100
108 12.5 13.4 20.0#
92 111.5 100.0 150.0









#52
Acenaphthene
Concen: 36.418 ng/ul
RT: 14.363 min Scan# 1951
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04




Instrument :
BNA_N
ClientSampleId :
E27H4MS


Tgt Ion:153 Resp: 1454204
Ion Ratio Lower Upper
153 100
152 46.5 36.5 54.7
154 86.6 69.3 103.9



#53
2,4-Dinitrophenol
Concen: 36.582 ng/ul
RT: 14.398 min Scan# 1957
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04

Tgt Ion:184 Resp: 106030
Ion Ratio Lower Upper
184 100
63 73.2 58.0 87.0
154 65.4 54.7 82.1









#55
4-Nitrophenol
Concen: 11.412 ng/ul
RT: 14.504 min Scan# 1975
Delta R.T. 0.006 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04




Instrument :
BNA_N
ClientSampleId :
E27H4MS


Tgt Ion:109 Resp: 66477
Ion Ratio Lower Upper
109 100
139 144.1 113.8 170.6
65 134.5 111.7 167.5



#56
Dibenzofuran
Concen: 35.767 ng/ul
RT: 14.698 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04

Tgt Ion:168 Resp: 1897827
Ion Ratio Lower Upper
168 100
139 37.9 29.7 44.5









#57
2,4-Dinitrotoluene
Concen: 41.518 ng/ul
RT: 14.657 min Scan# 2001
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04




Instrument :
BNA_N
ClientSampleId :
E27H4MS


Tgt Ion:165 Resp: 439945
Ion Ratio Lower Upper
165 100
63 47.1 37.1 55.7
182 2.9 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 41.908 ng/ul
RT: 14.922 min Scan# 2046
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04

Tgt Ion:232 Resp: 320555
Ion Ratio Lower Upper
232 100
131 53.0 43.5 65.3
130 2.8 2.2 3.4
166 34.9 29.4 44.0









#59
Diethylphthalate
Concen: 35.686 ng/ul
RT: 15.145 min Scan# 2083
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04




Instrument :
BNA_N
ClientSampleId :
E27H4MS


Tgt Ion:149 Resp: 1552952
Ion Ratio Lower Upper
149 100
177 22.5 16.5 24.7
150 12.5 9.8 14.6



#61
Fluorene
Concen: 35.527 ng/ul
RT: 15.351 min Scan# 2119
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04

Tgt Ion:166 Resp: 1518974
Ion Ratio Lower Upper
166 100
165 93.9 82.2 123.4
167 13.5 11.4 17.0









#62
4-Chlorophenyl-phenylether
Concen: 35.042 ng/ul
RT: 15.357 min Scan# 2119
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04




Instrument :
BNA_N
ClientSampleId :
E27H4MS


Tgt Ion:204 Resp: 672545
Ion Ratio Lower Upper
204 100
206 32.5 26.9 40.3
141 61.5 51.3 76.9



#63
4-Nitroaniline
Concen: 35.105 ng/ul
RT: 15.363 min Scan# 2121
Delta R.T. 0.006 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04

Tgt Ion:138 Resp: 331692
Ion Ratio Lower Upper
138 100
92 48.6 41.0 61.4
108 69.8 58.5 87.7









#66
 4,6-Dinitro-2-methylphenol
 Concen: 39.691 ng/ul
 RT: 15.422 min Scan# 2131
 Delta R.T. 0.000 min
 Lab File: BN034645.D
 Acq: 18 Oct 2024 11:04




Instrument :
 BNA_N
 ClientSampleId :
 E27H4MS


Tgt Ion:198 Resp: 175414
 Ion Ratio Lower Upper
 198 100
 182 4.4 3.4 5.2
 77 29.2 19.3 28.9#



#67
 N-Nitrosodiphenylamine
 Concen: 39.430 ng/ul
 RT: 15.563 min Scan# 2155
 Delta R.T. 0.000 min
 Lab File: BN034645.D
 Acq: 18 Oct 2024 11:04

Tgt Ion:169 Resp: 1266343
 Ion Ratio Lower Upper
 169 100
 168 68.5 52.4 78.6
 167 37.1 28.3 42.5









#68
4-Bromophenyl-phenylether
Concen: 39.928 ng/ul
RT: 16.245 min Scan# 2271
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04




Instrument :
BNA_N
ClientSampleId :
E27H4MS


Tgt Ion:248 Resp: 396153
Ion Ratio Lower Upper
248 100
250 98.9 78.4 117.6
141 85.9 68.3 102.5



#69
Hexachlorobenzene
Concen: 39.406 ng/ul
RT: 16.351 min Scan# 2289
Delta R.T. 0.006 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04

Tgt Ion:284 Resp: 448944
Ion Ratio Lower Upper
284 100
142 38.8 33.8 50.8
249 30.0 23.7 35.5





#70

Atrazine

Concen: 21.653 ng/ul

RT: 16.528 min Scan# 21

Delta R.T. 0.006 min

Lab File: BN034645.D

Acq: 18 Oct 2024 11:04

Instrument :

BNA_N

ClientSampleId :

E27H4MS





Tgt Ion:200 Resp: 232954

Ion Ratio Lower Upper

200 100

173 28.5 23.3 34.9

215 48.9 37.9 56.9



#71

Pentachlorophenol

Concen: 43.927 ng/ul

RT: 16.692 min Scan# 2347

Delta R.T. 0.000 min

Lab File: BN034645.D

Acq: 18 Oct 2024 11:04




Tgt Ion:266 Resp: 268222


Ion Ratio Lower Upper

266 100

264 62.3 47.8 71.8

268 62.8 51.8 77.6





#72

Phenanthrene

Concen: 38.982 ng/ul

RT: 17.086 min Scan# 2413

Delta R.T. 0.000 min

Lab File: BN034645.D

Acq: 18 Oct 2024 11:04

Instrument :





BNA_N

ClientSampleId :

E27H4MS

Tgt Ion:178 Resp: 2217468

Ion	Ratio	Lower	Upper
178	100		
179	15.4	12.2	18.4
176	19.0	15.5	23.3



#74

Anthracene

Concen: 38.144 ng/ul

RT: 17.175 min Scan# 2429




Delta R.T. 0.000 min


Lab File: BN034645.D

Acq: 18 Oct 2024 11:04

Tgt Ion:178 Resp: 2241987

Ion	Ratio	Lower	Upper
178	100		
179	15.3	12.9	19.3
176	18.0	14.7	22.1









#75
1,2,3,4-Tetrachlorobenzene
Concen: 43.098 ng/uL
RT: 13.087 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04




Instrument :
BNA_N
ClientSampleId :
E27H4MS


Tgt Ion:216 Resp: 614994
Ion Ratio Lower Upper
216 100
214 81.9 61.5 92.3
218 48.5 38.4 57.6



#76
Pentachlorobenzene
Concen: 39.924 ng/uL
RT: 14.622 min Scan# 1995
Delta R.T. 0.006 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04

Tgt Ion:250 Resp: 545201
Ion Ratio Lower Upper
250 100
248 62.2 49.0 73.4
252 62.8 50.6 75.8





#77

Carbazole

Concen: 39.036 ng/ul

RT: 17.439 min Scan# 2474

Delta R.T. 0.000 min

Lab File: BN034645.D

Acq: 18 Oct 2024 11:04

Instrument :





BNA_N

ClientSampleId :

E27H4MS

Tgt Ion:167 Resp: 2124586

Ion	Ratio	Lower	Upper
167	100		
166	22.1	17.2	25.8
139	13.4	10.6	16.0



#78

Di-n-butylphthalate

Concen: 39.803 ng/ul

RT: 18.039 min Scan# 2576




Delta R.T. 0.000 min


Lab File: BN034645.D

Acq: 18 Oct 2024 11:04

Tgt Ion:149 Resp: 2520599

Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.3	10.9
104	4.5	4.0	6.0









#80
 Fluoranthene
 Concen: 37.697 ng/u1
 RT: 19.104 min Scan# 2756
 Delta R.T. 0.000 min
 Lab File: BN034645.D
 Acq: 18 Oct 2024 11:04

Instrument :
 BNA_N
 ClientSampleId :
 E27H4MS

Tgt Ion:202 Resp: 2365233




Ion	Ratio	Lower	Upper
202	100		
101	13.3	10.6	15.8
100	10.5	8.2	12.2




#82
 Pyrene
 Concen: 38.439 ng/u1
 RT: 19.463 min Scan# 2818
 Delta R.T. 0.000 min
 Lab File: BN034645.D
 Acq: 18 Oct 2024 11:04

Tgt Ion:202 Resp: 2569128

Ion	Ratio	Lower	Upper
202	100		
101	16.0	12.9	19.3
100	13.2	10.7	16.1









#83
 Butylbenzylphthalate
 Concen: 46.825 ng/ul
 RT: 20.392 min Scan# 2975
 Delta R.T. -0.006 min
 Lab File: BN034645.D
 Acq: 18 Oct 2024 11:04




Instrument :
 BNA_N
 ClientSampleId :
 E27H4MS


Tgt Ion:149 Resp: 1160210
 Ion Ratio Lower Upper
 149 100
 91 70.5 55.4 83.0
 206 19.7 15.2 22.8



#84
 3,3'-Dichlorobenzidine
 Concen: 7.682 ng/ul
 RT: 21.180 min Scan# 3110
 Delta R.T. 0.000 min
 Lab File: BN034645.D
 Acq: 18 Oct 2024 11:04

Tgt Ion:252 Resp: 157403
 Ion Ratio Lower Upper
 252 100
 254 64.3 54.1 81.1
 126 14.6 11.4 17.0





#85

Benzo(a)anthracene

Concen: 39.522 ng/ul

RT: 21.251 min Scan# 3121

Delta R.T. 0.000 min

Lab File: BN034645.D

Acq: 18 Oct 2024 11:04

Instrument :

BNA_N

ClientSampleId :

E27H4MS





Tgt Ion:228 Resp: 2557634

Ion Ratio Lower Upper

228 100

229 19.2 15.7 23.5

226 26.4 21.3 31.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 43.530 ng/ul

RT: 21.221 min Scan# 3117

Delta R.T. -0.006 min

Lab File: BN034645.D

Acq: 18 Oct 2024 11:04




Tgt Ion:149 Resp: 1692345


Ion Ratio Lower Upper

149 100

167 27.2 21.6 32.4

279 3.9 2.9 4.3









#87
Chrysene
Concen: 39.097 ng/ul
RT: 21.310 min Scan# 3131
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04




Instrument :
BNA_N
ClientSampleId :
E27H4MS


Tgt Ion:228 Resp: 2401540
Ion Ratio Lower Upper
228 100
226 31.1 22.6 34.0
229 20.5 15.2 22.8



#89
Di-n-octyl phthalate
Concen: 41.412 ng/ul
RT: 22.333 min Scan# 3306
Delta R.T. -0.012 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04

Tgt Ion:149 Resp: 3068230









#90
Benzo(b)fluoranthene
Concen: 38.711 ng/ul
RT: 23.257 min Scan# 3462
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04




Instrument :
BNA_N
ClientSampleId :
E27H4MS


Tgt Ion:252 Resp: 2651702
Ion Ratio Lower Upper
252 100
253 20.9 17.3 25.9
125 12.7 10.3 15.5



#91
Benzo(k)fluoranthene
Concen: 38.543 ng/ul
RT: 23.310 min Scan# 3472
Delta R.T. -0.006 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04

Tgt Ion:252 Resp: 2697929
Ion Ratio Lower Upper
252 100
253 22.9 17.1 25.7
125 13.7 10.3 15.5









#93
Benzo(a)pyrene
Concen: 37.507 ng/u1
RT: 24.027 min Scan# 3594
Delta R.T. 0.000 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04




Instrument :
BNA_N
ClientSampleId :
E27H4MS


Tgt Ion:252 Resp: 2487471
Ion Ratio Lower Upper
252 100
253 20.5 17.7 26.5
125 13.9 11.7 17.5



#94
Indeno(1,2,3-cd)pyrene
Concen: 40.606 ng/u1
RT: 27.398 min Scan# 4167
Delta R.T. 0.006 min
Lab File: BN034645.D
Acq: 18 Oct 2024 11:04

Tgt Ion:276 Resp: 3359823
Ion Ratio Lower Upper
276 100
138 25.4 19.5 29.3
277 23.2 17.7 26.5





#95

Dibenzo(a,h)anthracene

Concen: 40.767 ng/ul

RT: 27.456 min Scan# 4176

Delta R.T. -0.012 min

Lab File: BN034645.D

Acq: 18 Oct 2024 11:04

Instrument :

BNA_N

ClientSampleId :

E27H4MS





Tgt Ion:278 Resp: 2741142

Ion Ratio Lower Upper

278 100

139 20.3 15.9 23.9

279 22.7 18.9 28.3



#96

Benzo(g,h,i)perylene

Concen: 40.281 ng/ul

RT: 28.398 min Scan# 4337

Delta R.T. 0.006 min

Lab File: BN034645.D

Acq: 18 Oct 2024 11:04

Tgt Ion:276 Resp: 2586998

Ion Ratio Lower Upper

276 100

138 29.1 21.9 32.9

277 24.0 18.3 27.5

