

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101821\
 Data File : BN016982.D
 Acq On : 18 Oct 2021 10:53
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Oct 18 13:22:58 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 18 13:18:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.598	152	16830	0.400	ng	0.00
7) Naphthalene-d8	10.370	136	75993	0.400	ng	0.00
13) Acenaphthene-d10	14.232	164	39857	0.400	ng	0.00
19) Phenanthrene-d10	16.982	188	76441	0.400	ng	0.00
29) Chrysene-d12	21.186	240	67811	0.400	ng	0.00
35) Perylene-d12	23.407	264	70497	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.210	112	3223	0.077	ng	0.00
5) Phenol-d6	6.786	99	3524	0.076	ng	0.00
8) Nitrobenzene-d5	8.748	82	4127	0.085	ng	0.00
11) 2-Methylnaphthalene-d10	11.964	152	10414	0.097	ng	0.00
14) 2,4,6-Tribromophenol	15.729	330	1047	0.058	ng	0.00
15) 2-Fluorobiphenyl	12.849	172	15043	0.099	ng	0.00
27) Fluoranthene-d10	19.020	212	16890	0.087	ng	0.00
31) Terphenyl-d14	19.635	244	14847	0.099	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.199	88	932	0.086	ng	# 90
3) n-Nitrosodimethylamine	3.531	42	1140	0.083	ng	98
6) bis(2-Chloroethyl)ether	7.044	93	4680	0.100	ng	99
9) Naphthalene	10.421	128	20308	0.095	ng	100
10) Hexachlorobutadiene	10.700	225	4028	0.104	ng	# 100
12) 2-Methylnaphthalene	12.040	142	12702	0.099	ng	99
16) Acenaphthylene	13.940	152	13318	0.081	ng	100
17) Acenaphthene	14.295	154	10983	0.099	ng	100
18) Fluorene	15.285	166	12505	0.094	ng	100
20) 4,6-Dinitro-2-methylph...	15.387	198	228	0.021	ng	# 37
21) 4-Bromophenyl-phenylether	16.179	248	3934	0.086	ng	97
22) Hexachlorobenzene	16.289	284	4982	0.095	ng	99
23) Atrazine	16.459	200	2061	0.065	ng	97
24) Pentachlorophenol	16.642	266	621	0.037	ng	99
25) Phenanthrene	17.019	178	21381	0.098	ng	100
26) Anthracene	17.116	178	15442	0.082	ng	100
28) Fluoranthene	19.049	202	21151	0.088	ng	99
30) Pyrene	19.412	202	21291	0.099	ng	100
32) Benzo(a)anthracene	21.165	228	17585	0.085	ng	100
33) Chrysene	21.218	228	23141	0.105	ng	100
34) Bis(2-ethylhexyl)phtha...	21.123	149	6468	0.066	ng	100
36) Indeno(1,2,3-cd)pyrene	25.649	276	21497	0.083	ng	99
37) Benzo(b)fluoranthene	22.740	252	20184	0.086	ng	97
38) Benzo(k)fluoranthene	22.784	252	21247	0.092	ng	# 95
39) Benzo(a)pyrene	23.308	252	17017	0.083	ng	95
40) Dibenzo(a,h)anthracene	25.670	278	16469	0.078	ng	96
41) Benzo(g,h,i)perylene	26.330	276	19880	0.086	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101821\
Data File : BN016982.D
Acq On : 18 Oct 2021 10:53
Operator : CG/JU
Sample : SSTDIC0.1
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Quant Time: Oct 18 13:22:58 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 18 13:18:40 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.598 min Scan# 501
Delta R.T. -0.009 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:152 Resp: 16830
Ion Ratio Lower Upper
152 100
150 156.0 123.0 184.4
115 59.7 46.3 69.5



1,4-Dioxane
Concen: 0.086 ng
RT: 3.199 min Scan# 24
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53

Tgt Ion: 88 Resp: 932
Ion Ratio Lower Upper
88 100
43 49.0 26.8 40.2#
58 81.3 63.0 94.4









#3
 n-Nitrosodimethylamine
 Concen: 0.083 ng
 RT: 3.531 min Scan# 60
 Delta R.T. 0.000 min
 Lab File: BN016982.D
 Acq: 18 Oct 2021 10:53




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:	42	Resp:	1140
Ion	Ratio	Lower	Upper
42	100		
74	145.6	114.7	172.1
44	4.2	3.8	5.6



#4
 2-Fluorophenol
 Concen: 0.077 ng
 RT: 5.210 min Scan# 242
 Delta R.T. 0.000 min
 Lab File: BN016982.D
 Acq: 18 Oct 2021 10:53

Tgt Ion:	112	Resp:	3223
Ion	Ratio	Lower	Upper
112	100		
64	57.2	45.5	68.3
63	28.9	23.0	34.6









Phenol-d6
 Concen: 0.076 ng
 RT: 6.786 min Scan# 413
 Delta R.T. 0.000 min
 Lab File: BN016982.D
 Acq: 18 Oct 2021 10:53




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:	99	Resp:	3524
Ion	Ratio	Lower	Upper
99	100		
42	18.4	14.6	21.8
71	31.4	24.8	37.2



bis(2-Chloroethyl)ether
 Concen: 0.100 ng
 RT: 7.044 min Scan# 441
 Delta R.T. 0.000 min
 Lab File: BN016982.D
 Acq: 18 Oct 2021 10:53

Tgt Ion:	93	Resp:	4680
Ion	Ratio	Lower	Upper
93	100		
63	77.0	60.9	91.3
95	33.2	26.4	39.6









Naphthalene-d8
 Concen: 0.400 ng
 RT: 10.370 min Scan# 741
 Delta R.T. 0.000 min
 Lab File: BN016982.D
 Acq: 18 Oct 2021 10:53




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:	136	Resp:	75993
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.8	13.2
54	5.8	5.2	7.8
68	5.0	4.4	6.6



Nitrobenzene-d5
 Concen: 0.085 ng
 RT: 8.748 min Scan# 613
 Delta R.T. 0.000 min
 Lab File: BN016982.D
 Acq: 18 Oct 2021 10:53

Tgt Ion:	82	Resp:	4127
Ion	Ratio	Lower	Upper
82	100		
128	44.6	33.9	50.9
54	48.8	38.3	57.5









Naphthalene
Concen: 0.095 ng
RT: 10.421 min Scan# 745
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	128	Resp:	20308
Ion	Ratio	Lower	Upper
128	100		
129	11.2	8.8	13.2
127	13.1	10.3	15.5



Hexachlorobutadiene
Concen: 0.104 ng
RT: 10.700 min Scan# 767
Delta R.T. -0.012 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53

Tgt Ion:	225	Resp:	4028
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.3	51.2	76.8









2-Methylnaphthalene-d10
Concen: 0.097 ng
RT: 11.964 min Scan# 933
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 10414
Ion Ratio Lower Upper
152 100
151 20.2 16.3 24.5



2-Methylnaphthalene
Concen: 0.099 ng
RT: 12.040 min Scan# 950
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53

Tgt Ion:142 Resp: 12702
Ion Ratio Lower Upper
142 100
141 88.7 70.4 105.6
115 33.2 26.4 39.6









Acenaphthene-d10
 Concen: 0.400 ng
 RT: 14.232 min Scan# 1191
 Delta R.T. 0.000 min
 Lab File: BN016982.D
 Acq: 18 Oct 2021 10:53




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Tgt Ion:	164	Resp:	39857
Ion Ratio	Lower	Upper	
164	100		
162	100.2	80.2	120.2
160	45.1	36.0	54.0



2,4,6-Tribromophenol
 Concen: 0.058 ng
 RT: 15.729 min Scan# 1309
 Delta R.T. 0.000 min
 Lab File: BN016982.D
 Acq: 18 Oct 2021 10:53



Tgt Ion:	330	Resp:	1047
Ion Ratio	Lower	Upper	
330	100		
332	0.0	0.0	0.0
141	32.3	25.6	38.4




Abundance Scan 1082 (12.849 min): BN016984.D\data.ms (-1074) (-)

2-Fluorobiphenyl
Concen: 0.099 ng
RT: 12.849 min Scan# 1082
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1



Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.2	28.7	43.1
170	24.7	19.6	29.4





Abundance Scan 1082 (12.849 min): BN016982.D\data.ms (-1074) (-)




Abundance Scan 1169 (13.953 min): BN016984.D\data.ms (-1153) (-)


Acenaphthylene
Concen: 0.081 ng
RT: 13.940 min Scan# 1168
Delta R.T. -0.012 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53



Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.4	15.4	23.2
153	13.0	10.4	15.6




Abundance Scan 1168 (13.940 min): BN016982.D\data.ms (-1153) (-)





Abundance Scan 1196 (14.295 min): BN016984.D\data.ms (-1197) (-)

Acenaphthene
Concen: 0.099 ng
RT: 14.295 min Scan# 1196
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1



Tgt Ion:	154	Resp:	10983
Ion	Ratio	Lower	Upper
154	100		
153	113.8	91.1	136.7
152	55.8	44.8	67.2




Abundance Scan 1196 (14.295 min): BN016982.D\data.ms (-1188) (-)





Abundance Scan 1274 (15.285 min): BN016984.D\data.ms (-1268) (-)


Fluorene
Concen: 0.094 ng
RT: 15.285 min Scan# 1274
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53




Tgt Ion:	166	Resp:	12505
Ion	Ratio	Lower	Upper
166	100		
165	97.9	78.5	117.7
167	13.3	10.8	16.2



Abundance Scan 1274 (15.285 min): BN016982.D\data.ms (-1258) (-)





Abundance Scan 1411 (16.982 min): BN016984.D\data.ms (-1411) (-)




Phenanthrene-d10
Concen: 0.400 ng
RT: 16.982 min Scan# 1411
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	188	Resp:	76441
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	7.3	6.2	9.4



Abundance Scan 1411 (16.982 min): BN016982.D\data.ms (-1395) (-)





Abundance Scan 1281 (15.374 min): BN016984.D\data.ms (-1281) (-)




4,6-Dinitro-2-methylphenol
Concen: 0.021 ng
RT: 15.387 min Scan# 1282
Delta R.T. 0.013 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53

Tgt Ion:	198	Resp:	228
Ion	Ratio	Lower	Upper
198	100		
182	42.2	12.5	18.7#
77	0.0	0.0	0.0




Abundance Scan 1282 (15.387 min): BN016982.D\data.ms (-1273) (-)






Abundance Scan 1345 (16.179 min): BN016984.D\data.ms (-1321) (-)

4-Bromophenyl-phenylether
Concen: 0.086 ng
RT: 16.179 min Scan# 1345
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:	248	Resp:	3934
Ion	Ratio	Lower	Upper
248	100		
250	90.7	74.3	111.5
141	74.3	57.0	85.4






Abundance Scan 1354 (16.289 min): BN016984.D\data.ms (-1322) (-)

Hexachlorobenzene
Concen: 0.095 ng
RT: 16.289 min Scan# 1354
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53




Tgt Ion:	284	Resp:	4982
Ion	Ratio	Lower	Upper
284	100		
142	32.7	26.6	39.8
249	29.3	23.8	35.6





Abundance Scan 1368 (16.459 min): BN016984.D\data.ms (-1324) (-)

Atrazine
Concen: 0.065 ng
RT: 16.459 min Scan# 1368
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1



Tgt Ion	Ratio	Lower	Upper
200	100		
173	29.9	23.5	35.3
215	49.8	37.9	56.9




Abundance Scan 1368 (16.459 min): BN016982.D\data.ms (-1327) (-)





Abundance Scan 1383 (16.641 min): BN016984.D\data.ms (-1324) (-)


Pentachlorophenol
Concen: 0.037 ng
RT: 16.642 min Scan# 1383
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53




Tgt Ion	Ratio	Lower	Upper
266	100		
264	63.4	50.0	75.0
268	63.0	50.9	76.3



Abundance Scan 1383 (16.642 min): BN016982.D\data.ms (-1375) (-)





Abundance Scan 1414 (17.019 min): BN016984.D\data.ms (-1425) (-)




Phenanthrene
Concen: 0.098 ng
RT: 17.019 min Scan# 1414
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53

Tgt Ion:	178	Resp:	21381
Ion	Ratio	Lower	Upper
178	100		
176	18.8	15.0	22.6
179	15.8	12.3	18.5


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1



Abundance Scan 1414 (17.019 min): BN016982.D\data.ms (-1389) (-)





Abundance Scan 1422 (17.116 min): BN016984.D\data.ms (-1425) (-)




Anthracene
Concen: 0.082 ng
RT: 17.116 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53

Tgt Ion:	178	Resp:	15442
Ion	Ratio	Lower	Upper
178	100		
176	18.3	14.6	21.8
179	15.4	12.3	18.5




Abundance Scan 1422 (17.116 min): BN016982.D\data.ms (-1414) (-)






Abundance Scan 1698 (19.019 min): BN016984.D\data.ms (-1625) (-)

Fluoranthene-d10
Concen: 0.087 ng
RT: 19.020 min Scan# 1698
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:	212	Resp:	16890
Ion Ratio	Lower	Upper	
212	100		
106	13.6	11.0	16.4
104	7.5	6.0	9.0






Abundance Scan 1704 (19.049 min): BN016984.D\data.ms (-1628) (-)

Fluoranthene
Concen: 0.088 ng
RT: 19.049 min Scan# 1704
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53






Tgt Ion:	202	Resp:	21151
Ion Ratio	Lower	Upper	
202	100		
101	11.7	9.3	13.9
203	16.9	13.8	20.8




Abundance Scan 2008 (21.186 min): BN016984.D\data.ms (-2029) (-)

Chrysene-d12
Concen: 0.400 ng
RT: 21.186 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1






Tgt Ion:	240	Resp:	67811
Ion Ratio	Lower	Upper	
240	100		
120	7.8	6.4	9.6
236	25.7	20.6	31.0




Abundance Scan 1777 (19.412 min): BN016984.D\data.ms (-1738) (-)

Pyrene
Concen: 0.099 ng
RT: 19.412 min Scan# 1777
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53




Tgt Ion:	202	Resp:	21291
Ion Ratio	Lower	Upper	
202	100		
200	20.4	16.5	24.7
203	17.5	14.0	21.0






Abundance Scan 1822 (19.635 min): BN016984.D\data.ms (-183) (-)

Terphenyl-d14
Concen: 0.099 ng
RT: 19.635 min Scan# 1822
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion	Ratio	Lower	Upper
244	100		
212	7.5	5.6	8.4
122	11.9	9.0	13.6






Abundance Scan 2006 (21.165 min): BN016984.D\data.ms (-193) (-)

Benzo(a)anthracene
Concen: 0.085 ng
RT: 21.165 min Scan# 2006
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53




Tgt Ion	Ratio	Lower	Upper
228	100		
226	27.8	22.3	33.5
229	19.4	15.4	23.0






Abundance Scan 2011 (21.218 min): BN016984.D\data.ms (-2003) (-)

Chrysene
Concen: 0.105 ng
RT: 21.218 min Scan# 2011
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:	228	Resp:	23141
Ion	Ratio	Lower	Upper
228	100		
226	30.2	24.2	36.4
229	19.5	15.4	23.2






Abundance Scan 2002 (21.122 min): BN016984.D\data.ms (-1994) (-)

Bis(2-ethylhexyl)phthalate
Concen: 0.066 ng
RT: 21.123 min Scan# 2002
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53




Tgt Ion:	149	Resp:	6468
Ion	Ratio	Lower	Upper
149	100		
167	27.1	21.8	32.6
279	4.3	3.4	5.0






Abundance Scan 2415 (23.407 min): BN016984.D\data.ms (-2395) (-)

Perylene-d12
Concen: 0.400 ng
RT: 23.407 min Scan# 2415
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.4	19.7	29.5
265	23.2	18.3	27.5






Abundance Scan 3070 (25.649 min): BN016984.D\data.ms (-3046) (-)


Indeno(1,2,3-cd)pyrene
Concen: 0.083 ng
RT: 25.649 min Scan# 3070
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53



Tgt Ion	Ratio	Lower	Upper
276	100		
138	25.7	19.6	29.4
277	25.3	20.2	30.2





Abundance Scan 2221 (22.743 min): BN016984.D\data.ms (-2162) (-)




Benzo(b)fluoranthene
Concen: 0.086 ng
RT: 22.740 min Scan# 2220
Delta R.T. -0.003 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53


Tgt Ion:	252	Resp:	20184
Ion	Ratio	Lower	Upper
252	100		
253	23.7	17.7	26.5
125	11.1	8.0	12.0



Abundance Scan 2220 (22.740 min): BN016982.D\data.ms (-2162) (-)





Abundance Scan 2233 (22.784 min): BN016984.D\data.ms (-2218) (-)




Benzo(k)fluoranthene
Concen: 0.092 ng
RT: 22.784 min Scan# 2233
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53

Tgt Ion:	252	Resp:	21247
Ion	Ratio	Lower	Upper
252	100		
253	24.2	17.8	26.8
125	13.8	8.5	12.7#




Abundance Scan 2233 (22.784 min): BN016982.D\data.ms (-2218) (-)





Abundance Scan 2387 (23.311 min): BN016984.D\data.ms (-2387) (-)

Benzo(a)pyrene
Concen: 0.083 ng
RT: 23.308 min Scan# 2386
Delta R.T. -0.003 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1



Tgt Ion	Ratio	Lower	Upper
252	100		
253	25.4	17.9	26.9
125	12.8	9.1	13.7




Abundance Scan 2386 (23.308 min): BN016982.D\data.ms (-2328) (-)





Abundance Scan 3076 (25.669 min): BN016984.D\data.ms (-3040) (-)


Dibenzo(a,h)anthracene
Concen: 0.078 ng
RT: 25.670 min Scan# 3076
Delta R.T. 0.000 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53




Tgt Ion	Ratio	Lower	Upper
278	100		
139	18.4	13.4	20.2
279	25.8	19.2	28.8



Abundance Scan 3076 (25.670 min): BN016982.D\data.ms (-3047) (-)




Abundance Scan 3270 (26.333 min): BN016984.D\data.ms (-3245) (-)



Benzo(g,h,i)perylene
Concen: 0.086 ng
RT: 26.330 min Scan# 3269
Delta R.T. -0.003 min
Lab File: BN016982.D
Acq: 18 Oct 2021 10:53

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1



Tgt Ion:	276	Resp:	19880
Ion	Ratio	Lower	Upper
276	100		
277	24.4	19.2	28.8
138	21.9	16.9	25.3

Abundance Scan 3269 (26.330 min): BN016982.D\data.ms (-3211) (-)

