

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101821\
 Data File : BN016994.D
 Acq On : 18 Oct 2021 19:03
 Operator : CG/JU
 Sample : M4251-05
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SW-RR-01-(2.0)

Manual Integrations
 APPROVED

mohammad
 10/19/2021 12:11:33 PM

Quant Time: Oct 19 04:12:08 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 18 15:14:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.607	152	19893	0.400	ng	0.00
7) Naphthalene-d8	10.370	136	91494	0.400	ng	0.00
13) Acenaphthene-d10	14.232	164	49824	0.400	ng	0.00
19) Phenanthrene-d10	16.982	188	98337	0.400	ng	0.00
29) Chrysene-d12	21.186	240	93604	0.400	ng	0.00
35) Perylene-d12	23.407	264	94643	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.237	112	6212	0.149	ng	0.03
5) Phenol-d6	6.795	99	3942	0.082	ng	0.00
8) Nitrobenzene-d5	8.748	82	17154	0.289	ng	0.00
11) 2-Methylnaphthalene-d10	11.964	152	44188	0.341	ng	0.00
14) 2,4,6-Tribromophenol	15.729	330	7127	0.398	ng	0.00
15) 2-Fluorobiphenyl	12.849	172	59776	0.320	ng	0.00
27) Fluoranthene-d10	19.020	212	109148	0.443	ng	0.00
31) Terphenyl-d14	19.630	244	92999	0.449	ng	0.00
Target Compounds						
24) Pentachlorophenol	16.642	266	141	0.243	ng	91
34) Bis(2-ethylhexyl)phtha...	21.123	149	27829	0.362	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101821\
Data File : BN016994.D
Acq On : 18 Oct 2021 19:03
Operator : CG/JU
Sample : M4251-05
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SW-RR-01-(2.0)

Manual Integrations
APPROVED

mohammad
10/19/2021 12:11:33 PM

Quant Time: Oct 19 04:12:08 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 18 15:14:53 2021
Response via : Initial Calibration

