


2,4,6-Tribromophenol

Response Ratio



$$\text{Response} = 2.200\text{e-}001 * \text{Amt} - 7.605\text{e-}002$$


Coef of Det (r^2) = 0.998879 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN101821.M

Calibration Table Last Updated: Mon Oct 18 15:14:53 2021

4,6-Dinitro-2-methylphenol


Response Ratio



$R = 3.017e-003 A^2 + 4.802e-002 A - 2.190e-002$
 Coef of Det (r^2) = 0.999181 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN101821.M
 Calibration Table Last Updated: Mon Oct 18 15:14:53 2021

Atrazine

Response Ratio



$$\text{Response} = 2.015\text{e-}001 * \text{Amt} - 6.022\text{e-}002$$


Coef of Det (r^2) = 0.999255 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN101821.M

Calibration Table Last Updated: Mon Oct 18 15:14:53 2021

Pentachlorophenol

Response Ratio



$$\text{Response} = 1.248\text{e-}001 * \text{Amt} - 7.435\text{e-}002$$


Coef of Det (r^2) = 0.994775 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN101821.M

Calibration Table Last Updated: Mon Oct 18 15:14:53 2021

Bis(2-ethylhexyl)phthalate

Response Ratio



$$\text{Response} = 7.695\text{e-}001 * \text{Amt} - 4.000\text{e-}001$$

Coef of Det (r^2) = 0.995534 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN101821.M

Calibration Table Last Updated: Mon Oct 18 15:14:53 2021