

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101821\
 Data File : BN016981.D
 Acq On : 18 Oct 2021 10:17
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Manual Integrations
 APPROVED

mohammad

Quant Time: Oct 18 13:21:39 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 18 13:18:40 2021
 Response via : Initial Calibration

10/19/2021 12:11:22 PM

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.598	152	17927	0.40	ng	0.00
7) Naphthalene-d8	10.370	136	80623	0.40	ng	0.00
13) Acenaphthene-d10	14.232	164	42794	0.40	ng	0.00
19) Phenanthrene-d10	16.982	188	79526	0.40	ng	0.00
29) Chrysene-d12	21.186	240	78398	0.40	ng	0.00
35) Perylene-d12	23.407	264	74762	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.209	112	14285	0.32	ng	0.00
5) Phenol-d6	6.786	99	16010	0.32	ng	0.00
8) Nitrobenzene-d5	8.748	82	18755	0.37	ng	0.00
11) 2-Methylnaphthalene-d10	11.964	152	43926	0.38	ng	0.00
14) 2,4,6-Tribromophenol	15.729	330	5780	0.30	ng	0.00
15) 2-Fluorobiphenyl	12.849	172	63412	0.39	ng	0.00
27) Fluoranthene-d10	19.020	212	75980	0.38	ng	0.00
31) Terphenyl-d14	19.630	244	67537	0.39	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.199	88	4786m	0.41	ng	
3) n-Nitrosodimethylamine	3.531	42	5423	0.37	ng	# 99
6) bis(2-Chloroethyl)ether	7.044	93	19327	0.39	ng	100
9) Naphthalene	10.421	128	81343	0.36	ng	100
10) Hexachlorobutadiene	10.712	225	15957	0.39	ng	# 100
12) 2-Methylnaphthalene	12.035	142	52745	0.39	ng	98
16) Acenaphthylene	13.940	152	64691	0.37	ng	100
17) Acenaphthene	14.295	154	46085	0.39	ng	99
18) Fluorene	15.285	166	55775	0.39	ng	100
20) 4,6-Dinitro-2-methylph...	15.374	198	2329	0.20	ng	98
21) 4-Bromophenyl-phenylether	16.179	248	17859	0.37	ng	96
22) Hexachlorobenzene	16.289	284	21079	0.39	ng	99
23) Atrazine	16.459	200	10449	0.32	ng	98
24) Pentachlorophenol	16.642	266	4702	0.27	ng	99
25) Phenanthrene	17.019	178	89512	0.39	ng	100
26) Anthracene	17.116	178	72544	0.37	ng	100
28) Fluoranthene	19.049	202	97746	0.39	ng	100
30) Pyrene	19.412	202	95179	0.38	ng	100
32) Benzo(a)anthracene	21.165	228	87233	0.36	ng	99
33) Chrysene	21.218	228	99334	0.39	ng	100
34) Bis(2-ethylhexyl)phtha...	21.123	149	30218	0.27	ng	100
36) Indeno(1,2,3-cd)pyrene	25.649	276	96833	0.35	ng	100
37) Benzo(b)fluoranthene	22.740	252	93901	0.38	ng	100
38) Benzo(k)fluoranthene	22.784	252	95995	0.39	ng	100
39) Benzo(a)pyrene	23.308	252	81687	0.37	ng	100
40) Dibenzo(a,h)anthracene	25.666	278	76845	0.34	ng	100
41) Benzo(g,h,i)perylene	26.327	276	88070	0.36	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101821\
Data File : BN016981.D
Acq On : 18 Oct 2021 10:17
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED


mohammad

Quant Time: Oct 18 13:21:39 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 18 13:18:40 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.598 min Scan# 501
Delta R.T. -0.009 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17





Tgt Ion:152 Resp: 17927
Ion Ratio Lower Upper
152 100
150 158.2 123.0 184.4
115 61.2 46.3 69.5



#2
1,4-Dioxane
Concen: 0.41 ng m
RT: 3.199 min Scan# 24
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17



Tgt Ion: 88 Resp: 4786
Ion Ratio Lower Upper
88 100
43 27.7 26.8 40.2
58 65.9 63.0 94.4




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad

10/19/2021 12:11:22 PM



#3
n-Nitrosodimethylamine
Concen: 0.37 ng
RT: 3.531 min Scan# 60
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17





Tgt Ion: 42 Resp: 5423
Ion Ratio Lower Upper
42 100
74 142.0 114.7 172.1
44 5.7 3.8 5.6#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED




mohammad


10/19/2021 12:11:22 PM



#4
2-Fluorophenol
Concen: 0.32 ng
RT: 5.209 min Scan# 242
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17

Tgt Ion: 112 Resp: 14285
Ion Ratio Lower Upper
112 100
64 56.8 45.5 68.3
63 28.7 23.0 34.6





#5
Phenol-d6
Concen: 0.32 ng
RT: 6.786 min Scan# 413
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17





Tgt Ion:	99	Resp:	16010
Ion	Ratio	Lower	Upper
99	100		
42	18.1	14.6	21.8
71	30.9	24.8	37.2

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED




mohammad


10/19/2021 12:11:22 PM



#6
bis(2-Chloroethyl)ether
Concen: 0.39 ng
RT: 7.044 min Scan# 441
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17


Tgt Ion:	93	Resp:	19327
Ion	Ratio	Lower	Upper
93	100		
63	76.6	60.9	91.3
95	33.2	26.4	39.6





#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.370 min Scan# 741
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:	136	Resp:	80623
Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.8	13.2
54	5.7	5.2	7.8
68	4.8	4.4	6.6

Manual Integrations
APPROVED


mohammad





10/19/2021 12:11:22 PM




#8
Nitrobenzene-d5
Concen: 0.37 ng
RT: 8.748 min Scan# 613
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17



Tgt Ion:	82	Resp:	18755
Ion	Ratio	Lower	Upper
82	100		
128	45.4	33.9	50.9
54	45.6	38.3	57.5





#9
Naphthalene
Concen: 0.36 ng
RT: 10.421 min Scan# 745
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17





Tgt Ion:	128	Resp:	81343
Ion Ratio	Lower	Upper	
128	100		
129	11.1	8.8	13.2
127	12.8	10.3	15.5

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED




mohammad


10/19/2021 12:11:22 PM



#10
Hexachlorobutadiene
Concen: 0.39 ng
RT: 10.712 min Scan# 768
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17

Tgt Ion:	225	Resp:	15957
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.9	51.2	76.8







#11
2-Methylnaphthalene-d10
Concen: 0.38 ng
RT: 11.964 min Scan# 933
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17

Tgt Ion:152 Resp: 43926
Ion Ratio Lower Upper
152 100
151 20.4 16.3 24.5



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







10/19/2021 12:11:22 PM




#12
2-Methylnaphthalene
Concen: 0.39 ng
RT: 12.035 min Scan# 949
Delta R.T. -0.004 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17

Tgt Ion:142 Resp: 52745
Ion Ratio Lower Upper
142 100
141 89.5 70.4 105.6
115 34.0 26.4 39.6






#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.232 min Scan# 1191
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17



Tgt Ion:164 Resp: 42794

Ion	Ratio	Lower	Upper
164	100		
162	100.3	80.2	120.2
160	44.5	36.0	54.0




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad

10/19/2021 12:11:22 PM






#14
2,4,6-Tribromophenol
Concen: 0.30 ng
RT: 15.729 min Scan# 1309
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17



Tgt Ion:330 Resp: 5780

Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	30.4	25.6	38.4





#15
2-Fluorobiphenyl
Concen: 0.39 ng
RT: 12.849 min Scan# 1082
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17





Tgt Ion:	172	Resp:	63412
Ion	Ratio	Lower	Upper
172	100		
171	35.1	28.7	43.1
170	23.5	19.6	29.4

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED




mohammad


10/19/2021 12:11:22 PM



#16
Acenaphthylene
Concen: 0.37 ng
RT: 13.940 min Scan# 1168
Delta R.T. -0.012 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17

Tgt Ion:	152	Resp:	64691
Ion	Ratio	Lower	Upper
152	100		
151	19.4	15.4	23.2
153	13.0	10.4	15.6





#17
Acenaphthene
Concen: 0.39 ng
RT: 14.295 min Scan# 1196
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17





Tgt Ion:	154	Resp:	46085
Ion	Ratio	Lower	Upper
154	100		
153	113.2	91.1	136.7
152	55.0	44.8	67.2

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED




mohammad


10/19/2021 12:11:22 PM



#18
Fluorene
Concen: 0.39 ng
RT: 15.285 min Scan# 1274
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17

Tgt Ion:	166	Resp:	55775
Ion	Ratio	Lower	Upper
166	100		
165	97.9	78.5	117.7
167	13.5	10.8	16.2









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 16.982 min Scan# 1411
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17

Tgt Ion:	188	Resp:	79526
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	6.9	6.2	9.4

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad



#20
4,6-Dinitro-2-methylphenol
Concen: 0.20 ng
RT: 15.374 min Scan# 1281
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17

Tgt Ion:	198	Resp:	2329
Ion Ratio	Lower	Upper	
198	100		
182	14.7	12.5	18.7
77	0.0	0.0	0.0





#21
4-Bromophenyl-phenylether
Concen: 0.37 ng
RT: 16.179 min Scan# 1345
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17





Tgt Ion:	248	Resp:	17859
Ion	Ratio	Lower	Upper
248	100		
250	93.6	74.3	111.5
141	64.6	57.0	85.4

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED




mohammad


10/19/2021 12:11:22 PM



#22
Hexachlorobenzene
Concen: 0.39 ng
RT: 16.289 min Scan# 1354
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17

Tgt Ion:	284	Resp:	21079
Ion	Ratio	Lower	Upper
284	100		
142	32.8	26.6	39.8
249	29.0	23.8	35.6





#23
Atrazine
Concen: 0.32 ng
RT: 16.459 min Scan# 1368
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17





Tgt Ion:	200	Resp:	10449
Ion Ratio	Lower	Upper	
200	100		
173	26.8	23.5	35.3
215	47.5	37.9	56.9

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED




mohammad


10/19/2021 12:11:22 PM



#24
Pentachlorophenol
Concen: 0.27 ng
RT: 16.642 min Scan# 1383
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17

Tgt Ion:	266	Resp:	4702
Ion Ratio	Lower	Upper	
266	100		
264	62.7	50.0	75.0
268	64.4	50.9	76.3





#25
Phenanthrene
Concen: 0.39 ng
RT: 17.019 min Scan# 1414
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17





Tgt Ion:178 Resp: 89512
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.6
179 15.5 12.3 18.5

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED




mohammad


10/19/2021 12:11:22 PM



#26
Anthracene
Concen: 0.37 ng
RT: 17.116 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17





Tgt Ion:178 Resp: 72544
Ion Ratio Lower Upper
178 100
176 18.1 14.6 21.8
179 15.3 12.3 18.5








#27
Fluoranthene-d10
Concen: 0.38 ng
RT: 19.020 min Scan# 1698
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17

Tgt Ion:	212	Resp:	75980
Ion	Ratio	Lower	Upper
212	100		
106	13.8	11.0	16.4
104	7.6	6.0	9.0



#28
Fluoranthene
Concen: 0.39 ng
RT: 19.049 min Scan# 1704
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17

Tgt Ion:	202	Resp:	97746
Ion	Ratio	Lower	Upper
202	100		
101	11.6	9.3	13.9
203	17.3	13.8	20.8




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad

10/19/2021 12:11:22 PM



#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.186 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17





Tgt Ion:240 Resp: 78398
Ion Ratio Lower Upper
240 100
120 7.2 6.4 9.6
236 25.5 20.6 31.0

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED




mohammad


10/19/2021 12:11:22 PM




#30
Pyrene
Concen: 0.38 ng
RT: 19.412 min Scan# 1777
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17

Tgt Ion:202 Resp: 95179
Ion Ratio Lower Upper
202 100
200 20.6 16.5 24.7
203 17.4 14.0 21.0







#31
Terphenyl-d14
Concen: 0.39 ng
RT: 19.630 min Scan# 1821
Delta R.T. -0.005 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17



Tgt Ion:244 Resp: 67537
Ion Ratio Lower Upper
244 100
212 7.2 5.6 8.4
122 13.0 9.0 13.6




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad



10/19/2021 12:11:22 PM




#32
Benzo(a)anthracene
Concen: 0.36 ng
RT: 21.165 min Scan# 2006
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17



Tgt Ion:228 Resp: 87233
Ion Ratio Lower Upper
228 100
226 27.5 22.3 33.5
229 19.2 15.4 23.0





#33
Chrysene
Concen: 0.39 ng
RT: 21.218 min Scan# 2011
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17





Tgt Ion:	228	Resp:	99334
Ion Ratio	Lower	Upper	
228	100		
226	29.9	24.2	36.4
229	19.3	15.4	23.2

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED




mohammad


10/19/2021 12:11:22 PM




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.27 ng
RT: 21.123 min Scan# 2002
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17

Tgt Ion:	149	Resp:	30218
Ion Ratio	Lower	Upper	
149	100		
167	27.3	21.8	32.6
279	4.2	3.4	5.0






#35
Perylene-d12
Concen: 0.40 ng
RT: 23.407 min Scan# 2415
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17



Tgt Ion: 264 Resp: 74762

Ion	Ratio	Lower	Upper
264	100		
260	24.5	19.7	29.5
265	23.1	18.3	27.5




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad

10/19/2021 12:11:22 PM






#36
Indeno(1,2,3-cd)pyrene
Concen: 0.35 ng
RT: 25.649 min Scan# 3070
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17



Tgt Ion: 276 Resp: 96833

Ion	Ratio	Lower	Upper
276	100		
138	24.5	19.6	29.4
277	25.1	20.2	30.2





#37
Benzo(b)fluoranthene
Concen: 0.38 ng
RT: 22.740 min Scan# 2220
Delta R.T. -0.003 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17





Tgt Ion:252	Resp:	93901
Ion Ratio	Lower	Upper
252	100	
253	22.1	17.7 26.5
125	10.2	8.0 12.0

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED




mohammad


10/19/2021 12:11:22 PM




#38
Benzo(k)fluoranthene
Concen: 0.39 ng
RT: 22.784 min Scan# 2233
Delta R.T. 0.000 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17

Tgt Ion:252	Resp:	95995
Ion Ratio	Lower	Upper
252	100	
253	22.4	17.8 26.8
125	10.6	8.5 12.7






#39
Benzo(a)pyrene
Concen: 0.37 ng
RT: 23.308 min Scan# 2386
Delta R.T. -0.003 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17



Tgt Ion:252 Resp: 81687
Ion Ratio Lower Upper
252 100
253 22.4 17.9 26.9
125 11.3 9.1 13.7




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad



10/19/2021 12:11:22 PM




#40
Dibenzo(a,h)anthracene
Concen: 0.34 ng
RT: 25.666 min Scan# 3075
Delta R.T. -0.003 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17



Tgt Ion:278 Resp: 76845
Ion Ratio Lower Upper
278 100
139 16.7 13.4 20.2
279 23.9 19.2 28.8








#41
Benzo(g,h,i)perylene
Concen: 0.36 ng
RT: 26.327 min Scan# 3268
Delta R.T. -0.007 min
Lab File: BN016981.D
Acq: 18 Oct 2021 10:17

Tgt Ion:	276	Resp:	88070
Ion Ratio	Lower	Upper	
276	100		
277	23.7	19.2	28.8
138	21.4	16.9	25.3

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad



10/19/2021 12:11:22 PM