

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101821\
 Data File : BN016987.D
 Acq On : 18 Oct 2021 13:54
 Operator : CG/JU
 Sample : SSTDIC3.2
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC3.2

Quant Time: Oct 18 14:24:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 18 13:18:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.607	152	19503	0.40	ng	0.00
7) Naphthalene-d8	10.370	136	82936	0.40	ng	0.00
13) Acenaphthene-d10	14.232	164	45807	0.40	ng	0.00
19) Phenanthrene-d10	16.982	188	86391	0.40	ng	0.00
29) Chrysene-d12	21.186	240	86856	0.40	ng	0.00
35) Perylene-d12	23.403	264	82850	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.209	112	138374	2.85	ng	0.00
5) Phenol-d6	6.786	99	163218	3.03	ng	0.00
8) Nitrobenzene-d5	8.747	82	193172	3.66	ng	0.00
11) 2-Methylnaphthalene-d10	11.964	152	376019	3.20	ng	0.00
14) 2,4,6-Tribromophenol	15.729	330	76255	3.70	ng	0.00
15) 2-Fluorobiphenyl	12.849	172	531288	3.05	ng	0.00
27) Fluoranthene-d10	19.019	212	728262	3.32	ng	0.00
31) Terphenyl-d14	19.630	244	611134	3.17	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.199	88	39303	3.12	ng	96
3) n-Nitrosodimethylamine	3.522	42	51421	3.23	ng	# 99
6) bis(2-Chloroethyl)ether	7.044	93	166685	3.08	ng	99
9) Naphthalene	10.420	128	665675	2.85	ng	100
10) Hexachlorobutadiene	10.712	225	132379	3.14	ng	# 100
12) 2-Methylnaphthalene	12.039	142	433076	3.11	ng	100
16) Acenaphthylene	13.953	152	648069	3.42	ng	100
17) Acenaphthene	14.295	154	400748	3.13	ng	98
18) Fluorene	15.285	166	493044	3.22	ng	99
20) 4,6-Dinitro-2-methylph...	15.374	198	49177	3.93	ng	# 80
21) 4-Bromophenyl-phenylether	16.179	248	168597	3.26	ng	99
22) Hexachlorobenzene	16.288	284	182847	3.08	ng	99
23) Atrazine	16.459	200	134538	3.75	ng	98
24) Pentachlorophenol	16.641	266	77345	4.08	ng	100
25) Phenanthrene	17.019	178	758553	3.08	ng	100
26) Anthracene	17.116	178	718872	3.39	ng	100
28) Fluoranthene	19.049	202	855125	3.17	ng	100
30) Pyrene	19.411	202	875833	3.17	ng	100
32) Benzo(a)anthracene	21.165	228	898738	3.39	ng	99
33) Chrysene	21.218	228	862455	3.06	ng	99
34) Bis(2-ethylhexyl)phtha...	21.122	149	494942	3.97	ng	99
36) Indeno(1,2,3-cd)pyrene	25.649	276	945700	3.09	ng	100
37) Benzo(b)fluoranthene	22.739	252	892419	3.25	ng	100
38) Benzo(k)fluoranthene	22.784	252	866858	3.19	ng	99
39) Benzo(a)pyrene	23.308	252	811631	3.35	ng	99
40) Dibenzo(a,h)anthracene	25.669	278	768284	3.10	ng	99
41) Benzo(g,h,i)perylene	26.330	276	828461	3.06	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101821\
Data File : BN016987.D
Acq On : 18 Oct 2021 13:54
Operator : CG/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Quant Time: Oct 18 14:24:29 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN101821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 18 13:18:40 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.607 min Scan# 502
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 19503
Ion Ratio Lower Upper
152 100
150 156.5 123.0 184.4
115 58.5 46.3 69.5



#2
1,4-Dioxane
Concen: 3.12 ng
RT: 3.199 min Scan# 24
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54

Tgt Ion: 88 Resp: 39303
Ion Ratio Lower Upper
88 100
43 28.7 26.8 40.2
58 76.4 63.0 94.4









#3
n-Nitrosodimethylamine
Concen: 3.23 ng
RT: 3.522 min Scan# 59
Delta R.T. -0.009 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 51421
Ion Ratio Lower Upper
42 100
74 144.7 114.7 172.1
44 6.4 3.8 5.6#



#4
2-Fluorophenol
Concen: 2.85 ng
RT: 5.209 min Scan# 242
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54

Tgt Ion: 112 Resp: 138374
Ion Ratio Lower Upper
112 100
64 56.6 45.5 68.3
63 28.5 23.0 34.6









#5
Phenol-d6
Concen: 3.03 ng
RT: 6.786 min Scan# 413
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 99 Resp: 163218
Ion Ratio Lower Upper
99 100
42 17.9 14.6 21.8
71 30.9 24.8 37.2



#6
bis(2-Chloroethyl)ether
Concen: 3.08 ng
RT: 7.044 min Scan# 441
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54

Tgt Ion: 93 Resp: 166685
Ion Ratio Lower Upper
93 100
63 75.0 60.9 91.3
95 32.5 26.4 39.6









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.370 min Scan# 741
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	82936
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.8	13.2
54	6.5	5.2	7.8
68	5.4	4.4	6.6



#8
Nitrobenzene-d5
Concen: 3.66 ng
RT: 8.747 min Scan# 613
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54

Tgt Ion:	82	Resp:	193172
Ion	Ratio	Lower	Upper
82	100		
128	45.3	33.9	50.9
54	45.6	38.3	57.5









#9
Naphthalene
Concen: 2.85 ng
RT: 10.420 min Scan# 745
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:128 Resp: 665675
Ion Ratio Lower Upper
128 100
129 11.0 8.8 13.2
127 12.8 10.3 15.5



#10
Hexachlorobutadiene
Concen: 3.14 ng
RT: 10.712 min Scan# 768
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54

Tgt Ion:225 Resp: 132379
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.1 51.2 76.8









#11
2-Methylnaphthalene-d10
Concen: 3.20 ng
RT: 11.964 min Scan# 933
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 376019
Ion Ratio Lower Upper
152 100
151 20.3 16.3 24.5



#12
2-Methylnaphthalene
Concen: 3.11 ng
RT: 12.039 min Scan# 950
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54

Tgt Ion:142 Resp: 433076
Ion Ratio Lower Upper
142 100
141 88.2 70.4 105.6
115 33.0 26.4 39.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.232 min Scan# 1191
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:164 Resp: 45807
Ion Ratio Lower Upper
164 100
162 101.6 80.2 120.2
160 45.4 36.0 54.0



#14
2,4,6-Tribromophenol
Concen: 3.70 ng
RT: 15.729 min Scan# 1309
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54

Tgt Ion:330 Resp: 76255
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 30.7 25.6 38.4









#15
2-Fluorobiphenyl
Concen: 3.05 ng
RT: 12.849 min Scan# 1082
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:172 Resp: 531288
Ion Ratio Lower Upper
172 100
171 35.9 28.7 43.1
170 24.4 19.6 29.4



#16
Acenaphthylene
Concen: 3.42 ng
RT: 13.953 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54

Tgt Ion:152 Resp: 648069
Ion Ratio Lower Upper
152 100
151 19.5 15.4 23.2
153 13.1 10.4 15.6









#17
Acenaphthene
Concen: 3.13 ng
RT: 14.295 min Scan# 1196
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:154 Resp: 400748
Ion Ratio Lower Upper
154 100
153 112.4 91.1 136.7
152 53.7 44.8 67.2



#18
Fluorene
Concen: 3.22 ng
RT: 15.285 min Scan# 1274
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54

Tgt Ion:166 Resp: 493044
Ion Ratio Lower Upper
166 100
165 97.1 78.5 117.7
167 13.5 10.8 16.2









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 16.982 min Scan# 1411
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:188 Resp: 86391




Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	7.8	6.2	9.4




#20
4,6-Dinitro-2-methylphenol
Concen: 3.93 ng
RT: 15.374 min Scan# 1281
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54

Tgt Ion:198 Resp: 49177

Ion	Ratio	Lower	Upper
198	100		
182	7.2	12.5	18.7#
77	0.0	0.0	0.0









#21
4-Bromophenyl-phenylether
Concen: 3.26 ng
RT: 16.179 min Scan# 1345
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:248 Resp: 168597




Ion	Ratio	Lower	Upper
248	100		
250	91.8	74.3	111.5
141	70.6	57.0	85.4




#22
Hexachlorobenzene
Concen: 3.08 ng
RT: 16.288 min Scan# 1354
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54

Tgt Ion:284 Resp: 182847

Ion	Ratio	Lower	Upper
284	100		
142	32.1	26.6	39.8
249	29.5	23.8	35.6









#23
Atrazine
Concen: 3.75 ng
RT: 16.459 min Scan# 1368
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:200 Resp: 134538
Ion Ratio Lower Upper
200 100
173 30.7 23.5 35.3
215 46.5 37.9 56.9



#24
Pentachlorophenol
Concen: 4.08 ng
RT: 16.641 min Scan# 1383
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54

Tgt Ion:266 Resp: 77345
Ion Ratio Lower Upper
266 100
264 62.5 50.0 75.0
268 64.1 50.9 76.3









#25
Phenanthrene
Concen: 3.08 ng
RT: 17.019 min Scan# 1414
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:178 Resp: 758553
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.6
179 15.4 12.3 18.5



#26
Anthracene
Concen: 3.39 ng
RT: 17.116 min Scan# 1422
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54

Tgt Ion:178 Resp: 718872
Ion Ratio Lower Upper
178 100
176 18.2 14.6 21.8
179 15.4 12.3 18.5









#27
Fluoranthene-d10
Concen: 3.32 ng
RT: 19.019 min Scan# 1698
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:212 Resp: 728262
Ion Ratio Lower Upper
212 100
106 13.8 11.0 16.4
104 7.6 6.0 9.0



#28
Fluoranthene
Concen: 3.17 ng
RT: 19.049 min Scan# 1704
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54

Tgt Ion:202 Resp: 855125
Ion Ratio Lower Upper
202 100
101 11.7 9.3 13.9
203 17.5 13.8 20.8









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.186 min Scan# 2008
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:240 Resp: 86856
Ion Ratio Lower Upper
240 100
120 6.7 6.4 9.6
236 25.3 20.6 31.0



#30
Pyrene
Concen: 3.17 ng
RT: 19.411 min Scan# 1777
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54

Tgt Ion:202 Resp: 875833
Ion Ratio Lower Upper
202 100
200 20.6 16.5 24.7
203 17.6 14.0 21.0









#31
Terphenyl-d14
Concen: 3.17 ng
RT: 19.630 min Scan# 1821
Delta R.T. -0.005 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:244 Resp: 611134
Ion Ratio Lower Upper
244 100
212 7.2 5.6 8.4
122 12.8 9.0 13.6



#32
Benzo(a)anthracene
Concen: 3.39 ng
RT: 21.165 min Scan# 2006
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54

Tgt Ion:228 Resp: 898738
Ion Ratio Lower Upper
228 100
226 27.1 22.3 33.5
229 19.6 15.4 23.0









#33
Chrysene
Concen: 3.06 ng
RT: 21.218 min Scan# 2011
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	228	Resp:	862455
Ion	Ratio	Lower	Upper
228	100		
226	29.8	24.2	36.4
229	19.7	15.4	23.2



#34
Bis(2-ethylhexyl)phthalate
Concen: 3.97 ng
RT: 21.122 min Scan# 2002
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54


Tgt Ion:	149	Resp:	494942
Ion	Ratio	Lower	Upper
149	100		
167	27.9	21.8	32.6
279	4.4	3.4	5.0








#35
Perylene-d12
Concen: 0.40 ng
RT: 23.403 min Scan# 2414
Delta R.T. -0.003 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2







Tgt Ion:264 Resp: 82850
Ion Ratio Lower Upper
264 100
260 24.5 19.7 29.5
265 22.7 18.3 27.5



#36
Indeno(1,2,3-cd)pyrene
Concen: 3.09 ng
RT: 25.649 min Scan# 3070
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54

Tgt Ion:276 Resp: 945700
Ion Ratio Lower Upper
276 100
138 24.5 19.6 29.4
277 25.2 20.2 30.2









#37
Benzo(b)fluoranthene
Concen: 3.25 ng
RT: 22.739 min Scan# 2220
Delta R.T. -0.003 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 892419
Ion Ratio Lower Upper
252 100
253 22.1 17.7 26.5
125 10.0 8.0 12.0



#38
Benzo(k)fluoranthene
Concen: 3.19 ng
RT: 22.784 min Scan# 2233
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54


Tgt Ion:252 Resp: 866858
Ion Ratio Lower Upper
252 100
253 22.2 17.8 26.8
125 10.0 8.5 12.7








#39
Benzo(a)pyrene
Concen: 3.35 ng
RT: 23.308 min Scan# 2386
Delta R.T. -0.003 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2







Tgt Ion:252 Resp: 811631
Ion Ratio Lower Upper
252 100
253 22.0 17.9 26.9
125 10.9 9.1 13.7



#40
Dibenzo(a,h)anthracene
Concen: 3.10 ng
RT: 25.669 min Scan# 3076
Delta R.T. -0.000 min
Lab File: BN016987.D
Acq: 18 Oct 2021 13:54

Tgt Ion:278 Resp: 768284
Ion Ratio Lower Upper
278 100
139 15.9 13.4 20.2
279 23.7 19.2 28.8





#41
 Benzo(g,h,i)perylene
 Concen: 3.06 ng
 RT: 26.330 min Scan# 3269
 Delta R.T. -0.003 min
 Lab File: BN016987.D
 Acq: 18 Oct 2021 13:54

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Tgt Ion:	276	Resp:	828461
Ion	Ratio	Lower	Upper
276	100		
277	23.9	19.2	28.8
138	21.3	16.9	25.3

