

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101822\
 Data File : BN022166.D
 Acq On : 18 Oct 2022 01:26
 Operator : CG/JU
 Sample : PB148279BL
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK279

Quant Time: Oct 18 02:27:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 13 03:08:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.952	152	142741	20.000	ng/u1	0.00
20) Naphthalene-d8	10.781	136	673439	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.628	164	441140	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.381	188	911054	20.000	ng/u1	-0.01
79) Chrysene-d12	21.580	240	698148	20.000	ng/u1	-0.01
88) Perylene-d12	24.045	264	603697	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.252	96	23649	6.279	ng/uL	0.00
4) Pyridine-d5	3.687	84	353471	30.173	ng/u1	0.00
7) Phenol-d5	7.105	99	462826	33.167	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.275	67	286057	32.714	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.475	132	339483	35.087	ng/u1	0.00
15) 4-Methylphenol-d8	8.669	113	370190	33.386	ng/u1	-0.01
21) Nitrobenzene-d5	9.128	128	181064	36.667	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.857	143	200424	39.599	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.399	165	347166	36.542	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.922	131	547520	35.250	ng/u1	-0.01
46) Dimethylphthalate-d6	14.040	166	1121852	36.482	ng/u1	0.00
49) Acenaphthylene-d8	14.322	160	1266453	36.880	ng/u1	0.00
54) 4-Nitrophenol-d4	14.828	143	197861	30.144	ng/u1	0.00
60) Fluorene-d10	15.622	176	953667	36.768	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	104748	20.154	ng/u1	0.00
73) Anthracene-d10	17.481	188	1495948	37.785	ng/u1	0.00
81) Pyrene-d10	19.780	212	1490346	42.876	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.886	264	1097643	37.074	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101822\
Data File : BN022166.D
Acq On : 18 Oct 2022 01:26
Operator : CG/JU
Sample : PB148279BL
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK279

Quant Time: Oct 18 02:27:21 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101222.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 13 03:08:08 2022
Response via : Initial Calibration

