

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101822\
 Data File : BN022173.D
 Acq On : 18 Oct 2022 12:27
 Operator : CG/JU
 Sample : PB148277BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS277

Quant Time: Oct 19 00:55:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 13 03:08:08 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
 Supervised By :mohammad ahmed 10/19/2022

10/19/2022
 Supervised By :mohammad
 ahmed

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.951	152	145949	20.000	ng/u1	0.00
20) Naphthalene-d8	10.781	136	682508	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.628	164	412727	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.386	188	841290	20.000	ng/u1	0.00
79) Chrysene-d12	21.580	240	691578	20.000	ng/u1	-0.01
88) Perylene-d12	24.051	264	585452	20.000	ng/u1	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.252	96	22117	5.743	ng/uL	0.00
4) Pyridine-d5	3.687	84	331885	27.707	ng/u1	0.00
7) Phenol-d5	7.110	99	462315	32.402	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.275	67	280916	31.419	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.475	132	345088	34.883	ng/u1	0.00
15) 4-Methylphenol-d8	8.675	113	366265	32.306	ng/u1	0.00
21) Nitrobenzene-d5	9.134	128	178064	35.580	ng/u1	0.00
24) 2-Nitrophenol-d4	9.857	143	195811	38.174	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.398	165	344498	35.779	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.928	131	469440	29.821	ng/u1	0.00
46) Dimethylphthalate-d6	14.039	166	986268	34.281	ng/u1	0.00
49) Acenaphthylene-d8	14.322	160	1137898	35.418	ng/u1	0.00
54) 4-Nitrophenol-d4	14.833	143	210620	34.297	ng/u1	0.00
60) Fluorene-d10	15.622	176	839982	34.615	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	158873	33.102	ng/u1	0.00
73) Anthracene-d10	17.480	188	1274823	34.870	ng/u1	0.00
81) Pyrene-d10	19.786	212	1433224	41.625	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.892	264	1067143	37.168	ng/u1	0.00

Target Compounds						Qvalue
2) 1,4-Dioxane	3.287	88	45976	10.955	ng/uL	96
5) Pyridine	3.710	79	327343	27.547	ng/u1	94
6) Benzaldehyde	7.081	77	279366	39.649	ng/u1	98
8) Phenol	7.134	94	462414	30.534	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.369	93	360416	29.915	ng/u1	100
12) 2-Chlorophenol	7.504	128	349611	32.005	ng/u1	96
13) 2-Methylphenol	8.404	108	340687	30.089	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.493	45	497271	28.844	ng/u1	99
16) Acetophenone	8.793	105	543582	30.009	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.775	70	282716	29.304	ng/u1	98
18) 4-Methylphenol	8.740	108	380960	30.048	ng/u1	97
19) Hexachloroethane	9.040	117	136385	31.980	ng/u1	99
22) Nitrobenzene	9.175	77	423170	32.022	ng/u1	100
23) Isophorone	9.704	82	801108	30.705	ng/u1	100
25) 2-Nitrophenol	9.893	139	204639	33.804	ng/u1	98
26) 2,4-Dimethylphenol	9.951	107	393109	30.799	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.192	93	507711	31.120	ng/u1	99
29) 2,4-Dichlorophenol	10.428	162	333715	33.233	ng/u1	99
30) Naphthalene	10.834	128	1128699	30.871	ng/u1	99
32) 4-Chloroaniline	10.951	127	452341	27.719	ng/u1	99
33) Hexachlorobutadiene	11.116	225	172469	31.302	ng/u1	96
34) Caprolactam	11.739	113	116854m	29.950	ng/u1	
35) 4-Chloro-3-methylphenol	12.081	107	386633	31.937	ng/u1	100
36) 2-Methylnaphthalene	12.445	142	766220	30.346	ng/u1	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101822\
 Data File : BN022173.D
 Acq On : 18 Oct 2022 12:27
 Operator : CG/JU
 Sample : PB148277BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :

SLCS277

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
 Supervised By :mohammad ahmed 10/19/2022

Quant Time: Oct 19 00:55:34 2022

Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101222.M

Quant Title : SVOA CALIBRATION

QLast Update : Thu Oct 13 03:08:08 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.669	142	768936	30.411	ng/ul	95
39) 1,2,4,5-Tetrachloroben...	12.816	216	348374	32.538	ng/ul	96
40) Hexachlorocyclopentadiene	12.786	237	157584	26.328	ng/ul	88
41) 2,4,6-Trichlorophenol	13.057	196	253388	34.708	ng/ul	98
42) 2,4,5-Trichlorophenol	13.134	196	283758	34.902	ng/ul	98
43) 1,1'-Biphenyl	13.457	154	999834	33.184	ng/ul	98
44) 2-Chloronaphthalene	13.504	162	768177	32.630	ng/ul	96
45) 2-Nitroaniline	13.716	65	258498	35.444	ng/ul	99
47) Dimethylphthalate	14.086	163	973071	31.465	ng/ul	99
48) 2,6-Dinitrotoluene	14.210	165	202957	35.036	ng/ul	98
50) Acenaphthylene	14.351	152	1212412	33.045	ng/ul	99
51) 3-Nitroaniline	14.539	138	226739	33.150	ng/ul	98
52) Acenaphthene	14.692	153	849957	32.667	ng/ul	98
53) 2,4-Dinitrophenol	14.745	184	96383	24.562	ng/ul	97
55) 4-Nitrophenol	14.851	109	157134	31.558	ng/ul	99
56) Dibenzofuran	15.028	168	1134881	31.965	ng/ul	100
57) 2,4-Dinitrotoluene	14.998	165	300140	35.368	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.251	232	222875	32.939	ng/ul	100
59) Diethylphthalate	15.451	149	999678	31.825	ng/ul	99
61) Fluorene	15.675	166	916973	32.127	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.669	204	425470	31.814	ng/ul	99
63) 4-Nitroaniline	15.710	138	246235	38.583	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.763	198	152591	29.614	ng/ul	98
67) N-Nitrosodiphenylamine	15.886	169	801453	32.942	ng/ul	99
68) 4-Bromophenyl-phenylether	16.569	248	266679	35.884	ng/ul	86
69) Hexachlorobenzene	16.680	284	297784	32.917	ng/ul	99
70) Atrazine	16.845	200	286466	34.964	ng/ul	98
71) Pentachlorophenol	17.027	266	179528	32.093	ng/ul	95
72) Phenanthrene	17.427	178	1471185	32.417	ng/ul	99
74) Anthracene	17.516	178	1446185	32.082	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.422	216	371977	35.153	ng/uL	96
76) Pentachlorobenzene	14.945	250	351307	30.931	ng/uL	91
77) Carbazole	17.792	167	1403915	33.672	ng/ul	98
78) Di-n-butylphthalate	18.351	149	1784402	34.533	ng/ul	99
80) Fluoranthene	19.451	202	1663147	38.479	ng/ul	97
82) Pyrene	19.816	202	1711908	37.859	ng/ul	98
83) Butylbenzylphthalate	20.710	149	806283	41.099	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.498	252	491606	32.869	ng/ul	94
85) Benzo(a)anthracene	21.568	228	1548537	35.042	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.486	149	1198722	42.246	ng/ul	100
87) Chrysene	21.621	228	1423538	33.521	ng/ul	97
89) Di-n-octyl phthalate	22.427	149	2002573	48.278	ng/ul	100
90) Benzo(b)fluoranthene	23.292	252	1343577	36.387	ng/ul	100
91) Benzo(k)fluoranthene	23.339	252	1275949	34.976	ng/ul	100
93) Benzo(a)pyrene	23.945	252	1125819	34.114	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.639	276	1295777	31.414	ng/ul	97
95) Dibenzo(a,h)anthracene	26.668	278	1102333	29.870	ng/ul	97
96) Benzo(g,h,i)perylene	27.439	276	1064452	28.684	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101822\
 Data File : BN022173.D
 Acq On : 18 Oct 2022 12:27
 Operator : CG/JU
 Sample : PB148277BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :

SLCS277


Manual Integrations


APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
 Supervised By :mohammad ahmed 10/19/2022

10/19/2022
 Supervised By :mohammad
 ahmed


10/19/2022





#2
1,4-Dioxane
Concen: 10.955 ng/uL
RT: 3.287 min Scan# 51
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Instrument :
BNA_N
ClientSampleId :
SLCS277






Tgt Ion: 88 Resp: 45970
Ion Ratio Lower Upper
88 100
43 26.9 22.6 34.0
58 70.9 59.9 89.9

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022




10/19/2022
Supervised By :mohammad
ahmed


10/19/2022



#5
Pyridine
Concen: 27.547 ng/uL
RT: 3.710 min Scan# 123
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27


Tgt Ion: 79 Resp: 327343
Ion Ratio Lower Upper
79 100
52 70.6 61.7 92.5
51 33.0 27.8 41.6





#6
Benzaldehyde
Concen: 39.649 ng/ul
RT: 7.081 min Scan# 696
Delta R.T. -0.012 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27



Instrument :
BNA_N
ClientSampleId :
SLCS277



Tgt Ion: 77 Resp: 279360
Ion Ratio Lower Upper
77 100
105 96.4 78.6 118.0
106 92.2 75.4 113.2


Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022




10/19/2022
Supervised By :mohammad
ahmed



10/19/2022




#8
Phenol
Concen: 30.534 ng/ul
RT: 7.134 min Scan# 705
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27




Tgt Ion: 94 Resp: 462414
Ion Ratio Lower Upper
94 100
65 27.1 22.1 33.1
66 35.3 29.0 43.6





#10
Bis(2-Chloroethyl)ether
Concen: 29.915 ng/ul
RT: 7.369 min Scan# 745
Delta R.T. -0.012 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Instrument :
BNA_N
ClientSampleId :
SLCS277






Tgt Ion: 93 Resp: 360410
Ion Ratio Lower Upper
93 100
63 76.3 61.1 91.7
95 32.6 26.4 39.6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022




10/19/2022
Supervised By :mohammad
ahmed


10/19/2022



#12
2-Chlorophenol
Concen: 32.005 ng/ul
RT: 7.504 min Scan# 768
Delta R.T. -0.012 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27


Tgt Ion:128 Resp: 349611
Ion Ratio Lower Upper
128 100
64 46.2 39.7 59.5
130 31.8 26.7 40.1





#13
2-Methylphenol
Concen: 30.089 ng/ul
RT: 8.404 min Scan# 91
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27



Instrument :
BNA_N
ClientSampleId :
SLCS277



Tgt Ion:108 Resp: 34068
Ion Ratio Lower Upper
108 100
107 92.3 72.4 108.6


Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022




10/19/2022
Supervised By :mohammad
ahmed



10/19/2022




#14
2,2'-oxybis(1-Chloropropane)
Concen: 28.844 ng/ul
RT: 8.493 min Scan# 936
Delta R.T. -0.012 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27




Tgt Ion: 45 Resp: 497271
Ion Ratio Lower Upper
45 100
77 14.4 11.7 17.5
79 11.2 8.7 13.1





#16
Acetophenone
Concen: 30.009 ng/ul
RT: 8.793 min Scan# 987
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Instrument :
BNA_N
ClientSampleId :
SLCS277






Tgt Ion: 105 Resp: 54358
Ion Ratio Lower Upper
105 100
77 78.5 64.9 97.3
51 26.4 21.8 32.6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022




10/19/2022
Supervised By :mohammad
ahmed


10/19/2022



#17
N-Nitroso-di-n-propylamine
Concen: 29.304 ng/ul
RT: 8.775 min Scan# 984
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27


Tgt Ion: 70 Resp: 282716
Ion Ratio Lower Upper
70 100
42 55.9 45.6 68.4
101 9.4 7.4 11.2
130 20.6 15.4 23.0





#18
4-Methylphenol
Concen: 30.048 ng/ul
RT: 8.740 min Scan# 91
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27


Instrument :
BNA_N
ClientSampleId :
SLCS277



Tgt Ion:108 Resp: 380960
Ion Ratio Lower Upper
108 100
107 112.3 87.2 130.8



Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022






10/19/2022
Supervised By :mohammad
ahmed


10/19/2022



#19
Hexachloroethane
Concen: 31.980 ng/ul
RT: 9.040 min Scan# 1029
Delta R.T. -0.012 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Tgt Ion:117 Resp: 136385
Ion Ratio Lower Upper
117 100
201 86.4 69.8 104.6
199 56.0 45.2 67.8





#22

Nitrobenzene

Concen: 32.022 ng/u1

RT: 9.175 min Scan# 1052

Delta R.T. -0.006 min

Lab File: BN022173.D

Acq: 18 Oct 2022 12:27

Instrument :

BNA_N

ClientSampleId :

SLCS277

Tgt Ion: 77 Resp: 423170

Ion	Ratio	Lower	Upper
77	100		
123	43.5	34.9	52.3
65	13.1	10.2	15.2

Manual Integrations





APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022

10/19/2022

Supervised By :mohammad
ahmed

10/19/2022



#23

Isophorone

Concen: 30.705 ng/u1

RT: 9.704 min Scan# 1142




Delta R.T. -0.006 min


Lab File: BN022173.D

Acq: 18 Oct 2022 12:27

Tgt Ion: 82 Resp: 801108

Ion	Ratio	Lower	Upper
82	100		
95	7.0	5.4	8.0
138	17.2	13.8	20.8





#25
2-Nitrophenol
Concen: 33.804 ng/ul
RT: 9.893 min Scan# 1174
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Instrument : BNA_N
ClientSampleId : SLCS277





Tgt Ion:139 Resp: 204639
Ion Ratio Lower Upper
139 100
65 51.2 42.7 64.1
109 29.0 23.8 35.6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022




10/19/2022
Supervised By :mohammad
ahmed


10/19/2022



#26
2,4-Dimethylphenol
Concen: 30.799 ng/ul
RT: 9.951 min Scan# 1184
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27


Tgt Ion:107 Resp: 393109
Ion Ratio Lower Upper
107 100
121 51.5 41.0 61.6
122 87.8 71.5 107.3





#27
 Bis(2-Chloroethoxy)methane
 Concen: 31.120 ng/ul
 RT: 10.192 min Scan# 1225
 Delta R.T. -0.006 min
 Lab File: BN022173.D
 Acq: 18 Oct 2022 12:27

Instrument :
 BNA_N
 ClientSampleId :
 SLCS277






Tgt Ion: 93 Resp: 507711
 Ion Ratio Lower Upper
 93 100
 95 32.0 25.2 37.8
 123 11.3 8.6 12.8

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
 Supervised By :mohammad ahmed 10/19/2022




10/19/2022
 Supervised By :mohammad
 ahmed


10/19/2022



#29
 2,4-Dichlorophenol
 Concen: 33.233 ng/ul
 RT: 10.428 min Scan# 1265
 Delta R.T. -0.006 min
 Lab File: BN022173.D
 Acq: 18 Oct 2022 12:27


Tgt Ion:162 Resp: 333715
 Ion Ratio Lower Upper
 162 100
 164 64.5 52.9 79.3
 98 38.0 30.5 45.7





#30
Naphthalene
Concen: 30.871 ng/uI
RT: 10.834 min Scan# 11
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Instrument :
BNA_N
ClientSampleId :
SLCS277






Tgt Ion:128 Resp: 112869
Ion Ratio Lower Upper
128 100
129 11.0 8.5 12.7
127 12.9 10.8 16.2

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022




10/19/2022
Supervised By :mohammad
ahmed


10/19/2022



#32
4-Chloroaniline
Concen: 27.719 ng/uI
RT: 10.951 min Scan# 1354
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27


Tgt Ion:127 Resp: 452341
Ion Ratio Lower Upper
127 100
129 32.3 26.2 39.4





#33
Hexachlorobutadiene
Concen: 31.302 ng/ul
RT: 11.116 min Scan# 1382
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Instrument :
BNA_N
ClientSampleId :
SLCS277






Tgt Ion:225 Resp: 172469
Ion Ratio Lower Upper
225 100
223 56.0 47.7 71.5
227 61.8 51.4 77.0

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022




10/19/2022
Supervised By :mohammad
ahmed


10/19/2022



#34
Caprolactam
Concen: 29.950 ng/ul m
RT: 11.739 min Scan# 1488
Delta R.T. -0.000 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27


Tgt Ion:113 Resp: 116854
Ion Ratio Lower Upper
113 100
55 166.6 138.7 208.1
56 126.0 101.4 152.0





#35
4-Chloro-3-methylphenol
Concen: 31.937 ng/ul
RT: 12.081 min Scan# 1545
Delta R.T. -0.000 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Instrument :
BNA_N
ClientSampleId :
SLCS277





Tgt Ion:107 Resp: 38663
Ion Ratio Lower Upper
107 100
144 25.2 20.3 30.5
142 80.4 64.4 96.6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022




10/19/2022
Supervised By :mohammad
ahmed


10/19/2022



#36
2-Methylnaphthalene
Concen: 30.346 ng/ul
RT: 12.445 min Scan# 1608
Delta R.T. -0.012 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27


Tgt Ion:142 Resp: 766220
Ion Ratio Lower Upper
142 100
141 90.5 69.7 104.5





#37
1-Methylnaphthalene
Concen: 30.411 ng/ul
RT: 12.669 min Scan# 1646
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Instrument :
BNA_N
ClientSampleId :
SLCS277






Tgt Ion:142 Resp: 768930
Ion Ratio Lower Upper
142 100
141 90.6 76.7 115.1
116 3.7 3.1 4.7

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022




10/19/2022
Supervised By :mohammad
ahmed


10/19/2022



#39
1,2,4,5-Tetrachlorobenzene
Concen: 32.538 ng/ul
RT: 12.816 min Scan# 1671
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27


Tgt Ion:216 Resp: 348374
Ion Ratio Lower Upper
216 100
214 85.3 65.4 98.2
179 24.7 17.4 26.0
108 22.6 16.4 24.6





#40
Hexachlorocyclopentadiene
Concen: 26.328 ng/ul
RT: 12.786 min Scan# 1666
Delta R.T. -0.012 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Instrument :
BNA_N
ClientSampleId :
SLCS277






Tgt Ion:237 Resp: 157584
Ion Ratio Lower Upper
237 100
235 64.8 43.4 65.0
272 13.7 10.2 15.2

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022




10/19/2022
Supervised By :mohammad
ahmed


10/19/2022



#41
2,4,6-Trichlorophenol
Concen: 34.708 ng/ul
RT: 13.057 min Scan# 1712
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27


Tgt Ion:196 Resp: 253388
Ion Ratio Lower Upper
196 100
198 93.1 73.4 110.2
200 31.5 24.3 36.5





#42
 2,4,5-Trichlorophenol
 Concen: 34.902 ng/ul
 RT: 13.134 min Scan# 1725
 Delta R.T. -0.000 min
 Lab File: BN022173.D
 Acq: 18 Oct 2022 12:27

Instrument : BNA_N
 ClientSampleId : SLCS277





Tgt Ion:196 Resp: 283758


Ion	Ratio	Lower	Upper
196	100		
198	94.0	77.0	115.6
200	29.5	24.7	37.1

Manual Integrations
 APPROVED


Reviewed By :Jagrut Upadhyay 10/19/2022
 Supervised By :mohammad ahmed 10/19/2022



10/19/2022
 Supervised By :mohammad
 ahmed
 10/19/2022






#43
 1,1'-Biphenyl
 Concen: 33.184 ng/ul
 RT: 13.457 min Scan# 1780
 Delta R.T. -0.012 min
 Lab File: BN022173.D
 Acq: 18 Oct 2022 12:27



Tgt Ion:154 Resp: 999834


Ion	Ratio	Lower	Upper
154	100		
153	39.7	30.9	46.3
76	14.9	11.6	17.4





#44
2-Chloronaphthalene
Concen: 32.630 ng/ul
RT: 13.504 min Scan# 1788
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Instrument : BNA_N
ClientSampleId : SLCS277






Tgt Ion:162 Resp: 76817
Ion Ratio Lower Upper
162 100
164 31.4 23.7 35.5
127 41.2 30.8 46.2

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022




10/19/2022
Supervised By :mohammad
ahmed


10/19/2022



#45
2-Nitroaniline
Concen: 35.444 ng/ul
RT: 13.716 min Scan# 1824
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27


Tgt Ion: 65 Resp: 258498
Ion Ratio Lower Upper
65 100
92 71.0 56.2 84.2
138 107.8 85.0 127.6





#47
Dimethylphthalate
Concen: 31.465 ng/ul
RT: 14.086 min Scan# 1887
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Instrument :
BNA_N
ClientSampleId :
SLCS277






Tgt Ion:163 Resp: 97307
Ion Ratio Lower Upper
163 100
194 4.1 3.1 4.7
164 9.7 8.2 12.2

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022




10/19/2022
Supervised By :mohammad
ahmed


10/19/2022



#48
2,6-Dinitrotoluene
Concen: 35.036 ng/ul
RT: 14.210 min Scan# 1908
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27


Tgt Ion:165 Resp: 202957
Ion Ratio Lower Upper
165 100
89 63.2 52.1 78.1
121 23.1 19.0 28.4





#50
Acenaphthylene
Concen: 33.045 ng/ul
RT: 14.351 min Scan# 1932
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Instrument :
BNA_N
ClientSampleId :
SLCS277





Tgt Ion:152 Resp: 121241
Ion Ratio Lower Upper
152 100
151 19.7 15.5 23.3
153 13.4 9.8 14.8

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022




10/19/2022
Supervised By :mohammad
ahmed


10/19/2022



#51
3-Nitroaniline
Concen: 33.150 ng/ul
RT: 14.539 min Scan# 1964
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27


Tgt Ion:138 Resp: 226739
Ion Ratio Lower Upper
138 100
108 10.5 9.0 13.6
92 117.9 96.4 144.6





#52
Acenaphthene
Concen: 32.667 ng/ul
RT: 14.692 min Scan# 1990
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Instrument :
BNA_N
ClientSampleId :
SLCS277






Tgt Ion:153 Resp: 84995
Ion Ratio Lower Upper
153 100
152 48.4 38.0 57.0
154 86.2 71.1 106.7

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022

10/19/2022
Supervised By :mohammad
ahmed



10/19/2022




#53
2,4-Dinitrophenol
Concen: 24.562 ng/ul
RT: 14.745 min Scan# 1999
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27




Tgt Ion:184 Resp: 96383
Ion Ratio Lower Upper
184 100
63 73.9 56.8 85.2
154 63.8 49.0 73.6





#55
4-Nitrophenol
Concen: 31.558 ng/ul
RT: 14.851 min Scan# 2016
Delta R.T. -0.000 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Instrument :
BNA_N
ClientSampleId :
SLCS277






Tgt Ion:109 Resp: 157134
Ion Ratio Lower Upper
109 100
139 135.1 108.3 162.5
65 133.9 107.9 161.9

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022




10/19/2022
Supervised By :mohammad
ahmed


10/19/2022



#56
Dibenzofuran
Concen: 31.965 ng/ul
RT: 15.028 min Scan# 2047
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27


Tgt Ion:168 Resp: 1134881
Ion Ratio Lower Upper
168 100
139 38.3 30.6 45.8





#57
2,4-Dinitrotoluene
Concen: 35.368 ng/ul
RT: 14.998 min Scan# 2042
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Instrument :
BNA_N
ClientSampleId :
SLCS277






Tgt Ion:165 Resp: 300140
Ion Ratio Lower Upper
165 100
63 46.8 36.2 54.4
182 2.5 2.2 3.2

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022

10/19/2022
Supervised By :mohammad
ahmed



10/19/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 32.939 ng/ul
RT: 15.251 min Scan# 2085
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27




Tgt Ion:232 Resp: 222875
Ion Ratio Lower Upper
232 100
131 55.7 44.4 66.6
130 2.9 2.2 3.4
166 35.9 28.8 43.2





#59
Diethylphthalate
Concen: 31.825 ng/ul
RT: 15.451 min Scan# 2119
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Instrument :
BNA_N
ClientSampleId :
SLCS277






Tgt Ion:149 Resp: 999678
Ion Ratio Lower Upper
149 100
177 21.2 17.5 26.3
150 12.4 9.8 14.8

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022




10/19/2022
Supervised By :mohammad
ahmed


10/19/2022



#61
Fluorene
Concen: 32.127 ng/ul
RT: 15.675 min Scan# 2157
Delta R.T. -0.012 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27


Tgt Ion:166 Resp: 916973
Ion Ratio Lower Upper
166 100
165 93.4 72.8 109.2
167 13.3 10.6 16.0





#62
4-Chlorophenyl-phenylether
Concen: 31.814 ng/ul
RT: 15.669 min Scan# 2156
Delta R.T. -0.012 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Instrument :
BNA_N
ClientSampleId :
SLCS277






Tgt Ion:204 Resp: 425470
Ion Ratio Lower Upper
204 100
206 34.3 26.2 39.2
141 65.0 52.0 78.0

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022




10/19/2022
Supervised By :mohammad
ahmed


10/19/2022



#63
4-Nitroaniline
Concen: 38.583 ng/ul
RT: 15.710 min Scan# 2163
Delta R.T. -0.000 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27


Tgt Ion:138 Resp: 246235
Ion Ratio Lower Upper
138 100
92 51.4 41.8 62.8
108 79.0 60.4 90.6





#66
4,6-Dinitro-2-methylphenol
Concen: 29.614 ng/ul
RT: 15.763 min Scan# 2172
Delta R.T. -0.000 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Instrument :
BNA_N
ClientSampleId :
SLCS277






Tgt Ion:198 Resp: 15259
Ion Ratio Lower Upper
198 100
182 4.2 3.8 5.8
77 29.1 24.4 36.6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022




10/19/2022
Supervised By :mohammad
ahmed


10/19/2022



#67
N-Nitrosodiphenylamine
Concen: 32.942 ng/ul
RT: 15.886 min Scan# 2193
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27


Tgt Ion:169 Resp: 801453
Ion Ratio Lower Upper
169 100
168 64.8 50.9 76.3
167 35.4 28.2 42.4





#68
4-Bromophenyl-phenylether
Concen: 35.884 ng/ul
RT: 16.569 min Scan# 2309
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Instrument :
BNA_N
ClientSampleId :
SLCS277






Tgt Ion:248 Resp: 266679
Ion Ratio Lower Upper
248 100
250 93.9 86.3 129.5
141 78.5 74.2 111.4

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022




10/19/2022
Supervised By :mohammad
ahmed


10/19/2022



#69
Hexachlorobenzene
Concen: 32.917 ng/ul
RT: 16.680 min Scan# 2328
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27


Tgt Ion:284 Resp: 297784
Ion Ratio Lower Upper
284 100
142 45.2 35.7 53.5
249 34.1 27.0 40.6





#70
Atrazine
Concen: 34.964 ng/ul
RT: 16.845 min Scan# 2356
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Instrument :
BNA_N
ClientSampleId :
SLCS277






Tgt Ion:200 Resp: 286460
Ion Ratio Lower Upper
200 100
173 28.1 22.8 34.2
215 46.0 38.1 57.1

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022




10/19/2022
Supervised By :mohammad
ahmed


10/19/2022



#71
Pentachlorophenol
Concen: 32.093 ng/ul
RT: 17.027 min Scan# 2387
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Tgt Ion:266 Resp: 179528
Ion Ratio Lower Upper
266 100
264 61.7 52.7 79.1
268 63.9 54.6 82.0






#72
 Phenanthrene
 Concen: 32.417 ng/u1
 RT: 17.427 min Scan# 2455
 Delta R.T. -0.006 min
 Lab File: BN022173.D
 Acq: 18 Oct 2022 12:27

Instrument :

BNA_N

ClientSampleId :

SLCS277





Tgt Ion:178 Resp: 1471185
 Ion Ratio Lower Upper
 178 100
 179 15.0 12.3 18.5
 176 20.1 16.5 24.7

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
 Supervised By :mohammad ahmed 10/19/2022




10/19/2022
 Supervised By :mohammad
 ahmed


10/19/2022



#74
 Anthracene
 Concen: 32.082 ng/u1
 RT: 17.516 min Scan# 2470
 Delta R.T. -0.006 min
 Lab File: BN022173.D
 Acq: 18 Oct 2022 12:27


Tgt Ion:178 Resp: 1446185
 Ion Ratio Lower Upper
 178 100
 179 15.5 13.0 19.6
 176 19.3 15.8 23.6





#75
1,2,3,4-Tetrachlorobenzene
Concen: 35.153 ng/uL
RT: 13.422 min Scan# 1774
Delta R.T. -0.012 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Instrument :
BNA_N
ClientSampleId :
SLCS277






Tgt Ion:216 Resp: 371977
Ion Ratio Lower Upper
216 100
214 81.9 61.2 91.8
218 48.4 39.4 59.0

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022

10/19/2022
Supervised By :mohammad
ahmed



10/19/2022




#76
Pentachlorobenzene
Concen: 30.931 ng/uL
RT: 14.945 min Scan# 2033
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27




Tgt Ion:250 Resp: 351307
Ion Ratio Lower Upper
250 100
248 67.8 46.8 70.2
252 62.9 53.9 80.9





#77
 Carbazole
 Concen: 33.672 ng/ul
 RT: 17.792 min Scan# 2517
 Delta R.T. -0.006 min
 Lab File: BN022173.D
 Acq: 18 Oct 2022 12:27

Instrument : BNA_N
 ClientSampleId : SLCS277



Tgt Ion:167 Resp: 1403915




Ion	Ratio	Lower	Upper
167	100		
166	20.5	17.3	25.9
139	13.2	10.7	16.1

Manual Integrations
 APPROVED


Reviewed By :Jagrut Upadhyay 10/19/2022
 Supervised By :mohammad ahmed 10/19/2022

10/19/2022
 Supervised By :mohammad
 ahmed

10/19/2022






#78
 Di-n-butylphthalate
 Concen: 34.533 ng/ul
 RT: 18.351 min Scan# 2612
 Delta R.T. -0.006 min
 Lab File: BN022173.D
 Acq: 18 Oct 2022 12:27



Tgt Ion:149 Resp: 1784402


Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.1	10.7
104	5.1	4.2	6.2





#80
Fluoranthene
Concen: 38.479 ng/u1
RT: 19.451 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Instrument :
BNA_N
ClientSampleId :
SLCS277





Tgt Ion:202 Resp: 166314
Ion Ratio Lower Upper
202 100
101 14.1 10.3 15.5
100 10.5 7.8 11.8

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022




10/19/2022
Supervised By :mohammad
ahmed


10/19/2022



#82
Pyrene
Concen: 37.859 ng/u1
RT: 19.816 min Scan# 2861
Delta R.T. -0.000 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27


Tgt Ion:202 Resp: 1711908
Ion Ratio Lower Upper
202 100
101 15.9 13.4 20.0
100 13.1 10.9 16.3





#83
Butylbenzylphthalate
Concen: 41.099 ng/ul
RT: 20.710 min Scan# 3013
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27



Instrument :
BNA_N
ClientSampleId :
SLCS277



Tgt Ion:149 Resp: 806283
Ion Ratio Lower Upper
149 100
91 76.5 61.7 92.5
206 18.1 14.7 22.1


Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022




10/19/2022
Supervised By :mohammad
ahmed



10/19/2022




#84
3,3'-Dichlorobenzidine
Concen: 32.869 ng/ul
RT: 21.498 min Scan# 3147
Delta R.T. -0.012 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27




Tgt Ion:252 Resp: 491606
Ion Ratio Lower Upper
252 100
254 60.8 52.9 79.3
126 14.5 11.2 16.8





#85
Benzo(a)anthracene
Concen: 35.042 ng/ul
RT: 21.568 min Scan# 3159
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Instrument :
BNA_N
ClientSampleId :
SLCS277






Tgt Ion:228 Resp: 154853
Ion Ratio Lower Upper
228 100
229 19.2 13.5 20.3
226 28.4 21.0 31.6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022




10/19/2022
Supervised By :mohammad
ahmed


10/19/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 42.246 ng/ul
RT: 21.486 min Scan# 3145
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Tgt Ion:149 Resp: 1198722
Ion Ratio Lower Upper
149 100
167 26.7 21.3 31.9
279 3.6 2.9 4.3





#87

Chrysene

Concen: 33.521 ng/ul

RT: 21.621 min Scan# 3168

Delta R.T. -0.006 min

Lab File: BN022173.D


Acq: 18 Oct 2022 12:27

Instrument :

BNA_N

ClientSampleId :

SLCS277



Tgt Ion:228 Resp: 142353

Ion Ratio Lower Upper

228 100

226 27.7 23.8 35.8

229 19.5 16.5 24.7

Manual Integrations



APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022

10/19/2022

Supervised By :mohammad
ahmed

10/19/2022



#89

Di-n-octyl phthalate

Concen: 48.278 ng/ul




RT: 22.427 min Scan# 3305


Delta R.T. -0.012 min

Lab File: BN022173.D

Acq: 18 Oct 2022 12:27


Tgt Ion:149 Resp: 2002573





#90
Benzo(b)fluoranthene
Concen: 36.387 ng/ul
RT: 23.292 min Scan# 3452
Delta R.T. -0.006 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Instrument :
BNA_N
ClientSampleId :
SLCS277






Tgt Ion:252 Resp: 134357
Ion Ratio Lower Upper
252 100
253 22.2 17.6 26.4
125 13.2 10.7 16.1

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022




10/19/2022
Supervised By :mohammad
ahmed


10/19/2022



#91
Benzo(k)fluoranthene
Concen: 34.976 ng/ul
RT: 23.339 min Scan# 3460
Delta R.T. -0.012 min
Lab File: BN022173.D
Acq: 18 Oct 2022 12:27

Tgt Ion:252 Resp: 1275949
Ion Ratio Lower Upper
252 100
253 22.3 17.7 26.5
125 13.3 10.6 15.8





#93

Benzo(a)pyrene

Concen: 34.114 ng/ul

RT: 23.945 min Scan# 3562

Delta R.T. -0.006 min

Lab File: BN022173.D

Acq: 18 Oct 2022 12:27

Instrument :

BNA_N

ClientSampleId :

SLCS277

Tgt Ion:252 Resp: 1125819

Ion Ratio Lower Upper

252 100

253 21.2 17.2 25.8

125 13.1 10.8 16.2

Manual Integrations





APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022

10/19/2022

Supervised By :mohammad
ahmed

10/19/2022



#94

Indeno(1,2,3-cd)pyrene

Concen: 31.414 ng/ul

RT: 26.639 min Scan# 4021

Delta R.T. -0.024 min

Lab File: BN022173.D

Acq: 18 Oct 2022 12:27




Tgt Ion:276 Resp: 1295777


Ion Ratio Lower Upper

276 100

138 29.0 24.6 36.8

277 21.9 18.4 27.6





#95

Dibenzo(a,h)anthracene

Concen: 29.870 ng/ul

RT: 26.668 min Scan# 4024

Delta R.T. -0.006 min

Lab File: BN022173.D

Acq: 18 Oct 2022 12:27

Instrument :

BNA_N

ClientSampleId :

SLCS277

Tgt Ion:278 Resp: 110233

Ion Ratio Lower Upper

278 100

139 18.6 15.7 23.5

279 22.0 16.5 24.7

Manual Integrations





APPROVED

Reviewed By :Jagrut Upadhyay 10/19/2022
Supervised By :mohammad ahmed 10/19/2022

10/19/2022

Supervised By :mohammad
ahmed

10/19/2022



#96

Benzo(g,h,i)perylene

Concen: 28.684 ng/ul

RT: 27.439 min Scan# 4157

Delta R.T. -0.018 min

Lab File: BN022173.D

Acq: 18 Oct 2022 12:27

Tgt Ion:276 Resp: 1064452

Ion Ratio Lower Upper

276 100

138 26.2 21.2 31.8

277 23.4 17.9 26.9

