

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN101920\  
 Data File : BN012348.D  
 Acq On : 19 Oct 2020 13:09  
 Operator : CG/JU  
 Sample : L4308-01  
 Misc :  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 C0AP6

Manual Integrations  
 APPROVED

Sohil  
 10/20/2020 9:35:43 AM

Quant Time: Oct 19 14:00:57 2020  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN101420.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Oct 19 11:52:04 2020  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units  | Dev(Min) |
|-----------------------------|-------|------|----------|--------|--------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 7.59  | 152  | 1153     | 0.40   | ng/ul  | 0.00     |
| 2) Naphthalene-d8           | 10.36 | 136  | 4908     | 0.40   | ng/ul  | 0.00     |
| 6) Acenaphthene-d10         | 14.24 | 164  | 2853     | 0.40   | ng/ul  | 0.00     |
| 10) Phenanthrene-d10        | 16.98 | 188  | 5968     | 0.40   | ng/ul  | 0.00     |
| 16) Chrysene-d12            | 21.19 | 240  | 6075     | 0.40   | ng/ul  | 0.00     |
| 20) Perylene-d12            | 23.36 | 264  | 6107     | 0.40   | ng/ul  | 0.00     |
| System Monitoring Compounds |       |      |          |        |        |          |
| 4) 2-Methylnaphthalene-d10  | 11.97 | 152  | 1383     | 0.19   | ng/ul  | 0.00     |
| 14) Fluoranthene-d10        | 19.02 | 212  | 3284     | 0.20   | ng/ul  | 0.00     |
| Target Compounds            |       |      |          | Ovalue |        |          |
| 3) Naphthalene              | 10.42 | 128  | 1831     | 0.129  | ng/ul# | 95       |
| 5) 2-Methylnaphthalene      | 12.04 | 142  | 556      | 0.062  | ng/ul  | 100      |
| 7) Acenaphthylene           | 13.95 | 152  | 4596     | 0.371  | ng/ul  | 99       |
| 8) Acenaphthene             | 14.30 | 153  | 5432     | 0.519  | ng/ul  | 99       |
| 9) Fluorene                 | 15.29 | 166  | 7933     | 0.681  | ng/ul  | 99       |
| 12) Phenanthrene            | 17.03 | 178  | 201353   | 10.584 | ng/ul  | 98       |
| 13) Anthracene              | 17.12 | 178  | 38381    | 2.352  | ng/ul  | 98       |
| 15) Fluoranthene            | 19.05 | 202  | 505658   | 22.731 | ng/ul  | 98       |
| 17) Pyrene                  | 19.42 | 202  | 384790   | 14.049 | ng/ul  | 98       |
| 18) Benzo(a)anthracene      | 21.17 | 228  | 218418   | 9.707  | ng/ul  | 99       |
| 19) Chrysene                | 21.22 | 228  | 249179   | 9.089  | ng/ul  | 98       |
| 21) Benzo(b)fluoranthene    | 22.71 | 252  | 351880m  | 14.265 | ng/ul  |          |
| 22) Benzo(k)fluoranthene    | 22.75 | 252  | 124849m  | 4.573  | ng/ul  |          |
| 23) Benzo(a)pyrene          | 23.27 | 252  | 200323   | 8.348  | ng/ul# | 81       |
| 24) Indeno(1,2,3-cd)pyrene  | 25.53 | 276  | 159066   | 5.160  | ng/ul# | 97       |
| 25) Dibenzo(a,h)anthracene  | 25.53 | 278  | 40601    | 1.646  | ng/ul  | 96       |
| 26) Benzo(a,h,i)perylene    | 26.19 | 276  | 133281   | 4.717  | ng/ul  | 98       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN101920\  
Data File : BN012348.D  
Acq On : 19 Oct 2020 13:09  
Operator : CG/JU  
Sample : L4308-01  
Misc :  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
C0AP6

Manual Integrations  
APPROVED

Sohil  
10/20/2020 9:35:43 AM

Quant Time: Oct 19 14:00:57 2020  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN101420.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Oct 19 11:52:04 2020  
Response via : Initial Calibration

