


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101921\
 Data File : BN016998.D
 Acq On : 19 Oct 2021 10:08
 Operator : CG/JU
 Sample : DFTPP207
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP207

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN101921.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Oct 20 10:49:43 2021



AutoFind: Scans 2453, 2454, 2455; Background Corrected with Scan 2447

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.8	114709	PASS
68	69	0.00	2	1.9	2249	PASS
69	198	0.00	100	35.5	120629	PASS
70	69	0.00	2	0.5	643	PASS
127	198	10	80	46.0	156133	PASS
197	198	0.00	2	0.2	773	PASS
198	198	100	100	100.0	339541	PASS
199	198	5	9	6.9	23376	PASS
275	198	10	60	28.0	95155	PASS
365	198	1	100	3.7	12616	PASS
441	198	0.01	100	15.2	51469	PASS
442	198	50	100	93.1	316096	PASS
443	442	15	24	18.5	58325	PASS

m/z	Abundance
39.90	543.0
41.00	327.0
41.90	390.0
43.00	1588.0
44.00	1576.0
45.00	1256.0
60.00	890.0

Instrument :
BNA_N
ClientSampleId :
DFTPP207

m/z	Abundance
37.00	469.0
38.00	1110.0
39.10	7108.0
40.00	742.0
41.00	447.0
43.00	1408.0
44.00	1417.0
45.00	1308.0
49.10	802.0
50.00	23704.0
51.00	94680.0
52.00	5012.0
54.90	566.0
56.00	2828.0
57.00	6389.0
60.00	934.0
61.00	1204.0
62.00	1343.0
63.00	3725.0
64.00	540.0
65.00	1929.0
68.00	1807.0
69.00	97368.0
70.00	481.0
73.00	1131.0
74.00	9734.0
75.00	15320.0
76.10	5311.0
77.00	115360.0
78.00	8287.0
79.00	7592.0
80.00	5511.0
81.00	8062.0
82.00	2064.0
83.00	1854.0
84.00	333.0
85.00	1485.0
86.00	2346.0
87.00	1054.0
87.90	415.0
91.00	1981.0
92.00	2113.0
93.00	12514.0
94.00	920.0
96.00	564.0
97.10	314.0
98.00	9752.0
99.00	8239.0
100.00	762.0
101.00	4486.0
103.00	1578.0
104.00	2827.0
105.00	2556.0
106.00	872.0
107.00	32952.0
108.00	5330.0
109.00	983.0
110.00	63128.0
111.00	9909.0
112.00	1154.0
113.00	392.0
116.00	1772.0
117.00	27016.0
118.00	2110.0
118.90	305.0
119.90	491.0
122.00	2118.0
123.00	3400.0
124.00	1687.0
125.00	1367.0
127.00	116592.0
128.00	9359.0
129.00	42904.0
130.00	3719.0
131.00	810.0
132.00	420.0
134.00	1189.0
135.00	3585.0
136.00	1472.0
137.00	2218.0
137.80	437.0
140.00	520.0
141.00	5609.0
142.00	1800.0
143.00	1360.0
144.00	379.0
145.00	334.0
145.90	884.0
147.00	2815.0
148.00	6574.0
149.00	1502.0
150.00	360.0

Instrument :
BNA_N
ClientSampleId :
DFTPP207

151.10	742.0
151.90	397.0
153.00	1822.0
154.00	1384.0
155.00	3186.0
156.00	4749.0
157.00	876.0
157.90	1047.0
159.00	749.0
160.00	1350.0
161.00	2514.0
162.00	814.0
164.00	340.0
165.00	2004.0
166.00	1670.0
167.00	11761.0
168.00	8073.0
169.00	1165.0
170.00	442.0
170.90	579.0
171.90	874.0
173.00	1316.0
174.00	2276.0
175.00	4517.0
176.10	1225.0
177.00	2135.0
178.00	670.0
179.00	8452.0
180.00	5986.0
181.00	2684.0
182.00	435.0
183.90	637.0
185.00	4074.0
186.00	32720.0
187.00	9211.0
188.00	1083.0
189.00	2189.0
189.90	406.0
191.00	922.0
192.00	2899.0
193.00	3016.0
194.00	733.0
194.90	383.0
196.00	7263.0
198.00	242432.0
199.00	16608.0
200.00	1247.0
201.60	1007.0
203.00	1556.0
204.00	8224.0
205.00	13398.0
206.00	58128.0
207.00	8326.0
208.00	2061.0
209.00	727.0
211.00	2050.0
211.80	561.0
215.00	675.0
216.00	1227.0
217.00	15274.0
218.00	1942.0
221.10	10489.0
221.90	2126.0
223.10	3332.0
224.10	31864.0
225.00	8585.0
226.00	1239.0
227.00	12418.0
228.00	1846.0
229.00	2907.0
230.00	385.0
231.10	967.0
233.90	845.0
235.00	1008.0
235.90	613.0
237.00	1333.0
239.00	484.0
240.00	461.0
241.00	657.0
242.00	1786.0
243.00	1963.0
244.10	25112.0
245.10	3599.0
246.00	4749.0
247.00	988.0
249.00	920.0
253.00	440.0
255.00	123528.0
256.00	19976.0
257.10	1790.0
258.00	6179.0
259.00	1294.0
265.00	2779.0
266.00	487.0
271.00	333.0
272.00	327.0

Instrument :
BNA_N
ClientSampleId :
DFTPP207

273.00	4125.0
274.00	9928.0
275.00	53912.0
276.10	8586.0
277.00	4747.0
278.00	932.0
283.00	519.0
284.00	373.0
285.00	868.0
293.00	1114.0
294.00	344.0
296.00	16293.0
297.00	2397.0
303.10	1891.0
304.10	447.0
314.00	678.0
315.00	1600.0
316.00	1018.0
321.10	451.0
322.10	324.0
323.10	4820.0
324.00	976.0
326.90	1089.0
328.00	498.0
332.00	358.0
333.00	532.0
334.00	3018.0
335.00	833.0
341.00	720.0
346.00	1192.0
346.90	318.0
352.00	1628.0
353.00	1018.0
354.00	1810.0
355.00	378.0
365.00	6695.0
366.00	1207.0
371.00	342.0
372.10	2917.0
373.00	682.0
382.90	762.0
390.00	323.0
402.00	1079.0
403.00	1770.0
404.00	667.0
421.00	1364.0
422.10	1704.0
423.10	9408.0
424.10	3558.0
425.00	472.0
441.10	24560.0
442.10	142016.0
443.10	30264.0
444.10	2892.0

Instrument :
BNA_N
ClientSampleId :
DFTPP207

m/z	Abundance
37.00	599.0
38.00	1362.0
39.00	9220.0
40.00	705.0
41.00	495.0
43.00	1373.0
44.00	1331.0
45.00	1401.0
49.00	1218.0
50.00	34448.0
51.00	131328.0
52.00	6650.0
55.00	740.0
56.00	3677.0
57.00	8916.0
58.00	423.0
60.00	851.0
61.00	1767.0
62.00	2051.0
63.00	5157.0
64.00	680.0
65.00	2741.0
68.10	2504.0
69.00	138880.0
70.00	756.0
73.10	1292.0
74.00	14570.0
75.00	22720.0
76.10	8083.0
77.10	164096.0
78.00	11551.0
79.00	10710.0
80.00	7818.0
81.00	11083.0
82.00	2764.0
83.00	2546.0
84.00	402.0
85.00	2246.0
86.00	3549.0
87.00	1475.0
88.00	608.0
91.00	2508.0
92.00	2876.0
93.00	18104.0
94.00	1314.0
95.00	369.0
96.00	779.0
96.90	315.0
98.00	14099.0
99.00	11907.0
100.00	935.0
101.00	6707.0
103.00	2130.0
104.00	3838.0
105.00	3826.0
106.10	1239.0
107.00	49680.0
108.00	7950.0
109.10	1591.0
110.00	94168.0
111.00	14437.0
112.00	1735.0
112.90	571.0
116.00	2631.0
117.00	41640.0
118.00	3113.0
119.00	419.0
120.00	607.0
122.00	3377.0
123.00	5241.0
124.00	2183.0
125.00	2084.0
127.00	182656.0
128.00	13696.0
129.00	68664.0
130.00	5832.0
131.00	1078.0
132.00	598.0
133.10	303.0
134.00	2174.0
135.00	6149.0
136.00	2223.0
137.00	3188.0
137.90	715.0
138.90	431.0
140.00	974.0
141.00	8951.0
142.00	3000.0
143.00	1972.0
144.00	544.0
145.00	588.0
146.00	1669.0

Instrument :
BNA_N
ClientSampleId :
DFTPP207

147.00	4586.0
148.00	11019.0
149.00	2264.0
150.10	777.0
151.10	1034.0
151.90	665.0
153.00	2850.0
154.00	2241.0
155.00	4904.0
156.00	7638.0
157.10	1389.0
158.00	1728.0
159.00	1290.0
160.00	2774.0
161.00	4186.0
162.00	1170.0
163.10	314.0
163.90	472.0
165.00	3087.0
166.00	2625.0
167.00	18600.0
168.00	10957.0
169.00	1634.0
170.00	581.0
170.90	885.0
172.00	1584.0
173.00	2128.0
174.00	3829.0
175.10	7422.0
176.10	1995.0
177.00	3741.0
177.90	1200.0
179.00	13254.0
180.00	9430.0
181.00	4545.0
182.00	780.0
183.00	413.0
184.00	1097.0
185.00	6665.0
186.00	50520.0
187.00	14846.0
188.00	1738.0
189.00	3038.0
190.00	602.0
191.00	1346.0
192.00	4587.0
193.00	4927.0
194.00	1011.0
194.90	478.0
196.00	11697.0
197.10	2320.0
198.00	394688.0
199.00	26400.0
200.10	2015.0
201.40	1514.0
203.00	2466.0
204.00	12912.0
205.10	22568.0
206.00	99200.0
207.10	12956.0
208.00	3255.0
209.00	1103.0
210.10	1598.0
211.00	3648.0
214.90	968.0
216.10	2167.0
217.00	24848.0
218.00	3339.0
221.10	16712.0
221.70	3758.0
223.00	5639.0
224.10	54984.0
225.10	13912.0
226.00	1683.0
227.00	21272.0
228.00	3217.0
229.00	4763.0
230.00	662.0
231.10	1960.0
232.00	402.0
233.00	496.0
233.90	1360.0
235.00	1476.0
236.00	946.0
237.00	2206.0
239.00	1068.0
240.00	603.0
240.90	1023.0
242.00	2892.0
243.10	3404.0
244.10	44136.0
245.10	6019.0
246.00	8249.0
247.00	1586.0
248.00	399.0
249.00	1665.0

Instrument :
BNA_N
ClientSampleId :
DFTPP207

250.00	328.0
251.90	424.0
253.10	853.0
255.00	230912.0
256.00	34512.0
257.00	2616.0
258.00	12644.0
259.00	2122.0
260.00	410.0
261.00	418.0
265.00	5467.0
266.00	708.0
270.90	550.0
272.00	614.0
273.00	7203.0
274.00	18528.0
275.00	113056.0
276.10	15590.0
277.00	9593.0
278.00	1479.0
279.00	361.0
283.00	976.0
284.00	612.0
285.00	1615.0
286.00	367.0
289.00	308.0
292.00	387.0
293.00	2190.0
294.00	572.0
295.10	644.0
296.00	28488.0
297.00	4049.0
298.00	323.0
300.90	413.0
302.00	492.0
303.10	3216.0
304.00	1154.0
307.90	424.0
310.00	395.0
314.00	1558.0
315.00	3257.0
316.00	1849.0
321.00	1047.0
322.00	544.0
323.00	10699.0
324.10	1797.0
327.00	1850.0
328.00	1083.0
332.10	687.0
333.10	1010.0
334.10	6644.0
335.10	1734.0
341.00	1140.0
342.00	400.0
346.00	2333.0
347.00	411.0
352.00	3163.0
353.00	2353.0
354.00	3878.0
355.00	706.0
365.00	13178.0
366.00	2224.0
371.00	842.0
372.00	6028.0
373.10	1704.0
383.00	1460.0
384.00	521.0
389.90	685.0
391.00	565.0
392.10	434.0
401.00	345.0
402.00	2666.0
403.00	3422.0
404.00	1476.0
421.00	3172.0
422.00	3325.0
423.00	22280.0
424.10	5599.0
425.00	864.0
441.10	55920.0
442.10	339584.0
443.10	64856.0
444.10	6484.0
445.00	444.0

Instrument :
BNA_N
ClientSampleId :
DFTPP207

m/z	Abundance
37.00	440.0
38.00	1349.0
39.00	8553.0
40.00	791.0
41.00	533.0
43.00	1438.0
44.00	1203.0
45.00	1152.0
49.00	1137.0
50.00	29976.0
51.00	118120.0
52.00	5913.0
55.00	751.0
56.00	3450.0
57.00	8328.0
57.90	451.0
60.00	913.0
61.00	1570.0
62.00	1802.0
63.00	5085.0
64.00	718.0
65.00	2438.0
68.00	2436.0
69.00	125640.0
70.00	693.0
73.00	1289.0
74.00	12952.0
75.00	20880.0
76.00	7266.0
77.00	149248.0
78.00	10669.0
79.00	9167.0
80.00	6974.0
81.00	10488.0
82.00	2679.0
83.00	2363.0
83.90	359.0
85.00	2042.0
86.00	3109.0
87.00	1455.0
87.90	521.0
91.00	2188.0
92.00	2705.0
93.00	16198.0
94.00	1214.0
95.00	332.0
96.10	817.0
97.00	385.0
98.00	13292.0
99.00	10579.0
100.00	930.0
101.00	6058.0
102.00	319.0
103.00	2075.0
104.00	3550.0
105.00	3502.0
106.00	1082.0
107.00	45528.0
108.00	7144.0
109.00	1406.0
110.00	90144.0
111.00	12562.0
112.00	1538.0
113.00	549.0
116.00	2514.0
117.00	39016.0
118.00	2801.0
119.00	405.0
119.90	591.0
122.00	3293.0
123.00	4862.0
124.00	2141.0
125.00	1988.0
126.20	1175.0
127.00	169152.0
128.00	13179.0
129.00	65600.0
130.00	5522.0
131.00	1119.0
132.00	677.0
133.00	311.0
134.00	1955.0
135.00	5443.0
136.00	2205.0
137.00	3137.0
137.90	580.0
138.90	426.0
140.00	795.0
141.00	8288.0
142.00	2543.0
143.00	1891.0
144.00	534.0

Instrument :
BNA_N
ClientSampleId :
DFTPP207

144.90	463.0
145.90	1486.0
147.00	4295.0
148.00	10594.0
149.00	2341.0
150.00	540.0
151.10	1328.0
152.00	789.0
153.00	2760.0
154.00	2129.0
155.00	5003.0
156.00	7188.0
157.00	1572.0
158.00	1666.0
159.00	1199.0
160.00	2776.0
161.00	3967.0
162.00	1212.0
163.00	494.0
164.00	508.0
165.00	3208.0
166.10	2561.0
167.00	18664.0
168.00	9873.0
169.00	1481.0
170.10	598.0
170.90	745.0
172.00	1439.0
173.00	2073.0
174.00	3742.0
175.00	7578.0
176.00	2105.0
177.00	3464.0
178.00	1011.0
179.00	13327.0
180.00	9321.0
181.00	4182.0
182.00	688.0
182.80	325.0
184.00	1086.0
185.00	6519.0
186.00	52592.0
187.00	14893.0
188.10	1598.0
189.00	3033.0
189.90	520.0
191.00	1556.0
192.00	4541.0
193.10	5019.0
194.00	935.0
195.10	532.0
196.00	12115.0
198.00	381504.0
199.00	27120.0
200.00	1878.0
201.60	1589.0
203.00	2496.0
204.10	13586.0
205.00	22832.0
206.00	102872.0
207.10	13238.0
208.00	3383.0
209.00	1082.0
210.20	1615.0
211.00	4176.0
215.00	1108.0
216.00	1883.0
217.00	25720.0
218.00	3393.0
219.00	335.0
221.10	17016.0
223.00	6133.0
224.10	57608.0
225.10	14936.0
226.10	1619.0
227.00	22664.0
228.00	2921.0
229.00	4810.0
230.00	590.0
231.00	2181.0
232.00	362.0
233.00	424.0
234.00	1666.0
235.00	1678.0
236.00	1104.0
237.00	2017.0
238.00	335.0
239.00	992.0
240.00	807.0
241.00	1254.0
242.00	3227.0
243.00	3462.0
244.00	46312.0
245.10	6847.0
246.00	8040.0
246.90	1749.0

Instrument :
BNA_N
ClientSampleId :
DFTPP207

248.00	360.0
249.00	1691.0
250.00	344.0
251.00	335.0
252.00	360.0
253.00	1024.0
255.00	241024.0
256.00	37248.0
257.10	2730.0
258.00	13136.0
259.00	1958.0
260.00	436.0
261.00	469.0
264.00	398.0
265.00	5244.0
266.00	815.0
270.90	498.0
272.10	597.0
273.00	7297.0
274.00	20920.0
275.00	118496.0
276.00	16357.0
277.10	9918.0
278.00	1712.0
283.00	1134.0
284.10	796.0
285.00	1534.0
285.90	391.0
289.00	403.0
289.90	338.0
292.00	431.0
293.00	2167.0
294.00	719.0
295.00	800.0
296.00	32880.0
297.00	4587.0
300.90	509.0
302.10	692.0
303.10	3874.0
304.00	1120.0
307.90	459.0
309.10	318.0
310.00	448.0
314.00	1612.0
315.00	3818.0
316.00	2387.0
317.00	363.0
321.00	1172.0
323.10	12351.0
324.00	1984.0
326.90	2118.0
328.00	1050.0
332.00	686.0
333.00	1108.0
334.00	7812.0
335.00	1907.0
341.00	1341.0
342.00	406.0
346.00	2986.0
347.00	570.0
352.00	3819.0
353.00	2555.0
354.00	4596.0
355.10	790.0
365.00	17976.0
366.00	2707.0
370.00	445.0
371.00	1124.0
372.00	6713.0
373.10	1850.0
383.00	1981.0
383.90	537.0
390.00	940.0
391.00	714.0
392.00	526.0
401.00	538.0
402.00	3449.0
403.10	4786.0
404.00	1781.0
421.10	3894.0
422.00	4196.0
423.00	26088.0
424.10	6700.0
425.10	746.0
441.10	73928.0
442.10	466688.0
443.10	79856.0
444.00	7798.0
445.00	449.0

Instrument :
BNA_N
ClientSampleId :
DFTPP207