

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101922\
 Data File : BN022183.D
 Acq On : 18 Oct 2022 18:41
 Operator : CG/JU
 Sample : N5149-14
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-RW5-GW-963-965

Quant Time: Oct 19 02:09:45 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 19 02:07:41 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 10/19/2022
 Supervised By :mohammad ahmed 10/20/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.956	152	4045	0.400	ng	0.00
7) Naphthalene-d8	10.784	136	13625	0.400	ng	0.00
13) Acenaphthene-d10	14.625	164	7899	0.400	ng	0.00
19) Phenanthrene-d10	17.378	188	16243	0.400	ng	# 0.00
29) Chrysene-d12	21.582	240	10450	0.400	ng	# 0.00
35) Perylene-d12	24.049	264	7838	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.486	112	4446	0.474	ng	0.00
5) Phenol-d6	7.111	99	6127	0.504	ng	0.00
8) Nitrobenzene-d5	9.129	82	3658m	0.334	ng	0.00
11) 2-Methylnaphthalene-d10	12.379	152	6263	0.250	ng	0.00
14) 2,4,6-Tribromophenol	16.125	330	730	0.242	ng	0.00
15) 2-Fluorobiphenyl	13.246	172	11081	0.321	ng	0.00
27) Fluoranthene-d10	19.419	212	12315	0.285	ng	0.00
31) Terphenyl-d14	20.014	244	8429	0.401	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.483	149	8352	0.443	ng	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101922\
Data File : BN022183.D
Acq On : 18 Oct 2022 18:41
Operator : CG/JU
Sample : N5149-14
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-RW5-GW-963-965

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 10/19/2022
Supervised By : mohammad ahmed 10/20/2022

Quant Time: Oct 19 02:09:45 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 19 02:07:41 2022
Response via : Initial Calibration

