

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101922\
 Data File : BN022188.D
 Acq On : 18 Oct 2022 21:45
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 19 02:10:51 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 19 02:07:41 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	104	0.00
2	1,4-Dioxane	0.400	0.373	6.8	102	0.00
3	n-Nitrosodimethylamine	0.400	0.384	4.0	110	0.00
4 S	2-Fluorophenol	0.400	0.474	-18.5	135	0.00
5 S	Phenol-d6	0.400	0.482	-20.5	140	0.00
6	bis(2-Chloroethyl)ether	0.400	0.405	-1.3	116	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	120	0.00
8 S	Nitrobenzene-d5	0.400	0.414	-3.5	136	0.00
9	Naphthalene	0.400	0.396	1.0	128	-0.01
10	Hexachlorobutadiene	0.400	0.373	6.8	119	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.343	14.2	112	0.00
12	2-Methylnaphthalene	0.400	0.517	-29.3#	185	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	136	0.00
14 S	2,4,6-Tribromophenol	0.400	0.386	3.5	154	0.00
15 S	2-Fluorobiphenyl	0.400	0.393	1.8	140	-0.01
16	Acenaphthylene	0.400	0.390	2.5	150	0.00
17	Acenaphthene	0.400	0.392	2.0	143	0.00
18	Fluorene	0.400	0.445	-11.2	167	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	133	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.289	27.8#	58	-0.01
21	4-Bromophenyl-phenylether	0.400	0.393	1.8	142	0.00
22	Hexachlorobenzene	0.400	0.398	0.5	143	0.00
23	Atrazine	0.400	0.385	3.8	149	0.00
24	Pentachlorophenol	0.400	0.263	34.3#	109	-0.01
25	Phenanthrene	0.400	0.391	2.3	140	0.00
26	Anthracene	0.400	0.387	3.3	146	0.00
27 SURR	Fluoranthene-d10	0.400	0.367	8.3	135	0.00
28	Fluoranthene	0.400	0.361	9.8	133	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	105	0.00
30	Pyrene	0.400	0.467	-16.8	130	0.00
31 S	Terphenyl-d14	0.400	0.455	-13.7	127	0.00
32	Benzo(a)anthracene	0.400	0.390	2.5	113	0.00
33	Chrysene	0.400	0.379	5.3	104	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.525	-31.3#	157	0.00
35 I	Perylene-d12	0.400	0.400	0.0	102	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.380	5.0	108	0.00
37	Benzo(b)fluoranthene	0.400	0.378	5.5	107	0.00
38	Benzo(k)fluoranthene	0.400	0.381	4.8	105	0.00
39 C	Benzo(a)pyrene	0.400	0.409	-2.2	116	0.00
40	Dibenzo(a,h)anthracene	0.400	0.376	6.0	106	0.00
41	Benzo(g,h,i)perylene	0.400	0.354	11.5	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4