

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102022\
 Data File : BN022206.D
 Acq On : 19 Oct 2022 21:13
 Operator : CG/JU
 Sample : N4982-10
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C1BB7

Quant Time: Oct 20 01:07:59 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 19 23:32:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.952	152	3796	0.400	ng/u1	0.00
4) Naphthalene-d8	10.777	136	12847	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.623	164	7826	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.378	188	17189	0.400	ng/u1	0.00
17) Chrysene-d12	21.579	240	13029	0.400	ng/u1	0.00
23) Perylene-d12	24.043	264	11370	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.253	96	26445	5.360	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.378	152	8161	0.420	ng/u1	0.00
18) Fluoranthene-d10	19.412	212	19682	0.450	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102022\
Data File : BN022206.D
Acq On : 19 Oct 2022 21:13
Operator : CG/JU
Sample : N4982-10
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C1BB7

Quant Time: Oct 20 01:07:59 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 19 23:32:34 2022
Response via : Initial Calibration

