




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102022\
Data File : BN022237.D
Acq On : 20 Oct 2022 17:17
Operator : CG/JU
Sample : N5060-15
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C1BB6

Manual IntegrationsAPPROVED

Quant Time: Oct 21 02:44:19 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 19 23:32:34 2022
Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 10/21/2022
Supervised By :mohammad ahmed 10/21/2022



TIC: BN022237.D\data.ms

(24) Benzo(b)fluoranthene

23.289min (-0.000) 0.03 ng/u1

response 1495

Ion	Exp%	Act%
252.00	100.00	100.00
253.00	24.50	77.57#
125.00	17.60	122.97#
0.00	0.00	0.00




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102022\
Data File : BN022237.D
Acq On : 20 Oct 2022 17:17
Operator : CG/JU
Sample : N5060-15
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C1BB6

Manual IntegrationsAPPROVED

Quant Time: Oct 21 02:44:19 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 19 23:32:34 2022
Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 10/21/2022
Supervised By :mohammad ahmed 10/21/2022



TIC: BN022237.D\data.ms

(24) Benzo(b)fluoranthene

23.289min (-0.000) 0.02 ng/ul m

response 1105

Ion	Exp%	Act%
252.00	100.00	100.00
253.00	24.50	77.57#
125.00	17.60	122.97#
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102022\
 Data File : BN022237.D
 Acq On : 20 Oct 2022 17:17
 Operator : CG/JU
 Sample : N5060-15
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C1BB6

Manual IntegrationsAPPROVED

Reviewed By :Jagrut Upadhyay 10/21/2022
 Supervised By :mohammad ahmed 10/21/2022

Quant Time: Oct 21 02:44:19 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 19 23:32:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.949	152	4244	0.400	ng/ul	0.00
4) Naphthalene-d8	10.778	136	14227	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.623	164	8196	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.379	188	17183	0.400	ng/ul	0.00
17) Chrysene-d12	21.582	240	11656	0.400	ng/ul	0.00
23) Perylene-d12	24.043	264	8876	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.253	96	20553	3.726	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.373	152	7022	0.326	ng/ul	0.00
18) Fluoranthene-d10	19.412	212	16125	0.412	ng/ul	0.00
Target Compounds						Qvalue
19) Fluoranthene	19.445	202	1123	0.021	ng/ul#	80
24) Benzo(b)fluoranthene	23.289	252	1105m	0.025	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102022\
Data File : BN022237.D
Acq On : 20 Oct 2022 17:17
Operator : CG/JU
Sample : N5060-15
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C1BB6

Manual IntegrationsAPPROVED

Quant Time: Oct 21 02:44:19 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 19 23:32:34 2022
Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 10/21/2022
Supervised By :mohammad ahmed 10/21/2022

