


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102022\
 Data File : BN022189.D
 Acq On : 19 Oct 2022 10:10
 Operator : CG/JU
 Sample : DFTPP288
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP288

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Oct 19 23:32:34 2022



AutoFind: Scans 2519, 2520, 2521; Background Corrected with Scan 2512

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.5	153836	PASS
68	69	0.00	2	1.7	2691	PASS
69	198	0.00	100	42.3	160747	PASS
70	69	0.00	2	0.5	799	PASS
127	198	10	80	50.9	193195	PASS
197	198	0.00	2	0.5	1957	PASS
198	198	100	100	100.0	379755	PASS
199	198	5	9	6.6	25027	PASS
275	198	10	60	23.5	89339	PASS
365	198	1	100	3.0	11419	PASS
441	198	0.01	100	10.9	41581	PASS
442	198	50	100	76.1	289088	PASS
443	442	15	24	19.1	55117	PASS

m/z	Abundance
39.00	214.0
39.90	1145.0
41.10	231.0
43.00	162.0
44.00	714.0
48.90	233.0
51.00	153.0
54.90	193.0
73.00	505.0
77.00	157.0
77.90	162.0
83.90	198.0
96.00	160.0
146.90	162.0
192.90	171.0
207.00	1259.0
208.00	240.0
209.00	190.0
281.10	392.0

Instrument :
BNA_N
ClientSampleId :
DFTPP288

m/z	Abundance
37.00	514.0
38.00	1506.0
39.00	9450.0
40.00	1265.0
41.00	381.0
43.00	174.0
44.00	862.0
44.90	267.0
49.00	1110.0
50.00	32376.0
51.00	130832.0
52.00	7081.0
53.00	287.0
55.00	818.0
56.00	3898.0
57.00	8311.0
57.90	401.0
61.00	1556.0
62.00	1826.0
63.00	4934.0
64.00	686.0
65.00	2597.0
65.90	256.0
66.90	308.0
68.00	2501.0
69.00	131648.0
70.00	825.0
73.00	1478.0
74.00	13844.0
75.00	21440.0
76.00	7615.0
77.10	159680.0
78.00	12256.0
79.00	9496.0
80.00	7721.0
81.00	10884.0
82.00	2749.0
82.90	2437.0
84.00	680.0
85.00	1843.0
86.00	3528.0
87.00	1459.0
88.00	594.0
88.90	295.0
91.00	2401.0
92.00	2808.0
93.00	17072.0
94.00	1137.0
95.00	276.0
96.00	972.0
97.10	364.0
98.00	13725.0
99.00	12496.0
100.00	1111.0
101.00	5816.0
101.90	326.0
103.00	1842.0
104.00	3685.0
105.00	3276.0
106.00	1184.0
107.00	44400.0
108.00	6701.0
109.00	1080.0
110.00	80424.0
111.00	12793.0
112.00	1600.0
113.00	483.0
114.90	287.0
116.00	1971.0
117.00	39488.0
118.00	3466.0
118.90	374.0
119.90	662.0
121.00	223.0
122.00	2748.0
122.90	4476.0
124.00	2237.0
124.90	1962.0
126.10	596.0
127.00	147648.0
128.00	11160.0
129.00	56928.0
130.00	5160.0
130.90	815.0
131.90	456.0
132.90	462.0
133.90	1751.0
135.00	4500.0
136.00	1972.0
137.00	3147.0
137.90	627.0
139.00	381.0

Instrument :
BNA_N
ClientSampleId :
DFTPP288

139.90	657.0
141.00	7261.0
142.00	2402.0
142.90	1653.0
144.00	457.0
145.00	343.0
146.00	1245.0
147.00	3843.0
148.00	9491.0
149.00	2176.0
150.00	552.0
151.10	1041.0
151.70	498.0
153.00	2249.0
154.00	1857.0
155.00	3781.0
156.00	5714.0
157.00	1238.0
158.00	1171.0
159.00	1077.0
160.00	2291.0
161.00	3096.0
162.00	1155.0
162.90	369.0
164.00	384.0
165.00	2448.0
166.10	2014.0
167.00	15095.0
168.00	11771.0
169.00	1483.0
170.10	504.0
170.90	581.0
172.00	1339.0
173.00	1555.0
174.00	2883.0
175.00	5677.0
176.00	1511.0
177.00	2554.0
177.90	867.0
179.00	10346.0
180.00	8882.0
181.00	2943.0
182.00	548.0
183.10	222.0
184.00	687.0
185.00	5143.0
186.00	37296.0
187.00	10519.0
188.00	1636.0
189.00	2317.0
190.00	421.0
191.00	1364.0
192.00	3135.0
193.00	3966.0
194.00	746.0
195.00	446.0
196.00	7491.0
196.90	1609.0
198.00	275904.0
199.00	19456.0
200.00	1454.0
201.50	1200.0
203.00	1635.0
204.00	9883.0
205.00	16279.0
206.10	66608.0
207.10	11082.0
208.00	2514.0
209.00	1120.0
210.00	985.0
211.00	2626.0
213.00	237.0
215.00	729.0
216.00	1375.0
217.00	17664.0
218.00	2447.0
219.00	222.0
221.00	10346.0
221.70	2502.0
223.00	3445.0
224.00	36760.0
225.00	9380.0
226.00	1458.0
227.00	14911.0
228.00	2085.0
229.00	3434.0
230.00	431.0
231.00	1148.0
231.90	246.0
233.00	282.0
234.00	966.0
235.00	1197.0
235.90	747.0
237.00	2034.0
237.90	284.0
239.00	634.0

Instrument :
BNA_N
ClientSampleId :
DFTPP288

239.90	474.0
240.00	476.0
240.90	757.0
242.10	1884.0
243.00	2234.0
244.10	28440.0
245.10	3809.0
246.00	5304.0
247.00	1048.0
249.00	1154.0
250.00	254.0
250.90	238.0
251.90	239.0
253.00	740.0
254.00	899.0
255.00	130264.0
256.00	20464.0
257.00	2115.0
258.00	8339.0
259.00	1415.0
260.00	241.0
261.00	218.0
264.00	188.0
265.00	3146.0
265.90	461.0
270.00	193.0
270.90	407.0
272.00	457.0
273.00	4423.0
274.00	10595.0
275.10	58088.0
276.00	8261.0
277.00	5301.0
278.00	1166.0
278.80	161.0
281.00	248.0
282.00	287.0
283.00	663.0
284.00	464.0
285.00	824.0
289.90	207.0
290.90	158.0
291.90	203.0
293.00	1363.0
294.00	328.0
295.00	294.0
296.00	16848.0
297.00	2326.0
301.00	243.0
302.00	217.0
303.00	1661.0
304.00	557.0
308.00	240.0
310.00	228.0
312.80	199.0
314.00	698.0
315.00	1830.0
316.10	1197.0
316.90	221.0
321.00	575.0
322.10	305.0
323.10	4713.0
324.00	1013.0
325.90	190.0
327.00	1007.0
328.00	620.0
332.00	284.0
332.90	394.0
334.00	2960.0
335.00	809.0
341.10	604.0
341.90	163.0
346.00	1162.0
347.00	219.0
352.00	1412.0
353.00	1085.0
354.00	1332.0
355.00	410.0
365.00	5226.0
366.00	1005.0
371.00	337.0
372.10	1961.0
373.00	664.0
383.00	691.0
383.90	206.0
390.00	359.0
391.00	225.0
391.80	224.0
400.90	169.0
402.00	1036.0
403.00	1643.0
404.00	600.0
404.90	239.0
420.90	1199.0
422.00	1474.0
423.00	9437.0

Instrument :
BNA_N
ClientSampleId :
DFTPP288

424.00	3482.0
425.10	527.0
441.00	22536.0
442.00	135360.0
443.00	28752.0
444.00	2907.0

Instrument :
BNA_N
ClientSampleId :
DFTPP288

m/z	Abundance
37.00	816.0
38.00	1993.0
39.00	12420.0
40.00	1398.0
41.00	462.0
43.00	173.0
44.00	837.0
45.00	384.0
49.10	1424.0
50.00	44072.0
51.00	175104.0
52.10	9266.0
53.00	298.0
55.00	1041.0
56.00	4776.0
57.00	11952.0
57.90	586.0
59.00	161.0
61.00	2369.0
62.00	2689.0
63.00	6851.0
64.00	981.0
65.00	3583.0
65.90	245.0
66.90	328.0
67.90	2938.0
69.00	183424.0
70.00	899.0
71.00	170.0
73.00	2007.0
74.00	19464.0
75.00	30368.0
76.00	10619.0
77.00	221568.0
78.00	16472.0
79.00	13886.0
80.00	10330.0
81.00	15336.0
82.00	3903.0
83.00	3685.0
84.00	556.0
85.00	2784.0
86.00	4314.0
87.00	2196.0
88.00	791.0
88.90	341.0
91.00	3190.0
92.00	3780.0
93.00	24544.0
94.00	1587.0
95.00	401.0
96.00	1115.0
97.00	464.0
98.00	19184.0
99.00	16341.0
100.00	1314.0
101.00	8963.0
101.90	351.0
103.00	2574.0
104.00	5067.0
105.00	4899.0
106.00	1529.0
107.00	65816.0
108.00	10359.0
109.00	1828.0
110.00	120032.0
111.00	18464.0
112.00	2408.0
113.00	752.0
114.30	169.0
114.90	323.0
116.00	3230.0
117.00	58776.0
118.00	4734.0
118.90	528.0
120.00	826.0
121.10	253.0
122.00	3923.0
123.00	6794.0
124.00	3201.0
125.00	2809.0
126.10	967.0
127.00	224448.0
128.00	17424.0
129.00	85520.0
130.00	7304.0
131.00	1428.0
132.00	534.0
133.00	470.0
134.00	2519.0
135.00	7043.0
136.00	2732.0

Instrument :
BNA_N
ClientSampleId :
DFTPP288

137.00	4138.0
137.90	717.0
138.90	433.0
139.90	1107.0
141.00	11560.0
142.00	3681.0
143.00	2378.0
144.00	641.0
144.90	491.0
146.00	2163.0
147.00	5727.0
148.00	14539.0
149.00	3074.0
150.00	919.0
151.10	1469.0
153.00	3419.0
154.00	2588.0
155.00	5917.0
156.00	9992.0
157.10	1747.0
157.90	1980.0
159.00	1722.0
160.00	3340.0
161.00	5116.0
162.00	1422.0
163.00	401.0
164.10	635.0
165.00	3726.0
166.00	3283.0
167.00	22768.0
168.00	14736.0
169.00	2036.0
170.00	704.0
171.00	975.0
172.00	1840.0
173.00	2546.0
174.00	4479.0
175.00	8847.0
176.00	2218.0
176.90	3796.0
178.00	1169.0
179.00	15974.0
180.00	10701.0
181.00	5484.0
181.90	828.0
183.00	439.0
184.00	1164.0
185.00	8116.0
186.10	60232.0
187.00	18392.0
188.00	2004.0
189.00	3352.0
190.00	602.0
191.00	2179.0
192.00	5246.0
193.00	5992.0
194.00	1247.0
195.00	612.0
196.00	12043.0
196.90	2148.0
198.00	444352.0
199.00	27752.0
200.00	2275.0
201.30	1718.0
203.00	3129.0
204.00	15304.0
205.10	27696.0
206.00	114160.0
207.00	15972.0
208.00	3901.0
209.00	1338.0
210.20	1647.0
211.10	4137.0
213.00	313.0
215.00	1182.0
216.00	2346.0
217.00	29792.0
218.00	3683.0
219.10	424.0
221.00	17960.0
221.80	4350.0
223.10	6245.0
224.00	58624.0
225.00	15890.0
226.10	1919.0
227.00	26040.0
228.00	3698.0
229.00	5149.0
230.00	766.0
231.10	1739.0
232.00	409.0
233.00	470.0
234.00	1610.0
235.00	1709.0
235.90	1294.0
237.00	2504.0

Instrument :
BNA_N
ClientSampleId :
DFTPP288

238.00	444.0
239.00	1218.0
240.00	874.0
241.00	1521.0
242.00	3227.0
243.10	3443.0
244.10	47352.0
245.00	6147.0
246.00	9069.0
247.00	1852.0
248.00	437.0
249.00	1842.0
249.90	414.0
250.90	341.0
251.90	344.0
253.00	1015.0
254.10	1795.0
255.00	223104.0
256.00	34920.0
257.00	2840.0
258.00	13636.0
259.00	2430.0
259.90	509.0
261.00	459.0
262.90	169.0
263.90	241.0
265.00	5617.0
266.00	854.0
266.90	170.0
270.00	265.0
271.00	553.0
271.90	695.0
273.00	6948.0
274.10	18160.0
275.00	101176.0
276.00	12961.0
277.10	8949.0
278.00	1567.0
279.00	267.0
281.00	333.0
282.20	308.0
283.00	1133.0
284.00	749.0
285.00	1714.0
285.90	313.0
289.00	359.0
290.10	332.0
291.00	265.0
292.00	548.0
293.00	2037.0
294.00	677.0
295.00	428.0
296.00	29760.0
297.00	4658.0
298.00	295.0
300.90	360.0
302.10	498.0
303.00	2909.0
304.00	1005.0
308.00	442.0
309.00	295.0
309.90	419.0
313.00	296.0
314.00	1385.0
315.00	3562.0
316.10	1562.0
317.00	373.0
321.00	962.0
322.00	451.0
323.10	8371.0
324.00	1666.0
325.00	185.0
325.90	225.0
327.00	2028.0
328.00	926.0
328.80	280.0
332.00	661.0
333.00	1034.0
334.00	6227.0
335.00	1590.0
336.10	179.0
341.00	1039.0
342.00	266.0
346.00	2130.0
346.90	418.0
352.10	3011.0
353.00	2060.0
354.10	3167.0
355.00	663.0
359.10	212.0
365.00	13874.0
366.10	1813.0
366.90	234.0
369.90	255.0
371.00	711.0
372.00	5282.0

Instrument :
BNA_N
ClientSampleId :
DFTPP288

373.10	1409.0
373.80	158.0
383.00	1301.0
384.00	409.0
384.90	180.0
390.00	684.0
391.00	501.0
392.10	332.0
401.00	404.0
402.00	2245.0
403.00	3002.0
404.10	1473.0
405.10	236.0
421.00	2392.0
422.00	2808.0
423.10	16736.0
424.10	5244.0
425.10	752.0
440.00	174.0
441.10	43592.0
442.10	333632.0
443.00	62360.0
444.10	5026.0
445.00	316.0

Instrument :
BNA_N
ClientSampleId :
DFTPP288

m/z	Abundance
35.00	188.0
37.00	567.0
38.10	1818.0
39.00	11167.0
40.00	1359.0
41.00	513.0
43.00	314.0
44.00	802.0
45.00	362.0
49.00	1232.0
50.00	40808.0
51.00	156032.0
52.00	8053.0
53.00	376.0
55.00	1058.0
56.00	4486.0
57.00	10502.0
58.00	498.0
60.00	194.0
61.00	1996.0
62.00	2344.0
63.00	6251.0
64.00	867.0
65.00	3349.0
66.00	247.0
67.00	332.0
68.00	2633.0
69.00	167168.0
70.00	674.0
70.90	208.0
73.00	1752.0
74.00	17576.0
75.00	27088.0
76.00	9266.0
77.00	204032.0
78.00	14377.0
79.00	12440.0
80.00	9601.0
81.00	13552.0
82.00	3442.0
83.00	3222.0
83.90	517.0
85.00	2607.0
86.00	4250.0
87.00	1886.0
88.00	839.0
89.00	370.0
91.00	3152.0
92.00	3752.0
93.00	22584.0
94.00	1665.0
95.00	206.0
96.00	1333.0
96.90	349.0
97.10	377.0
98.00	17840.0
99.00	14299.0
100.00	1315.0
101.00	7575.0
102.00	465.0
103.00	2514.0
104.00	4788.0
105.00	4770.0
106.00	1314.0
107.00	58488.0
108.00	8966.0
109.00	1783.0
110.00	108416.0
111.00	16231.0
112.00	2265.0
113.00	677.0
114.10	249.0
115.00	228.0
116.00	3307.0
117.00	53272.0
118.00	4037.0
118.90	548.0
120.00	792.0
121.00	252.0
122.00	3701.0
123.00	5844.0
124.00	2884.0
125.00	2798.0
126.10	875.0
127.00	207488.0
128.00	15101.0
129.00	79992.0
130.00	6766.0
131.00	1304.0
132.00	602.0
133.00	570.0
134.00	2292.0

Instrument :
BNA_N
ClientSampleId :
DFTPP288

135.00	6438.0
136.00	2691.0
137.00	3518.0
138.00	783.0
139.00	407.0
140.00	978.0
141.00	10161.0
142.00	3446.0
143.00	2507.0
144.00	584.0
145.00	613.0
146.00	1746.0
147.00	5462.0
148.00	13406.0
149.00	2906.0
150.00	875.0
151.10	1446.0
151.80	870.0
153.00	3181.0
154.00	2118.0
155.00	5995.0
156.00	8612.0
157.00	1416.0
158.00	1750.0
159.00	1363.0
160.00	3524.0
161.00	4806.0
162.00	1423.0
163.00	478.0
164.00	560.0
165.00	4001.0
166.00	3076.0
167.00	23528.0
168.00	12470.0
169.00	1898.0
170.10	763.0
171.00	802.0
172.00	1812.0
173.00	2518.0
174.00	4492.0
175.00	9115.0
176.00	2348.0
177.00	3730.0
177.90	1326.0
179.00	16166.0
180.00	10799.0
181.00	5047.0
182.00	892.0
183.00	467.0
183.90	1130.0
185.00	7628.0
186.00	61352.0
187.00	17448.0
188.00	1658.0
189.00	3722.0
190.00	659.0
191.00	1615.0
192.00	4923.0
193.00	5815.0
194.10	1059.0
195.10	538.0
196.00	12360.0
196.80	2115.0
198.00	419008.0
199.00	27872.0
200.00	2033.0
201.50	1662.0
203.00	3012.0
204.00	15414.0
205.10	26176.0
206.10	105776.0
207.10	15780.0
208.00	3586.0
209.00	1395.0
210.10	1638.0
211.00	4227.0
213.10	349.0
215.00	1228.0
216.00	2266.0
217.00	28304.0
218.00	3863.0
219.00	314.0
221.10	15935.0
221.80	4126.0
223.00	6604.0
224.10	58704.0
225.00	16149.0
226.00	1774.0
227.00	23320.0
228.00	3540.0
229.00	5582.0
230.00	667.0
231.00	1911.0
232.00	370.0
233.00	491.0
234.00	1725.0

Instrument :
BNA_N
ClientSampleId :
DFTPP288

234.90	1685.0
235.90	1128.0
237.00	2640.0
238.00	345.0
239.00	1032.0
239.90	778.0
241.00	1464.0
242.00	3531.0
243.10	3627.0
244.10	50400.0
245.00	6684.0
246.00	9183.0
247.00	1839.0
248.00	397.0
249.00	1616.0
250.00	278.0
251.00	402.0
252.00	484.0
253.00	1072.0
254.10	1583.0
255.00	249216.0
256.00	37376.0
257.00	2876.0
258.00	14750.0
259.00	2225.0
260.00	394.0
260.90	373.0
264.00	371.0
265.00	6119.0
266.00	964.0
267.00	207.0
270.00	264.0
271.00	591.0
272.00	607.0
273.10	6858.0
274.00	20008.0
275.10	108752.0
276.00	13346.0
277.00	9655.0
278.00	1862.0
279.00	317.0
281.00	330.0
282.00	368.0
283.00	1025.0
284.00	599.0
285.00	1834.0
286.00	302.0
289.00	321.0
290.00	361.0
291.10	181.0
292.10	388.0
293.00	2317.0
294.00	728.0
295.00	808.0
296.00	29200.0
297.00	4690.0
298.00	307.0
300.90	425.0
301.90	626.0
303.00	3611.0
304.00	1016.0
304.90	184.0
308.00	486.0
309.00	366.0
310.00	483.0
313.00	316.0
314.00	1699.0
315.00	3950.0
316.10	2145.0
317.00	306.0
320.00	206.0
321.10	1086.0
322.00	534.0
323.10	9198.0
324.10	1983.0
325.00	195.0
326.00	259.0
327.10	2055.0
328.00	983.0
328.80	202.0
332.00	797.0
333.00	1156.0
334.10	6504.0
335.00	1983.0
336.10	366.0
339.00	156.0
339.90	234.0
341.00	1418.0
342.00	350.0
346.00	2606.0
346.90	383.0
352.00	2955.0
353.00	2446.0
354.00	3992.0
355.10	744.0
358.90	384.0

Instrument :
BNA_N
ClientSampleId :
DFTPP288

365.00	15156.0
366.00	2534.0
369.90	372.0
371.00	837.0
372.10	5955.0
373.00	1748.0
374.00	252.0
377.00	161.0
383.00	1374.0
384.00	619.0
385.10	211.0
389.90	685.0
391.00	571.0
391.90	420.0
401.00	423.0
402.00	2664.0
403.00	3483.0
404.00	1176.0
405.00	338.0
415.10	197.0
421.10	2842.0
422.10	3554.0
423.10	22528.0
424.00	7119.0
425.00	708.0
440.10	291.0
441.10	58616.0
442.10	398272.0
443.10	74240.0
444.00	6320.0
445.00	524.0

Instrument :
BNA_N
ClientSampleId :
DFTPP288