




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102224\
Data File : BN034672.D
Acq On : 22 Oct 2024 10:10
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 23 02:40:57 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 23 02:40:52 2024
Response via : Initial Calibration






TIC: BN034672.D\data.ms

(70) Pentachlorophenol (C)**16.687min (0.000) 17591.97 ng****response 356188**

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	62.20
264.00	61.60	62.69
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102224\
Data File : BN034672.D
Acq On : 22 Oct 2024 10:10
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 23 02:40:57 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 23 02:40:52 2024
Response via : Initial Calibration



TIC: BN034672.D\data.ms

(77) Benzidine

19.286min (0.000) 0.00 ng

response 2070557

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.42
183.00	13.20	11.51
0.00	0.00	0.00