

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102524\
 Data File : BN034738.D
 Acq On : 26 Oct 2024 12:41
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 28 00:53:46 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 28 00:44:15 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	129	0.00
2	1,4-Dioxane	0.400	0.355	11.3	125	0.00
3	n-Nitrosodimethylamine	0.400	0.379	5.3	134	0.00
4 S	2-Fluorophenol	0.400	0.413	-3.2	148	0.00
5 S	Phenol-d6	0.400	0.420	-5.0	153	0.00
6	bis(2-Chloroethyl)ether	0.400	0.380	5.0	136	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	133	0.00
8 S	Nitrobenzene-d5	0.400	0.362	9.5	137	0.00
9	Naphthalene	0.400	0.366	8.5	136	0.00
10	Hexachlorobutadiene	0.400	0.359	10.3	133	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.360	10.0	137	0.00
12	2-Methylnaphthalene	0.400	0.365	8.8	138	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	135	0.00
14 S	2,4,6-Tribromophenol	0.400	0.379	5.3	165	0.00
15 S	2-Fluorobiphenyl	0.400	0.344	14.0	134	0.00
16	Acenaphthylene	0.400	0.335	16.3	138	0.00
17	Acenaphthene	0.400	1.171	-192.8#	468	0.00
18	Fluorene	0.400	0.350	12.5	139	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	137	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.311	22.3	140	0.00
21	4-Bromophenyl-phenylether	0.400	0.351	12.3	135	0.00
22	Hexachlorobenzene	0.400	0.357	10.8	134	0.00
23	Atrazine	0.400	0.363	9.3	149	0.00
24	Pentachlorophenol	0.400	2.614	-553.5#	1144	0.00
25	Phenanthrene	0.400	0.361	9.8	139	0.00
26	Anthracene	0.400	0.353	11.8	138	0.00
27 SURR	Fluoranthene-d10	0.400	0.348	13.0	137	0.00
28	Fluoranthene	0.400	0.356	11.0	140	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	127	0.00
30	Pyrene	0.400	1.638	-309.5#	571	0.00
31 S	Terphenyl-d14	0.400	0.391	2.3	138	0.00
32	Benzo(a)anthracene	0.400	0.367	8.3	130	0.00
33	Chrysene	0.400	0.362	9.5	127	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.579	-44.7#	221	0.00
35 I	Perylene-d12	0.400	0.400	0.0	118	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.353	11.8	113	0.00
37	Benzo(b)fluoranthene	0.400	0.368	8.0	117	0.00
38	Benzo(k)fluoranthene	0.400	0.374	6.5	125	0.00
39 C	Benzo(a)pyrene	0.400	0.361	9.8	118	0.00
40	Dibenzo(a,h)anthracene	0.400	0.364	9.0	117	0.00
41	Benzo(g,h,i)perylene	0.400	0.355	11.3	111	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4